





Method Name: Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG112219.M
Calibration Table Last Updated: Sat Nov 23 00:54:00 2019

4,6-Dinitro-2-methylphenol




$R = 6.35e-003 A^2 + 6.37e-002 A - 1.34e-002$
Coef of Det (r^2) = 0.994 Curve Fit: Quadratic

Method Name: Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG112219.M
Calibration Table Last Updated: Sat Nov 23 00:54:00 2019

Carbazole

Response Ratio



$R = -6.64e-002 A^2 + 1.06e+000 A + 1.46e-001$
Coef of Det (r^2) = 0.990 Curve Fit: Quadratic

Method Name: Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG112219.M
Calibration Table Last Updated: Sat Nov 23 00:54:00 2019