




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG112223\
Data File : BG059837.D
Acq On : 23 Nov 2023 4:31
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 23 05:59:02 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 17 01:51:38 2023
Response via : Initial Calibration



TIC: BG059837.D\data.ms

(70) Pentachlorophenol (C)

17.336min (-0.003) 83730.90 ng

response 110168

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	57.10	65.82
264.00	62.30	68.39
0.00	0.00	0.00

Abundance



Ion 184.00 (183.70 to 184.70): BG059837.D\data.ms
Ion 185.00 (184.70 to 185.70): BG059837.D\data.ms
Ion 183.00 (182.70 to 183.70): BG059837.D\data.ms

800000
600000
400000
200000
0

19.938
19.938
Tail = 0.96

S E

me--> 18.80 19.00 19.20 19.40 19.60 19.80 20.00 20.20 20.40 20.60 20.80 21.00



TIC: BG059837.D\data.ms

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.20	13.70
183.00	11.00	11.57
0.00	0.00	0.00