

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG112321\
 Data File : BG051230.D
 Acq On : 25 Nov 2021 4:24
 Operator : CG/JU
 Sample : PB140965BS
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SLCS965

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 11/30/2021
 Supervised By :Sohil Jodhani 11/30/2021

Quant Time: Nov 25 07:40:21 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG112321.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 24 06:04:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.200	152	34090	20.000	ng/u1	0.00
20) Naphthalene-d8	11.026	136	151745	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.833	164	99124	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.583	188	209695	20.000	ng/u1	0.00
79) Chrysene-d12	21.883	240	179635	20.000	ng/u1	0.00
88) Perylene-d12	25.285	264	178214	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.540	96	5242	5.344	ng/uL	0.00
4) Pyridine-d5	3.969	84	69784	24.242	ng/u1	0.00
7) Phenol-d5	7.359	99	98235	29.156	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.512	67	61050	28.851	ng/u1	0.00
11) 2-Chlorophenol-d4	7.729	132	72081	29.710	ng/u1	0.00
15) 4-Methylphenol-d8	8.916	113	77663	28.564	ng/u1	0.00
21) Nitrobenzene-d5	9.380	128	37357	29.164	ng/u1	0.00
24) 2-Nitrophenol-d4	10.103	143	43630	30.195	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.650	165	75003	30.593	ng/u1	0.00
31) 4-Chloroaniline-d4	11.167	131	95106	26.512	ng/u1	0.00
46) Dimethylphthalate-d6	14.228	166	221356	29.023	ng/u1	0.00
49) Acenaphthylene-d8	14.527	160	288587	30.006	ng/u1	0.00
54) 4-Nitrophenol-d4	15.050	143	33187	26.882	ng/u1	0.00
60) Fluorene-d10	15.820	176	199535	29.052	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.955	200	35370	27.335	ng/u1	0.00
73) Anthracene-d10	17.683	188	292874	29.203	ng/u1	0.00
81) Pyrene-d10	19.962	212	327952	30.172	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.056	264	286108	30.060	ng/u1	0.01
Target Compounds						Qvalue
2) 1,4-Dioxane	3.576	88	10967	9.913	ng/uL	93
5) Pyridine	3.987	79	75675	25.264	ng/u1	96
6) Benzaldehyde	7.330	77	64551m	30.084	ng/u1	
8) Phenol	7.389	94	103170	29.559	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.606	93	76098	28.819	ng/u1	97
12) 2-Chlorophenol	7.765	128	74258	30.035	ng/u1	96
13) 2-Methylphenol	8.646	108	77871	29.952	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.722	45	113024	29.662	ng/u1#	95
16) Acetophenone	9.028	105	121109	28.798	ng/u1	99
17) N-Nitroso-di-n-propyla...	9.004	70	70150	29.028	ng/u1	98
18) 4-Methylphenol	8.975	108	82089	29.528	ng/u1	99
19) Hexachloroethane	9.286	117	30074	28.799	ng/u1	97
22) Nitrobenzene	9.422	77	102871	30.627	ng/u1	97
23) Isophorone	9.939	82	191587	29.360	ng/u1	100
25) 2-Nitrophenol	10.138	139	45540	30.427	ng/u1	100
26) 2,4-Dimethylphenol	10.185	107	91289	29.833	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.415	93	106962	29.691	ng/u1	99
29) 2,4-Dichlorophenol	10.679	162	73102	30.291	ng/u1	96
30) Naphthalene	11.079	128	247810	30.013	ng/u1	98
32) 4-Chloroaniline	11.190	127	98508	27.353	ng/u1	100
33) Hexachlorobutadiene	11.343	225	47566	28.575	ng/u1	96
34) Caprolactam	11.960	113	27648m	29.141	ng/u1	
35) 4-Chloro-3-methylphenol	12.306	107	89718	30.948	ng/u1	97
36) 2-Methylnaphthalene	12.671	142	167803	29.879	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG112321\
 Data File : BG051230.D
 Acq On : 25 Nov 2021 4:24
 Operator : CG/JU
 Sample : PB140965BS
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SLCS965


Quant Time: Nov 25 07:40:21 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG112321.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 24 06:04:50 2021
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 11/30/2021
 Supervised By :Sohil Jodhani 11/30/2021


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.888	142	171097	29.612	ng/ul	96
39) 1,2,4,5-Tetrachloroben...	13.035	216	98285	31.584	ng/ul	97
40) Hexachlorocyclopentadiene	13.000	237	16716	13.290	ng/ul	92
41) 2,4,6-Trichlorophenol	13.276	196	63832	32.687	ng/ul	99
42) 2,4,5-Trichlorophenol	13.358	196	66467	32.502	ng/ul	98
43) 1,1'-Biphenyl	13.664	154	226577	30.604	ng/ul	98
44) 2-Chloronaphthalene	13.717	162	182537	30.995	ng/ul	99
45) 2-Nitroaniline	13.922	65	65492	32.132	ng/ul	92
47) Dimethylphthalate	14.275	163	229061	29.671	ng/ul	100
48) 2,6-Dinitrotoluene	14.410	165	50841	31.352	ng/ul	90
50) Acenaphthylene	14.557	152	287812	30.290	ng/ul	98
51) 3-Nitroaniline	14.745	138	51039	31.841	ng/ul	98
52) Acenaphthene	14.898	153	190790	30.446	ng/ul	98
53) 2,4-Dinitrophenol	14.968	184	19023	21.223	ng/ul	89
55) 4-Nitrophenol	15.062	109	29823	27.847	ng/ul	96
56) Dibenzofuran	15.232	168	270013	29.873	ng/ul	100
57) 2,4-Dinitrotoluene	15.203	165	69392	29.960	ng/ul	94
58) 2,3,4,6-Tetrachlorophenol	15.462	232	48655	30.298	ng/ul#	96
59) Diethylphthalate	15.626	149	237811	29.347	ng/ul	99
61) Fluorene	15.879	166	217238	30.005	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.861	204	114533	29.354	ng/ul	96
63) 4-Nitroaniline	15.908	138	51807	33.212	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.967	198	35454	28.411	ng/ul#	98
67) N-Nitrosodiphenylamine	16.079	169	189325	31.537	ng/ul	99
68) 4-Bromophenyl-phenylether	16.754	248	70921	31.556	ng/ul	94
69) Hexachlorobenzene	16.883	284	71015	30.989	ng/ul	96
70) Atrazine	17.019	200	73058	28.957	ng/ul	99
71) Pentachlorophenol	17.236	266	24165m	23.797	ng/ul	
72) Phenanthrene	17.624	178	353665	30.546	ng/ul	100
74) Anthracene	17.718	178	344751	29.982	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.634	216	102366	33.468	ng/ul	99
76) Pentachlorobenzene	15.150	250	89019	31.236	ng/ul	99
77) Carbazole	17.988	167	309850	30.699	ng/ul	99
78) Di-n-butylphthalate	18.511	149	396699	30.482	ng/ul	100
80) Fluoranthene	19.627	202	414015	31.012	ng/ul	97
82) Pyrene	19.992	202	401963	30.781	ng/ul	97
83) Butylbenzylphthalate	20.849	149	171031	31.503	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.766	252	122239	29.227	ng/ul	99
85) Benzo(a)anthracene	21.860	228	372842	30.601	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.719	149	245784	31.461	ng/ul	99
87) Chrysene	21.930	228	355590	30.380	ng/ul	99
89) Di-n-octyl phthalate	22.988	149	415385	32.173	ng/ul	100
90) Benzo(b)fluoranthene	24.198	252	375978	31.261	ng/ul	99
91) Benzo(k)fluoranthene	24.269	252	341773	30.282	ng/ul	99
93) Benzo(a)pyrene	25.127	252	354592	30.904	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	29.198	276	398316	31.022	ng/ul	98
95) Dibenzo(a,h)anthracene	29.257	278	337303	30.966	ng/ul	98
96) Benzo(g,h,i)perylene	30.438	276	331143	30.653	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 9.913 ng/uL
RT: 3.576 min Scan# 11
Delta R.T. -0.009 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Instrument :
BNA_G
ClientSampleId :
SLCS965



Tgt Ion	Ratio	Lower	Upper
88	100		
43	29.4	23.0	34.4
58	70.6	62.4	93.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#5
Pyridine
Concen: 25.264 ng/uL
RT: 3.987 min Scan# 85
Delta R.T. -0.015 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24

Tgt Ion	Ratio	Lower	Upper
79	100		
52	79.0	60.8	91.2
51	34.8	30.5	45.7









#6
Benzaldehyde
Concen: 30.084 ng/ul m
RT: 7.330 min Scan# 61
Delta R.T. -0.003 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24

Instrument :
BNA_G
ClientSampleId :
SLCS965

Tgt Ion: 77 Resp: 6455
Ion Ratio Lower Upper
77 100
105 85.1 70.4 105.6
106 82.7 61.2 91.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#8
Phenol
Concen: 29.559 ng/ul
RT: 7.389 min Scan# 664
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24


Tgt Ion: 94 Resp: 103170
Ion Ratio Lower Upper
94 100
65 29.3 25.1 37.7
66 39.2 31.6 47.4





#10
Bis(2-Chloroethyl)ether
Concen: 28.819 ng/ul
RT: 7.606 min Scan# 701
Delta R.T. -0.004 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Instrument :
BNA_G
ClientSampleId :
SLCS965




Tgt Ion: 93 Resp: 76098
Ion Ratio Lower Upper
93 100
63 82.6 68.6 103.0
95 32.0 26.1 39.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#12
2-Chlorophenol
Concen: 30.035 ng/ul
RT: 7.765 min Scan# 728
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24



Tgt Ion:128 Resp: 74258
Ion Ratio Lower Upper
128 100
64 54.2 45.3 67.9
130 30.1 27.1 40.7





#13

2-Methylphenol

Concen: 29.952 ng/ul

RT: 8.646 min Scan# 81

Delta R.T. 0.002 min

Lab File: BG051230.D

Acq: 25 Nov 2021 4:24

Instrument :

BNA_G

ClientSampleId :

SLCS965

Tgt Ion:108 Resp: 7787

Ion Ratio Lower Upper

108 100





107 90.5 71.8 107.6

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/30/2021

Supervised By :Sohil Jodhani 11/30/2021



#14

2,2'-oxybis(1-Chloropropane)

Concen: 29.662 ng/ul

RT: 8.722 min Scan# 891

Delta R.T. -0.004 min

Lab File: BG051230.D

Acq: 25 Nov 2021 4:24




Tgt Ion: 45 Resp: 113024


Ion Ratio Lower Upper

45 100

77 16.2 10.7 16.1#


79 11.4 8.5 12.7





#16
Acetophenone
Concen: 28.798 ng/ul
RT: 9.028 min Scan# 943
Delta R.T. -0.003 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Instrument :
BNA_G
ClientSampleId :
SLCS965



Tgt Ion: 105 Resp: 121109
Ion Ratio Lower Upper
105 100
77 84.2 67.3 100.9
51 27.5 24.0 36.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#17
N-Nitroso-di-n-propylamine
Concen: 29.028 ng/ul
RT: 9.004 min Scan# 939
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24

Tgt Ion: 70 Resp: 70150
Ion Ratio Lower Upper
70 100
42 56.0 46.0 69.0
101 8.3 7.0 10.4
130 17.0 13.4 20.2





#18

4-Methylphenol

Concen: 29.528 ng/ul

RT: 8.975 min Scan# 91

Delta R.T. -0.003 min

Lab File: BG051230.D

Acq: 25 Nov 2021 4:24

Instrument :

BNA_G

ClientSampleId :

SLCS965

Tgt Ion:108 Resp: 82089

Ion Ratio Lower Upper

108 100





107 119.7 95.0 142.4

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/30/2021

Supervised By :Sohil Jodhani 11/30/2021



#19

Hexachloroethane

Concen: 28.799 ng/ul

RT: 9.286 min Scan# 987

Delta R.T. -0.004 min

Lab File: BG051230.D

Acq: 25 Nov 2021 4:24




Tgt Ion:117 Resp: 30074


Ion Ratio Lower Upper

117 100

201 101.7 77.2 115.8


199 62.9 49.9 74.9





#22
Nitrobenzene
Concen: 30.627 ng/ul
RT: 9.422 min Scan# 1010
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Instrument :
BNA_G
ClientSampleId :
SLCS965




Tgt Ion: 77 Resp: 10287
Ion Ratio Lower Upper
77 100
123 39.7 30.1 45.1
65 13.2 9.6 14.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#23
Isophorone
Concen: 29.360 ng/ul
RT: 9.939 min Scan# 1098
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Tgt Ion: 82 Resp: 191587
Ion Ratio Lower Upper
82 100
95 7.1 5.6 8.4
138 14.7 11.8 17.8





#25
2-Nitrophenol
Concen: 30.427 ng/ul
RT: 10.138 min Scan# 1131
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Instrument :
BNA_G
ClientSampleId :
SLCS965




Tgt Ion:139 Resp: 45540
Ion Ratio Lower Upper
139 100
65 60.1 48.0 72.0
109 35.1 28.6 42.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#26
2,4-Dimethylphenol
Concen: 29.833 ng/ul
RT: 10.185 min Scan# 1140
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24



Tgt Ion:107 Resp: 91289
Ion Ratio Lower Upper
107 100
121 46.8 39.3 58.9
122 78.0 63.7 95.5





#27

Bis(2-Chloroethoxy)methane

Concen: 29.691 ng/ul

RT: 10.415 min Scan# 1179

Delta R.T. 0.002 min

Lab File: BG051230.D

Acq: 25 Nov 2021 4:24

Instrument :

BNA_G

ClientSampleId :

SLCS965

Tgt Ion: 93 Resp: 106962

Ion Ratio Lower Upper

93 100

95 32.5 26.5 39.7





123 9.4 8.1 12.1

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/30/2021

Supervised By :Sohil Jodhani 11/30/2021



#29

2,4-Dichlorophenol

Concen: 30.291 ng/ul

RT: 10.679 min Scan# 1224

Delta R.T. 0.002 min

Lab File: BG051230.D

Acq: 25 Nov 2021 4:24




Tgt Ion:162 Resp: 73102


Ion Ratio Lower Upper

162 100

164 66.4 49.4 74.0


98 39.0 30.4 45.6





#30
Naphthalene
Concen: 30.013 ng/ul
RT: 11.079 min Scan# 1292
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Instrument :
BNA_G
ClientSampleId :
SLCS965




Tgt Ion:128 Resp: 247810
Ion Ratio Lower Upper
128 100
129 11.3 8.5 12.7
127 13.8 11.8 17.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#32
4-Chloroaniline
Concen: 27.353 ng/ul
RT: 11.190 min Scan# 1311
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Tgt Ion:127 Resp: 98508
Ion Ratio Lower Upper
127 100
129 33.0 26.2 39.2





#33
Hexachlorobutadiene
Concen: 28.575 ng/ul
RT: 11.343 min Scan# 11337
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Instrument :
BNA_G
ClientSampleId :
SLCS965



Tgt Ion:225 Resp: 47560
Ion Ratio Lower Upper
225 100
223 66.6 50.6 75.8
227 64.8 49.8 74.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#34
Caprolactam
Concen: 29.141 ng/ul m
RT: 11.960 min Scan# 1442
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24

Tgt Ion:113 Resp: 27648
Ion Ratio Lower Upper
113 100
55 187.9 147.0 220.6
56 139.7 109.2 163.8





#35

4-Chloro-3-methylphenol

Concen: 30.948 ng/ul

RT: 12.306 min Scan# 1501

Delta R.T. 0.002 min

Lab File: BG051230.D

Acq: 25 Nov 2021 4:24

Instrument :

BNA_G

ClientSampleId :





SLCS965

Tgt Ion:	107	Resp:	89718
Ion Ratio	Lower	Upper	
107	100		
144	24.9	19.8	29.6
142	73.9	56.2	84.4

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#36

2-Methylnaphthalene

Concen: 29.879 ng/ul




RT: 12.671 min Scan# 1563


Delta R.T. 0.002 min

Lab File: BG051230.D

Acq: 25 Nov 2021 4:24

Tgt Ion:	142	Resp:	167803
Ion Ratio	Lower	Upper	
142	100		
141	93.5	73.3	109.9





#37

1-Methylnaphthalene

Concen: 29.612 ng/ul

RT: 12.888 min Scan# 1

Delta R.T. 0.002 min

Lab File: BG051230.D

Acq: 25 Nov 2021 4:24

Instrument :

BNA_G

ClientSampleId :





SLCS965

Tgt Ion:142	Resp:	17109
Ion Ratio	Lower	Upper
142	100	
141	93.4	77.8 116.8
116	4.3	3.3 4.9

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#39

1,2,4,5-Tetrachlorobenzene

Concen: 31.584 ng/ul




RT: 13.035 min Scan# 1625


Delta R.T. 0.002 min

Lab File: BG051230.D

Acq: 25 Nov 2021 4:24


Tgt Ion:216	Resp:	98285
Ion Ratio	Lower	Upper
216	100	
214	79.0	61.0 91.4
179	21.0	16.2 24.2
108	16.6	13.1 19.7





#40
Hexachlorocyclopentadiene
Concen: 13.290 ng/ul
RT: 13.000 min Scan# 1619
Delta R.T. -0.003 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Instrument :
BNA_G
ClientSampleId :
SLCS965




Tgt Ion:237 Resp: 16710
Ion Ratio Lower Upper
237 100
235 68.3 49.4 74.0
272 12.6 11.1 16.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#41
2,4,6-Trichlorophenol
Concen: 32.687 ng/ul
RT: 13.276 min Scan# 1666
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24



Tgt Ion:196 Resp: 63832
Ion Ratio Lower Upper
196 100
198 91.4 74.3 111.5
200 30.2 23.8 35.6






#42
2,4,5-Trichlorophenol
Concen: 32.502 ng/ul
RT: 13.358 min Scan# 1680
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24

Instrument :

BNA_G

ClientSampleId :

SLCS965






Tgt Ion:196 Resp: 6646
Ion Ratio Lower Upper
196 100
198 95.9 77.0 115.4
200 28.2 25.4 38.2


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#43
1,1'-Biphenyl
Concen: 30.604 ng/ul
RT: 13.664 min Scan# 1732
Delta R.T. -0.003 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Tgt Ion:154 Resp: 226577
Ion Ratio Lower Upper
154 100
153 42.4 34.5 51.7
76 15.3 13.1 19.7





#44
2-Chloronaphthalene
Concen: 30.995 ng/ul
RT: 13.717 min Scan# 1741
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Instrument :
BNA_G
ClientSampleId :
SLCS965



Tgt Ion:162 Resp: 18253
Ion Ratio Lower Upper
162 100
164 32.3 26.1 39.1
127 39.6 31.0 46.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#45
2-Nitroaniline
Concen: 32.132 ng/ul
RT: 13.922 min Scan# 1776
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24


Tgt Ion: 65 Resp: 65492
Ion Ratio Lower Upper
65 100
92 66.6 53.0 79.4
138 96.3 67.4 101.0





#47
Dimethylphthalate
Concen: 29.671 ng/ul
RT: 14.275 min Scan# 1835
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Instrument :
BNA_G
ClientSampleId :
SLCS965




Tgt Ion:163 Resp: 22906
Ion Ratio Lower Upper
163 100
194 3.8 3.0 4.4
164 10.3 8.2 12.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#48
2,6-Dinitrotoluene
Concen: 31.352 ng/ul
RT: 14.410 min Scan# 1859
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Tgt Ion:165 Resp: 50841
Ion Ratio Lower Upper
165 100
89 61.7 57.4 86.0
121 21.8 18.8 28.2





#50
Acenaphthylene
Concen: 30.290 ng/ul
RT: 14.557 min Scan# 1884
Delta R.T. -0.003 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Instrument :
BNA_G
ClientSampleId :
SLCS965



Tgt Ion:152 Resp: 28781
Ion Ratio Lower Upper
152 100
151 20.4 17.4 26.0
153 13.2 10.3 15.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#51
3-Nitroaniline
Concen: 31.841 ng/ul
RT: 14.745 min Scan# 1916
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24


Tgt Ion:138 Resp: 51039
Ion Ratio Lower Upper
138 100
108 9.6 8.8 13.2
92 121.0 98.0 147.0





#52
Acenaphthene
Concen: 30.446 ng/ul
RT: 14.898 min Scan# 1942
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Instrument :
BNA_G
ClientSampleId :
SLCS965




Tgt Ion:153 Resp: 190790
Ion Ratio Lower Upper
153 100
152 47.1 36.5 54.7
154 84.1 65.8 98.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#53
2,4-Dinitrophenol
Concen: 21.223 ng/ul
RT: 14.968 min Scan# 1954
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24



Tgt Ion:184 Resp: 19023
Ion Ratio Lower Upper
184 100
63 69.3 66.2 99.2
154 62.3 53.6 80.4









#55
4-Nitrophenol
Concen: 27.847 ng/ul
RT: 15.062 min Scan# 1970
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24

Instrument :
BNA_G
ClientSampleId :
SLCS965

Tgt Ion:109 Resp: 2982
Ion Ratio Lower Upper
109 100
139 110.5 88.7 133.1
65 133.3 113.6 170.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#56
Dibenzofuran
Concen: 29.873 ng/ul
RT: 15.232 min Scan# 1999
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24

Tgt Ion:168 Resp: 270013
Ion Ratio Lower Upper
168 100
139 39.0 31.2 46.8









#57
2,4-Dinitrotoluene
Concen: 29.960 ng/ul
RT: 15.203 min Scan# 1994
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24

Instrument :
BNA_G
ClientSampleId :
SLCS965

Tgt Ion:165 Resp: 69392
Ion Ratio Lower Upper
165 100
63 48.3 42.1 63.1
182 2.3 1.8 2.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#58
2,3,4,6-Tetrachlorophenol
Concen: 30.298 ng/ul
RT: 15.462 min Scan# 2038
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24


Tgt Ion:232 Resp: 48655
Ion Ratio Lower Upper
232 100
131 45.5 38.9 58.3
130 2.4 2.4 3.6#
166 30.8 25.8 38.8





#59
Diethylphthalate
Concen: 29.347 ng/ul
RT: 15.626 min Scan# 2066
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Instrument :
BNA_G
ClientSampleId :
SLCS965




Tgt Ion:149 Resp: 23781
Ion Ratio Lower Upper
149 100
177 21.3 16.8 25.2
150 11.9 10.1 15.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#61
Fluorene
Concen: 30.005 ng/ul
RT: 15.879 min Scan# 2109
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24



Tgt Ion:166 Resp: 217238
Ion Ratio Lower Upper
166 100
165 99.4 81.0 121.6
167 13.8 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 29.354 ng/ul
RT: 15.861 min Scan# 2106
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24

Instrument :
BNA_G
ClientSampleId :
SLCS965

Tgt Ion:204 Resp: 11453
Ion Ratio Lower Upper
204 100
206 32.8 27.4 41.2
141 56.8 48.6 73.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#63
4-Nitroaniline
Concen: 33.212 ng/ul
RT: 15.908 min Scan# 2114
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24


Tgt Ion:138 Resp: 51807
Ion Ratio Lower Upper
138 100
92 57.4 49.3 73.9
108 93.9 72.6 108.8





#66
4,6-Dinitro-2-methylphenol
Concen: 28.411 ng/ul
RT: 15.967 min Scan# 2124
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Instrument :
BNA_G
ClientSampleId :
SLCS965




Tgt Ion:198 Resp: 35454
Ion Ratio Lower Upper
198 100
182 3.4 3.8 5.8
77 28.2 23.2 34.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#67
N-Nitrosodiphenylamine
Concen: 31.537 ng/ul
RT: 16.079 min Scan# 2143
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Tgt Ion:169 Resp: 189325
Ion Ratio Lower Upper
169 100
168 71.8 58.5 87.7
167 39.6 31.3 46.9





#68
4-Bromophenyl-phenylether
Concen: 31.556 ng/ul
RT: 16.754 min Scan# 21
Delta R.T. -0.003 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Instrument :
BNA_G
ClientSampleId :
SLCS965




Tgt Ion:248 Resp: 7092
Ion Ratio Lower Upper
248 100
250 95.4 75.5 113.3
141 73.7 67.6 101.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#69
Hexachlorobenzene
Concen: 30.989 ng/ul
RT: 16.883 min Scan# 2280
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24



Tgt Ion:284 Resp: 71015
Ion Ratio Lower Upper
284 100
142 36.9 31.9 47.9
249 34.4 25.8 38.6





#70
Atrazine
Concen: 28.957 ng/ul
RT: 17.019 min Scan# 21
Delta R.T. -0.003 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24

Instrument :

BNA_G

ClientSampleId :





SLCS965

Tgt Ion:200 Resp: 73058
Ion Ratio Lower Upper
200 100
173 24.9 20.2 30.2
215 46.8 38.1 57.1

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#71
Pentachlorophenol
Concen: 23.797 ng/ul m
RT: 17.236 min Scan# 2340
Delta R.T. -0.003 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24


Tgt Ion:266 Resp: 24165
Ion Ratio Lower Upper
266 100
264 66.5 54.3 81.5
268 62.6 51.0 76.6





#72
Phenanthrene
Concen: 30.546 ng/u1
RT: 17.624 min Scan# 2406
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Instrument :
BNA_G
ClientSampleId :
SLCS965




Tgt Ion:178 Resp: 35366
Ion Ratio Lower Upper
178 100
179 15.3 12.3 18.5
176 19.5 15.5 23.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#74
Anthracene
Concen: 29.982 ng/u1
RT: 17.718 min Scan# 2422
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Tgt Ion:178 Resp: 344751
Ion Ratio Lower Upper
178 100
179 16.0 13.0 19.6
176 18.5 15.6 23.4





#75
1,2,3,4-Tetrachlorobenzene
Concen: 33.468 ng/uL
RT: 13.634 min Scan# 1727
Delta R.T. -0.003 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Instrument :
BNA_G
ClientSampleId :
SLCS965




Tgt Ion:216 Resp: 102360
Ion Ratio Lower Upper
216 100
214 77.9 61.3 91.9
218 46.8 37.0 55.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#76
Pentachlorobenzene
Concen: 31.236 ng/uL
RT: 15.150 min Scan# 1985
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Tgt Ion:250 Resp: 89019
Ion Ratio Lower Upper
250 100
248 62.0 50.5 75.7
252 63.9 51.2 76.8





#77
 Carbazole
 Concen: 30.699 ng/ul
 RT: 17.988 min Scan# 2468
 Delta R.T. 0.002 min
 Lab File: BG051230.D
 Acq: 25 Nov 2021 4:24




Instrument :
 BNA_G
 ClientSampleId :
 SLCS965




Tgt Ion:167 Resp: 309850
 Ion Ratio Lower Upper
 167 100
 166 22.7 17.3 25.9
 139 13.4 10.6 16.0

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
 Supervised By :Sohil Jodhani 11/30/2021




#78
 Di-n-butylphthalate
 Concen: 30.482 ng/ul
 RT: 18.511 min Scan# 2557
 Delta R.T. -0.003 min
 Lab File: BG051230.D
 Acq: 25 Nov 2021 4:24




Tgt Ion:149 Resp: 396699
 Ion Ratio Lower Upper
 149 100
 150 9.5 7.5 11.3
 104 6.0 5.0 7.6





#80
Fluoranthene
Concen: 31.012 ng/u1
RT: 19.627 min Scan# 21
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Instrument :
BNA_G
ClientSampleId :
SLCS965




Tgt Ion:202 Resp: 414015
Ion Ratio Lower Upper
202 100
101 10.1 9.3 13.9
100 8.3 7.0 10.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#82
Pyrene
Concen: 30.781 ng/u1
RT: 19.992 min Scan# 2809
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24



Tgt Ion:202 Resp: 401963
Ion Ratio Lower Upper
202 100
101 12.0 10.7 16.1
100 10.1 9.0 13.6





#83
Butylbenzylphthalate
Concen: 31.503 ng/ul
RT: 20.849 min Scan# 2955
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24

Instrument :

BNA_G





ClientSampleId :

SLCS965

Tgt Ion:149 Resp: 17103
Ion Ratio Lower Upper
149 100
91 76.7 65.2 97.8
206 21.1 16.2 24.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#84
3,3'-Dichlorobenzidine
Concen: 29.227 ng/ul
RT: 21.766 min Scan# 3111
Delta R.T. -0.003 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24


Tgt Ion:252 Resp: 122239
Ion Ratio Lower Upper
252 100
254 64.6 50.8 76.2
126 10.7 9.0 13.4





#85
Benzo(a)anthracene
Concen: 30.601 ng/ul
RT: 21.860 min Scan# 3127
Delta R.T. -0.003 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Instrument :
BNA_G
ClientSampleId :
SLCS965




Tgt Ion:228 Resp: 37284
Ion Ratio Lower Upper
228 100
229 20.4 15.7 23.5
226 27.7 22.3 33.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#86
Bis(2-ethylhexyl)phthalate
Concen: 31.461 ng/ul
RT: 21.719 min Scan# 3103
Delta R.T. -0.003 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Tgt Ion:149 Resp: 245784
Ion Ratio Lower Upper
149 100
167 26.1 21.1 31.7
279 3.6 2.7 4.1





#87
Chrysene
Concen: 30.380 ng/ul
RT: 21.930 min Scan# 3139
Delta R.T. -0.003 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Instrument :
BNA_G
ClientSampleId :
SLCS965



Tgt Ion:228 Resp: 355590
Ion Ratio Lower Upper
228 100
226 30.9 24.8 37.2
229 20.5 15.8 23.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#89
Di-n-octyl phthalate
Concen: 32.173 ng/ul
RT: 22.988 min Scan# 3319
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24


Tgt Ion:149 Resp: 415385





#90
Benzo(b)fluoranthene
Concen: 31.261 ng/ul
RT: 24.198 min Scan# 3525
Delta R.T. 0.008 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Instrument :
BNA_G
ClientSampleId :
SLCS965




Tgt Ion:252 Resp: 375978
Ion Ratio Lower Upper
252 100
253 22.4 17.8 26.6
125 9.4 8.5 12.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#91
Benzo(k)fluoranthene
Concen: 30.282 ng/ul
RT: 24.269 min Scan# 3537
Delta R.T. 0.008 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Tgt Ion:252 Resp: 341773
Ion Ratio Lower Upper
252 100
253 22.5 17.3 25.9
125 9.4 7.8 11.6





#93
Benzo(a)pyrene
Concen: 30.904 ng/ul
RT: 25.127 min Scan# 30
Delta R.T. 0.002 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Instrument :
BNA_G
ClientSampleId :
SLCS965




Tgt Ion:252 Resp: 354592
Ion Ratio Lower Upper
252 100
253 21.3 17.4 26.2
125 10.1 9.1 13.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021




#94
Indeno(1,2,3-cd)pyrene
Concen: 31.022 ng/ul
RT: 29.198 min Scan# 4376
Delta R.T. 0.008 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Tgt Ion:276 Resp: 398316
Ion Ratio Lower Upper
276 100
138 18.0 15.5 23.3
277 24.7 20.5 30.7





#95
Dibenzo(a,h)anthracene
Concen: 30.966 ng/ul
RT: 29.257 min Scan# 41
Delta R.T. 0.014 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24




Instrument :
BNA_G
ClientSampleId :
SLCS965




Tgt Ion:278 Resp: 33730
Ion Ratio Lower Upper
278 100
139 13.6 12.2 18.4
279 24.2 19.5 29.3

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 11/30/2021
Supervised By :Sohil Jodhani 11/30/2021



#96
Benzo(g,h,i)perylene
Concen: 30.653 ng/ul
RT: 30.438 min Scan# 4587
Delta R.T. 0.026 min
Lab File: BG051230.D
Acq: 25 Nov 2021 4:24



Tgt Ion:276 Resp: 331143
Ion Ratio Lower Upper
276 100
138 18.8 16.6 24.8
277 23.4 17.6 26.4

