

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG112423\
 Data File : BG059870.D
 Acq On : 24 Nov 2023 23:04
 Operator : MA/JU
 Sample : 05412-05
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 MH-2

Quant Time: Nov 25 01:03:52 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 24 15:11:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.318	152	30782	20.000	ng	0.00
21) Naphthalene-d8	11.167	136	140632	20.000	ng	# 0.00
39) Acenaphthene-d10	14.951	164	120177	20.000	ng	0.00
64) Phenanthrene-d10	17.701	188	287963	20.000	ng	0.00
76) Chrysene-d12	22.019	240	220491	20.000	ng	-0.01
86) Perylene-d12	25.592	264	234128	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.815	112	140531	76.841	ng	0.00
7) Phenol-d6	7.442	99	198128	73.105	ng	0.00
23) Nitrobenzene-d5	9.516	82	231336	62.825	ng	0.00
42) 2,4,6-Tribromophenol	16.438	330	155849	114.296	ng	0.00
45) 2-Fluorobiphenyl	13.570	172	525771	58.889	ng	0.00
79) Terphenyl-d14	20.274	244	968456	64.841	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG112423\
Data File : BG059870.D
Acq On : 24 Nov 2023 23:04
Operator : MA/JU
Sample : 05412-05
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
MH-2

Quant Time: Nov 25 01:03:52 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 24 15:11:07 2023
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.318 min Scan# 81
Delta R.T. 0.003 min
Lab File: BG059870.D
Acq: 24 Nov 2023 23:04




Instrument :
BNA_G
ClientSampleId :
MH-2


Tgt Ion:152 Resp: 30782
Ion Ratio Lower Upper
152 100
150 163.3 122.2 183.4
115 65.7 54.2 81.4



#5
2-Fluorophenol
Concen: 76.841 ng
RT: 5.815 min Scan# 413
Delta R.T. 0.003 min
Lab File: BG059870.D
Acq: 24 Nov 2023 23:04

Tgt Ion:112 Resp: 140531
Ion Ratio Lower Upper
112 100
64 51.0 49.5 74.3
63 44.6 37.4 56.2









#7
Phenol-d6
Concen: 73.105 ng
RT: 7.442 min Scan# 691
Delta R.T. -0.003 min
Lab File: BG059870.D
Acq: 24 Nov 2023 23:04




Instrument :
BNA_G
ClientSampleId :
MH-2


Tgt Ion: 99 Resp: 198128
Ion Ratio Lower Upper
99 100
42 21.3 14.3 21.5
71 83.3 59.3 88.9



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.167 min Scan# 1324
Delta R.T. -0.003 min
Lab File: BG059870.D
Acq: 24 Nov 2023 23:04

Tgt Ion: 136 Resp: 140632
Ion Ratio Lower Upper
136 100
137 11.0 7.4 11.0
54 6.0 3.1 4.7#
68 1.3 1.8 2.8#





#23

Nitrobenzene-d5

Concen: 62.825 ng

RT: 9.516 min Scan# 1043

Delta R.T. -0.009 min

Lab File: BG059870.D

Acq: 24 Nov 2023 23:04

Instrument :

BNA_G

ClientSampleId :

MH-2





Tgt Ion: 82 Resp: 231336

Ion Ratio Lower Upper

82 100

128 31.2 23.0 34.6

54 36.7 28.7 43.1



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.951 min Scan# 1968

Delta R.T. -0.003 min

Lab File: BG059870.D

Acq: 24 Nov 2023 23:04




Tgt Ion:164 Resp: 120177


Ion Ratio Lower Upper

164 100

162 90.8 73.3 109.9

160 45.2 34.1 51.1









#42
2,4,6-Tribromophenol
Concen: 114.296 ng
RT: 16.438 min Scan# 21
Delta R.T. -0.003 min
Lab File: BG059870.D
Acq: 24 Nov 2023 23:04




Instrument :
BNA_G
ClientSampleId :
MH-2


Tgt Ion:330 Resp: 155849
Ion Ratio Lower Upper
330 100
332 90.5 73.3 109.9
141 63.9 52.1 78.1



#45
2-Fluorobiphenyl
Concen: 58.889 ng
RT: 13.570 min Scan# 1733
Delta R.T. -0.009 min
Lab File: BG059870.D
Acq: 24 Nov 2023 23:04

Tgt Ion:172 Resp: 525771
Ion Ratio Lower Upper
172 100
171 33.0 27.7 41.5
170 23.4 18.4 27.6





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.701 min Scan# 2439

Delta R.T. -0.009 min

Lab File: BG059870.D

Acq: 24 Nov 2023 23:04

Instrument :

BNA_G

ClientSampleId :

MH-2





Tgt Ion:188 Resp: 287963

Ion Ratio Lower Upper

188 100

94 3.3 3.0 4.4

80 5.1 4.2 6.4



#76

Chrysene-d12

Concen: 20.000 ng

RT: 22.019 min Scan# 3171

Delta R.T. -0.015 min

Lab File: BG059870.D

Acq: 24 Nov 2023 23:04




Tgt Ion:240 Resp: 220491


Ion Ratio Lower Upper

240 100

120 2.6 2.2 3.2

236 27.8 19.4 29.2









#79
 Terphenyl-d14
 Concen: 64.841 ng
 RT: 20.274 min Scan# 21
 Delta R.T. -0.009 min
 Lab File: BG059870.D
 Acq: 24 Nov 2023 23:04

Instrument :
 BNA_G
 ClientSampleId :
 MH-2

Tgt Ion:244 Resp: 968456
 Ion Ratio Lower Upper
 244 100
 212 9.3 6.6 10.0
 122 5.8 4.4 6.6



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 25.592 min Scan# 3779
 Delta R.T. -0.020 min
 Lab File: BG059870.D
 Acq: 24 Nov 2023 23:04

Tgt Ion:264 Resp: 234128
 Ion Ratio Lower Upper
 264 100
 260 24.7 18.5 27.7
 265 22.3 17.0 25.6

