

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG112423\
 Data File : BG059877.D
 Acq On : 25 Nov 2023 3:49
 Operator : MA/JU
 Sample : 05470-02
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SB-2

Quant Time: Nov 25 04:31:23 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111623.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 24 15:11:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.323	152	22349	20.000	ng	# 0.00
21) Naphthalene-d8	11.161	136	111599	20.000	ng	# 0.00
39) Acenaphthene-d10	14.951	164	78053	20.000	ng	# 0.00
64) Phenanthrene-d10	17.701	188	180209	20.000	ng	# 0.00
76) Chrysene-d12	22.019	240	137272	20.000	ng	-0.01
86) Perylene-d12	25.586	264	159502	20.000	ng	-0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.815	112	153870	115.882	ng	0.00
7) Phenol-d6	7.448	99	211523	107.498	ng	0.00
23) Nitrobenzene-d5	9.516	82	166484	56.975	ng	0.00
42) 2,4,6-Tribromophenol	16.432	330	101654	114.784	ng	0.00
45) 2-Fluorobiphenyl	13.570	172	350458	60.437	ng	0.00
79) Terphenyl-d14	20.274	244	596045	64.100	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG112423\
Data File : BG059877.D
Acq On : 25 Nov 2023 3:49
Operator : MA/JU
Sample : 05470-02
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SB-2

Quant Time: Nov 25 04:31:23 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111623.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 24 15:11:07 2023
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.323 min Scan# 843
Delta R.T. 0.008 min
Lab File: BG059877.D
Acq: 25 Nov 2023 3:49




Instrument :
BNA_G
ClientSampleId :
SB-2

Tgt Ion:152 Resp: 22349
Ion Ratio Lower Upper
152 100
150 138.7 122.2 183.4
115 82.5 54.2 81.4#



#5
2-Fluorophenol
Concen: 115.882 ng
RT: 5.815 min Scan# 413
Delta R.T. 0.003 min
Lab File: BG059877.D
Acq: 25 Nov 2023 3:49

Tgt Ion:112 Resp: 153870
Ion Ratio Lower Upper
112 100
64 48.2 49.5 74.3#
63 42.6 37.4 56.2









#7
Phenol-d6
Concen: 107.498 ng
RT: 7.448 min Scan# 691
Delta R.T. 0.003 min
Lab File: BG059877.D
Acq: 25 Nov 2023 3:49




Instrument :
BNA_G
ClientSampleId :
SB-2


Tgt Ion: 99 Resp: 211523
Ion Ratio Lower Upper
99 100
42 17.3 14.3 21.5
71 65.9 59.3 88.9



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 11.161 min Scan# 1323
Delta R.T. -0.009 min
Lab File: BG059877.D
Acq: 25 Nov 2023 3:49

Tgt Ion: 136 Resp: 111599
Ion Ratio Lower Upper
136 100
137 8.2 7.4 11.0
54 3.4 3.1 4.7
68 1.6 1.8 2.8





#23

Nitrobenzene-d5

Concen: 56.975 ng

RT: 9.516 min Scan# 1043

Delta R.T. -0.009 min

Lab File: BG059877.D

Acq: 25 Nov 2023 3:49

Instrument :

BNA_G

ClientSampleId :

SB-2





Tgt Ion: 82 Resp: 166484

Ion Ratio Lower Upper

82 100

128 39.7 23.0 34.6#

54 37.4 28.7 43.1



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.951 min Scan# 1968

Delta R.T. -0.003 min

Lab File: BG059877.D

Acq: 25 Nov 2023 3:49




Tgt Ion:164 Resp: 78053


Ion Ratio Lower Upper

164 100

162 100.5 73.3 109.9

160 54.9 34.1 51.1#









#42
2,4,6-Tribromophenol
Concen: 114.784 ng
RT: 16.432 min Scan# 21
Delta R.T. -0.009 min
Lab File: BG059877.D
Acq: 25 Nov 2023 3:49




Instrument :
BNA_G
ClientSampleId :
SB-2


Tgt Ion:330 Resp: 101654
Ion Ratio Lower Upper
330 100
332 94.9 73.3 109.9
141 59.9 52.1 78.1



#45
2-Fluorobiphenyl
Concen: 60.437 ng
RT: 13.570 min Scan# 1733
Delta R.T. -0.009 min
Lab File: BG059877.D
Acq: 25 Nov 2023 3:49

Tgt Ion:172 Resp: 350458
Ion Ratio Lower Upper
172 100
171 33.5 27.7 41.5
170 25.6 18.4 27.6





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.701 min Scan# 2439

Delta R.T. -0.009 min

Lab File: BG059877.D

Acq: 25 Nov 2023 3:49

Instrument :

BNA_G

ClientSampleId :

SB-2





Tgt Ion:188 Resp: 180209

Ion Ratio Lower Upper

188 100

94 2.7 3.0 4.4#

80 3.7 4.2 6.4#



#76

Chrysene-d12

Concen: 20.000 ng

RT: 22.019 min Scan# 3171

Delta R.T. -0.015 min

Lab File: BG059877.D

Acq: 25 Nov 2023 3:49




Tgt Ion:240 Resp: 137272


Ion Ratio Lower Upper

240 100

120 3.0 2.2 3.2

236 23.3 19.4 29.2









#79
 Terphenyl-d14
 Concen: 64.100 ng
 RT: 20.274 min Scan# 21
 Delta R.T. -0.009 min
 Lab File: BG059877.D
 Acq: 25 Nov 2023 3:49

Instrument :
 BNA_G
 ClientSampleId :
 SB-2

Tgt Ion:244 Resp: 596045
 Ion Ratio Lower Upper
 244 100
 212 7.9 6.6 10.0
 122 5.2 4.4 6.6



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 25.586 min Scan# 3778
 Delta R.T. -0.026 min
 Lab File: BG059877.D
 Acq: 25 Nov 2023 3:49

Tgt Ion:264 Resp: 159502
 Ion Ratio Lower Upper
 264 100
 260 22.6 18.5 27.7
 265 21.7 17.0 25.6

