


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG112425\
Data File : BG064847.D
Acq On : 24 Nov 2025 13:23
Operator : RC/JU
Sample : DFTPP460
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP460

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG112425.MA.M
Title : SVOA CALIBRATION
Last Update : Tue Nov 25 10:23:09 2025



AutoFind: Scans 2523, 2524, 2525; Background Corrected with Scan 2518

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.8	26755	PASS
68	69	0.00	2	1.2	274	PASS
69	198	0.00	100	27.6	22501	PASS
70	69	0.00	2	0.2	55	PASS
127	198	10	80	35.6	29053	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	81531	PASS
199	198	5	9	7.3	5956	PASS
275	198	10	60	36.3	29555	PASS
365	198	1	100	5.3	4301	PASS
441	198	0.01	100	12.8	10404	PASS
442	198	50	100	83.4	68011	PASS
443	442	15	24	19.8	13491	PASS

m/z	Abundance
42.90	161.0
44.00	965.0
48.90	605.0
54.90	154.0
83.90	350.0
85.90	313.0
207.10	343.0

Instrument :
BNA_G
ClientSampleId :
DFTPP460

m/z	Abundance
37.00	161.0
37.90	250.0
39.00	1259.0
39.90	293.0
44.00	922.0
45.00	205.0
48.90	743.0
50.00	5803.0
51.00	23384.0
52.00	1374.0
55.90	593.0
57.00	1580.0
61.00	250.0
62.10	202.0
63.00	1093.0
65.00	442.0
68.90	20032.0
73.00	250.0
74.00	1847.0
74.90	2695.0
76.10	1159.0
77.00	21328.0
78.00	1226.0
78.90	1094.0
79.90	1377.0
81.00	1726.0
81.90	398.0
83.00	517.0
83.90	160.0
85.10	298.0
86.00	592.0
86.70	169.0
91.00	479.0
92.00	652.0
93.00	2268.0
94.00	193.0
97.90	2638.0
99.00	1771.0
100.90	834.0
102.90	503.0
104.00	578.0
105.00	375.0
107.00	7399.0
108.00	1356.0
110.00	12822.0
110.90	2268.0
111.90	490.0
116.10	477.0
117.00	6151.0
118.00	607.0
120.10	150.0
121.90	538.0
122.90	551.0
124.00	346.0
124.80	336.0
125.60	209.0
127.00	21944.0
127.90	2094.0
128.90	9128.0
129.90	657.0
134.00	361.0
135.00	1054.0
135.80	301.0
137.00	592.0
139.80	255.0
140.90	1165.0
142.00	412.0
142.90	179.0
146.00	302.0
146.90	807.0
147.90	1228.0
148.90	260.0
151.10	245.0
152.80	428.0
153.90	411.0
155.00	560.0
156.00	989.0
157.90	311.0
159.00	328.0
159.80	488.0
161.00	655.0
165.00	580.0
165.80	501.0
167.00	3048.0
167.90	1922.0
168.90	194.0
170.80	211.0
171.90	263.0
173.00	347.0
173.90	668.0
175.00	1302.0
176.00	287.0

Instrument :
BNA_G
ClientSampleId :
DFTPP460

176.80	533.0
178.90	2640.0
180.00	1718.0
180.90	834.0
181.90	249.0
182.70	173.0
183.90	202.0
184.90	1014.0
186.00	8812.0
186.90	3114.0
187.90	308.0
189.00	711.0
190.90	270.0
191.90	780.0
192.90	809.0
194.00	266.0
195.90	2001.0
197.90	64296.0
198.90	4647.0
199.80	265.0
201.10	268.0
202.80	847.0
203.90	2751.0
205.00	4166.0
206.00	18136.0
207.00	2604.0
207.90	572.0
209.00	312.0
209.90	293.0
210.90	708.0
211.60	238.0
214.90	263.0
216.00	231.0
216.90	6134.0
217.80	798.0
220.90	2017.0
222.80	1484.0
224.00	10939.0
224.90	2922.0
226.00	307.0
226.90	4456.0
227.90	652.0
228.90	1006.0
230.00	176.0
230.80	430.0
232.80	174.0
233.90	352.0
235.00	309.0
235.90	178.0
237.00	405.0
239.00	214.0
241.00	316.0
242.00	808.0
242.90	881.0
243.90	8569.0
245.00	1184.0
245.90	1689.0
246.90	370.0
248.90	562.0
250.80	175.0
252.90	165.0
254.90	46120.0
255.90	7472.0
256.90	611.0
257.90	3087.0
259.00	489.0
264.90	1175.0
265.90	223.0
270.70	261.0
272.80	1307.0
273.90	3259.0
274.90	22520.0
275.80	2777.0
276.80	1760.0
278.10	296.0
282.80	260.0
284.90	172.0
290.00	177.0
292.80	353.0
295.90	5713.0
296.90	746.0
300.90	158.0
302.90	589.0
304.00	171.0
312.70	162.0
313.80	297.0
314.70	445.0
315.90	482.0
320.90	239.0
322.80	2171.0
324.00	431.0
326.80	382.0
327.80	277.0
332.70	247.0
333.00	259.0
333.90	1212.0

Instrument :
BNA_G
ClientSampleId :
DFTPP460

335.00	290.0
340.80	212.0
345.70	415.0
351.80	605.0
352.60	374.0
353.90	477.0
364.80	3029.0
365.90	608.0
371.10	174.0
371.80	902.0
372.80	323.0
382.90	241.0
389.70	156.0
400.70	158.0
401.90	607.0
402.80	551.0
403.80	302.0
420.80	450.0
421.80	395.0
422.80	2558.0
423.80	804.0
440.90	6726.0
441.80	43400.0
442.80	7949.0
443.90	742.0

Instrument :
BNA_G
ClientSampleId :
DFTPP460

m/z	Abundance
37.90	295.0
39.00	1886.0
40.00	157.0
43.10	174.0
43.90	860.0
48.90	574.0
50.00	7246.0
51.00	31920.0
52.00	1694.0
54.90	292.0
56.00	1035.0
57.00	2168.0
61.00	576.0
62.00	406.0
63.00	1126.0
63.90	203.0
64.90	531.0
68.00	821.0
68.90	25992.0
72.90	222.0
74.00	2727.0
75.00	4055.0
76.00	1436.0
77.00	30312.0
78.00	2164.0
78.90	1866.0
80.00	1384.0
81.00	2072.0
82.00	656.0
82.90	355.0
84.00	329.0
84.90	378.0
85.90	983.0
86.90	233.0
87.80	284.0
91.00	481.0
91.90	662.0
93.00	3852.0
93.90	403.0
94.90	150.0
95.90	246.0
97.00	224.0
98.00	2250.0
98.90	2112.0
99.90	160.0
100.90	1441.0
102.80	382.0
104.00	897.0
104.90	655.0
105.70	247.0
107.00	10836.0
107.90	1706.0
109.90	17112.0
111.00	2734.0
112.00	254.0
116.00	786.0
116.90	8247.0
118.00	512.0
119.90	156.0
121.90	696.0
122.90	1140.0
123.90	483.0
124.90	501.0
127.00	34328.0
128.00	2542.0
129.00	14925.0
130.00	1452.0
130.90	207.0
133.90	426.0
135.00	1247.0
136.00	435.0
137.00	610.0
140.00	166.0
140.90	2071.0
141.90	564.0
142.90	636.0
146.00	411.0
146.90	1227.0
147.90	2288.0
148.90	451.0
150.00	244.0
151.00	225.0
153.00	712.0
154.00	548.0
154.90	1410.0
156.00	1807.0
157.00	248.0
157.90	296.0
159.00	262.0
160.00	794.0
160.90	1349.0
162.00	292.0

Instrument :
BNA_G
ClientSampleId :
DFTPP460

163.90	174.0
165.00	672.0
165.90	875.0
167.00	3911.0
167.90	2009.0
169.00	547.0
170.90	205.0
171.90	341.0
172.90	527.0
174.00	1077.0
175.00	2154.0
175.80	497.0
176.90	947.0
177.90	442.0
178.90	3123.0
179.90	2413.0
181.00	1640.0
181.90	426.0
183.00	224.0
183.70	250.0
185.00	1686.0
185.90	13812.0
187.00	3871.0
188.00	599.0
188.90	1136.0
190.00	213.0
190.80	425.0
191.90	1162.0
193.00	1452.0
193.90	320.0
196.00	2666.0
197.90	93464.0
198.90	6928.0
199.90	425.0
201.50	318.0
202.90	777.0
203.90	3579.0
205.00	7277.0
206.00	28808.0
206.90	3818.0
207.90	1078.0
208.90	278.0
209.90	418.0
210.90	1147.0
211.80	241.0
215.00	367.0
215.90	713.0
216.90	7211.0
217.90	989.0
220.90	3635.0
222.90	1767.0
223.90	16496.0
224.90	3903.0
226.10	495.0
226.90	7307.0
227.90	1178.0
228.90	1665.0
229.80	339.0
230.90	566.0
231.90	279.0
234.00	513.0
234.70	564.0
235.80	461.0
236.90	480.0
238.70	249.0
239.80	168.0
240.90	556.0
242.00	1000.0
242.90	1168.0
244.00	12977.0
244.90	1728.0
245.90	2874.0
247.00	701.0
247.80	253.0
248.90	370.0
253.00	410.0
254.90	74136.0
255.90	9610.0
256.90	904.0
257.90	4640.0
258.90	703.0
264.90	1385.0
265.90	376.0
270.90	258.0
272.00	215.0
272.90	2145.0
273.90	5038.0
274.90	33240.0
275.90	4575.0
276.90	2530.0
277.70	366.0
278.70	162.0
281.70	170.0
282.90	381.0
283.70	219.0
284.90	459.0

Instrument :
BNA_G
ClientSampleId :
DFTPP460

289.80	159.0
292.80	630.0
294.80	413.0
295.80	9465.0
296.90	1348.0
298.00	166.0
301.90	161.0
302.80	1033.0
303.90	329.0
312.90	166.0
313.80	613.0
314.90	1227.0
315.80	669.0
320.80	300.0
321.80	182.0
322.90	4003.0
323.90	648.0
326.80	731.0
328.00	424.0
331.90	345.0
333.00	635.0
333.80	2292.0
334.80	693.0
340.80	308.0
345.80	633.0
351.90	945.0
352.80	887.0
353.80	1118.0
354.90	415.0
364.80	4598.0
365.70	699.0
370.80	319.0
371.80	2003.0
372.80	414.0
382.80	518.0
389.70	339.0
401.80	750.0
402.80	1137.0
403.90	419.0
414.90	204.0
420.80	1199.0
422.80	4908.0
423.80	896.0
440.80	10980.0
441.80	72512.0
442.80	14677.0
443.80	997.0

Instrument :
BNA_G
ClientSampleId :
DFTPP460

m/z	Abundance
37.90	246.0
39.00	1598.0
40.90	184.0
43.90	722.0
49.00	918.0
50.00	5816.0
51.00	24960.0
52.00	1141.0
54.80	213.0
55.90	751.0
57.00	1828.0
60.90	407.0
61.90	498.0
63.00	1300.0
65.00	576.0
68.90	21480.0
69.90	165.0
73.00	270.0
74.00	2558.0
75.00	3094.0
76.00	1380.0
77.00	24176.0
78.00	1552.0
79.00	1398.0
80.00	1237.0
81.00	2072.0
82.00	574.0
82.90	502.0
83.90	417.0
85.00	354.0
85.90	795.0
86.80	209.0
91.00	424.0
91.90	609.0
92.90	3502.0
93.90	225.0
95.90	168.0
97.90	2457.0
99.00	2093.0
100.10	223.0
100.90	1295.0
102.80	528.0
104.00	936.0
104.90	797.0
105.90	328.0
107.00	8391.0
107.90	1192.0
110.00	15177.0
111.00	2454.0
112.10	221.0
112.90	206.0
116.00	643.0
116.90	7085.0
117.90	431.0
121.90	659.0
123.00	1086.0
124.00	485.0
124.80	397.0
127.00	30888.0
128.00	2143.0
129.00	12013.0
129.90	1213.0
131.00	224.0
131.80	161.0
133.90	255.0
134.90	1188.0
135.90	317.0
137.00	633.0
140.90	1931.0
142.00	625.0
142.80	565.0
145.90	609.0
147.00	1274.0
148.00	2442.0
149.00	393.0
149.80	164.0
151.00	166.0
152.90	535.0
153.90	505.0
154.90	1163.0
155.90	1583.0
157.00	368.0
157.90	339.0
159.00	280.0
160.00	711.0
160.90	1140.0
161.80	306.0
164.00	176.0
164.90	610.0
166.00	765.0
167.00	3675.0
167.90	1875.0

Instrument :
BNA_G
ClientSampleId :
DFTPP460

169.00	459.0
170.00	177.0
170.90	196.0
172.00	417.0
172.90	436.0
174.00	1162.0
175.00	1875.0
175.80	598.0
176.90	778.0
177.90	424.0
178.90	3327.0
180.00	2364.0
180.90	1437.0
181.90	301.0
182.10	310.0
183.90	276.0
184.90	1794.0
185.90	12797.0
187.00	3720.0
187.90	453.0
188.90	945.0
189.80	160.0
190.80	480.0
192.00	1017.0
192.90	1338.0
194.00	292.0
195.90	2902.0
197.90	86832.0
198.90	6293.0
200.00	598.0
201.40	557.0
202.80	745.0
203.90	4083.0
205.00	6546.0
206.00	26856.0
207.00	3764.0
207.90	1130.0
208.90	384.0
210.10	383.0
210.90	904.0
211.50	351.0
212.10	190.0
214.90	316.0
216.00	533.0
216.90	6767.0
217.90	1007.0
219.10	178.0
220.10	416.0
220.90	3288.0
222.00	909.0
222.90	1839.0
224.00	16227.0
225.00	3487.0
226.90	7788.0
227.90	1095.0
228.90	1560.0
230.00	231.0
230.90	389.0
233.80	577.0
234.80	656.0
235.80	388.0
236.80	765.0
238.80	351.0
239.90	233.0
240.90	348.0
242.00	759.0
243.00	817.0
243.90	12545.0
244.90	1771.0
245.90	2967.0
246.80	293.0
247.90	170.0
249.00	382.0
249.80	151.0
251.00	151.0
251.90	159.0
252.90	400.0
254.90	75256.0
255.90	10748.0
256.90	874.0
257.90	4609.0
258.90	771.0
259.90	252.0
264.90	1607.0
271.00	220.0
271.70	249.0
272.00	292.0
272.90	2311.0
273.90	6268.0
274.90	32904.0
275.90	4885.0
276.90	2401.0
277.80	613.0
281.80	361.0
282.90	315.0
284.00	310.0

Instrument :
BNA_G
ClientSampleId :
DFTPP460

284.90	576.0
289.80	266.0
290.70	162.0
291.90	244.0
293.00	631.0
293.90	198.0
295.90	11160.0
296.90	1447.0
297.80	150.0
301.00	200.0
301.90	371.0
302.90	1287.0
303.80	355.0
307.90	252.0
308.70	232.0
312.80	219.0
313.70	542.0
314.80	1242.0
315.90	629.0
316.90	163.0
320.90	434.0
321.90	178.0
322.90	2983.0
323.90	550.0
326.90	668.0
327.60	354.0
331.90	181.0
332.90	320.0
333.90	2318.0
334.90	507.0
340.90	437.0
345.80	595.0
351.80	1115.0
352.90	669.0
353.90	873.0
355.00	310.0
364.90	5276.0
365.90	680.0
370.70	253.0
371.00	246.0
371.80	2029.0
372.80	615.0
382.80	853.0
383.70	191.0
389.90	182.0
390.70	206.0
401.70	643.0
402.70	1159.0
403.70	467.0
420.90	1011.0
421.80	713.0
422.80	5913.0
423.80	1119.0
425.00	182.0
440.80	13505.0
441.80	88120.0
442.80	17848.0
443.80	1458.0

Instrument :
BNA_G
ClientSampleId :
DFTPP460