

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG120118\
 Data File : BG038289.D
 Acq On : 1 Dec 2018 16:07
 Operator : JU/SJ
 Sample : PB115178TB
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB115178TB

Quant Time: Dec 03 02:48:01 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG111318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 13 16:39:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.08	152	26861	20.00	ng	-0.01
21) Naphthalene-d8	10.90	136	108667	20.00	ng	-0.01
39) Acenaphthene-d10	14.71	164	69215	20.00	ng	-0.01
64) Phenanthrene-d10	17.46	188	187811	20.00	ng	-0.01
76) Chrysene-d12	21.74	240	193146	20.00	ng	-0.01
87) Perylene-d12	25.06	264	194188	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.62	112	219741	141.24	ng	-0.01
7) Phenol-d6	7.22	99	302289	136.12	ng	0.00
23) Nitrobenzene-d5	9.25	82	192312	94.73	ng	-0.01
42) 2,4,6-Tribromophenol	16.20	330	128276	162.50	ng	-0.01
45) 2-Fluorobiphenyl	13.34	172	442501	93.14	ng	-0.01
79) Terphenyl-d14	20.06	244	899586	109.59	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG120118\
Data File : BG038289.D
Acq On : 1 Dec 2018 16:07
Operator : JU/SJ
Sample : PB115178TB
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB115178TB

Quant Time: Dec 03 02:48:01 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG111318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 13 16:39:28 2018
Response via : Initial Calibration

