

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG120123\
 Data File : BG059932.D
 Acq On : 2 Dec 2023 00:36
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG120123

Quant Time: Dec 02 03:37:46 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG120123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 02 03:28:09 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	85	0.00
2	1,4-Dioxane	40.000	48.404	-21.0	93	0.00
3	Pyridine	40.000	42.884	-7.2	77	0.00
4	n-Nitrosodimethylamine	40.000	55.439	-38.6#	97	0.00
5 S	2-Fluorophenol	80.000	85.960	-7.4	79	0.00
6	Aniline	40.000	41.402	-3.5	77	0.00
7 S	Phenol-d6	80.000	89.874	-12.3	84	0.00
8	2-Chlorophenol	40.000	46.450	-16.1	88	0.00
9	Benzaldehyde	40.000	48.154	-20.4	91	0.00
10 C	Phenol	40.000	43.987	-10.0	89	0.00
11	bis(2-Chloroethyl)ether	40.000	43.603	-9.0	82	0.00
12	1,3-Dichlorobenzene	40.000	45.512	-13.8	87	0.00
13 C	1,4-Dichlorobenzene	40.000	45.025	-12.6	84	0.00
14	1,2-Dichlorobenzene	40.000	45.301	-13.3	94	0.00
15	Benzyl Alcohol	40.000	48.910	-22.3	97	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	47.744	-19.4	92	0.00
17	2-Methylphenol	40.000	45.056	-12.6	95	0.00
18	Hexachloroethane	40.000	48.769	-21.9	92	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	44.332	-10.8	89	0.00
20	3+4-Methylphenols	40.000	46.100	-15.3	89	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	82	0.00
22	Acetophenone	40.000	43.578	-8.9	88	0.00
23 S	Nitrobenzene-d5	80.000	86.362	-8.0	86	0.00
24	Nitrobenzene	40.000	45.648	-14.1	92	0.00
25	Isophorone	40.000	42.659	-6.6	87	0.00
26 C	2-Nitrophenol	40.000	42.676	-6.7	86	0.00
27	2,4-Dimethylphenol	40.000	44.641	-11.6	89	0.00
28	bis(2-Chloroethoxy)methane	40.000	42.256	-5.6	88	0.00
29 C	2,4-Dichlorophenol	40.000	43.785	-9.5	92	0.00
30	1,2,4-Trichlorobenzene	40.000	44.993	-12.5	92	0.00
31	Naphthalene	40.000	40.582	-1.5	83	0.00
32	Benzoic acid	40.000	42.286	-5.7	83	0.00
33	4-Chloroaniline	40.000	42.535	-6.3	89	0.00
34 C	Hexachlorobutadiene	40.000	47.409	-18.5	95	0.00
35	Caprolactam	40.000	42.503	-6.3	90	0.00
36 C	4-Chloro-3-methylphenol	40.000	46.862	-17.2	95	0.00
37	2-Methylnaphthalene	40.000	42.258	-5.6	87	0.00
38	1-Methylnaphthalene	40.000	43.254	-8.1	86	-0.22
39 I	Acenaphthene-d10	20.000	20.000	0.0	90	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	44.453	-11.1	95	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.245	-8.1	89	0.00
42 S	2,4,6-Tribromophenol	80.000	93.359	-16.7	99	0.00
43 C	2,4,6-Trichlorophenol	40.000	44.648	-11.6	94	0.00
44	2,4,5-Trichlorophenol	40.000	44.284	-10.7	94	0.00
45 S	2-Fluorobiphenyl	80.000	85.729	-7.2	91	0.00
46	1,1'-Biphenyl	40.000	40.978	-2.4	89	0.00
47	2-Chloronaphthalene	40.000	40.908	-2.3	86	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG120123\
 Data File : BG059932.D
 Acq On : 2 Dec 2023 00:36
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG120123

Quant Time: Dec 02 03:37:46 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG120123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 02 03:28:09 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.507	-11.3	98	0.00
49	Acenaphthylene	40.000	41.405	-3.5	89	0.00
50	Dimethylphthalate	40.000	43.247	-8.1	93	0.00
51	2,6-Dinitrotoluene	40.000	39.031	2.4	87	0.00
52 C	Acenaphthene	40.000	42.538	-6.3	91	0.00
53	3-Nitroaniline	40.000	39.041	2.4	86	0.00
54 P	2,4-Dinitrophenol	40.000	42.004	-5.0	85	0.00
55	Dibenzofuran	40.000	41.811	-4.5	90	0.00
56 P	4-Nitrophenol	40.000	38.790	3.0	84	0.00
57	2,4-Dinitrotoluene	40.000	40.870	-2.2	88	0.00
58	Fluorene	40.000	41.405	-3.5	91	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	43.111	-7.8	93	0.00
60	Diethylphthalate	40.000	44.403	-11.0	94	0.00
61	4-Chlorophenyl-phenylether	40.000	44.359	-10.9	97	0.00
62	4-Nitroaniline	40.000	40.485	-1.2	90	0.00
63	Azobenzene	40.000	43.952	-9.9	93	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	92	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.723	-4.3	89	0.00
66 c	n-Nitrosodiphenylamine	40.000	42.174	-5.4	91	0.00
67	4-Bromophenyl-phenylether	40.000	43.738	-9.3	97	0.00
68	Hexachlorobenzene	40.000	44.822	-12.1	97	0.00
69	Atrazine	40.000	42.865	-7.2	100	0.00
70 C	Pentachlorophenol	40.000	47.155	-17.9	98	0.00
71	Phenanthrene	40.000	41.067	-2.7	91	0.00
72	Anthracene	40.000	41.547	-3.9	93	0.00
73	Carbazole	40.000	41.535	-3.8	91	0.00
74	Di-n-butylphthalate	40.000	43.385	-8.5	100	0.00
75 C	Fluoranthene	40.000	42.322	-5.8	92	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	98	0.00
77	Benzydine	40.000	38.394	4.0	80	0.00
78	Pyrene	40.000	41.477	-3.7	98	0.00
79 S	Terphenyl-d14	80.000	84.919	-6.1	100	0.00
80	Butylbenzylphthalate	40.000	38.824	2.9	94	0.00
81	Benzo(a)anthracene	40.000	42.996	-7.5	101	0.00
82	3,3'-Dichlorobenzidine	40.000	43.576	-8.9	104	0.00
83	Chrysene	40.000	41.891	-4.7	99	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.415	-1.0	97	0.00
85 c	Di-n-octyl phthalate	40.000	40.994	-2.5	100	0.00
86 I	Perylene-d12	20.000	20.000	0.0	106	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.553	1.1	101	0.00
88	Benzo(b)fluoranthene	40.000	40.496	-1.2	102	0.00
89	Benzo(k)fluoranthene	40.000	42.242	-5.6	108	0.00
90 C	Benzo(a)pyrene	40.000	42.161	-5.4	108	0.00
91	Dibenzo(a,h)anthracene	40.000	40.113	-0.3	101	0.00
92	Benzo(g,h,i)perylene	40.000	38.888	2.8	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG120123\
Data File : BG059932.D
Acq On : 2 Dec 2023 00:36
Operator : MA/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG120123

Quant Time: Dec 02 03:37:46 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG120123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 02 03:28:09 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0