

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG120221\
 Data File : BG051340.D
 Acq On : 3 Dec 2021 22:10
 Operator : CG/JU
 Sample : M4833-22
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ESQN9

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 12/06/2021
 Supervised By :mohammad ahmed 12/07/2021

Quant Time: Dec 03 23:38:42 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG112321.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Dec 03 15:23:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.191	152	29853	20.000	ng/u1	-0.01
20) Naphthalene-d8	11.017	136	131286	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.825	164	86045	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.574	188	184596	20.000	ng/u1	0.00
79) Chrysene-d12	21.875	240	154252	20.000	ng/u1	0.00
88) Perylene-d12	25.277	264	152968	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.532	96	3839	4.469	ng/uL	-0.01
4) Pyridine-d5	3.967	84	19893	7.891	ng/u1	-0.01
7) Phenol-d5	7.351	99	74260	25.169	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.504	67	49545	26.737	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.721	132	55597	26.168	ng/u1	-0.01
15) 4-Methylphenol-d8	8.908	113	46130	19.375	ng/u1	0.00
21) Nitrobenzene-d5	9.366	128	30598	27.609	ng/u1	-0.01
24) 2-Nitrophenol-d4	10.095	143	34488	27.587	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.647	165	53741	25.336	ng/u1	0.00
31) 4-Chloroaniline-d4	11.158	131	54169	17.454	ng/u1	0.00
46) Dimethylphthalate-d6	14.214	166	174661	26.381	ng/u1	-0.01
49) Acenaphthylene-d8	14.519	160	233738	27.997	ng/u1	-0.01
54) 4-Nitrophenol-d4	15.054	143	20724m	19.338	ng/u1	0.00
60) Fluorene-d10	15.812	176	162292	27.221	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.947	200	14945	13.120	ng/u1	0.00
73) Anthracene-d10	17.674	188	243511	27.582	ng/u1	0.00
81) Pyrene-d10	19.954	212	288099	30.867	ng/u1	0.00
92) Benzo(a)pyrene-d12	25.042	264	237781	29.106	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG120221\
Data File : BG051340.D
Acq On : 3 Dec 2021 22:10
Operator : CG/JU
Sample : M4833-22
Misc :
ALS Vial : 45 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ESQN9

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 12/06/2021
Supervised By :mohammad ahmed 12/07/2021

Quant Time: Dec 03 23:38:42 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG112321.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Dec 03 15:23:09 2021
Response via : Initial Calibration

