


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG120221\
 Data File : BG051296.D
 Acq On : 2 Dec 2021 9:21
 Operator : CG/JU
 Sample : DFTPP437
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP437

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG112321.M
 Title : SVOA CALIBRATION
 Last Update : Fri Dec 03 15:23:09 2021



AutoFind: Scans 2465, 2466, 2467; Background Corrected with Scan 2459

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.8	30763	PASS
68	69	0.00	2	0.8	269	PASS
69	198	0.00	100	42.0	33267	PASS
70	69	0.00	2	0.5	172	PASS
127	198	10	80	49.6	39331	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	79301	PASS
199	198	5	9	6.9	5451	PASS
275	198	10	60	26.4	20942	PASS
365	198	1	100	3.5	2757	PASS
441	198	0.01	100	9.4	7431	PASS
442	198	50	100	58.3	46253	PASS
443	442	15	24	18.9	8738	PASS

m/z	Abundance
40.00	370.0
44.00	752.0

Instrument :
BNA_G
ClientSampleId :
DFTPP437

m/z	Abundance
38.00	672.0
39.00	3091.0
40.00	389.0
41.10	155.0
44.00	1034.0
44.90	161.0
48.90	305.0
50.00	9547.0
51.00	38640.0
52.00	1975.0
52.90	167.0
55.00	226.0
56.00	1245.0
57.00	2506.0
61.10	416.0
62.00	579.0
63.00	1653.0
64.00	299.0
65.00	903.0
67.10	160.0
67.90	808.0
69.00	41560.0
70.00	218.0
73.00	449.0
74.00	4138.0
75.00	7004.0
76.00	2223.0
77.00	48976.0
78.00	3142.0
79.00	3184.0
80.00	2472.0
81.00	3551.0
82.00	677.0
82.90	713.0
85.00	413.0
86.00	879.0
87.00	527.0
90.90	641.0
92.00	681.0
93.00	5274.0
94.00	456.0
96.00	241.0
98.00	4114.0
99.00	2932.0
99.90	426.0
101.00	1939.0
103.00	481.0
104.00	1345.0
105.00	1006.0
106.10	384.0
107.00	14666.0
108.00	2013.0
109.00	537.0
110.00	23728.0
111.00	4068.0
112.00	450.0
116.10	821.0
117.00	11673.0
118.00	813.0
120.00	221.0
121.90	793.0
123.00	1427.0
124.00	799.0
125.00	663.0
127.00	47064.0
128.00	3515.0
129.00	19072.0
130.00	1907.0
131.00	266.0
134.00	666.0
135.00	1404.0
136.00	613.0
137.00	695.0
138.20	188.0
140.00	254.0
141.00	2431.0
142.00	691.0
142.90	487.0
144.10	200.0
146.00	486.0
147.00	1437.0
148.00	2911.0
149.00	531.0
150.10	281.0
151.10	365.0
151.90	197.0
152.90	847.0
154.00	566.0
155.00	1517.0
156.00	1959.0
157.00	330.0
158.00	577.0

Instrument :
BNA_G
ClientSampleId :
DFTPP437

159.00	353.0
160.00	1102.0
161.00	1008.0
162.00	328.0
165.00	927.0
166.10	901.0
167.00	4959.0
168.00	2291.0
169.00	435.0
170.90	297.0
172.00	397.0
173.10	616.0
174.00	1138.0
175.00	1949.0
176.00	548.0
176.90	692.0
178.00	306.0
179.00	3775.0
180.00	2704.0
181.00	1336.0
182.00	156.0
183.90	290.0
185.00	1824.0
186.00	13052.0
187.00	4091.0
188.00	437.0
189.00	801.0
191.00	449.0
192.00	1133.0
193.00	1461.0
193.80	189.0
196.00	2965.0
198.00	87736.0
199.00	6256.0
200.00	601.0
201.20	267.0
203.00	652.0
204.00	3889.0
205.00	6484.0
206.00	24296.0
207.00	3724.0
208.00	1194.0
209.00	312.0
209.90	354.0
211.00	1093.0
212.00	161.0
215.10	242.0
216.00	530.0
217.00	6955.0
218.00	957.0
221.00	3115.0
222.00	803.0
223.00	1710.0
224.00	13795.0
225.00	3696.0
226.10	576.0
227.00	6469.0
228.00	678.0
229.00	1187.0
229.90	208.0
231.00	565.0
234.00	558.0
235.00	437.0
236.00	352.0
237.00	457.0
239.00	277.0
240.00	195.0
241.00	292.0
242.00	731.0
243.10	791.0
244.10	9885.0
245.00	1402.0
246.00	2238.0
247.00	556.0
249.00	271.0
250.90	174.0
252.90	276.0
255.00	53440.0
256.00	8318.0
257.00	788.0
258.00	3387.0
259.10	580.0
265.00	1560.0
265.90	212.0
271.00	180.0
272.10	154.0
273.00	1532.0
274.00	4191.0
275.00	21392.0
276.10	3016.0
277.00	1976.0
278.00	286.0
279.10	156.0
283.10	355.0
284.10	205.0
285.00	293.0

Instrument :
BNA_G
ClientSampleId :
DFTPP437

293.00	463.0
296.00	6793.0
297.00	931.0
303.10	764.0
304.10	198.0
314.10	319.0
315.00	625.0
316.10	569.0
323.10	1965.0
324.00	423.0
327.10	315.0
328.10	163.0
331.90	236.0
333.10	219.0
334.10	1389.0
335.10	441.0
341.10	251.0
346.10	388.0
352.10	466.0
353.10	401.0
354.00	574.0
365.10	2563.0
366.10	467.0
372.00	934.0
373.00	179.0
383.00	215.0
402.20	333.0
403.10	403.0
421.10	362.0
422.00	405.0
423.10	2281.0
424.10	393.0
441.10	6809.0
442.10	41072.0
443.10	8339.0
444.10	1053.0

Instrument :
BNA_G
ClientSampleId :
DFTPP437

m/z	Abundance
38.00	605.0
39.00	2321.0
39.90	432.0
44.00	797.0
49.00	351.0
50.00	8888.0
51.00	36352.0
52.00	1636.0
54.90	212.0
56.00	1061.0
57.00	2472.0
61.00	481.0
62.10	560.0
63.00	1671.0
63.90	292.0
65.00	815.0
69.00	38416.0
70.00	297.0
73.00	396.0
74.00	4083.0
75.00	6549.0
76.10	2398.0
77.00	44952.0
78.00	3145.0
79.00	3199.0
80.00	2284.0
81.00	3136.0
82.00	734.0
83.00	804.0
85.00	735.0
85.90	997.0
87.00	437.0
88.00	159.0
91.00	865.0
91.90	779.0
93.00	5683.0
94.00	340.0
96.00	262.0
98.00	3700.0
99.00	3052.0
99.90	333.0
101.00	1843.0
102.90	675.0
104.00	1440.0
105.00	1138.0
107.00	13120.0
108.00	2063.0
108.90	385.0
110.00	23952.0
111.00	3570.0
112.00	440.0
113.00	155.0
116.00	884.0
117.00	11445.0
118.00	768.0
120.00	169.0
122.00	916.0
123.00	1362.0
123.90	456.0
125.00	612.0
127.00	45608.0
128.00	3695.0
129.00	18456.0
130.00	1829.0
130.90	480.0
132.00	195.0
134.00	558.0
135.00	1482.0
136.00	543.0
137.00	775.0
138.00	184.0
140.00	331.0
141.00	2529.0
142.00	713.0
143.00	641.0
144.00	154.0
144.90	230.0
146.00	410.0
147.00	1054.0
148.00	2870.0
149.00	627.0
150.00	191.0
151.00	335.0
151.90	228.0
153.00	740.0
154.00	650.0
155.00	1568.0
156.00	1969.0
157.00	383.0
157.90	413.0
159.00	318.0
159.90	726.0

Instrument :

BNA_G

ClientSampleId :

DFTPP437

161.00	1000.0
162.00	349.0
165.00	940.0
166.00	731.0
167.00	4794.0
168.00	2380.0
169.00	363.0
171.00	221.0
171.90	527.0
173.00	681.0
174.00	1018.0
175.00	2425.0
176.00	681.0
177.00	769.0
178.00	341.0
179.00	3733.0
180.00	2619.0
181.00	1257.0
181.90	234.0
184.00	367.0
185.00	1944.0
186.00	14065.0
187.00	4379.0
188.00	299.0
189.00	845.0
191.00	390.0
192.00	1051.0
193.00	1355.0
194.00	238.0
195.20	233.0
196.00	2844.0
198.00	94304.0
199.00	6424.0
200.00	450.0
201.50	434.0
202.00	219.0
203.00	664.0
204.00	3790.0
205.00	6353.0
206.10	27128.0
207.00	3623.0
208.00	881.0
209.00	343.0
211.00	1000.0
212.00	207.0
215.00	340.0
216.00	619.0
217.00	7353.0
218.00	968.0
221.00	3671.0
222.00	934.0
223.00	1879.0
224.00	14743.0
225.00	3726.0
226.10	457.0
227.00	6936.0
228.00	1114.0
229.00	1457.0
230.00	217.0
231.00	511.0
233.00	155.0
234.00	504.0
235.00	491.0
235.90	394.0
237.00	507.0
239.00	221.0
241.10	417.0
242.00	906.0
243.00	792.0
244.00	11286.0
245.10	1447.0
246.00	2498.0
247.00	640.0
248.90	430.0
253.00	301.0
255.00	60024.0
256.00	8440.0
257.10	797.0
258.00	3652.0
259.00	596.0
259.90	150.0
264.00	241.0
265.00	1533.0
266.10	183.0
271.00	224.0
272.00	184.0
273.00	1864.0
274.00	4760.0
275.10	25224.0
276.00	3405.0
277.00	2183.0
278.00	416.0
279.00	153.0
283.00	337.0
284.00	156.0
285.10	461.0

Instrument :
BNA_G
ClientSampleId :
DFTPP437

293.00	560.0
296.10	7496.0
297.10	1102.0
303.10	823.0
304.00	309.0
314.10	419.0
315.00	793.0
316.00	509.0
321.00	287.0
323.10	2134.0
323.90	444.0
327.00	497.0
328.00	221.0
332.10	226.0
333.00	207.0
334.10	1491.0
335.10	382.0
341.00	290.0
346.00	589.0
352.00	571.0
353.10	437.0
354.10	631.0
365.00	3265.0
366.00	510.0
370.00	169.0
371.20	154.0
372.10	1031.0
373.10	259.0
383.00	220.0
402.10	420.0
403.00	671.0
404.10	224.0
421.10	379.0
422.10	553.0
423.10	2823.0
424.10	701.0
441.10	8192.0
442.10	53912.0
443.10	10219.0
444.20	1031.0

Instrument :
BNA_G
ClientSampleId :
DFTPP437

m/z	Abundance
38.00	225.0
39.00	1404.0
40.00	403.0
44.00	751.0
50.00	4796.0
51.00	17296.0
52.00	846.0
56.00	586.0
57.00	1384.0
60.90	194.0
61.90	216.0
63.00	705.0
65.10	296.0
69.00	19824.0
74.00	2088.0
75.00	3458.0
76.00	1198.0
77.00	24232.0
78.10	1667.0
79.00	1536.0
80.00	1238.0
81.00	1806.0
82.00	471.0
82.90	362.0
85.10	334.0
86.00	419.0
87.00	203.0
91.00	453.0
92.00	488.0
93.00	2735.0
93.90	182.0
98.00	2086.0
99.00	1611.0
99.90	155.0
101.00	1071.0
102.80	186.0
104.00	649.0
105.00	641.0
105.80	178.0
107.00	7544.0
108.00	1223.0
109.00	245.0
110.00	12513.0
111.00	1833.0
112.00	353.0
116.00	384.0
117.00	6593.0
118.00	451.0
122.00	552.0
122.90	747.0
124.00	346.0
124.90	392.0
127.00	25320.0
128.00	1757.0
129.00	10241.0
130.00	696.0
131.00	150.0
134.00	358.0
135.00	917.0
135.90	301.0
137.00	495.0
139.90	184.0
141.00	1446.0
142.00	440.0
142.90	310.0
146.00	250.0
147.00	906.0
148.00	1624.0
148.80	249.0
153.00	570.0
154.00	356.0
155.00	790.0
156.00	1301.0
157.00	182.0
157.90	233.0
158.70	213.0
160.00	501.0
161.00	645.0
162.00	222.0
164.90	692.0
166.00	446.0
167.00	2538.0
168.00	1220.0
169.00	212.0
171.10	178.0
172.00	272.0
173.00	419.0
174.10	562.0
175.10	1156.0
176.00	281.0
177.00	504.0
177.90	223.0

Instrument :
BNA_G
ClientSampleId :
DFTPP437

179.00	2255.0
180.00	1402.0
181.00	687.0
183.90	187.0
185.10	1131.0
186.10	8295.0
187.00	2389.0
188.00	354.0
189.00	573.0
191.00	385.0
192.00	896.0
193.00	1095.0
196.00	1573.0
198.00	55864.0
199.00	3672.0
200.00	233.0
201.40	282.0
203.00	543.0
204.00	2377.0
205.00	3888.0
206.00	16269.0
207.00	1690.0
208.00	639.0
208.90	234.0
211.00	812.0
215.00	230.0
216.00	510.0
217.00	4526.0
218.00	508.0
221.00	2320.0
222.00	559.0
223.00	1143.0
224.10	8846.0
225.10	2534.0
226.00	233.0
227.00	4253.0
228.00	567.0
229.00	772.0
231.00	350.0
234.00	310.0
235.10	257.0
235.90	236.0
237.00	343.0
238.90	259.0
240.00	261.0
241.00	195.0
242.00	470.0
243.10	509.0
244.10	7230.0
245.10	979.0
246.00	1353.0
247.00	374.0
249.10	267.0
255.00	38648.0
256.00	5819.0
257.10	456.0
258.10	2073.0
259.10	433.0
265.00	1133.0
266.10	204.0
273.00	1404.0
274.00	2814.0
275.10	16210.0
276.10	1987.0
277.00	1788.0
278.10	219.0
283.00	225.0
284.10	152.0
285.00	220.0
293.00	422.0
296.10	4861.0
297.00	715.0
303.10	541.0
314.10	337.0
315.00	698.0
316.10	293.0
321.00	211.0
323.10	1639.0
324.00	355.0
327.00	343.0
332.10	162.0
333.10	214.0
334.00	1023.0
335.10	266.0
341.00	232.0
346.10	358.0
352.00	425.0
353.20	310.0
354.00	469.0
365.10	2444.0
366.00	310.0
371.00	194.0
372.10	761.0
373.10	208.0
383.10	273.0
402.00	206.0

Instrument :
BNA_G
ClientSampleId :
DFTPP437

403.10	402.0
404.10	291.0
421.10	423.0
422.10	413.0
423.10	2179.0
424.10	341.0
441.10	7291.0
442.10	43776.0
443.10	7656.0
444.20	655.0

Instrument :
BNA_G
ClientSampleId :
DFTPP437