


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG120518\  
 Data File : BG038365.D  
 Acq On : 5 Dec 2018 18:29  
 Operator : JU/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_G\METHODS\8270-BG120418.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Tue Dec 04 17:15:40 2018



AutoFind: Scans 2478, 2479, 2480; Background Corrected with Scan 2471

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 42.3      | 17643   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.8       | 284     | PASS             |
| 69          | 198          | 0.00         | 100          | 36.9      | 15419   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 47.1      | 19665   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 41752   | PASS             |
| 199         | 198          | 5            | 9            | 7.1       | 2970    | PASS             |
| 275         | 198          | 10           | 60           | 26.7      | 11144   | PASS             |
| 365         | 198          | 1            | 100          | 3.9       | 1626    | PASS             |
| 441         | 198          | 0.01         | 100          | 10.4      | 4333    | PASS             |
| 442         | 198          | 50           | 100          | 61.8      | 25789   | PASS             |
| 443         | 442          | 15           | 24           | 20.0      | 5148    | PASS             |