

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG120618\
 Data File : BG038388.D
 Acq On : 6 Dec 2018 18:33
 Operator : JU/SJ
 Sample : J6225-01
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 120318-JETT-E-D-B

Quant Time: Dec 06 21:17:00 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG120418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 04 17:15:40 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.07	152	27677	20.00	ng	0.00
21) Naphthalene-d8	10.89	136	135964	20.00	ng	0.00
39) Acenaphthene-d10	14.71	164	77439	20.00	ng	0.00
64) Phenanthrene-d10	17.46	188	198179	20.00	ng	0.00
76) Chrysene-d12	21.73	240	194528	20.00	ng	-0.01
87) Perylene-d12	25.04	264	205922	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.62	112	60458	38.66	ng	0.00
7) Phenol-d6	7.21	99	48816	21.60	ng	-0.01
23) Nitrobenzene-d5	9.25	82	186485	68.10	ng	0.00
42) 2,4,6-Tribromophenol	16.19	330	79783	83.92	ng	-0.01
45) 2-Fluorobiphenyl	13.33	172	417651	76.85	ng	0.00
79) Terphenyl-d14	20.05	244	719709	79.26	ng	0.00
Target Compounds						
10) Phenol	7.24	94	8695	3.641	ng	Qvalue 91
50) Dimethylphthalate	14.15	163	15895	2.559	ng	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG120618\
Data File : BG038388.D
Acq On : 6 Dec 2018 18:33
Operator : JU/SJ
Sample : J6225-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
120318-JETT-E-D-B

Quant Time: Dec 06 21:17:00 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG120418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 04 17:15:40 2018
Response via : Initial Calibration

