


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG120718\
 Data File : BG038422.D
 Acq On : 8 Dec 2018 4:39
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG120418.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Dec 04 17:15:40 2018



AutoFind: Scans 2477, 2478, 2479; Background Corrected with Scan 2470

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.7	24597	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	33.9	20036	PASS
70	69	0.00	2	0.5	108	PASS
127	198	10	80	45.7	26994	PASS
197	198	0.00	2	0.3	176	PASS
198	198	100	100	100.0	59045	PASS
199	198	5	9	7.2	4224	PASS
275	198	10	60	29.8	17601	PASS
365	198	1	100	4.2	2493	PASS
441	198	0.01	100	12.0	7086	PASS
442	198	50	100	77.4	45674	PASS
443	442	15	24	19.2	8750	PASS