

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG120720\
 Data File : BG047635.D
 Acq On : 8 Dec 2020 2:51
 Operator : CG/JU
 Sample : L4862-20
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C0BH3

Manual Integrations
 APPROVED

mohammad
 12/9/2020 12:19:13 PM

Quant Time: Dec 08 05:27:41 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG120720MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Dec 08 04:12:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.21	152	23539	20.00	ng/ul	0.00
18) Naphthalene-d8	11.04	136	95057	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.84	164	73470	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.57	188	190316	20.00	ng/ul	0.00
78) Chrysene-d12	21.85	240	192169	20.00	ng/ul	0.00
86) Perylene-d12	25.21	264	202686	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.53	96	3514m	7.95	ng/uL	0.00
5) Phenol-d5	7.35	99	94029	46.68	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.52	67	51132	49.37	ng/ul	0.00
9) 2-Chlorophenol-d4	7.74	132	72261	47.29	ng/ul	0.00
13) 4-Methylphenol-d8	8.91	113	71213	44.53	ng/ul	0.00
19) Nitrobenzene-d5	9.39	128	38526	49.71	ng/ul	0.00
22) 2-Nitrophenol-d4	10.11	143	43967	47.06	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.65	165	89405	44.07	ng/ul	0.00
29) 4-Chloroaniline-d4	11.17	131	90477	48.57	ng/ul	0.00
44) Dimethylphthalate-d6	14.23	166	302522	47.41	ng/ul	0.00
47) Acenaphthylene-d8	14.53	160	349147	47.90	ng/ul	0.00
52) 4-Nitrophenol-d4	15.00	143	42545	44.95	ng/ul	0.00
58) Fluorene-d10	15.82	176	279332	47.55	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.93	200	56392	43.77	ng/ul	0.00
71) Anthracene-d10	17.67	188	469760m	49.49	ng/ul	0.00
79) Pyrene-d10	19.94	212	544771	47.96	ng/ul	0.00
90) Benzo(a)pyrene-d12	24.99	264	549560	46.42	ng/ul	0.01

Target Compounds

				Ovalue	
6) Phenol	7.38	94	9510	4.997	ng/ul# 85
45) Dimethylphthalate	14.27	163	55730	8.621	ng/ul 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG120720\
Data File : BG047635.D
Acq On : 8 Dec 2020 2:51
Operator : CG/JU
Sample : L4862-20
Misc :
ALS Vial : 13 Sample Multiplier: 1


Instrument :
BNA_G
ClientSampleId :
C0BH3

Manual Integrations
APPROVED

mohammad
12/9/2020 12:19:13 PM

Quant Time: Dec 08 05:27:41 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG120720MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Dec 08 04:12:44 2020
Response via : Initial Calibration





#6

Phenol

Concen: 4.997 ng/ul

RT: 7.38 min Scan# 687

Delta R.T. -0.00 min

Lab File: BG047635.D

Acq: 8 Dec 2020 2:51

Instrument :

BNA_G





ClientSampleId :

C0BH3

Tot Ion:	94	Resp:	9510
Ion Ratio	Lower	Upper	
94	100		
65	53.0	35.9	53.9
66	76.2	50.1	75.1#

Manual Integrations
APPROVED

mohammad
12/9/2020 12:19:13 PM



#45

Dimethylphthalate

Concen: 8.621 ng/ul

RT: 14.27 min Scan# 1861

Delta R.T. 0.00 min

Lab File: BG047635.D

Acq: 8 Dec 2020 2:51

Tgt Ion:	163	Resp:	55730
Ion Ratio	Lower	Upper	
163	100		
194	4.8	3.9	5.9
164	10.3	7.4	11.2

