


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG120820\
 Data File : BG047649.D
 Acq On : 8 Dec 2020 15:30
 Operator : CG/JU
 Sample : DFTPP67
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP67

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG120720MA.M
 Title : SVOA CALIBRATION
 Last Update : Tue Dec 08 17:39:41 2020



AutoFind: Scans 2491, 2492, 2493; Background Corrected with Scan 2486

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.9	22226	PASS
68	69	0.00	2	0.6	160	PASS
69	198	0.00	100	44.6	24828	PASS
70	69	0.00	2	0.2	60	PASS
127	198	10	80	45.3	25213	PASS
197	198	0.00	2	0.3	192	PASS
198	198	100	100	100.0	55696	PASS
199	198	5	9	7.9	4405	PASS
275	198	10	60	33.0	18392	PASS
365	198	1	100	5.2	2887	PASS
441	198	0.01	100	14.0	7801	PASS
442	198	50	100	82.0	45690	PASS
443	442	15	24	18.7	8537	PASS

m/z	Abundance
41.00	181.0
45.00	224.0
55.10	190.0
69.20	153.0
71.50	154.0
73.00	302.0
83.20	173.0
84.90	213.0
111.30	160.0
113.10	228.0
115.10	212.0
132.90	181.0
207.10	181.0
207.80	156.0
232.50	292.0
268.20	226.0
281.00	178.0
286.70	170.0
287.30	213.0
523.70	235.0

Instrument :

BNA_G

ClientSampleId :

DFTPP67

m/z	Abundance
39.10	1447.0
40.90	151.0
50.10	4318.0
51.10	20184.0
52.10	1053.0
52.90	168.0
54.90	158.0
56.00	460.0
57.00	1200.0
61.10	422.0
62.10	374.0
63.00	800.0
63.90	291.0
65.00	354.0
67.60	213.0
69.00	20832.0
73.00	358.0
74.10	2276.0
75.00	3433.0
76.00	2293.0
77.00	18080.0
78.10	1086.0
79.00	1409.0
79.90	1514.0
81.10	1256.0
82.00	553.0
83.20	497.0
84.20	196.0
85.00	254.0
86.00	562.0
87.00	397.0
90.90	888.0
92.00	530.0
93.10	2976.0
95.40	299.0
98.00	1987.0
99.10	2174.0
100.90	1078.0
102.90	440.0
104.10	1036.0
105.00	1070.0
107.00	7573.0
108.00	1366.0
108.90	330.0
110.00	9617.0
111.00	1775.0
111.90	281.0
112.90	222.0
115.90	322.0
117.00	6825.0
118.00	647.0
121.90	634.0
123.20	802.0
124.10	586.0
125.00	537.0
127.00	20400.0
128.00	1753.0
129.00	11528.0
130.00	985.0
133.00	295.0
134.10	253.0
135.10	650.0
135.80	446.0
136.80	494.0
138.10	179.0
140.20	331.0
140.90	1458.0
141.90	267.0
142.90	534.0
147.00	969.0
148.00	1876.0
149.00	298.0
150.90	170.0
152.80	455.0
153.90	158.0
155.10	1195.0
156.10	859.0
158.00	567.0
158.90	241.0
160.00	384.0
161.10	636.0
165.00	370.0
166.00	424.0
167.00	2658.0
168.10	992.0
170.10	218.0
172.20	465.0
172.80	485.0
174.00	709.0
175.00	937.0
176.30	567.0
176.90	410.0

Instrument :
BNA_G
ClientSampleId :
DFTPP67

177.10	421.0
177.90	379.0
179.00	2068.0
179.90	1200.0
181.00	731.0
182.10	315.0
184.00	470.0
184.90	1672.0
186.00	7652.0
187.10	2162.0
187.90	276.0
188.90	1090.0
190.90	343.0
192.00	833.0
193.00	857.0
193.70	250.0
196.20	1453.0
197.90	40760.0
198.90	3013.0
199.90	390.0
201.60	412.0
203.00	423.0
204.00	2366.0
205.00	3292.0
206.00	14397.0
206.90	1770.0
207.80	605.0
208.10	609.0
208.90	281.0
209.90	611.0
210.90	369.0
211.20	393.0
215.00	249.0
216.00	609.0
216.90	4506.0
218.00	864.0
220.90	1609.0
221.90	945.0
223.10	942.0
224.00	7854.0
225.10	2473.0
226.20	290.0
227.00	3844.0
227.90	921.0
228.90	1043.0
230.20	161.0
231.00	463.0
233.90	166.0
235.10	286.0
235.90	160.0
236.90	382.0
238.90	300.0
240.20	245.0
242.00	466.0
244.00	5363.0
244.90	1300.0
246.00	1311.0
246.70	382.0
248.90	268.0
251.90	151.0
253.10	221.0
255.00	33568.0
255.90	5107.0
256.90	396.0
258.00	1935.0
259.00	400.0
264.90	1023.0
273.00	736.0
274.00	2067.0
275.00	11521.0
276.00	2121.0
276.90	1234.0
277.90	191.0
280.90	268.0
282.90	276.0
285.10	413.0
293.10	495.0
296.00	4144.0
296.90	888.0
301.20	221.0
303.20	794.0
313.80	162.0
314.90	368.0
315.90	208.0
323.00	1569.0
323.90	373.0
326.80	471.0
334.00	1034.0
343.10	210.0
346.00	166.0
352.00	589.0
353.10	345.0
354.10	424.0
355.10	225.0
364.90	1672.0
366.00	272.0

Instrument :
BNA_G
ClientSampleId :
DFTPP67

366.70	153.0
372.00	618.0
383.00	166.0
402.90	240.0
403.30	250.0
403.90	246.0
421.00	353.0
421.70	211.0
422.90	1682.0
424.10	199.0
438.10	210.0
441.00	4276.0
442.00	25728.0
443.00	5368.0
444.00	369.0
489.00	152.0
535.50	170.0

Instrument :
BNA_G
ClientSampleId :
DFTPP67

m/z	Abundance
38.00	824.0
39.10	1701.0
41.10	333.0
49.20	367.0
50.10	6523.0
51.10	26600.0
51.90	681.0
53.20	264.0
55.00	443.0
56.20	1076.0
57.00	2128.0
60.80	485.0
61.90	326.0
63.00	907.0
64.20	320.0
65.10	316.0
65.80	166.0
66.40	163.0
69.00	30264.0
70.10	181.0
73.00	539.0
74.10	2967.0
75.10	5062.0
76.00	1259.0
77.10	29024.0
78.10	1630.0
79.00	3270.0
80.00	1972.0
81.00	1671.0
82.10	515.0
83.10	487.0
83.90	224.0
85.10	719.0
86.00	1080.0
87.00	310.0
91.00	869.0
92.10	615.0
93.00	5449.0
94.10	407.0
94.90	180.0
96.20	196.0
98.00	4260.0
99.10	2645.0
100.20	337.0
101.00	1515.0
102.90	359.0
104.00	979.0
105.00	1173.0
107.00	11337.0
108.00	2050.0
109.10	991.0
110.00	12807.0
111.00	3557.0
111.90	375.0
116.10	668.0
117.00	11672.0
118.00	664.0
120.30	171.0
121.10	171.0
122.20	661.0
123.00	1100.0
124.00	853.0
125.10	366.0
125.70	283.0
127.00	29144.0
128.10	1794.0
129.00	13148.0
129.90	983.0
130.90	305.0
131.80	308.0
134.00	468.0
135.10	1255.0
136.00	316.0
136.90	336.0
138.10	293.0
139.90	264.0
141.10	2146.0
142.20	489.0
142.80	409.0
144.70	167.0
145.70	215.0
147.00	1574.0
148.00	2727.0
149.00	341.0
150.10	456.0
151.00	420.0
152.10	197.0
152.80	527.0
154.10	292.0
155.10	1018.0
156.00	1839.0
157.00	578.0

Instrument :
BNA_G
ClientSampleId :
DFTPP67

157.80	408.0
158.20	398.0
158.90	652.0
160.00	796.0
161.00	1097.0
162.00	163.0
163.90	204.0
164.90	1013.0
166.00	485.0
167.10	3541.0
168.10	1316.0
168.90	394.0
170.10	248.0
170.90	314.0
172.00	502.0
173.10	318.0
174.00	835.0
175.10	2091.0
175.90	525.0
177.10	694.0
178.10	311.0
179.00	3727.0
179.90	2082.0
181.00	796.0
182.10	209.0
184.00	396.0
185.00	1284.0
186.00	12277.0
187.00	3434.0
187.80	507.0
188.90	849.0
189.80	157.0
190.90	553.0
191.90	1150.0
193.10	1428.0
196.00	1787.0
196.80	577.0
198.00	67408.0
198.90	5492.0
199.90	589.0
203.10	641.0
204.00	3684.0
205.00	6001.0
206.00	21520.0
207.00	2847.0
207.90	1356.0
209.20	357.0
210.00	416.0
211.00	791.0
212.30	161.0
215.20	270.0
216.00	602.0
217.00	6871.0
218.00	603.0
220.00	208.0
220.90	2149.0
222.10	800.0
222.90	2113.0
224.00	14772.0
225.00	3950.0
225.90	612.0
227.00	9677.0
228.00	288.0
228.90	1499.0
229.80	163.0
231.10	394.0
231.90	251.0
233.10	382.0
234.00	540.0
235.10	544.0
235.90	563.0
237.00	451.0
239.00	368.0
241.10	776.0
242.10	299.0
243.00	869.0
244.00	8224.0
244.80	883.0
246.00	1968.0
246.80	658.0
247.70	158.0
249.00	616.0
249.80	427.0
250.70	158.0
253.00	318.0
255.00	57080.0
256.00	6706.0
257.10	588.0
258.00	3773.0
258.90	748.0
259.90	393.0
260.70	282.0
263.90	225.0
265.00	2000.0
266.00	254.0
270.00	352.0

Instrument :
BNA_G
ClientSampleId :
DFTPP67

270.90	401.0
271.90	295.0
272.90	1640.0
274.00	3336.0
275.00	22376.0
276.00	2610.0
277.00	2421.0
277.90	499.0
282.20	266.0
284.90	552.0
289.00	193.0
293.00	338.0
294.10	404.0
294.90	227.0
296.00	7363.0
297.00	1178.0
298.20	155.0
298.80	161.0
302.20	238.0
303.10	659.0
304.10	244.0
314.10	595.0
314.90	628.0
315.90	442.0
321.80	258.0
323.00	1589.0
324.10	433.0
326.90	319.0
328.00	496.0
332.10	208.0
334.00	1545.0
335.00	442.0
340.90	207.0
341.90	198.0
346.00	550.0
352.00	680.0
353.00	692.0
353.90	577.0
354.90	175.0
364.90	3208.0
366.20	393.0
369.80	184.0
371.00	277.0
372.10	1148.0
373.00	271.0
383.00	431.0
389.90	160.0
390.20	180.0
390.70	321.0
400.90	212.0
401.90	407.0
402.90	166.0
404.10	240.0
414.90	151.0
420.80	345.0
421.90	597.0
423.00	2546.0
424.00	498.0
425.00	182.0
441.00	8495.0
442.00	48440.0
443.00	9402.0
444.10	940.0
445.00	153.0

Instrument :
BNA_G
ClientSampleId :
DFTPP67

m/z	Abundance
37.90	306.0
39.10	1466.0
41.00	259.0
43.10	205.0
44.00	257.0
44.80	281.0
47.90	193.0
49.10	253.0
50.10	6250.0
51.10	19896.0
52.10	1275.0
52.90	197.0
54.90	386.0
56.00	516.0
57.00	1227.0
60.00	267.0
60.90	416.0
62.10	406.0
63.10	1377.0
64.20	186.0
65.00	556.0
67.00	245.0
68.00	482.0
69.00	23848.0
73.10	458.0
74.10	3557.0
75.00	4593.0
76.00	1520.0
77.10	25672.0
78.10	1400.0
79.00	3256.0
80.10	1384.0
81.00	2477.0
82.10	385.0
83.00	500.0
84.90	533.0
85.80	1021.0
87.00	217.0
90.80	253.0
92.10	456.0
93.00	3186.0
97.20	406.0
98.00	3189.0
99.10	2283.0
100.10	360.0
101.10	1439.0
101.90	151.0
102.90	260.0
104.00	1008.0
105.10	933.0
106.10	503.0
107.00	10282.0
108.00	1373.0
108.80	330.0
110.00	12674.0
111.00	2310.0
112.00	484.0
113.00	219.0
116.00	1019.0
117.00	9265.0
117.80	906.0
120.10	229.0
121.90	590.0
123.10	901.0
123.80	558.0
124.80	426.0
125.90	714.0
127.00	26096.0
128.00	2384.0
129.00	12755.0
129.90	1164.0
131.00	354.0
132.40	248.0
134.00	421.0
135.00	1072.0
136.10	575.0
137.10	554.0
137.90	345.0
140.90	2091.0
141.90	1005.0
143.10	652.0
146.10	400.0
147.00	1092.0
148.00	2710.0
149.10	615.0
149.80	213.0
150.20	156.0
151.30	345.0
152.90	513.0
153.80	299.0
154.20	194.0
154.90	1335.0

Instrument :
BNA_G
ClientSampleId :
DFTPP67

155.90	1423.0
157.00	452.0
157.60	352.0
158.00	407.0
159.00	251.0
160.10	361.0
160.90	837.0
165.00	683.0
165.70	382.0
166.00	548.0
167.00	3133.0
168.00	2291.0
169.00	623.0
169.90	184.0
171.00	291.0
171.80	601.0
173.20	456.0
173.90	608.0
175.00	1415.0
176.00	1030.0
177.00	668.0
179.00	3051.0
180.00	2478.0
181.00	1114.0
181.80	205.0
183.30	257.0
184.30	289.0
184.90	1581.0
186.00	11352.0
187.00	4580.0
188.10	316.0
189.10	736.0
191.10	746.0
192.00	1208.0
193.00	981.0
195.90	1733.0
197.90	58920.0
198.90	4712.0
199.90	408.0
200.70	405.0
203.00	657.0
204.00	3438.0
205.00	6366.0
206.00	21952.0
207.00	3232.0
208.00	893.0
209.00	350.0
210.10	582.0
210.90	490.0
211.20	541.0
213.40	159.0
214.90	299.0
216.00	881.0
217.00	7041.0
217.80	1258.0
218.90	184.0
220.10	332.0
221.00	1970.0
221.90	1150.0
223.10	1579.0
224.00	13709.0
225.10	3667.0
226.10	515.0
227.00	8064.0
227.90	697.0
229.10	1042.0
231.00	892.0
231.90	190.0
234.00	330.0
234.90	386.0
236.00	224.0
236.90	445.0
237.80	194.0
240.10	221.0
240.60	209.0
241.10	521.0
242.00	1127.0
243.00	326.0
244.00	8250.0
245.00	1509.0
245.90	3100.0
246.90	353.0
248.90	332.0
250.80	164.0
252.00	178.0
253.00	425.0
255.00	54296.0
256.00	7397.0
257.00	565.0
257.90	4035.0
259.00	735.0
261.10	224.0
265.00	2223.0
266.00	379.0
273.00	1715.0
274.00	3581.0

Instrument :
BNA_G
ClientSampleId :
DFTPP67

275.00	21280.0
276.00	3724.0
277.00	2829.0
277.90	275.0
280.70	187.0
282.70	194.0
283.10	224.0
284.00	196.0
285.00	364.0
285.20	358.0
287.20	176.0
290.80	164.0
293.00	300.0
295.00	206.0
296.00	7813.0
297.10	1278.0
303.10	771.0
303.90	620.0
314.10	352.0
315.10	874.0
315.90	506.0
322.10	264.0
323.00	2345.0
323.90	769.0
327.00	581.0
328.00	432.0
329.00	217.0
334.10	1625.0
335.10	315.0
341.10	375.0
341.80	156.0
345.90	585.0
351.60	373.0
352.10	660.0
352.90	615.0
354.20	184.0
355.00	226.0
365.00	3782.0
366.10	412.0
371.30	161.0
372.10	1194.0
373.10	316.0
383.00	524.0
395.10	183.0
401.00	230.0
401.80	277.0
403.00	749.0
416.30	203.0
420.80	662.0
421.80	586.0
423.00	3065.0
424.00	629.0
426.90	173.0
441.00	10634.0
442.00	62904.0
443.10	10842.0
444.00	1358.0
445.10	202.0

Instrument :
BNA_G
ClientSampleId :
DFTPP67