

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG120822\
 Data File : BG055934.D
 Acq On : 8 Dec 2022 19:27
 Operator : CG/JU
 Sample : N5938-01
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PIER-STOCK-PILE

Quant Time: Dec 09 08:11:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 09 07:57:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.137	152	26030	20.000	ng	0.00
21) Naphthalene-d8	10.969	136	110984	20.000	ng	0.00
39) Acenaphthene-d10	14.776	164	85043	20.000	ng	0.00
64) Phenanthrene-d10	17.520	188	232238	20.000	ng	0.00
76) Chrysene-d12	21.797	240	210370	20.000	ng	0.00
86) Perylene-d12	25.123	264	236989	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.675	112	25944	18.015	ng	0.00
7) Phenol-d6	7.297	99	146340	76.338	ng	0.00
23) Nitrobenzene-d5	9.318	82	165130	78.641	ng	0.00
42) 2,4,6-Tribromophenol	16.292	330	501	0.418	ng	0.03
45) 2-Fluorobiphenyl	13.395	172	449085	79.111	ng	0.00
79) Terphenyl-d14	20.105	244	894291	85.026	ng	0.00
Target Compounds						
25) Isophorone	9.876	82	11071	2.783	ng	Qvalue # 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG120822\
Data File : BG055934.D
Acq On : 8 Dec 2022 19:27
Operator : CG/JU
Sample : N5938-01
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PIER-STOCK-PILE

Quant Time: Dec 09 08:11:59 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 09 07:57:14 2022
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.137 min Scan# 81
Delta R.T. -0.003 min
Lab File: BG055934.D
Acq: 8 Dec 2022 19:27




Instrument :
BNA_G
ClientSampleId :
PIER-STOCK-PILE

Tgt Ion:152 Resp: 26030
Ion Ratio Lower Upper
152 100
150 153.4 123.0 184.6
115 58.9 47.0 70.4



#5
2-Fluorophenol
Concen: 18.015 ng
RT: 5.675 min Scan# 406
Delta R.T. 0.002 min
Lab File: BG055934.D
Acq: 8 Dec 2022 19:27

Tgt Ion:112 Resp: 25944
Ion Ratio Lower Upper
112 100
64 65.8 50.6 76.0
63 34.6 27.5 41.3









#7
Phenol-d6
Concen: 76.338 ng
RT: 7.297 min Scan# 682
Delta R.T. -0.004 min
Lab File: BG055934.D
Acq: 8 Dec 2022 19:27




Instrument :
BNA_G
ClientSampleId :
PIER-STOCK-PILE


Tgt Ion: 99 Resp: 146340
Ion Ratio Lower Upper
99 100
42 23.5 18.0 27.0
71 37.2 29.0 43.6



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.969 min Scan# 1307
Delta R.T. -0.004 min
Lab File: BG055934.D
Acq: 8 Dec 2022 19:27

Tgt Ion:136 Resp: 110984
Ion Ratio Lower Upper
136 100
137 11.4 9.0 13.6
54 8.9 7.8 11.6
68 5.1 5.0 7.4





#23

Nitrobenzene-d5

Concen: 78.641 ng

RT: 9.318 min Scan# 1045

Delta R.T. -0.004 min

Lab File: BG055934.D

Acq: 8 Dec 2022 19:27

Instrument :

BNA_G

ClientSampleId :

PIER-STOCK-PILE





Tgt Ion: 82 Resp: 165130

Ion Ratio Lower Upper

82 100

128 43.1 34.1 51.1

54 58.0 47.3 70.9



#25

Isophorone

Concen: 2.783 ng

RT: 9.876 min Scan# 1121

Delta R.T. -0.010 min

Lab File: BG055934.D

Acq: 8 Dec 2022 19:27




Tgt Ion: 82 Resp: 11071


Ion Ratio Lower Upper

82 100

95 8.7 5.5 8.3#

138 17.1 14.2 21.4









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.776 min Scan# 1972
Delta R.T. -0.004 min
Lab File: BG055934.D
Acq: 8 Dec 2022 19:27




Instrument :
BNA_G
ClientSampleId :
PIER-STOCK-PILE


Tgt Ion:164 Resp: 85043
Ion Ratio Lower Upper
164 100
162 103.5 81.1 121.7
160 46.9 34.8 52.2



#42
2,4,6-Tribromophenol
Concen: 0.418 ng
RT: 16.292 min Scan# 2213
Delta R.T. 0.026 min
Lab File: BG055934.D
Acq: 8 Dec 2022 19:27

Tgt Ion:330 Resp: 501
Ion Ratio Lower Upper
330 100
332 94.2 77.8 116.8
141 0.0 23.9 35.9#









#45
2-Fluorobiphenyl
Concen: 79.111 ng
RT: 13.395 min Scan# 11
Delta R.T. -0.010 min
Lab File: BG055934.D
Acq: 8 Dec 2022 19:27




Instrument :
BNA_G
ClientSampleId :
PIER-STOCK-PILE


Tgt Ion:172 Resp: 449085
Ion Ratio Lower Upper
172 100
171 36.8 29.8 44.6
170 24.6 20.2 30.4



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.520 min Scan# 2422
Delta R.T. -0.004 min
Lab File: BG055934.D
Acq: 8 Dec 2022 19:27

Tgt Ion:188 Resp: 232238
Ion Ratio Lower Upper
188 100
94 7.3 5.4 8.2
80 8.4 6.6 10.0









#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.797 min Scan# 3171
Delta R.T. -0.004 min
Lab File: BG055934.D
Acq: 8 Dec 2022 19:27




Instrument :
BNA_G
ClientSampleId :
PIER-STOCK-PILE


Tgt Ion:240 Resp: 210370
Ion Ratio Lower Upper
240 100
120 7.9 6.3 9.5
236 27.0 21.9 32.9



#79
Terphenyl-d14
Concen: 85.026 ng
RT: 20.105 min Scan# 2862
Delta R.T. -0.004 min
Lab File: BG055934.D
Acq: 8 Dec 2022 19:27

Tgt Ion:244 Resp: 894291
Ion Ratio Lower Upper
244 100
212 7.8 6.3 9.5
122 9.2 6.9 10.3





#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 25.123 min Scan# 31
 Delta R.T. -0.010 min
 Lab File: BG055934.D
 Acq: 8 Dec 2022 19:27

Instrument :
 BNA_G
 ClientSampleId :
 PIER-STOCK-PILE

Tgt Ion:264 Resp: 236989
 Ion Ratio Lower Upper
 264 100
 260 23.6 19.5 29.3
 265 21.0 17.1 25.7

