

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121125\
 Data File : BG064935.D
 Acq On : 11 Dec 2025 15:33
 Operator : RC/JU
 Sample : PB170874TB
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB170874TB

Quant Time: Dec 11 16:31:01 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 12 17:10:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.150	152	17348	20.000	ng	-0.02
21) Naphthalene-d8	10.970	136	65024	20.000	ng	#-0.02
39) Acenaphthene-d10	14.766	164	56650	20.000	ng	-0.01
64) Phenanthrene-d10	17.498	188	149344	20.000	ng	# 0.00
76) Chrysene-d12	21.746	240	196781	20.000	ng	#-0.01
86) Perylene-d12	25.007	264	223958	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.694	112	145100	146.055	ng	0.00
7) Phenol-d6	7.304	99	181931	137.686	ng	0.00
23) Nitrobenzene-d5	9.313	82	120233	91.423	ng	-0.02
42) 2,4,6-Tribromophenol	16.240	330	174144	157.618	ng	-0.01
45) 2-Fluorobiphenyl	13.397	172	398720	98.208	ng	-0.01
79) Terphenyl-d14	20.083	244	1149733	114.931	ng	-0.01


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121125\
Data File : BG064935.D
Acq On : 11 Dec 2025 15:33
Operator : RC/JU
Sample : PB170874TB
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB170874TB

Quant Time: Dec 11 16:31:01 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG111225.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 12 17:10:56 2025
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 8.150 min Scan# 864
 Delta R.T. -0.017 min
 Lab File: BG064935.D
 Acq: 11 Dec 2025 15:33




Instrument :
 BNA_G
 ClientSampleId :
 PB170874TB


Tgt Ion:152 Resp: 17348
 Ion Ratio Lower Upper
 152 100
 150 146.2 129.2 193.8
 115 66.1 53.7 80.5



#5
 2-Fluorophenol
 Concen: 146.055 ng
 RT: 5.694 min Scan# 443
 Delta R.T. -0.006 min
 Lab File: BG064935.D
 Acq: 11 Dec 2025 15:33

Tgt Ion:112 Resp: 145100
 Ion Ratio Lower Upper
 112 100
 64 67.8 61.0 91.4
 63 40.4 33.3 49.9









#7
Phenol-d6
Concen: 137.686 ng
RT: 7.304 min Scan# 718
Delta R.T. -0.006 min
Lab File: BG064935.D
Acq: 11 Dec 2025 15:33




Instrument :
BNA_G
ClientSampleId :
PB170874TB

Tgt Ion: 99 Resp: 181931
Ion Ratio Lower Upper
99 100
42 23.1 16.8 25.2
71 45.4 31.1 46.7



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.970 min Scan# 1341
Delta R.T. -0.017 min
Lab File: BG064935.D
Acq: 11 Dec 2025 15:33

Tgt Ion: 136 Resp: 65024
Ion Ratio Lower Upper
136 100
137 11.6 9.3 13.9
54 9.3 9.3 13.9
68 3.5 5.0 7.6#









#23
Nitrobenzene-d5
Concen: 91.423 ng
RT: 9.313 min Scan# 1062
Delta R.T. -0.017 min
Lab File: BG064935.D
Acq: 11 Dec 2025 15:33




Instrument :
BNA_G
ClientSampleId :
PB170874TB


Tgt Ion: 82 Resp: 120233
Ion Ratio Lower Upper
82 100
128 41.1 34.1 51.1
54 68.5 54.3 81.5



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.766 min Scan# 1987
Delta R.T. -0.012 min
Lab File: BG064935.D
Acq: 11 Dec 2025 15:33

Tgt Ion:164 Resp: 56650
Ion Ratio Lower Upper
164 100
162 101.4 79.8 119.6
160 43.1 34.9 52.3





#42

2,4,6-Tribromophenol

Concen: 157.618 ng

RT: 16.240 min Scan# 21

Delta R.T. -0.012 min

Lab File: BG064935.D


Acq: 11 Dec 2025 15:33

Instrument :

BNA_G

ClientSampleId :

PB170874TB






Tgt Ion:330 Resp: 174144

Ion Ratio Lower Upper

330 100

332 97.9 77.0 115.6

141 19.8 18.2 27.2



#45

2-Fluorobiphenyl

Concen: 98.208 ng

RT: 13.397 min Scan# 1754

Delta R.T. -0.012 min

Lab File: BG064935.D

Acq: 11 Dec 2025 15:33




Tgt Ion:172 Resp: 398720


Ion Ratio Lower Upper

172 100

171 35.8 29.9 44.9

170 25.0 19.8 29.6









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.498 min Scan# 2453
Delta R.T. -0.006 min
Lab File: BG064935.D
Acq: 11 Dec 2025 15:33




Instrument :
BNA_G
ClientSampleId :
PB170874TB


Tgt Ion:188 Resp: 149344
Ion Ratio Lower Upper
188 100
94 3.7 4.7 7.1#
80 4.4 5.4 8.2#



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.746 min Scan# 3175
Delta R.T. -0.012 min
Lab File: BG064935.D
Acq: 11 Dec 2025 15:33

Tgt Ion:240 Resp: 196781
Ion Ratio Lower Upper
240 100
120 3.5 4.2 6.4#
236 24.5 20.8 31.2









#79
 Terphenyl-d14
 Concen: 114.931 ng
 RT: 20.083 min Scan# 2892
 Delta R.T. -0.012 min
 Lab File: BG064935.D
 Acq: 11 Dec 2025 15:33

Instrument :
 BNA_G
 ClientSampleId :
 PB170874TB

Tgt Ion:244 Resp: 1149733
 Ion Ratio Lower Upper
 244 100
 212 7.0 5.8 8.6
 122 4.2 4.3 6.5#



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 25.007 min Scan# 3730
 Delta R.T. -0.017 min
 Lab File: BG064935.D
 Acq: 11 Dec 2025 15:33

Tgt Ion:264 Resp: 223958
 Ion Ratio Lower Upper
 264 100
 260 23.0 17.2 25.8
 265 20.0 16.6 25.0

