




Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG121318\
Data File : BG038535.D
Acq On : 13 Dec 2018 16:46
Operator : JU/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 14 08:52:25 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG121218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 12 15:26:27 2018
Response via : Initial Calibration



TIC: BG038535.D

(70) Pentachlorophenol (C)




17.089min (-0.001) 28216.59ng

response 36131

Ion	Exp%	Act%
265.70	100	100
268.00	68.80	69.09
264.00	70.90	69.48
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG121318\
Data File : BG038535.D
Acq On : 13 Dec 2018 16:46
Operator : JU/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 14 08:52:25 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG121218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 12 15:26:27 2018
Response via : Initial Calibration



TIC: BG038535.D

(77) Benzidine

19.680min (-0.001) 0.00ng

response 264808

Ion	Exp%	Act%
184.00	100	100
185.00	15.20	14.66
183.00	12.00	11.98
0.00	0.00	0.00