

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121323\
 Data File : BG060000.D
 Acq On : 13 Dec 2023 17:23
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG121323

Quant Time: Dec 14 01:48:17 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG121323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 14 01:47:12 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2	1,4-Dioxane	0.553	0.536	3.1	98	0.00
3	Pyridine	1.548	1.558	-0.6	102	0.00
4	n-Nitrosodimethylamine	0.762	0.769	-0.9	107	0.00
5 S	2-Fluorophenol	1.168	1.229	-5.2	118	0.00
6	Aniline	1.674	1.671	0.2	101	0.00
7 S	Phenol-d6	1.643	1.660	-1.0	104	0.00
8	2-Chlorophenol	1.107	1.088	1.7	102	0.00
9	Benzaldehyde	0.999	1.024	-2.5	104	0.00
10 C	Phenol	1.651	1.667	-1.0	103	0.00
11	bis(2-Chloroethyl)ether	1.182	1.147	3.0	101	0.00
12	1,3-Dichlorobenzene	1.481	1.487	-0.4	102	0.00
13 C	1,4-Dichlorobenzene	1.541	1.519	1.4	105	0.00
14	1,2-Dichlorobenzene	1.482	1.431	3.4	102	0.00
15	Benzyl Alcohol	1.321	1.304	1.3	102	0.00
16	2,2'-oxybis(1-Chloropropane	2.115	2.118	-0.1	104	0.00
17	2-Methylphenol	1.101	1.108	-0.6	100	0.00
18	Hexachloroethane	0.467	0.461	1.3	101	0.00
19 P	n-Nitroso-di-n-propylamine	1.060	1.060	0.0	102	0.00
20	3+4-Methylphenols	1.537	1.514	1.5	103	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
22	Acetophenone	0.583	0.581	0.3	102	0.00
23 S	Nitrobenzene-d5	0.440	0.453	-3.0	104	0.00
24	Nitrobenzene	0.447	0.475	-6.3	107	0.00
25	Isophorone	0.843	0.846	-0.4	100	0.00
26 C	2-Nitrophenol	0.164	0.177	-7.9	110	0.00
27	2,4-Dimethylphenol	0.225	0.217	3.6	96	0.00
28	bis(2-Chloroethoxy)methane	0.435	0.437	-0.5	102	0.00
29 C	2,4-Dichlorophenol	0.405	0.426	-5.2	108	0.00
30	1,2,4-Trichlorobenzene	0.548	0.564	-2.9	106	0.00
31	Naphthalene	1.041	1.049	-0.8	103	0.00
32	Benzoic acid	0.205	0.213	-3.9	103	0.00
33	4-Chloroaniline	0.396	0.398	-0.5	102	0.00
34 C	Hexachlorobutadiene	0.502	0.504	-0.4	102	0.00
35	Caprolactam	0.103	0.103	0.0	102	0.00
36 C	4-Chloro-3-methylphenol	0.368	0.372	-1.1	101	0.00
37	2-Methylnaphthalene	0.817	0.829	-1.5	103	0.00
38	1-Methylnaphthalene	0.807	0.817	-1.2	104	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
40	1,2,4,5-Tetrachlorobenzene	1.043	1.092	-4.7	105	0.00
41 P	Hexachlorocyclopentadiene	0.431	0.462	-7.2	107	0.00
42 S	2,4,6-Tribromophenol	0.466	0.486	-4.3	106	0.00
43 C	2,4,6-Trichlorophenol	0.585	0.617	-5.5	105	0.00
44	2,4,5-Trichlorophenol	0.636	0.664	-4.4	102	0.00
45 S	2-Fluorobiphenyl	1.634	1.703	-4.2	105	0.00
46	1,1'-Biphenyl	1.483	1.529	-3.1	103	0.00
47	2-Chloronaphthalene	1.285	1.325	-3.1	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121323\
 Data File : BG060000.D
 Acq On : 13 Dec 2023 17:23
 Operator : MA/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG121323

Quant Time: Dec 14 01:48:17 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG121323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 14 01:47:12 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.277	0.312	-12.6	112	0.00
49	Acenaphthylene	1.757	1.819	-3.5	102	0.00
50	Dimethylphthalate	1.552	1.598	-3.0	104	0.00
51	2,6-Dinitrotoluene	0.322	0.357	-10.9	110	0.00
52 C	Acenaphthene	1.210	1.264	-4.5	107	0.00
53	3-Nitroaniline	0.257	0.278	-8.2	111	0.00
54 P	2,4-Dinitrophenol	0.237	0.260	-9.7	107	0.00
55	Dibenzofuran	1.969	2.062	-4.7	105	0.00
56 P	4-Nitrophenol	0.216	0.224	-3.7	104	0.00
57	2,4-Dinitrotoluene	0.452	0.491	-8.6	108	0.00
58	Fluorene	1.660	1.732	-4.3	105	0.00
59	2,3,4,6-Tetrachlorophenol	0.662	0.707	-6.8	110	0.00
60	Diethylphthalate	1.428	1.475	-3.3	103	0.00
61	4-Chlorophenyl-phenylether	1.196	1.257	-5.1	106	0.00
62	4-Nitroaniline	0.280	0.293	-4.6	106	0.00
63	Azobenzene	1.408	1.438	-2.1	102	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	0.125	0.141	-12.8	111	0.00
66 c	n-Nitrosodiphenylamine	0.489	0.516	-5.5	107	0.00
67	4-Bromophenyl-phenylether	0.282	0.298	-5.7	106	0.00
68	Hexachlorobenzene	0.330	0.342	-3.6	104	0.00
69	Atrazine	0.187	0.179	4.3	107	0.00
70 C	Pentachlorophenol	0.214	0.230	-7.5	106	0.00
71	Phenanthrene	0.969	1.000	-3.2	105	0.00
72	Anthracene	0.974	1.013	-4.0	106	0.00
73	Carbazole	0.828	0.850	-2.7	105	0.00
74	Di-n-butylphthalate	0.755	0.801	-6.1	108	0.00
75 C	Fluoranthene	1.414	1.472	-4.1	106	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
77	Benzidine	0.257	0.356	-38.5#	91	0.00
78	Pyrene	1.212	1.298	-7.1	106	0.00
79 S	Terphenyl-d14	1.095	1.174	-7.2	104	0.00
80	Butylbenzylphthalate	0.218	0.240	-10.1	108	0.00
81	Benzo(a)anthracene	1.343	1.402	-4.4	103	0.00
82	3,3'-Dichlorobenzidine	0.442	0.467	-5.7	102	0.00
83	Chrysene	1.273	1.338	-5.1	104	0.00
84	Bis(2-ethylhexyl)phthalate	0.330	0.364	-10.3	111	0.00
85 c	Di-n-octyl phthalate	0.571	0.622	-8.9	108	0.00
86 I	Perylene-d12	1.000	1.000	0.0	104	0.00
87	Indeno(1,2,3-cd)pyrene	1.427	1.468	-2.9	104	0.00
88	Benzo(b)fluoranthene	1.235	1.262	-2.2	102	0.00
89	Benzo(k)fluoranthene	1.198	1.235	-3.1	104	0.00
90 C	Benzo(a)pyrene	1.094	1.117	-2.1	103	0.00
91	Dibenzo(a,h)anthracene	1.194	1.233	-3.3	105	0.00
92	Benzo(g,h,i)perylene	1.187	1.212	-2.1	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121323\
Data File : BG060000.D
Acq On : 13 Dec 2023 17:23
Operator : MA/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG121323

Quant Time: Dec 14 01:48:17 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG121323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 14 01:47:12 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0