

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG121418\
 Data File : BG038588.D
 Acq On : 15 Dec 2018 4:20
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 15 05:50:19 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG121218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 12 15:26:27 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	0.00
2	1,4-Dioxane	0.525	0.508	3.2	111	0.00
3	Pyridine	1.445	1.412	2.3	95	0.00
4	n-Nitrosodimethylamine	0.708	0.763	-7.8	115	0.00
5 S	2-Fluorophenol	1.128	1.168	-3.5	120	0.00
6	Aniline	1.976	2.034	-2.9	118	0.00
7 S	Phenol-d6	1.548	1.611	-4.1	121	0.00
8	2-Chlorophenol	1.305	1.341	-2.8	119	0.00
9	Benzaldehyde	1.004	0.980	2.4	123	0.00
10 C	Phenol	1.645	1.691	-2.8	120	0.00
11	bis(2-Chloroethyl)ether	1.309	1.342	-2.5	122	0.00
12	1,3-Dichlorobenzene	1.543	1.572	-1.9	120	0.00
13 C	1,4-Dichlorobenzene	1.580	1.570	0.6	118	0.00
14	1,2-Dichlorobenzene	1.490	1.486	0.3	119	0.00
15	Benzyl Alcohol	1.208	1.299	-7.5	120	0.00
16	2,2'-oxybis(1-Chloropropane	2.999	3.020	-0.7	119	0.00
17	2-Methylphenol	1.102	1.160	-5.3	120	0.00
18	Hexachloroethane	0.577	0.587	-1.7	120	0.00
19 P	n-Nitroso-di-n-propylamine	1.087	1.115	-2.6	118	0.00
20	3+4-Methylphenols	1.522	1.640	-7.8	123	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	120	0.00
22	Acetophenone	0.500	0.491	1.8	117	0.00
23 S	Nitrobenzene-d5	0.391	0.402	-2.8	122	0.00
24	Nitrobenzene	0.396	0.400	-1.0	121	0.00
25	Isophorone	0.703	0.669	4.8	97	0.00
26 C	2-Nitrophenol	0.203	0.207	-2.0	122	0.00
27	2,4-Dimethylphenol	0.291	0.300	-3.1	123	0.00
28	bis(2-Chloroethoxy)methane	0.397	0.402	-1.3	119	0.00
29 C	2,4-Dichlorophenol	0.323	0.345	-6.8	122	0.00
30	1,2,4-Trichlorobenzene	0.390	0.399	-2.3	123	0.00
31	Naphthalene	0.985	0.973	1.2	118	0.00
32	Benzoic acid	0.154	0.161	-4.5	113	0.00
33	4-Chloroaniline	0.428	0.439	-2.6	119	0.00
34 C	Hexachlorobutadiene	0.264	0.270	-2.3	120	0.00
35	Caprolactam	0.134	0.125	6.7	117	0.02
36 C	4-Chloro-3-methylphenol	0.369	0.365	1.1	119	0.00
37	2-Methylnaphthalene	0.703	0.714	-1.6	122	0.00
38	1-Methylnaphthalene	0.682	0.678	0.6	120	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	121	0.00
40	1,2,4,5-Tetrachlorobenzene	0.748	0.758	-1.3	121	0.00
41 P	Hexachlorocyclopentadiene	0.411	0.399	2.9	114	0.00
42 S	2,4,6-Tribromophenol	0.276	0.277	-0.4	118	0.00
43 C	2,4,6-Trichlorophenol	0.460	0.476	-3.5	117	0.00
44	2,4,5-Trichlorophenol	0.487	0.495	-1.6	116	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG121418\
 Data File : BG038588.D
 Acq On : 15 Dec 2018 4:20
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 15 05:50:19 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG121218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 12 15:26:27 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.458	1.439	1.3	119	0.00
46	1,1'-Biphenyl	1.677	1.675	0.1	118	0.00
47	2-Chloronaphthalene	1.295	1.287	0.6	120	0.00
48	2-Nitroaniline	0.413	0.435	-5.3	123	0.00
49	Acenaphthylene	1.936	1.926	0.5	118	0.00
50	Dimethylphthalate	1.633	1.632	0.1	119	0.00
51	2,6-Dinitrotoluene	0.370	0.374	-1.1	120	0.00
52 C	Acenaphthene	1.158	1.133	2.2	117	0.00
53	3-Nitroaniline	0.377	0.388	-2.9	120	0.00
54 P	2,4-Dinitrophenol	0.193	0.174	9.8	104	0.00
55	Dibenzofuran	1.985	1.962	1.2	120	0.00
56 P	4-Nitrophenol	0.286	0.284	0.7	113	0.00
57	2,4-Dinitrotoluene	0.521	0.520	0.2	116	0.00
58	Fluorene	1.470	1.445	1.7	117	0.00
59	2,3,4,6-Tetrachlorophenol	0.438	0.436	0.5	114	0.00
60	Diethylphthalate	1.793	1.729	3.6	117	0.00
61	4-Chlorophenyl-phenylether	0.917	0.916	0.1	119	0.00
62	4-Nitroaniline	0.388	0.398	-2.6	117	0.00
63	Azobenzene	1.498	1.469	1.9	116	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	118	0.00
65	4,6-Dinitro-2-methylphenol	0.143	0.142	0.7	116	0.00
66 c	n-Nitrosodiphenylamine	0.588	0.592	-0.7	119	0.00
67	4-Bromophenyl-phenylether	0.244	0.247	-1.2	118	0.00
68	Hexachlorobenzene	0.253	0.255	-0.8	116	0.00
69	Atrazine	0.218	0.223	-2.3	117	0.00
70 C	Pentachlorophenol	0.150	0.153	-2.0	116	0.00
71	Phenanthrene	1.037	1.030	0.7	118	0.00
72	Anthracene	1.024	1.031	-0.7	117	0.00
73	Carbazole	1.013	1.021	-0.8	118	0.00
74	Di-n-butylphthalate	1.162	1.177	-1.3	117	0.00
75 C	Fluoranthene	1.283	1.269	1.1	119	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	116	0.00
77	Benzidine	0.591	0.654	-10.7	107	0.00
78	Pyrene	1.321	1.295	2.0	116	0.00
79 S	Terphenyl-d14	0.919	0.926	-0.8	117	0.00
80	Butylbenzylphthalate	0.516	0.534	-3.5	117	0.00
81	Benzo(a)anthracene	1.219	1.239	-1.6	115	0.00
82	3,3'-Dichlorobenzidine	0.483	0.503	-4.1	115	0.00
83	Chrysene	1.174	1.173	0.1	115	0.00
84	Bis(2-ethylhexyl)phthalate	0.732	0.761	-4.0	115	0.00
85 c	Di-n-octyl phthalate	1.226	1.265	-3.2	114	0.00
86	Indeno(1,2,3-cd)pyrene	1.380	1.399	-1.4	113	0.00
87 I	Perylene-d12	1.000	1.000	0.0	112	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG121418\
Data File : BG038588.D
Acq On : 15 Dec 2018 4:20
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Dec 15 05:50:19 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG121218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 12 15:26:27 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.098	1.128	-2.7	113	0.00
89	Benzo(k)fluoranthene	1.093	1.112	-1.7	111	0.00
90 C	Benzo(a)pyrene	1.059	1.099	-3.8	114	0.00
91	Dibenzo(a,h)anthracene	1.055	1.091	-3.4	113	0.00
92	Benzo(q,h,i)perylene	1.039	1.061	-2.1	112	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0