

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121522\
 Data File : BG055975.D
 Acq On : 15 Dec 2022 13:38
 Operator : CG/JU
 Sample : PB149554BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB149554BL

Quant Time: Dec 16 04:01:24 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG121322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 14 02:24:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.372	152	6116	20.000	ng	0.00
21) Naphthalene-d8	11.204	136	23899	20.000	ng	# 0.00
39) Acenaphthene-d10	14.982	164	21467	20.000	ng	0.00
64) Phenanthrene-d10	17.720	188	66195	20.000	ng	# 0.00
76) Chrysene-d12	22.027	240	69966	20.000	ng	0.00
86) Perylene-d12	25.523	264	85436	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.887	112	44242	169.646	ng	0.00
7) Phenol-d6	7.503	99	53373	159.126	ng	0.00
23) Nitrobenzene-d5	9.547	82	34095	99.696	ng	0.00
42) 2,4,6-Tribromophenol	16.457	330	83997	162.271	ng	0.00
45) 2-Fluorobiphenyl	13.613	172	141585	92.500	ng	0.00
79) Terphenyl-d14	20.294	244	402872	101.768	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121522\
Data File : BG055975.D
Acq On : 15 Dec 2022 13:38
Operator : CG/JU
Sample : PB149554BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB149554BL

Quant Time: Dec 16 04:01:24 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG121322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 14 02:24:06 2022
Response via : Initial Calibration

