

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121522\
 Data File : BG055981.D
 Acq On : 15 Dec 2022 17:48
 Operator : CG/JU
 Sample : N6048-04
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 TP-R

Quant Time: Dec 16 04:04:52 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG121322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 14 02:24:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.378	152	8204	20.000	ng	0.00
21) Naphthalene-d8	11.204	136	32263	20.000	ng	# 0.00
39) Acenaphthene-d10	14.982	164	27602	20.000	ng	0.00
64) Phenanthrene-d10	17.714	188	69811	20.000	ng	0.00
76) Chrysene-d12	22.021	240	72394	20.000	ng	0.00
86) Perylene-d12	25.523	264	88688	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.893	112	48245	137.912	ng	0.00
7) Phenol-d6	7.503	99	60004	133.365	ng	0.00
23) Nitrobenzene-d5	9.547	82	38110	82.547	ng	0.00
42) 2,4,6-Tribromophenol	16.457	330	82677	124.220	ng	0.00
45) 2-Fluorobiphenyl	13.613	172	152448	77.460	ng	0.00
79) Terphenyl-d14	20.294	244	325125	79.374	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121522\
Data File : BG055981.D
Acq On : 15 Dec 2022 17:48
Operator : CG/JU
Sample : N6048-04
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
TP-R

Quant Time: Dec 16 04:04:52 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG121322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 14 02:24:06 2022
Response via : Initial Calibration

