

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG121618\
 Data File : BG038638.D
 Acq On : 16 Dec 2018 13:25
 Operator : JU/SJ
 Sample : PB115713BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB115713BL

Quant Time: Dec 17 06:00:13 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG121218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 12 15:26:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.06	152	22194	20.00	ng	0.00
21) Naphthalene-d8	10.88	136	89631	20.00	ng	0.00
39) Acenaphthene-d10	14.69	164	60509	20.00	ng	0.00
64) Phenanthrene-d10	17.44	188	156371	20.00	ng	0.00
76) Chrysene-d12	21.72	240	163226	20.00	ng	0.00
87) Perylene-d12	25.01	264	173780	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.61	112	137474	109.86	ng	0.00
7) Phenol-d6	7.21	99	185023	107.71	ng	0.00
23) Nitrobenzene-d5	9.23	82	120065	68.43	ng	0.00
42) 2,4,6-Tribromophenol	16.18	330	82924	99.44	ng	0.00
45) 2-Fluorobiphenyl	13.32	172	289956	65.74	ng	0.00
79) Terphenyl-d14	20.04	244	569058	75.87	ng	0.00


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG121618\
Data File : BG038638.D
Acq On : 16 Dec 2018 13:25
Operator : JU/SJ
Sample : PB115713BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB115713BL

Quant Time: Dec 17 06:00:13 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG121218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 12 15:26:27 2018
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.06 min Scan# 846
Delta R.T. -0.00 min
Lab File: BG038638.D
Acq: 16 Dec 2018 13:25

Instrument :
BNA_G
ClientSampleId :
PB115713BL

Tot Ion:152 Resp: 22194
Ion Ratio Lower Upper
152 100
150 158.7 119.5 179.3
115 61.4 49.6 74.4




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 109.863 ng
RT: 5.61 min Scan# 429
Delta R.T. 0.01 min
Lab File: BG038638.D
Acq: 16 Dec 2018 13:25

Tgt Ion:112 Resp: 137474
Ion Ratio Lower Upper
112 100
64 68.0 53.4 80.2
63 38.4 29.3 43.9



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BG
Ion 63.00 (62.70 to 63.70): BG





#7

Phenol-d6

Concen: 107.708 ng

RT: 7.21 min Scan# 702

Delta R.T. 0.00 min

Lab File: BG038638.D

Acq: 16 Dec 2018 13:25

Instrument :

BNA_G

ClientSampleId :

PB115713BL


Tot Ion: 99 Resp: 185023

Ion Ratio Lower Upper




99 100

42 26.1 18.5 27.7

71 40.6 31.1 46.7



Abundance Ion 99.00 (98.70 to 99.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 71.00 (70.70 to 71.70): BGC



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 10.88 min Scan# 1326

Delta R.T. -0.00 min

Lab File: BG038638.D

Acq: 16 Dec 2018 13:25

Tgt Ion: 136 Resp: 89631


Ion Ratio Lower Upper

136 100



137 12.8 9.3 13.9


54 10.3 8.5 12.7

68 5.0 3.5 5.3



Abundance Ion 136.00 (135.70 to 136.70): BGC
Ion 137.00 (136.70 to 137.70): BGC
Ion 54.00 (53.70 to 54.70): BGC
Ion 68.00 (67.70 to 68.70): BGC









#23
Nitrobenzene-d5
Concen: 68.434 ng
RT: 9.23 min Scan# 1046
Delta R.T. -0.01 min
Lab File: BG038638.D
Acq: 16 Dec 2018 13:25




Instrument :
BNA_G
ClientSampleId :
PB115713BL


Tot Ion: 82 Resp: 120065
Ion Ratio Lower Upper
82 100
128 41.8 34.6 52.0
54 64.6 50.6 76.0



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.69 min Scan# 1975
Delta R.T. -0.01 min
Lab File: BG038638.D
Acq: 16 Dec 2018 13:25

Tgt Ion:164 Resp: 60509
Ion Ratio Lower Upper
164 100
162 103.2 86.6 130.0
160 42.1 34.2 51.2





#42

2,4,6-Tribromophenol

Concen: 99.438 ng

RT: 16.18 min Scan# 2229

Delta R.T. -0.01 min

Lab File: BG038638.D

Acq: 16 Dec 2018 13:25

Instrument :

BNA_G

ClientSampleId :

PB115713BL


Tot Ion:330 Resp: 82924

Ion Ratio Lower Upper

330 100

332 94.2 76.9 115.3

141 36.6 28.8 43.2






Abundance

Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E



#45

2-Fluorobiphenyl

Concen: 65.738 ng

RT: 13.32 min Scan# 1741

Delta R.T. -0.00 min

Lab File: BG038638.D

Acq: 16 Dec 2018 13:25


Tgt Ion:172 Resp: 289956

Ion Ratio Lower Upper

172 100

171 35.6 29.8 44.6

170 25.0 19.5 29.3






Abundance

Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.44 min Scan# 2443
Delta R.T. -0.00 min
Lab File: BG038638.D
Acq: 16 Dec 2018 13:25

Instrument :
BNA_G
ClientSampleId :
PB115713BL

Tot Ion:188 Resp: 156371
Ion Ratio Lower Upper
188 100
94 6.2 5.1 7.7
80 7.4 6.1 9.1




Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BG
Ion 80.00 (79.70 to 80.70): BG






#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.72 min Scan# 3172
Delta R.T. -0.00 min
Lab File: BG038638.D
Acq: 16 Dec 2018 13:25

Tgt Ion:240 Resp: 163226
Ion Ratio Lower Upper
240 100
120 7.1 5.3 7.9
236 26.5 21.4 32.2



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E





#79

Terphenyl-d14

Concen: 75.874 ng

RT: 20.04 min Scan# 2886

Delta R.T. -0.00 min

Lab File: BG038638.D

Acq: 16 Dec 2018 13:25

Instrument :

BNA_G

ClientSampleId :

PB115713BL


Tot Ion:244 Resp: 569058

Ion Ratio Lower Upper

244 100

212 7.3 6.2 9.2




122 7.2 6.3 9.5



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#87

Perylene-d12

Concen: 20.000 ng

RT: 25.01 min Scan# 3732

Delta R.T. -0.01 min

Lab File: BG038638.D

Acq: 16 Dec 2018 13:25


Tgt Ion:264 Resp: 173780

Ion Ratio Lower Upper

264 100

260 21.5 18.9 28.3

265 19.7 17.0 25.4



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E

