

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121720\
 Data File : BG047767.D
 Acq On : 17 Dec 2020 15:12
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG121720

Quant Time: Dec 17 15:57:16 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG121720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 17 15:55:16 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	100	0.00
2	1,4-Dioxane	40.000	33.485	16.3	106	0.00
3	Pyridine	40.000	36.142	9.6	103	0.00
4	n-Nitrosodimethylamine	40.000	37.084	7.3	106	0.00
5 S	2-Fluorophenol	80.000	71.157	11.1	106	0.00
6	Aniline	40.000	35.878	10.3	107	0.00
7 S	Phenol-d6	80.000	71.366	10.8	103	0.00
8	2-Chlorophenol	40.000	35.657	10.9	104	0.00
9	Benzaldehyde	40.000	33.763	15.6	106	0.00
10 C	Phenol	40.000	36.017	10.0	103	0.00
11	bis(2-Chloroethyl)ether	40.000	35.134	12.2	106	0.00
12	1,3-Dichlorobenzene	40.000	35.068	12.3	103	0.00
13 C	1,4-Dichlorobenzene	40.000	34.720	13.2	103	0.00
14	1,2-Dichlorobenzene	40.000	34.597	13.5	104	0.00
15	Benzyl Alcohol	40.000	37.636	5.9	107	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	34.056	14.9	103	0.00
17	2-Methylphenol	40.000	34.368	14.1	103	0.00
18	Hexachloroethane	40.000	36.138	9.7	104	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	35.016	12.5	103	0.00
20	3+4-Methylphenols	40.000	35.432	11.4	105	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	104	0.00
22	Acetophenone	40.000	36.289	9.3	108	0.00
23 S	Nitrobenzene-d5	80.000	70.588	11.8	104	0.00
24	Nitrobenzene	40.000	35.317	11.7	105	0.00
25	Isophorone	40.000	35.989	10.0	107	0.00
26 C	2-Nitrophenol	40.000	36.068	9.8	105	0.00
27	2,4-Dimethylphenol	40.000	34.131	14.7	104	0.00
28	bis(2-Chloroethoxy)methane	40.000	35.233	11.9	109	0.00
29 C	2,4-Dichlorophenol	40.000	34.854	12.9	104	0.00
30	1,2,4-Trichlorobenzene	40.000	36.100	9.7	103	0.00
31	Naphthalene	40.000	35.656	10.9	104	0.00
32	Benzoic acid	40.000	21.617	46.0#	117	0.00
33	4-Chloroaniline	40.000	35.241	11.9	101	0.00
34 C	Hexachlorobutadiene	40.000	35.654	10.9	106	0.00
35	Caprolactam	40.000	35.716	10.7	109	0.00
36 C	4-Chloro-3-methylphenol	40.000	35.923	10.2	105	0.00
37	2-Methylnaphthalene	40.000	36.213	9.5	107	0.00
38	1-Methylnaphthalene	40.000	35.475	11.3	105	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	112	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	34.119	14.7	106	0.00
41 P	Hexachlorocyclopentadiene	40.000	36.789	8.0	103	0.00
42 S	2,4,6-Tribromophenol	80.000	72.616	9.2	110	0.00
43 C	2,4,6-Trichlorophenol	40.000	35.397	11.5	109	0.00
44	2,4,5-Trichlorophenol	40.000	35.704	10.7	109	0.00
45 S	2-Fluorobiphenyl	80.000	69.187	13.5	106	0.00
46	1,1'-Biphenyl	40.000	34.623	13.4	107	0.00
47	2-Chloronaphthalene	40.000	34.439	13.9	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121720\
 Data File : BG047767.D
 Acq On : 17 Dec 2020 15:12
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG121720

Quant Time: Dec 17 15:57:16 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG121720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 17 15:55:16 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	37.189	7.0	113	0.00
49	Acenaphthylene	40.000	35.195	12.0	107	0.00
50	Dimethylphthalate	40.000	34.607	13.5	106	0.00
51	2,6-Dinitrotoluene	40.000	35.561	11.1	110	0.00
52 C	Acenaphthene	40.000	35.850	10.4	107	0.00
53	3-Nitroaniline	40.000	35.467	11.3	107	0.00
54 P	2,4-Dinitrophenol	40.000	37.462	6.3	111	0.00
55	Dibenzofuran	40.000	34.672	13.3	108	0.00
56 P	4-Nitrophenol	40.000	38.457	3.9	115	0.00
57	2,4-Dinitrotoluene	40.000	34.709	13.2	109	0.00
58	Fluorene	40.000	34.147	14.6	107	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	35.547	11.1	109	0.00
60	Diethylphthalate	40.000	34.555	13.6	107	0.00
61	4-Chlorophenyl-phenylether	40.000	35.209	12.0	109	0.00
62	4-Nitroaniline	40.000	34.993	12.5	106	0.00
63	Azobenzene	40.000	34.999	12.5	108	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	106	0.00
65	4,6-Dinitro-2-methylphenol	40.000	36.214	9.5	107	0.00
66 c	n-Nitrosodiphenylamine	40.000	34.628	13.4	104	0.00
67	4-Bromophenyl-phenylether	40.000	34.651	13.4	103	0.00
68	Hexachlorobenzene	40.000	36.132	9.7	108	0.00
69	Atrazine	40.000	35.954	10.1	109	0.00
70 C	Pentachlorophenol	40.000	39.734	0.7	114	0.00
71	Phenanthrene	40.000	34.749	13.1	107	0.00
72	Anthracene	40.000	34.354	14.1	106	0.00
73	Carbazole	40.000	34.347	14.1	108	0.00
74	Di-n-butylphthalate	40.000	34.951	12.6	106	0.00
75 C	Fluoranthene	40.000	34.511	13.7	107	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	107	0.00
77	Benzydine	40.000	35.460	11.3	98	0.00
78	Pyrene	40.000	34.243	14.4	106	0.00
79 S	Terphenyl-d14	80.000	68.543	14.3	105	0.00
80	Butylbenzylphthalate	40.000	35.804	10.5	109	0.00
81	Benzo(a)anthracene	40.000	34.943	12.6	106	0.00
82	3,3'-Dichlorobenzidine	40.000	34.813	13.0	104	0.00
83	Chrysene	40.000	33.904	15.2	104	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	35.335	11.7	107	0.00
85 c	Di-n-octyl phthalate	40.000	35.206	12.0	107	0.00
86 I	Perylene-d12	20.000	20.000	0.0	106	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	35.038	12.4	104	0.02
88	Benzo(b)fluoranthene	40.000	35.480	11.3	105	0.00
89	Benzo(k)fluoranthene	40.000	35.112	12.2	107	0.00
90 C	Benzo(a)pyrene	40.000	35.049	12.4	105	0.00
91	Dibenzo(a,h)anthracene	40.000	34.734	13.2	104	0.00
92	Benzo(g,h,i)perylene	40.000	34.448	13.9	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121720\
Data File : BG047767.D
Acq On : 17 Dec 2020 15:12
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG121720

Quant Time: Dec 17 15:57:16 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG121720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 17 15:55:16 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0