

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121720\
 Data File : BG047768.D
 Acq On : 17 Dec 2020 16:33
 Operator : CG/JU
 Sample : PB133427BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB133427BL

Quant Time: Dec 17 17:10:10 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG121720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 17 15:55:16 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.200	152	30749	20.00	ng	0.00
21) Naphthalene-d8	11.026	136	117757	20.00	ng	# 0.00
39) Acenaphthene-d10	14.822	164	92461	20.00	ng	0.00
64) Phenanthrene-d10	17.560	188	238408	20.00	ng	# 0.00
76) Chrysene-d12	21.831	240	271576	20.00	ng	# 0.00
86) Perylene-d12	25.180	264	271304	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.733	112	258180	145.25	ng	0.00
7) Phenol-d6	7.343	99	343788	132.31	ng	0.00
23) Nitrobenzene-d5	9.375	82	269795	84.27	ng	0.00
42) 2,4,6-Tribromophenol	16.303	330	199702	126.52	ng	0.00
45) 2-Fluorobiphenyl	13.453	172	587980	82.29	ng	0.00
79) Terphenyl-d14	20.145	244	1208905	73.35	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.641	88	71	0.09	ng	# 1
3) Pyridine	3.976	79	56	0.03	ng	# 20
4) n-Nitrosodimethylamine	3.888	42	119	0.08	ng	# 1
6) Aniline	7.501	93	88	0.03	ng	# 45
9) Benzaldehyde	7.337	77	232	0.14	ng	# 1
10) Phenol	7.677	94	184	0.08	ng	# 44
11) bis(2-Chloroethyl)ether	7.501	93	88	0.05	ng	# 20
12) 1,3-Dichlorobenzene	8.253	146	69	0.03	ng	# 28
13) 1,4-Dichlorobenzene	8.253	146	69	0.03	ng	# 25
14) 1,2-Dichlorobenzene	8.641	146	61	0.03	ng	# 24
15) Benzyl Alcohol	8.195	79	1789	0.78	ng	# 43
16) 2,2'-oxybis(1-Chloropr...	8.835	45	54	0.03	ng	69
17) 2-Methylphenol	8.835	107	88	0.05	ng	# 13
18) Hexachloroethane	9.440	117	68	0.07	ng	# 1
20) 3+4-Methylphenols	8.835	107	88	0.04	ng	# 20
22) Acetophenone	8.864	105	66	0.02	ng	# 52
24) Nitrobenzene	9.364	77	963	0.32	ng	# 36
25) Isophorone	9.916	82	90	0.02	ng	# 63
27) 2,4-Dimethylphenol	10.410	122	84	0.05	ng	# 4
28) bis(2-Chloroethoxy)met...	10.368	93	78	0.03	ng	# 49
29) 2,4-Dichlorophenol	10.897	162	72	0.03	ng	# 26
30) 1,2,4-Trichlorobenzene	10.921	180	65	0.02	ng	# 12
31) Naphthalene	11.379	128	61	0.01	ng	# 68
32) Benzoic acid	10.410	122	84	0.09	ng	# 1
33) 4-Chloroaniline	11.214	127	82	0.03	ng	# 47
35) Caprolactam	12.019	113	71	0.10	ng	# 1
36) 4-Chloro-3-methylphenol	12.360	107	61	0.02	ng	# 15
37) 2-Methylnaphthalene	12.395	142	82	0.02	ng	# 15
43) 2,4,6-Trichlorophenol	13.306	196	63	0.03	ng	# 13
44) 2,4,5-Trichlorophenol	13.306	196	63	0.03	ng	# 17
46) 1,1'-Biphenyl	13.665	154	537	0.08	ng	# 20
47) 2-Chloronaphthalene	13.888	162	54	0.01	ng	# 40
48) 2-Nitroaniline	13.794	65	56	0.03	ng	# 1
49) Acenaphthylene	14.857	152	119	0.01	ng	# 1
50) Dimethylphthalate	14.405	163	60	0.01	ng	# 75
52) Acenaphthene	14.828	154	226	0.04	ng	# 50

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121720\
 Data File : BG047768.D
 Acq On : 17 Dec 2020 16:33
 Operator : CG/JU
 Sample : PB133427BL
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB133427BL

Quant Time: Dec 17 17:10:10 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG121720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 17 15:55:16 2020
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
53) 3-Nitroaniline	14.869	138	76	0.05	ng	# 3
54) 2,4-Dinitrophenol	14.981	184	57	0.05	ng	# 31
55) Dibenzofuran	15.474	168	74	0.01	ng	# 46
56) 4-Nitrophenol	14.998	139	59	0.05	ng	# 11
58) Fluorene	16.303	166	3750	0.51	ng	# 95
59) 2,3,4,6-Tetrachlorophenol	15.251	232	85	0.03	ng	# 46
60) Diethylphthalate	15.815	149	76	0.01	ng	# 54
61) 4-Chlorophenyl-phenyle...	15.856	204	74	0.02	ng	# 32
62) 4-Nitroaniline	15.715	138	56	0.03	ng	# 21
63) Azobenzene	16.138	77	55	0.01	ng	# 46
65) 4,6-Dinitro-2-methylph...	16.297	198	475	0.31	ng	# 1
66) n-Nitrosodiphenylamine	16.303	169	7091	1.04	ng	# 42
69) Atrazine	17.061	200	81	0.03	ng	# 35
71) Phenanthrene	17.560	178	201	0.02	ng	# 1
72) Anthracene	17.560	178	201	0.02	ng	# 1
73) Carbazole	17.965	167	55	0.01	ng	# 58
74) Di-n-butylphthalate	18.430	149	56	0.00	ng	# 77
75) Fluoranthene	19.564	202	135	0.01	ng	# 63
78) Pyrene	20.145	202	5529	0.28	ng	# 67
80) Butylbenzylphthalate	20.833	149	66	0.01	ng	# 27
81) Benzo(a)anthracene	21.831	228	656	0.03	ng	# 90
82) 3,3'-Dichlorobenzidine	21.814	252	53	0.01	ng	# 23
83) Chrysene	21.831	228	656	0.04	ng	# 92
84) Bis(2-ethylhexyl)phtha...	21.696	149	103	0.01	ng	# 60
85) Di-n-octyl phthalate	22.960	149	158	0.01	ng	# 1
87) Indeno(1,2,3-cd)pyrene	29.035	276	61	0.00	ng	# 53
88) Benzo(b)fluoranthene	24.111	252	62	0.00	ng	# 59
89) Benzo(k)fluoranthene	24.223	252	54	0.00	ng	# 1
90) Benzo(a)pyrene	25.057	252	90	0.01	ng	# 60
91) Dibenzo(a,h)anthracene	29.064	278	106	0.01	ng	# 57


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG121720\
Data File : BG047768.D
Acq On : 17 Dec 2020 16:33
Operator : CG/JU
Sample : PB133427BL
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB133427BL

Quant Time: Dec 17 17:10:10 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG121720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 17 15:55:16 2020
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 8.200 min Scan# 827
Delta R.T. -0.003 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion:152 Resp: 30749
Ion Ratio Lower Upper
152 100
150 161.1 125.0 187.6
115 62.0 41.8 62.6



#2
1,4-Dioxane
Concen: 0.09 ng
RT: 3.641 min Scan# 51
Delta R.T. 0.079 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion: 88 Resp: 71
Ion Ratio Lower Upper
88 100
58 208.5 61.6 92.4#
43 0.0 21.3 31.9#









#3
Pyridine
Concen: 0.03 ng
RT: 3.976 min Scan# 108
Delta R.T. -0.003 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion: 79	Resp: 56
Ion Ratio	Lower Upper
79 100	
52 0.0	61.3 91.9#
51 0.0	28.4 42.6#



#4
n-Nitrosodimethylamine
Concen: 0.08 ng
RT: 3.888 min Scan# 93
Delta R.T. 0.003 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion: 42	Resp: 119
Ion Ratio	Lower Upper
42 100	
74 0.0	120.4 180.6#
44 0.0	6.2 9.2#









#5
2-Fluorophenol
Concen: 145.25 ng
RT: 5.733 min Scan# 407
Delta R.T. 0.003 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion:112	Resp:	258180
Ion Ratio	Lower	Upper
112	100	
64	64.4	46.7 70.1
63	44.9	22.8 34.2#



#6
Aniline
Concen: 0.03 ng
RT: 7.501 min Scan# 708
Delta R.T. -0.020 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion: 93	Resp:	88
Ion Ratio	Lower	Upper
93	100	
66	0.0	29.5 44.3#
65	0.0	15.5 23.3#









#7
Phenol-d6
Concen: 132.31 ng
RT: 7.343 min Scan# 681
Delta R.T. -0.003 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion: 99 Resp: 343788
Ion Ratio Lower Upper
99 100
42 25.6 12.8 19.2#
71 63.2 24.6 36.8#



#9
Benzaldehyde
Concen: 0.14 ng
RT: 7.337 min Scan# 680
Delta R.T. 0.009 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion: 77 Resp: 232
Ion Ratio Lower Upper
77 100
105 0.0 85.9 125.9#
106 0.0 80.9 120.9#









#10
Phenol
Concen: 0.08 ng
RT: 7.677 min Scan# 738
Delta R.T. 0.309 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion: 94	Resp:	184
Ion	Ratio	Lower Upper
94	100	
65	0.0	6.6 46.6#
66	0.0	13.9 53.9#



#11
bis(2-Chloroethyl)ether
Concen: 0.05 ng
RT: 7.501 min Scan# 708
Delta R.T. -0.109 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion: 93	Resp:	88
Ion	Ratio	Lower Upper
93	100	
63	0.0	56.4 96.4#
95	0.0	13.8 53.8#









#12
1,3-Dichlorobenzene
Concen: 0.03 ng
RT: 8.253 min Scan# 836
Delta R.T. 0.162 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion:146 Resp: 69
Ion Ratio Lower Upper
146 100
148 0.0 50.1 75.1#
75 0.0 20.1 30.1#



#13
1,4-Dichlorobenzene
Concen: 0.03 ng
RT: 8.253 min Scan# 836
Delta R.T. 0.015 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion:146 Resp: 69
Ion Ratio Lower Upper
146 100
148 0.0 51.1 76.7#
111 0.0 29.7 44.5#









#14
1,2-Dichlorobenzene
Concen: 0.03 ng
RT: 8.641 min Scan# 902
Delta R.T. 0.079 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion:146 Resp: 61
Ion Ratio Lower Upper
146 100
148 0.0 53.0 79.6#
111 0.0 30.7 46.1#



#15
Benzyl Alcohol
Concen: 0.78 ng
RT: 8.195 min Scan# 826
Delta R.T. -0.244 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion: 79 Resp: 1789
Ion Ratio Lower Upper
79 100
108 0.0 65.5 98.3#
77 71.8 50.1 75.1









#16
2,2'-oxybis(1-Chloropropane)
Concen: 0.03 ng
RT: 8.835 min Scan# 935
Delta R.T. 0.109 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion:	45	Resp:	54
Ion	Ratio	Lower	Upper
45	100		
77	0.0	0.0	33.4
79	0.0	0.0	30.8



#17
2-Methylphenol
Concen: 0.05 ng
RT: 8.835 min Scan# 935
Delta R.T. 0.197 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33


Tgt Ion:	107	Resp:	88
Ion	Ratio	Lower	Upper
107	100		
108	0.0	87.2	130.8#
77	0.0	33.8	50.8#
79	0.0	32.7	49.1#








#18
Hexachloroethane
Concen: 0.07 ng
RT: 9.440 min Scan# 1038
Delta R.T. 0.138 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Instrument :
BNA_G
ClientSampleId :
PB133427BL







Tgt Ion:117 Resp: 68
Ion Ratio Lower Upper
117 100
119 0.0 77.5 116.3#
201 0.0 104.9 157.3#



#20
3+4-Methylphenols
Concen: 0.04 ng
RT: 8.835 min Scan# 935
Delta R.T. -0.132 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion:107 Resp: 88
Ion Ratio Lower Upper
107 100
108 0.0 70.6 110.6#
77 0.0 10.2 50.2#
79 0.0 6.8 46.8#









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 11.026 min Scan# 1308
Delta R.T. -0.003 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Instrument :
BNA_G
ClientSampleId :
PB133427BL

Tgt Ion:136 Resp: 117757




Ion	Ratio	Lower	Upper
136	100		
137	11.8	9.1	13.7
54	5.4	6.2	9.4#
68	2.3	5.0	7.4#




#22
Acetophenone
Concen: 0.02 ng
RT: 8.864 min Scan# 940
Delta R.T. -0.161 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion:105 Resp: 66

Ion	Ratio	Lower	Upper
105	100		
71	0.0	2.7	4.1#
51	0.0	21.0	31.6#
120	0.0	19.3	28.9#









#23
Nitrobenzene-d5
Concen: 84.27 ng
RT: 9.375 min Scan# 1027
Delta R.T. 0.003 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion: 82 Resp: 269795
Ion Ratio Lower Upper
82 100
128 31.5 37.0 55.6#
54 51.0 39.0 58.4



#24
Nitrobenzene
Concen: 0.32 ng
RT: 9.364 min Scan# 1025
Delta R.T. -0.056 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion: 77 Resp: 963
Ion Ratio Lower Upper
77 100
123 0.0 39.7 59.5#
65 0.0 10.4 15.6#









#25
Isophorone
Concen: 0.02 ng
RT: 9.916 min Scan# 1119
Delta R.T. -0.020 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion: 82	Resp: 90
Ion Ratio	Lower Upper
82 100	
95 0.0	5.6 8.4#
138 0.0	15.8 23.6#



#27
2,4-Dimethylphenol
Concen: 0.05 ng
RT: 10.410 min Scan# 1203
Delta R.T. 0.232 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion: 122	Resp: 84
Ion Ratio	Lower Upper
122 100	
107 0.0	89.2 133.8#
121 0.0	46.9 70.3#









#28
bis(2-Chloroethoxy)methane
Concen: 0.03 ng
RT: 10.368 min Scan# 1196
Delta R.T. -0.050 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion: 93 Resp: 78
Ion Ratio Lower Upper
93 100
95 0.0 26.5 39.7#
123 0.0 9.4 14.0#



#29
2,4-Dichlorophenol
Concen: 0.03 ng
RT: 10.897 min Scan# 1286
Delta R.T. 0.226 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33


Tgt Ion:162 Resp: 72
Ion Ratio Lower Upper
162 100
164 0.0 45.6 85.6#
98 0.0 11.0 51.0#








#30
1,2,4-Trichlorobenzene
Concen: 0.02 ng
RT: 10.921 min Scan# 1290
Delta R.T. 0.032 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33


Instrument :
BNA_G
ClientSampleId :
PB133427BL





Tgt Ion:180 Resp: 65
Ion Ratio Lower Upper
180 100
182 0.0 76.9 115.3#
145 0.0 21.4 32.0#




#31
Naphthalene
Concen: 0.01 ng
RT: 11.379 min Scan# 1368
Delta R.T. 0.297 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33



Tgt Ion:128 Resp: 61
Ion Ratio Lower Upper
128 100
129 0.0 9.0 13.6#
127 0.0 11.0 16.4#









#32
Benzoic acid
Concen: 0.09 ng
RT: 10.410 min Scan# 1203
Delta R.T. 0.109 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion:122	Resp:	84
Ion	Ratio	Lower Upper
122	100	
105	0.0	103.7 143.7#
77	0.0	63.0 103.0#



#33
4-Chloroaniline
Concen: 0.03 ng
RT: 11.214 min Scan# 1340
Delta R.T. 0.038 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33


Tgt Ion:127	Resp:	82
Ion	Ratio	Lower Upper
127	100	
129	0.0	26.8 40.2#
65	0.0	23.8 35.6#
92	0.0	14.5 21.7#








#35
Caprolactam
Concen: 0.10 ng
RT: 12.019 min Scan# 1477
Delta R.T. 0.074 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33


Instrument :
BNA_G
ClientSampleId :
PB133427BL





Tgt Ion:113 Resp: 71
Ion Ratio Lower Upper
113 100
55 0.0 149.2 189.2#
56 0.0 107.2 147.2#




#36
4-Chloro-3-methylphenol
Concen: 0.02 ng
RT: 12.360 min Scan# 1535
Delta R.T. 0.079 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33



Tgt Ion:107 Resp: 61
Ion Ratio Lower Upper
107 100
144 0.0 22.5 33.7#
142 0.0 70.8 106.2#









#37
2-Methylnaphthalene
Concen: 0.02 ng
RT: 12.395 min Scan# 1541
Delta R.T. -0.273 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion:142	Resp:	82
Ion Ratio	Lower	Upper
142	100	
141	0.0	71.6 107.4#
115	0.0	25.4 38.2#



#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.822 min Scan# 1954
Delta R.T. -0.003 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion:164	Resp:	92461
Ion Ratio	Lower	Upper
164	100	
162	100.1	82.3 123.5
160	44.7	35.3 52.9









#42
2,4,6-Tribromophenol
Concen: 126.52 ng
RT: 16.303 min Scan# 2206
Delta R.T. -0.003 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion:330 Resp: 199702
Ion Ratio Lower Upper
330 100
332 96.5 76.5 114.7
141 37.7 26.0 39.0



#43
2,4,6-Trichlorophenol
Concen: 0.03 ng
RT: 13.306 min Scan# 1696
Delta R.T. 0.044 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion:196 Resp: 63
Ion Ratio Lower Upper
196 100
198 0.0 74.9 112.3#
200 0.0 25.3 37.9#









#44
2,4,5-Trichlorophenol
Concen: 0.03 ng
RT: 13.306 min Scan# 1696
Delta R.T. -0.026 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion:196	Resp:	63
Ion Ratio	Lower	Upper
196	100	
198	0.0	77.6 116.4#
97	0.0	28.6 43.0#
132	0.0	17.8 26.8#



#45
2-Fluorobiphenyl
Concen: 82.29 ng
RT: 13.453 min Scan# 1721
Delta R.T. 0.003 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion:172	Resp:	587980
Ion Ratio	Lower	Upper
172	100	
171	37.0	31.3 46.9
170	24.0	20.8 31.2









#46
1,1'-Biphenyl
Concen: 0.08 ng
RT: 13.665 min Scan# 1757
Delta R.T. 0.003 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion:	154	Resp:	537
Ion	Ratio	Lower	Upper
154	100		
153	102.4	22.6	62.6#
76	0.0	0.0	32.6



#47
2-Chloronaphthalene
Concen: 0.01 ng
RT: 13.888 min Scan# 1795
Delta R.T. 0.173 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion:	162	Resp:	54
Ion	Ratio	Lower	Upper
162	100		
127	0.0	27.8	41.6#
164	0.0	27.0	40.4#









#48
2-Nitroaniline
Concen: 0.03 ng
RT: 13.794 min Scan# 1779
Delta R.T. -0.109 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion: 65 Resp: 56
Ion Ratio Lower Upper
65 100
92 0.0 58.6 88.0#
138 0.0 98.2 147.4#



#49
Acenaphthylene
Concen: 0.01 ng
RT: 14.857 min Scan# 1960
Delta R.T. 0.309 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33


Tgt Ion: 152 Resp: 119
Ion Ratio Lower Upper
152 100
151 100.0 17.4 26.0#
153 0.0 11.5 17.3#








#50
Dimethylphthalate
Concen: 0.01 ng
RT: 14.405 min Scan# 1883
Delta R.T. 0.138 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Instrument :
BNA_G
ClientSampleId :
PB133427BL







Tgt Ion:163 Resp: 60
Ion Ratio Lower Upper
163 100
194 0.0 3.8 5.8#
164 0.0 8.8 13.2#



#52
Acenaphthene
Concen: 0.04 ng
RT: 14.828 min Scan# 1955
Delta R.T. -0.062 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion:154 Resp: 226
Ion Ratio Lower Upper
154 100
153 72.5 90.3 135.5#
152 0.0 43.8 65.8#









#53
3-Nitroaniline
Concen: 0.05 ng
RT: 14.869 min Scan# 1962
Delta R.T. 0.150 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion:138 Resp: 76
Ion Ratio Lower Upper
138 100
108 0.0 7.4 11.0#
92 0.0 85.2 127.8#



#54
2,4-Dinitrophenol
Concen: 0.05 ng
RT: 14.981 min Scan# 1981
Delta R.T. 0.056 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion:184 Resp: 57
Ion Ratio Lower Upper
184 100
63 0.0 32.3 48.5#
154 0.0 40.4 60.6#









#55
Dibenzofuran
Concen: 0.01 ng
RT: 15.474 min Scan# 2065
Delta R.T. 0.256 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Instrument :
BNA_G
ClientSampleId :
PB133427BL

Tgt Ion:168 Resp: 74




Ion	Ratio	Lower	Upper
168	100		
139	0.0	28.6	43.0#
169	0.0	11.0	16.6#




#56
4-Nitrophenol
Concen: 0.05 ng
RT: 14.998 min Scan# 1984
Delta R.T. -0.015 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion:139 Resp: 59

Ion	Ratio	Lower	Upper
139	100		
109	0.0	45.9	85.9#
65	0.0	65.4	105.4#









#58
Fluorene
Concen: 0.51 ng
RT: 16.303 min Scan# 2206
Delta R.T. 0.438 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion:166 Resp: 3750
Ion Ratio Lower Upper
166 100
165 96.9 77.2 115.8
167 28.4 11.4 17.2#



#59
2,3,4,6-Tetrachlorophenol
Concen: 0.03 ng
RT: 15.251 min Scan# 2027
Delta R.T. -0.191 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion:232 Resp: 85
Ion Ratio Lower Upper
232 100
131 0.0 26.6 40.0#
130 0.0 1.4 2.0#
166 0.0 21.0 31.6#









#60
Diethylphthalate
Concen: 0.01 ng
RT: 15.815 min Scan# 2123
Delta R.T. 0.197 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion:149 Resp: 76
Ion Ratio Lower Upper
149 100
177 0.0 21.4 32.2#
150 0.0 10.7 16.1#



#61
4-Chlorophenyl-phenylether
Concen: 0.02 ng
RT: 15.856 min Scan# 2130
Delta R.T. 0.003 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion:204 Resp: 74
Ion Ratio Lower Upper
204 100
206 0.0 27.8 41.8#
141 0.0 41.5 62.3#









#62
4-Nitroaniline
Concen: 0.03 ng
RT: 15.715 min Scan# 2106
Delta R.T. -0.161 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion:138 Resp: 56
Ion Ratio Lower Upper
138 100
92 0.0 26.9 66.9#
108 0.0 49.4 89.4#



#63
Azobenzene
Concen: 0.01 ng
RT: 16.138 min Scan# 2178
Delta R.T. -0.009 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion: 77 Resp: 55
Ion Ratio Lower Upper
77 100
182 0.0 14.0 54.0#
105 0.0 3.0 43.0#
51 0.0 8.1 48.1#









#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.560 min Scan# 2420
Delta R.T. -0.003 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion:188 Resp: 238408
Ion Ratio Lower Upper
188 100
94 3.3 6.3 9.5#
80 4.8 7.0 10.6#



#65
4,6-Dinitro-2-methylphenol
Concen: 0.31 ng
RT: 16.297 min Scan# 2205
Delta R.T. 0.361 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion:198 Resp: 475
Ion Ratio Lower Upper
198 100
51 112.6 10.4 50.4#
105 151.4 12.9 52.9#









#66
n-Nitrosodiphenylamine
Concen: 1.04 ng
RT: 16.303 min Scan# 2206
Delta R.T. 0.238 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion:169 Resp: 7091
Ion Ratio Lower Upper
169 100
168 7.2 55.5 83.3#
167 23.5 30.5 45.7#



#69
Atrazine
Concen: 0.03 ng
RT: 17.061 min Scan# 2335
Delta R.T. 0.062 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion:200 Resp: 81
Ion Ratio Lower Upper
200 100
173 0.0 4.2 44.2#
215 0.0 30.0 70.0#









#71
Phenanthrene
Concen: 0.02 ng
RT: 17.560 min Scan# 2420
Delta R.T. -0.044 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion:178 Resp: 201
Ion Ratio Lower Upper
178 100
176 0.0 16.7 25.1#
179 115.6 14.0 21.0#



#72
Anthracene
Concen: 0.02 ng
RT: 17.560 min Scan# 2420
Delta R.T. -0.138 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion:178 Resp: 201
Ion Ratio Lower Upper
178 100
176 0.0 16.4 24.6#
179 115.6 13.8 20.6#









#73
 Carbazole
 Concen: 0.01 ng
 RT: 17.965 min Scan# 2489
 Delta R.T. 0.009 min
 Lab File: BG047768.D
 Acq: 17 Dec 2020 16:33




Instrument :
 BNA_G
 ClientSampleId :
 PB133427BL


Tgt Ion:	167	Resp:	55
Ion	Ratio	Lower	Upper
167	100		
166	0.0	18.7	28.1#
139	0.0	9.6	14.4#



#74
 Di-n-butylphthalate
 Concen: 0.00 ng
 RT: 18.430 min Scan# 2568
 Delta R.T. -0.067 min
 Lab File: BG047768.D
 Acq: 17 Dec 2020 16:33

Tgt Ion:	149	Resp:	56
Ion	Ratio	Lower	Upper
149	100		
150	0.0	8.2	12.4#
104	0.0	3.8	5.8#









#75
Fluoranthene
Concen: 0.01 ng
RT: 19.564 min Scan# 2761
Delta R.T. -0.032 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion:	202	Resp:	135
Ion	Ratio	Lower	Upper
202	100		
101	0.0	0.0	29.7
203	0.0	0.0	39.2



#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.831 min Scan# 3147
Delta R.T. -0.003 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion:	240	Resp:	271576
Ion	Ratio	Lower	Upper
240	100		
120	3.0	6.2	9.4#
236	27.0	20.5	30.7









#78
Pyrene
Concen: 0.28 ng
RT: 20.145 min Scan# 2860
Delta R.T. 0.185 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion:	202	Resp:	5529
Ion Ratio	Lower	Upper	
202	100		
200	49.2	18.6	27.8#
203	15.1	15.3	22.9#



#79
Terphenyl-d14
Concen: 73.35 ng
RT: 20.145 min Scan# 2860
Delta R.T. -0.003 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion:	244	Resp:	1208905
Ion Ratio	Lower	Upper	
244	100		
212	8.8	6.7	10.1
122	5.1	7.8	11.6#









#80
Butylbenzylphthalate
Concen: 0.01 ng
RT: 20.833 min Scan# 2977
Delta R.T. 0.009 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion:149 Resp: 66
Ion Ratio Lower Upper
149 100
91 0.0 51.5 77.3#
206 0.0 21.2 31.8#



#81
Benzo(a)anthracene
Concen: 0.03 ng
RT: 21.831 min Scan# 3147
Delta R.T. 0.015 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33


Tgt Ion:228 Resp: 656
Ion Ratio Lower Upper
228 100
226 31.0 23.3 34.9
229 30.1 17.4 26.0#








#82
3,3'-Dichlorobenzidine
Concen: 0.01 ng
RT: 21.814 min Scan# 3144
Delta R.T. 0.091 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Instrument :
BNA_G
ClientSampleId :
PB133427BL







Tgt Ion:252 Resp: 53
Ion Ratio Lower Upper
252 100
254 0.0 53.2 79.8#
126 0.0 6.8 10.2#



#83
Chrysene
Concen: 0.04 ng
RT: 21.831 min Scan# 3147
Delta R.T. -0.056 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion:228 Resp: 656
Ion Ratio Lower Upper
228 100
226 31.0 25.0 37.4
229 30.1 17.3 25.9#









#84
Bis(2-ethylhexyl)phthalate
Concen: 0.01 ng
RT: 21.696 min Scan# 3124
Delta R.T. -0.003 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion:	149	Resp:	103
Ion	Ratio	Lower	Upper
149	100		
167	55.6	24.8	37.2#
279	0.0	4.6	7.0#



#85
Di-n-octyl phthalate
Concen: 0.01 ng
RT: 22.960 min Scan# 3339
Delta R.T. 0.009 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion:	149	Resp:	158
Ion	Ratio	Lower	Upper
149	100		
167	0.0	1.4	2.2#
43	62.0	6.8	10.2#









#86
Perylene-d12
Concen: 20.00 ng
RT: 25.180 min Scan# 3717
Delta R.T. -0.003 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion:264 Resp: 271304
Ion Ratio Lower Upper
264 100
260 23.4 18.8 28.2
265 23.4 17.8 26.8



#87
Indeno(1,2,3-cd)pyrene
Concen: 0.00 ng
RT: 29.035 min Scan# 4373
Delta R.T. 0.009 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion:276 Resp: 61
Ion Ratio Lower Upper
276 100
138 0.0 16.2 24.2#
277 0.0 20.3 30.5#









#88
Benzo(b)fluoranthene
Concen: 0.00 ng
RT: 24.111 min Scan# 3535
Delta R.T. -0.009 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33




Instrument :
BNA_G
ClientSampleId :
PB133427BL


Tgt Ion:252 Resp: 62
Ion Ratio Lower Upper
252 100
253 0.0 18.7 28.1#
125 0.0 6.5 9.7#



#89
Benzo(k)fluoranthene
Concen: 0.00 ng
RT: 24.223 min Scan# 3554
Delta R.T. 0.032 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion:252 Resp: 54
Ion Ratio Lower Upper
252 100
253 0.0 18.6 27.8#
125 177.8 6.8 10.2#









#90
Benzo(a)pyrene
Concen: 0.01 ng
RT: 25.057 min Scan# 3696
Delta R.T. 0.021 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Instrument :
BNA_G
ClientSampleId :
PB133427BL

Tgt Ion:252	Resp:	90
Ion Ratio	Lower	Upper
252	100	
253	0.0	17.9 26.9#
125	0.0	7.0 10.4#



#91
Dibenzo(a,h)anthracene
Concen: 0.01 ng
RT: 29.064 min Scan# 4378
Delta R.T. -0.038 min
Lab File: BG047768.D
Acq: 17 Dec 2020 16:33

Tgt Ion:278	Resp:	106
Ion Ratio	Lower	Upper
278	100	
139	0.0	10.5 15.7#
279	0.0	19.4 29.0#

