

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG121818\
 Data File : BG038709.D
 Acq On : 19 Dec 2018 1:57
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 19 07:30:47 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG121218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 12 15:26:27 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.525	0.447	14.9	84	0.00
3	Pyridine	1.445	1.262	12.7	73	0.00
4	n-Nitrosodimethylamine	0.708	0.666	5.9	86	0.00
5 S	2-Fluorophenol	1.128	1.101	2.4	97	0.00
6	Aniline	1.976	1.895	4.1	94	0.00
7 S	Phenol-d6	1.548	1.563	-1.0	101	0.00
8	2-Chlorophenol	1.305	1.251	4.1	95	0.00
9	Benzaldehyde	1.004	0.894	11.0	96	0.00
10 C	Phenol	1.645	1.642	0.2	99	0.00
11	bis(2-Chloroethyl)ether	1.309	1.185	9.5	92	0.00
12	1,3-Dichlorobenzene	1.543	1.480	4.1	97	0.00
13 C	1,4-Dichlorobenzene	1.580	1.485	6.0	95	0.00
14	1,2-Dichlorobenzene	1.490	1.408	5.5	97	-0.01
15	Benzyl Alcohol	1.208	1.220	-1.0	96	0.00
16	2,2'-oxybis(1-Chloropropane	2.999	2.550	15.0	86	0.00
17	2-Methylphenol	1.102	1.083	1.7	96	0.00
18	Hexachloroethane	0.577	0.515	10.7	90	0.00
19 P	n-Nitroso-di-n-propylamine	1.087	1.003	7.7	91	0.00
20	3+4-Methylphenols	1.522	1.550	-1.8	100	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
22	Acetophenone	0.500	0.464	7.2	97	0.00
23 S	Nitrobenzene-d5	0.391	0.355	9.2	95	0.00
24	Nitrobenzene	0.396	0.359	9.3	95	0.00
25	Isophorone	0.703	0.596	15.2	76	0.00
26 C	2-Nitrophenol	0.203	0.185	8.9	96	-0.01
27	2,4-Dimethylphenol	0.291	0.275	5.5	98	0.00
28	bis(2-Chloroethoxy)methane	0.397	0.355	10.6	92	-0.01
29 C	2,4-Dichlorophenol	0.323	0.337	-4.3	105	0.00
30	1,2,4-Trichlorobenzene	0.390	0.376	3.6	101	0.00
31	Naphthalene	0.985	0.931	5.5	99	0.00
32	Benzoic acid	0.154	0.148	3.9	91	0.00
33	4-Chloroaniline	0.428	0.430	-0.5	102	0.00
34 C	Hexachlorobutadiene	0.264	0.250	5.3	97	0.00
35	Caprolactam	0.134	0.123	8.2	101	0.00
36 C	4-Chloro-3-methylphenol	0.369	0.359	2.7	103	0.00
37	2-Methylnaphthalene	0.703	0.687	2.3	103	0.00
38	1-Methylnaphthalene	0.682	0.670	1.8	104	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	116	0.00
40	1,2,4,5-Tetrachlorobenzene	0.748	0.686	8.3	106	0.00
41 P	Hexachlorocyclopentadiene	0.411	0.358	12.9	99	0.00
42 S	2,4,6-Tribromophenol	0.276	0.279	-1.1	115	0.00
43 C	2,4,6-Trichlorophenol	0.460	0.437	5.0	104	0.00
44	2,4,5-Trichlorophenol	0.487	0.478	1.8	108	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG121818\
 Data File : BG038709.D
 Acq On : 19 Dec 2018 1:57
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 19 07:30:47 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG121218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 12 15:26:27 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.458	1.326	9.1	106	0.00
46	1,1'-Biphenyl	1.677	1.515	9.7	103	0.00
47	2-Chloronaphthalene	1.295	1.173	9.4	105	0.00
48	2-Nitroaniline	0.413	0.382	7.5	104	0.00
49	Acenaphthylene	1.936	1.775	8.3	105	0.00
50	Dimethylphthalate	1.633	1.473	9.8	103	0.00
51	2,6-Dinitrotoluene	0.370	0.345	6.8	107	0.00
52 C	Acenaphthene	1.158	1.068	7.8	107	0.00
53	3-Nitroaniline	0.377	0.361	4.2	108	0.00
54 P	2,4-Dinitrophenol	0.193	0.227	-17.6	132	0.00
55	Dibenzofuran	1.985	1.843	7.2	109	0.00
56 P	4-Nitrophenol	0.286	0.238	16.8	91	0.00
57	2,4-Dinitrotoluene	0.521	0.480	7.9	103	0.00
58	Fluorene	1.470	1.396	5.0	109	0.00
59	2,3,4,6-Tetrachlorophenol	0.438	0.406	7.3	103	0.00
60	Diethylphthalate	1.793	1.565	12.7	102	0.00
61	4-Chlorophenyl-phenylether	0.917	0.865	5.7	108	0.00
62	4-Nitroaniline	0.388	0.380	2.1	108	0.00
63	Azobenzene	1.498	1.356	9.5	104	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	118	0.00
65	4,6-Dinitro-2-methylphenol	0.143	0.118	17.5	96	0.00
66 c	n-Nitrosodiphenylamine	0.588	0.552	6.1	111	0.00
67	4-Bromophenyl-phenylether	0.244	0.232	4.9	111	0.00
68	Hexachlorobenzene	0.253	0.239	5.5	109	0.00
69	Atrazine	0.218	0.208	4.6	109	0.00
70 C	Pentachlorophenol	0.150	0.136	9.3	104	0.00
71	Phenanthrene	1.037	0.967	6.8	111	0.00
72	Anthracene	1.024	0.953	6.9	109	0.00
73	Carbazole	1.013	0.941	7.1	109	0.00
74	Di-n-butylphthalate	1.162	1.040	10.5	104	0.00
75 C	Fluoranthene	1.283	1.165	9.2	110	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	120	0.00
77	Benzidine	0.591	0.511	13.5	86	0.00
78	Pyrene	1.321	1.175	11.1	109	0.00
79 S	Terphenyl-d14	0.919	0.849	7.6	111	0.00
80	Butylbenzylphthalate	0.516	0.460	10.9	104	0.00
81	Benzo(a)anthracene	1.219	1.132	7.1	109	0.00
82	3,3'-Dichlorobenzidine	0.483	0.452	6.4	106	0.00
83	Chrysene	1.174	1.065	9.3	108	0.00
84	Bis(2-ethylhexyl)phthalate	0.732	0.648	11.5	102	0.00
85 c	Di-n-octyl phthalate	1.226	1.078	12.1	100	0.00
86	Indeno(1,2,3-cd)pyrene	1.380	1.280	7.2	107	-0.01
87 I	Perylene-d12	1.000	1.000	0.0	118	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG121818\
Data File : BG038709.D
Acq On : 19 Dec 2018 1:57
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Dec 19 07:30:47 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG121218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 12 15:26:27 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.098	1.007	8.3	106	0.00
89	Benzo(k)fluoranthene	1.093	1.025	6.2	108	0.00
90 C	Benzo(a)pyrene	1.059	0.987	6.8	108	0.00
91	Dibenzo(a,h)anthracene	1.055	0.967	8.3	106	0.00
92	Benzo(q,h,i)perylene	1.039	0.950	8.6	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0