

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG122018\
 Data File : BG038768.D
 Acq On : 20 Dec 2018 18:52
 Operator : JU/SJ
 Sample : J6426-09
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 P001-SS013-3638-01

Quant Time: Dec 21 00:44:10 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG121218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 12 15:26:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.05	152	24036	20.00	ng	0.00
21) Naphthalene-d8	10.87	136	93335	20.00	ng	0.00
39) Acenaphthene-d10	14.69	164	61409	20.00	ng	0.00
64) Phenanthrene-d10	17.44	188	144694	20.00	ng	0.00
76) Chrysene-d12	21.72	240	152320	20.00	ng	0.00
87) Perylene-d12	25.02	264	166774	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.61	112	168248	124.15	ng	0.00
7) Phenol-d6	7.21	99	228376	122.76	ng	0.00
23) Nitrobenzene-d5	9.23	82	142074	77.76	ng	0.00
42) 2,4,6-Tribromophenol	16.18	330	102048	120.58	ng	0.00
45) 2-Fluorobiphenyl	13.31	172	312589	69.83	ng	0.00
79) Terphenyl-d14	20.04	244	545068	77.88	ng	0.00
Target Compounds						
10) Phenol	7.24	94	18670	9.446	ng	95
17) 2-Methylphenol	8.49	107	3596	2.716	ng	98
20) 3+4-Methylphenols	8.83	107	4302	2.352	ng	91
30) 1,2,4-Trichlorobenzene	10.73	180	15928	8.743	ng	91
31) Naphthalene	10.93	128	76756	16.691	ng	97
37) 2-Methylnaphthalene	12.52	142	46095	14.050	ng	98
38) 1-Methylnaphthalene	12.74	142	26872	8.441	ng	96
50) Dimethylphthalate	14.13	163	38005	7.578	ng	97
71) Phenanthrene	17.48	178	21727	2.896	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG122018\
Data File : BG038768.D
Acq On : 20 Dec 2018 18:52
Operator : JU/SJ
Sample : J6426-09
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
P001-SS013-3638-01

Quant Time: Dec 21 00:44:10 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG121218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 12 15:26:27 2018
Response via : Initial Calibration

