


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122120\
 Data File : BG047778.D
 Acq On : 21 Dec 2020 14:19
 Operator : CG/JU
 Sample : DFTPP74
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP74

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG120720MA.M
 Title : SVOA CALIBRATION
 Last Update : Mon Dec 21 16:48:34 2020



AutoFind: Scans 2489, 2490, 2491; Background Corrected with Scan 2484

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.4	36173	PASS
68	69	0.00	2	0.7	279	PASS
69	198	0.00	100	44.6	41976	PASS
70	69	0.00	2	0.5	219	PASS
127	198	10	80	43.2	40682	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	94088	PASS
199	198	5	9	7.1	6639	PASS
275	198	10	60	31.0	29208	PASS
365	198	1	100	4.8	4558	PASS
441	198	0.01	100	12.3	11553	PASS
442	198	50	100	75.8	71274	PASS
443	442	15	24	20.0	14258	PASS

m/z	Abundance
41.00	166.0
56.30	155.0
73.00	253.0
97.20	227.0
168.00	188.0
193.00	446.0
199.30	151.0
206.90	193.0
208.00	231.0
209.20	177.0
245.00	153.0
285.10	259.0
415.70	189.0

Instrument :
BNA_G
ClientSampleId :
DFTPP74

m/z	Abundance
37.50	279.0
38.10	408.0
39.00	2247.0
41.10	189.0
44.10	235.0
49.00	318.0
50.10	8154.0
51.10	31432.0
52.10	1850.0
54.90	459.0
56.10	818.0
57.10	2212.0
61.00	494.0
62.00	870.0
63.10	1633.0
64.00	228.0
65.10	690.0
68.00	838.0
69.00	35552.0
73.00	574.0
74.10	4019.0
75.00	5849.0
76.00	1977.0
77.10	33152.0
78.00	2142.0
79.10	2861.0
80.00	2680.0
81.10	2579.0
82.00	481.0
83.00	482.0
84.90	903.0
86.00	1282.0
87.00	542.0
87.80	282.0
91.00	583.0
92.00	631.0
93.00	5592.0
93.90	604.0
95.20	172.0
98.00	3864.0
99.10	3296.0
100.00	348.0
101.10	2031.0
102.90	585.0
104.00	1280.0
105.10	559.0
107.10	12542.0
108.00	1346.0
109.00	497.0
110.00	17112.0
111.00	2343.0
112.10	338.0
112.90	212.0
115.10	244.0
116.00	996.0
117.00	11599.0
118.10	957.0
119.80	164.0
121.90	1035.0
123.10	1820.0
124.10	187.0
125.00	540.0
127.00	30560.0
128.00	3319.0
129.00	14804.0
130.00	1647.0
133.90	502.0
135.00	1575.0
136.10	624.0
137.20	723.0
137.70	443.0
141.00	2810.0
142.10	721.0
142.80	428.0
143.10	472.0
146.10	418.0
147.00	1007.0
148.00	2589.0
148.90	653.0
150.00	264.0
151.10	342.0
153.10	551.0
154.00	532.0
155.10	1444.0
156.00	1369.0
158.10	431.0
159.10	227.0
159.80	724.0
161.10	1265.0
161.90	589.0
165.00	1068.0
166.00	462.0

Instrument :
BNA_G
ClientSampleId :
DFTPP74

167.00	3681.0
167.80	1550.0
168.90	168.0
170.10	180.0
171.00	177.0
172.00	493.0
172.60	501.0
173.90	557.0
174.20	554.0
175.00	1991.0
175.90	229.0
177.00	1148.0
177.90	601.0
178.90	2732.0
180.10	2530.0
181.00	828.0
182.00	243.0
183.80	181.0
185.10	1422.0
186.00	11588.0
187.10	3794.0
188.00	202.0
189.00	557.0
189.90	200.0
190.90	963.0
192.00	1312.0
193.00	1663.0
194.10	357.0
196.00	1894.0
197.90	68504.0
198.90	5790.0
200.00	457.0
203.10	445.0
204.10	3809.0
205.00	6426.0
206.00	22640.0
207.00	2752.0
207.90	1328.0
209.00	572.0
209.90	502.0
211.00	789.0
212.00	204.0
215.10	160.0
216.00	385.0
216.90	6079.0
218.10	981.0
219.10	339.0
220.00	465.0
221.00	2609.0
221.80	569.0
223.10	1303.0
224.00	11797.0
225.00	3170.0
225.90	555.0
227.00	7350.0
227.90	890.0
228.90	1145.0
231.10	803.0
234.00	718.0
234.80	658.0
235.80	606.0
236.90	728.0
240.10	310.0
240.60	223.0
241.80	947.0
242.90	856.0
244.00	8404.0
245.10	2057.0
245.90	2709.0
247.00	560.0
248.90	420.0
251.00	251.0
251.80	322.0
252.90	365.0
255.00	52944.0
256.00	7971.0
256.80	451.0
258.00	3815.0
259.00	764.0
264.90	2047.0
266.00	492.0
270.00	153.0
271.80	228.0
273.00	1517.0
274.10	3263.0
275.00	18976.0
276.00	2651.0
277.00	2562.0
278.00	537.0
281.10	245.0
282.00	213.0
282.90	199.0
284.00	210.0
284.90	509.0
292.00	294.0
293.10	610.0

Instrument :
BNA_G
ClientSampleId :
DFTPP74

293.60	218.0
294.20	258.0
296.00	6471.0
297.00	1109.0
298.00	270.0
303.00	869.0
310.00	219.0
313.00	227.0
313.80	308.0
315.00	683.0
315.90	542.0
320.70	249.0
323.00	1795.0
324.20	529.0
326.70	331.0
331.90	292.0
334.00	1320.0
335.00	484.0
346.10	512.0
352.20	452.0
353.00	285.0
353.30	242.0
353.90	245.0
364.90	2887.0
370.70	180.0
371.00	220.0
371.30	236.0
372.00	649.0
373.00	151.0
383.00	534.0
401.80	273.0
403.00	366.0
404.00	400.0
421.00	292.0
422.00	557.0
423.00	2045.0
424.10	619.0
436.60	177.0
439.10	173.0
441.00	6057.0
442.00	38664.0
443.00	7696.0
444.00	846.0
472.90	171.0

Instrument :
BNA_G
ClientSampleId :
DFTPP74

m/z	Abundance
37.30	185.0
38.10	802.0
39.10	2654.0
40.00	213.0
43.20	210.0
43.90	274.0
49.00	498.0
50.10	14353.0
51.10	39952.0
52.10	3172.0
53.00	198.0
55.20	216.0
56.00	1243.0
57.00	2839.0
58.10	519.0
61.10	732.0
62.00	909.0
63.00	2487.0
64.00	192.0
65.20	891.0
69.00	48328.0
69.90	461.0
73.00	512.0
74.10	5531.0
75.00	7960.0
76.10	2915.0
77.10	46952.0
78.00	3184.0
79.00	4109.0
80.00	3516.0
81.00	4807.0
82.10	1286.0
82.90	924.0
85.00	1110.0
86.00	1401.0
87.10	838.0
87.90	346.0
91.10	978.0
92.10	1918.0
93.10	7879.0
94.00	795.0
98.00	6076.0
99.00	4483.0
99.90	235.0
101.00	1981.0
101.90	204.0
103.00	978.0
104.00	1981.0
105.10	1951.0
107.00	19344.0
108.00	2793.0
109.00	529.0
110.00	24104.0
111.00	4373.0
112.00	632.0
113.10	261.0
116.10	1204.0
117.00	18240.0
118.00	975.0
119.20	301.0
119.90	195.0
122.00	1413.0
123.10	1428.0
124.00	1094.0
125.00	934.0
125.70	435.0
127.00	48328.0
128.00	4180.0
129.00	26584.0
130.00	2095.0
130.90	552.0
132.00	280.0
133.90	1047.0
135.10	1502.0
136.00	932.0
137.00	1017.0
138.00	230.0
138.90	371.0
140.10	593.0
141.00	3193.0
141.80	1330.0
143.00	898.0
144.80	177.0
145.80	633.0
147.00	2392.0
148.00	4459.0
149.00	664.0
150.00	244.0
151.20	325.0
151.60	329.0
153.00	1421.0
153.90	728.0

Instrument :
BNA_G
ClientSampleId :
DFTPP74

155.10	1110.0
156.10	3012.0
158.00	517.0
158.90	515.0
160.10	972.0
161.00	1715.0
162.00	794.0
163.70	376.0
165.10	1829.0
166.00	1192.0
167.00	7045.0
168.10	3253.0
169.10	409.0
170.30	172.0
171.10	599.0
171.90	1125.0
172.90	743.0
174.00	1219.0
175.00	2596.0
176.10	552.0
176.80	1609.0
178.00	586.0
178.90	6037.0
180.00	3162.0
181.00	1371.0
182.10	251.0
183.90	528.0
185.10	3592.0
186.00	21224.0
187.00	5520.0
188.10	504.0
189.00	1870.0
191.10	814.0
192.10	1807.0
193.00	1562.0
194.20	576.0
195.00	284.0
196.10	3103.0
197.90	108088.0
198.90	7889.0
200.10	411.0
201.40	428.0
202.90	1654.0
204.00	6635.0
205.00	9941.0
206.00	36128.0
207.00	5010.0
208.00	2478.0
209.00	239.0
210.00	431.0
210.90	1388.0
211.80	443.0
212.90	264.0
214.10	160.0
214.90	715.0
216.00	1162.0
216.90	12583.0
218.00	1073.0
219.00	183.0
219.90	186.0
220.90	4070.0
221.90	2060.0
223.00	2579.0
224.00	23408.0
225.00	5064.0
226.10	850.0
227.00	12698.0
228.00	1749.0
229.00	2186.0
229.80	403.0
230.90	923.0
233.80	846.0
235.10	596.0
236.00	568.0
237.10	897.0
238.90	325.0
240.00	378.0
240.90	1096.0
242.00	1428.0
243.10	1211.0
244.00	14184.0
244.90	1299.0
246.00	4416.0
247.00	609.0
247.70	154.0
248.90	690.0
250.00	298.0
252.10	253.0
253.30	559.0
253.60	555.0
255.00	86352.0
256.00	13668.0
257.00	1134.0
258.00	6201.0
259.00	1050.0
260.90	405.0

Instrument :
BNA_G
ClientSampleId :
DFTPP74

265.00	2407.0
266.00	300.0
270.70	252.0
271.70	266.0
272.10	477.0
273.00	3097.0
274.00	6946.0
275.00	31888.0
276.00	4650.0
277.00	3181.0
278.00	701.0
278.90	188.0
281.10	192.0
281.70	174.0
283.00	531.0
284.00	356.0
284.90	869.0
291.00	218.0
293.00	844.0
295.10	749.0
295.90	11714.0
296.90	2080.0
303.00	1435.0
304.10	604.0
308.10	286.0
309.10	241.0
310.00	294.0
313.20	160.0
314.00	297.0
315.00	1866.0
316.00	558.0
317.00	268.0
318.30	165.0
320.00	184.0
321.10	425.0
321.90	485.0
323.00	4378.0
324.10	720.0
326.90	318.0
327.90	571.0
331.90	245.0
332.90	599.0
334.00	2632.0
335.00	515.0
335.90	256.0
341.20	273.0
346.00	889.0
347.00	534.0
351.50	193.0
352.00	704.0
352.90	546.0
354.10	857.0
365.00	5473.0
365.80	527.0
371.90	968.0
372.80	605.0
382.70	401.0
391.10	215.0
402.00	974.0
403.00	1005.0
404.10	259.0
421.00	592.0
422.00	896.0
423.00	4368.0
424.10	893.0
437.50	185.0
441.10	12835.0
442.00	81584.0
443.10	15240.0
444.00	1879.0
445.20	188.0

Instrument :
BNA_G
ClientSampleId :
DFTPP74

m/z	Abundance
37.00	221.0
38.20	426.0
39.10	3605.0
39.90	179.0
43.00	198.0
49.00	413.0
50.10	10649.0
51.10	37136.0
52.00	1264.0
55.10	267.0
56.00	1009.0
57.00	2398.0
58.30	191.0
59.80	151.0
61.00	410.0
62.00	630.0
63.10	1890.0
63.90	251.0
65.00	725.0
65.90	214.0
67.20	157.0
69.00	42048.0
69.90	198.0
73.00	475.0
74.00	6036.0
75.10	7224.0
76.00	2471.0
77.10	43448.0
78.10	2770.0
79.10	4962.0
80.10	3181.0
81.10	3169.0
82.10	1238.0
83.00	1054.0
84.00	301.0
84.90	859.0
86.10	1594.0
86.90	839.0
90.90	1386.0
92.10	483.0
93.00	7744.0
94.00	468.0
95.90	452.0
97.00	236.0
98.10	4148.0
99.10	2992.0
99.90	557.0
101.00	2337.0
102.00	246.0
102.90	575.0
104.00	1943.0
105.00	2398.0
105.90	595.0
107.00	18128.0
108.00	1920.0
110.00	22240.0
111.00	5021.0
112.10	880.0
113.10	249.0
115.90	911.0
117.00	14851.0
118.10	1055.0
120.10	368.0
122.10	939.0
122.90	1437.0
124.00	799.0
125.10	653.0
127.00	43160.0
128.00	3592.0
129.00	22600.0
130.00	2057.0
131.00	325.0
133.10	182.0
134.00	817.0
135.00	2108.0
136.00	603.0
137.10	1128.0
137.90	189.0
139.30	205.0
140.00	495.0
141.00	3295.0
142.00	770.0
143.00	774.0
144.10	352.0
146.10	718.0
147.00	2073.0
148.00	4300.0
149.10	1285.0
151.00	345.0
151.90	349.0
153.10	874.0
154.10	1308.0

Instrument :
BNA_G
ClientSampleId :
DFTPP74

155.10	1905.0
156.10	2995.0
157.00	833.0
157.90	582.0
158.80	369.0
159.10	524.0
160.00	1333.0
161.10	1145.0
162.10	287.0
163.70	184.0
165.00	1144.0
166.00	1025.0
167.00	5983.0
168.00	2626.0
169.00	688.0
170.10	438.0
171.10	334.0
171.80	386.0
173.00	901.0
174.00	1609.0
175.00	2725.0
176.10	415.0
176.80	1464.0
178.10	374.0
179.00	5425.0
180.00	2890.0
181.00	1683.0
182.10	338.0
184.00	456.0
185.10	2647.0
186.00	19408.0
187.00	4896.0
188.00	500.0
189.10	1776.0
190.00	511.0
191.10	519.0
192.00	1700.0
192.90	2116.0
194.10	530.0
196.10	3378.0
197.90	105672.0
198.90	6240.0
200.00	727.0
201.00	156.0
201.80	504.0
203.00	1796.0
204.00	5274.0
205.00	9239.0
206.00	35736.0
207.00	4390.0
208.00	1365.0
209.10	451.0
209.80	472.0
211.10	1229.0
212.90	152.0
214.80	870.0
216.10	1183.0
217.00	12014.0
218.00	1314.0
219.30	301.0
221.00	3918.0
222.00	1470.0
223.00	2483.0
224.00	22296.0
225.00	6099.0
226.00	1410.0
227.00	11796.0
227.90	2290.0
228.90	2574.0
229.90	414.0
231.00	695.0
232.90	337.0
234.00	864.0
234.90	917.0
236.00	1115.0
237.10	527.0
238.00	214.0
239.00	455.0
240.00	156.0
241.00	751.0
242.00	1376.0
243.00	1821.0
244.00	15121.0
245.00	1923.0
246.00	3401.0
247.00	814.0
248.00	586.0
248.90	661.0
251.00	165.0
251.80	160.0
252.90	1016.0
255.00	88056.0
256.00	13395.0
257.10	960.0
258.00	6832.0
258.90	1548.0

Instrument :
BNA_G
ClientSampleId :
DFTPP74

260.80	270.0
262.00	231.0
264.00	226.0
265.00	2377.0
266.00	740.0
268.10	150.0
270.20	171.0
271.10	224.0
272.00	559.0
272.90	3064.0
274.00	6903.0
275.00	36760.0
276.00	5073.0
276.90	4242.0
278.10	795.0
279.00	240.0
280.80	198.0
283.10	548.0
283.90	640.0
285.10	829.0
287.80	215.0
289.10	300.0
292.00	300.0
293.10	666.0
293.80	332.0
295.10	695.0
296.00	15297.0
297.00	1407.0
302.00	434.0
302.90	1126.0
303.90	549.0
308.90	220.0
314.10	320.0
315.00	1843.0
316.10	945.0
317.20	178.0
320.80	323.0
322.30	422.0
323.00	3525.0
324.10	724.0
326.90	982.0
328.00	481.0
331.90	367.0
333.00	282.0
334.10	2025.0
335.00	866.0
341.00	816.0
346.00	720.0
351.00	166.0
352.10	810.0
352.90	395.0
354.10	1307.0
355.10	330.0
365.00	5316.0
366.10	890.0
369.70	299.0
370.80	216.0
372.10	1980.0
373.00	200.0
383.00	651.0
390.10	300.0
390.90	238.0
402.00	644.0
403.00	1069.0
404.00	584.0
410.90	268.0
421.00	750.0
422.00	1290.0
423.00	5265.0
424.10	2172.0
425.10	190.0
431.10	209.0
439.00	167.0
439.80	392.0
441.00	15769.0
442.00	93576.0
443.00	19840.0
444.00	1326.0

Instrument :
BNA_G
ClientSampleId :
DFTPP74