


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122221\
 Data File : BG051750.D
 Acq On : 22 Dec 2021 14:36
 Operator : CG/JU
 Sample : DFTPP445
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP445

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121421.M
 Title : SVOA CALIBRATION
 Last Update : Tue Dec 14 14:19:57 2021



AutoFind: Scans 2483, 2484, 2485; Background Corrected with Scan 2477

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
51	198	10	80	38.6	39371	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	41.3	42152	PASS
70	69	0.00	2	0.3	146	PASS
127	198	10	80	46.6	47624	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	102128	PASS
199	198	5	9	6.9	7047	PASS
275	198	10	60	31.0	31693	PASS
365	198	1	100	3.7	3811	PASS
441	198	0.01	100	13.5	13826	PASS
442	198	50	100	80.4	82067	PASS
443	442	15	24	19.5	15963	PASS

Scan 2477 (18.039 min): BG051750.D\data.ms
DFTPP445

m/z	Abundance
40.05	192.0
44.05	170.0
48.95	194.0
207.05	221.0
208.05	165.0

Instrument :
BNA_G
ClientSampleId :
DFTPP445

m/z	Abundance
37.15	168.0
38.25	499.0
39.15	3555.0
40.15	359.0
41.25	159.0
44.15	420.0
49.15	414.0
50.15	11792.0
51.15	41760.0
52.15	2315.0
55.15	243.0
56.15	1085.0
57.05	3302.0
58.05	152.0
61.05	723.0
62.15	623.0
63.15	1827.0
64.15	249.0
65.05	842.0
65.95	179.0
67.05	164.0
69.05	44832.0
70.05	233.0
74.15	4898.0
75.05	7656.0
76.25	2857.0
77.15	48200.0
78.15	3484.0
79.15	2932.0
80.05	3045.0
81.05	3720.0
82.05	1168.0
82.95	812.0
84.05	157.0
85.05	640.0
86.05	1045.0
87.05	589.0
87.95	282.0
91.05	1068.0
92.05	963.0
93.05	6840.0
94.05	588.0
96.25	168.0
98.05	4688.0
99.05	2994.0
100.05	222.0
100.95	1876.0
103.15	715.0
104.05	1326.0
105.05	1558.0
107.05	16062.0
108.15	2571.0
110.05	24784.0
111.05	4273.0
112.05	523.0
113.25	208.0
116.15	947.0
117.05	13144.0
118.05	1179.0
119.85	274.0
122.05	572.0
123.05	1638.0
124.05	570.0
125.15	672.0
127.05	46880.0
128.05	3248.0
129.05	21704.0
130.05	1616.0
131.25	277.0
131.95	197.0
132.85	171.0
134.05	770.0
135.05	1760.0
136.05	586.0
137.05	711.0
140.05	401.0
141.05	2965.0
142.05	794.0
142.95	613.0
144.95	210.0
145.95	331.0
147.05	1164.0
147.95	3141.0
148.95	639.0
149.85	215.0
151.25	401.0
153.05	782.0
154.15	901.0
155.05	1418.0
156.05	2304.0
156.95	564.0
157.95	561.0

Instrument :
BNA_G
ClientSampleId :
DFTPP445

159.05	725.0
159.95	1009.0
161.05	1210.0
161.75	272.0
161.95	264.0
165.05	929.0
165.95	911.0
167.05	4942.0
168.05	2428.0
169.15	492.0
170.05	184.0
171.15	230.0
171.95	551.0
173.05	718.0
174.05	1204.0
175.05	1866.0
176.05	431.0
176.95	909.0
178.15	535.0
179.05	4229.0
180.05	2899.0
181.05	1442.0
182.05	249.0
184.05	417.0
185.05	2359.0
186.05	14897.0
187.05	5137.0
188.05	400.0
189.05	1180.0
189.95	322.0
191.05	675.0
192.05	1033.0
193.05	1638.0
194.05	388.0
195.15	154.0
196.05	2802.0
197.95	96968.0
198.95	7369.0
199.95	771.0
201.55	545.0
202.15	178.0
203.05	921.0
204.05	4498.0
205.05	6463.0
206.05	29608.0
207.05	3684.0
207.95	909.0
209.05	413.0
209.85	512.0
211.05	1078.0
215.05	347.0
216.05	649.0
216.95	7405.0
218.05	1007.0
220.05	258.0
221.05	3309.0
221.95	1118.0
222.95	1942.0
224.05	17360.0
224.95	4915.0
225.95	447.0
227.05	6395.0
227.95	995.0
228.95	1420.0
230.15	229.0
230.95	560.0
232.05	154.0
233.85	252.0
234.95	639.0
235.85	443.0
236.95	620.0
238.85	336.0
240.25	263.0
241.05	388.0
241.95	828.0
243.05	961.0
244.05	12639.0
245.15	1533.0
245.95	2378.0
246.95	584.0
248.95	285.0
250.85	258.0
251.95	193.0
253.05	220.0
255.05	64248.0
256.05	9716.0
256.95	896.0
258.05	3573.0
259.05	559.0
265.05	1804.0
265.95	275.0
270.95	205.0
272.05	281.0
272.95	2100.0
274.05	4914.0
274.95	27864.0

Instrument :
BNA_G
ClientSampleId :
DFTPP445

275.95	4053.0
277.05	2554.0
277.95	408.0
283.05	312.0
283.85	233.0
285.05	356.0
293.05	707.0
294.15	152.0
295.05	244.0
296.05	7434.0
296.85	798.0
301.05	183.0
303.05	723.0
304.05	161.0
308.75	165.0
314.05	292.0
314.95	1024.0
315.95	469.0
320.95	239.0
322.95	2802.0
323.95	585.0
326.85	248.0
328.05	246.0
331.95	223.0
332.95	241.0
334.05	1660.0
335.05	578.0
340.95	265.0
345.95	636.0
352.05	703.0
352.95	473.0
354.05	906.0
364.95	2736.0
365.85	508.0
369.85	163.0
370.75	181.0
371.95	918.0
372.95	294.0
382.95	564.0
390.95	160.0
402.05	627.0
402.95	727.0
403.85	249.0
420.95	508.0
421.95	625.0
422.95	3835.0
423.95	780.0
424.85	205.0
440.95	9891.0
441.95	63336.0
442.95	12184.0
443.95	1383.0

Instrument :
BNA_G
ClientSampleId :
DFTPP445

m/z	Abundance
36.95	161.0
38.15	603.0
39.15	2977.0
40.15	326.0
44.05	394.0
44.85	152.0
50.15	12264.0
51.15	46728.0
52.15	2275.0
55.15	185.0
56.15	1281.0
57.05	3162.0
61.05	839.0
62.05	636.0
63.15	2258.0
64.05	361.0
65.15	1021.0
69.05	48992.0
70.05	206.0
73.25	519.0
74.15	5439.0
75.15	8011.0
76.15	2686.0
77.15	54696.0
78.15	3747.0
79.05	3890.0
80.15	3155.0
81.05	4641.0
82.05	1036.0
83.05	895.0
85.05	581.0
86.05	1215.0
87.15	575.0
88.15	372.0
91.15	996.0
92.05	1248.0
93.05	6812.0
94.15	488.0
95.05	162.0
96.05	192.0
98.05	5510.0
99.05	4347.0
100.15	412.0
101.05	2558.0
103.05	1018.0
104.05	1666.0
105.05	1444.0
106.05	591.0
107.05	17536.0
108.05	2690.0
110.05	27456.0
111.05	4048.0
111.95	717.0
112.95	227.0
116.05	763.0
117.05	15929.0
118.15	969.0
120.05	354.0
122.05	955.0
123.05	1222.0
124.05	1068.0
124.95	547.0
127.05	55192.0
128.05	3809.0
129.05	23264.0
130.05	2218.0
131.05	566.0
132.15	190.0
134.05	853.0
135.05	1496.0
135.95	769.0
137.05	841.0
138.15	305.0
140.05	346.0
141.05	3277.0
142.15	1035.0
142.95	597.0
143.95	206.0
146.05	736.0
147.05	1869.0
148.05	3129.0
149.05	1066.0
149.95	252.0
151.15	426.0
153.05	1290.0
154.05	632.0
155.05	2082.0
156.15	2490.0
157.05	298.0
157.95	552.0
159.15	419.0
160.05	771.0

Instrument :
BNA_G
ClientSampleId :
DFTPP445

161.05	1720.0
162.05	558.0
164.05	161.0
165.05	1456.0
166.05	1083.0
167.05	5854.0
168.05	2423.0
169.05	559.0
170.15	320.0
171.05	250.0
171.85	832.0
172.95	839.0
174.05	1426.0
175.05	2429.0
175.85	495.0
176.15	493.0
177.05	1390.0
178.05	422.0
178.95	5417.0
180.05	3856.0
181.05	1714.0
182.05	296.0
182.95	360.0
183.85	519.0
185.05	2395.0
186.05	17864.0
187.05	4969.0
187.95	494.0
189.05	1453.0
190.05	190.0
191.05	665.0
192.05	1092.0
193.05	2000.0
194.05	351.0
196.05	3767.0
197.95	117992.0
199.05	7555.0
200.05	764.0
201.35	538.0
202.15	161.0
203.05	1048.0
204.05	4799.0
205.05	8564.0
206.05	35840.0
207.05	4210.0
207.95	1678.0
209.05	398.0
210.05	872.0
211.05	1357.0
211.95	323.0
214.95	457.0
216.15	789.0
216.95	9793.0
218.05	1570.0
220.95	5458.0
222.05	1082.0
223.05	1974.0
224.05	20528.0
225.05	5615.0
226.95	8269.0
228.05	1151.0
229.05	1868.0
229.95	323.0
231.05	575.0
232.05	223.0
233.95	405.0
235.05	675.0
236.15	427.0
237.05	536.0
238.95	258.0
240.15	200.0
241.05	614.0
242.05	1437.0
243.05	1364.0
244.05	15384.0
245.05	2255.0
246.05	3628.0
247.05	663.0
248.95	502.0
253.05	330.0
255.05	78784.0
255.95	11753.0
257.05	819.0
257.95	4723.0
258.95	658.0
263.95	267.0
264.95	2262.0
265.95	260.0
269.85	170.0
271.05	192.0
271.95	219.0
272.95	2812.0
274.05	6260.0
275.05	37552.0
276.05	4953.0
276.95	2923.0

Instrument :
BNA_G
ClientSampleId :
DFTPP445

277.95	555.0
278.95	164.0
282.95	580.0
283.95	393.0
284.95	661.0
288.75	241.0
292.95	655.0
294.65	224.0
295.95	10500.0
296.95	1476.0
303.05	1207.0
304.05	323.0
307.95	219.0
310.15	163.0
313.95	861.0
315.05	1118.0
315.85	565.0
320.95	422.0
323.05	3312.0
324.05	599.0
326.95	757.0
327.95	343.0
332.95	249.0
333.95	2323.0
335.05	680.0
340.95	352.0
345.95	555.0
346.95	200.0
351.95	1447.0
352.95	675.0
353.95	873.0
355.05	233.0
361.05	184.0
364.95	4766.0
365.95	571.0
370.95	351.0
371.95	1806.0
372.85	427.0
382.95	769.0
383.85	169.0
389.85	332.0
390.95	255.0
392.05	206.0
400.75	271.0
401.95	816.0
402.95	1071.0
403.95	417.0
415.05	162.0
420.95	541.0
421.85	763.0
423.05	5732.0
424.05	1233.0
440.95	15887.0
441.95	93608.0
442.95	18608.0
443.95	1919.0
444.95	157.0

Instrument :
BNA_G
ClientSampleId :
DFTPP445

m/z	Abundance
38.05	391.0
39.15	2273.0
40.15	306.0
44.15	493.0
49.05	313.0
50.15	8370.0
51.15	29624.0
52.15	1515.0
55.25	173.0
56.05	1074.0
57.05	2154.0
61.05	328.0
62.05	483.0
63.15	1614.0
64.15	219.0
65.15	695.0
66.85	171.0
69.05	32632.0
73.05	251.0
74.15	3003.0
75.05	6196.0
76.15	1721.0
77.15	36072.0
78.15	2332.0
79.15	2847.0
80.05	2084.0
81.05	3505.0
82.05	634.0
83.05	847.0
85.05	384.0
86.15	659.0
87.05	306.0
87.85	153.0
91.05	720.0
92.05	641.0
93.05	4647.0
93.85	270.0
96.15	255.0
98.05	3640.0
99.05	2736.0
100.15	283.0
101.05	1749.0
102.95	623.0
104.05	1222.0
104.95	1207.0
107.05	12515.0
108.05	1611.0
110.05	20688.0
111.05	3374.0
112.05	516.0
116.05	704.0
117.05	11730.0
118.05	972.0
119.95	163.0
121.95	769.0
123.05	1244.0
123.95	791.0
125.15	651.0
125.85	178.0
127.05	40800.0
128.15	2620.0
129.05	16528.0
130.05	1439.0
131.15	321.0
131.95	183.0
134.05	406.0
135.05	1372.0
135.85	393.0
136.15	382.0
137.05	873.0
138.05	292.0
139.95	466.0
141.05	2511.0
141.95	667.0
143.05	461.0
144.05	191.0
146.05	613.0
147.05	1249.0
148.05	3010.0
149.05	708.0
151.65	338.0
153.05	539.0
153.95	722.0
154.95	1449.0
156.05	1794.0
157.35	294.0
157.85	433.0
158.95	453.0
160.05	744.0
161.05	1017.0
163.95	185.0
164.95	986.0

Instrument :
BNA_G
ClientSampleId :
DFTPP445

166.05	608.0
167.05	4441.0
168.05	1938.0
169.05	395.0
169.85	216.0
170.45	234.0
171.15	213.0
172.15	418.0
173.05	654.0
174.05	1136.0
175.15	1735.0
175.85	254.0
177.05	969.0
177.95	481.0
178.95	4593.0
180.05	2831.0
180.95	1255.0
182.95	228.0
184.05	543.0
185.05	1690.0
186.05	13259.0
187.05	3342.0
188.15	536.0
189.05	943.0
189.95	169.0
191.05	456.0
191.95	1125.0
193.05	1364.0
194.05	509.0
195.95	2697.0
196.65	846.0
197.95	91424.0
199.05	6217.0
200.05	524.0
200.85	155.0
201.55	642.0
203.05	973.0
204.05	4034.0
205.05	5929.0
206.05	27744.0
207.05	3579.0
208.05	986.0
209.15	352.0
210.95	1073.0
215.15	380.0
215.95	845.0
216.95	7022.0
217.95	936.0
221.05	3921.0
221.85	851.0
223.05	1951.0
224.05	16808.0
225.05	4436.0
226.05	425.0
227.05	7757.0
228.05	960.0
228.95	1797.0
229.85	221.0
230.95	830.0
231.95	156.0
232.85	192.0
233.95	263.0
234.95	332.0
236.05	296.0
236.95	566.0
238.95	340.0
240.05	297.0
241.05	372.0
241.95	959.0
243.05	960.0
244.05	13512.0
245.05	1474.0
246.05	2847.0
247.05	706.0
248.95	652.0
253.05	430.0
255.05	67032.0
255.95	9457.0
256.95	919.0
258.05	3834.0
258.95	599.0
260.95	254.0
264.95	1997.0
266.05	287.0
270.95	210.0
271.95	362.0
272.95	2203.0
274.05	5279.0
274.95	29664.0
276.05	3205.0
276.95	2882.0
278.05	434.0
283.05	289.0
284.05	162.0
285.05	413.0
289.05	154.0

Instrument :
BNA_G
ClientSampleId :
DFTPP445

290.05	166.0
293.05	620.0
296.05	9170.0
296.95	1477.0
301.05	226.0
302.05	190.0
303.05	1334.0
303.85	479.0
313.95	378.0
314.95	900.0
315.95	510.0
321.05	353.0
321.95	205.0
322.95	2739.0
324.05	493.0
326.95	506.0
328.05	398.0
331.85	245.0
332.95	286.0
334.05	1582.0
334.95	453.0
336.05	191.0
341.05	397.0
345.85	667.0
351.95	986.0
352.85	467.0
353.85	957.0
364.95	3931.0
365.85	584.0
370.95	345.0
372.05	1836.0
373.05	528.0
383.05	386.0
392.05	216.0
401.95	680.0
402.95	1028.0
403.75	325.0
414.85	156.0
420.95	747.0
421.95	785.0
422.95	5016.0
423.95	1227.0
440.95	15700.0
441.95	89256.0
442.95	17096.0
444.05	1793.0

Instrument :
BNA_G
ClientSampleId :
DFTPP445