

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122422\
 Data File : BG056089.D
 Acq On : 24 Dec 2022 14:24
 Operator : CG/JU
 Sample : SST00541
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :

Quant Time: Dec 24 21:47:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG122422.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Dec 24 18:26:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.347	152	20439	20.000	ng/ul	0.00
20) Naphthalene-d8	11.179	136	75633	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.957	164	75516	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.689	188	230842	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.990	240	276295	20.000	ng/ul	# 0.00
88) Perylene-d12	25.462	264	303860	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.653	96	855	1.751	ng/uL	0.00
4) Pyridine-d5	4.093	84	7951	4.994	ng/ul	0.00
7) Phenol-d5	7.466	99	8089	4.363	ng/ul	-0.01
9) Bis-(2-Chloroethyl)eth...	7.648	67	3666	5.017	ng/ul	0.00
11) 2-Chlorophenol-d4	7.871	132	5130	4.682	ng/ul	0.00
15) 4-Methylphenol-d8	9.029	113	6693	4.797	ng/ul	0.00
21) Nitrobenzene-d5	9.516	128	2240	4.066	ng/ul	0.00
24) 2-Nitrophenol-d4	10.245	143	3425	4.759	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.780	165	7451	4.659	ng/ul	0.00
31) 4-Chloroaniline-d4	11.302	131	8130	4.494	ng/ul	0.00
46) Dimethylphthalate-d6	14.340	166	27875	4.769	ng/ul	0.00
49) Acenaphthylene-d8	14.651	160	32542	4.930	ng/ul	0.00
54) 4-Nitrophenol-d4	15.110	143	3877	4.407	ng/ul	0.00
60) Fluorene-d10	15.938	176	29895	5.182	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	16.038	200	7332	4.941	ng/ul	0.00
73) Anthracene-d10	17.789	188	52617	4.932	ng/ul	0.00
81) Pyrene-d10	20.051	212	73770	4.834	ng/ul	0.00
92) Benzo(a)pyrene-d12	25.221	264	75536	4.891	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.688	88	854	1.625	ng/uL#	58
5) Pyridine	4.111	79	7007	4.566	ng/ul#	93
6) Benzaldehyde	7.466	77	3246	4.769	ng/ul	96
8) Phenol	7.501	94	8108	4.394	ng/ul#	88
10) Bis(2-Chloroethyl)ether	7.748	93	5612	4.621	ng/ul#	88
12) 2-Chlorophenol	7.901	128	4977	4.522	ng/ul#	90
13) 2-Methylphenol	8.770	108	6481	4.678	ng/ul	94
14) 2,2'-oxybis(1-Chloropr...	8.864	45	110	0.705	ng/ul#	64
16) Acetophenone	9.170	105	11616	5.026	ng/ul	93
17) N-Nitroso-di-n-propyla...	9.146	70	4942	5.057	ng/ul#	96
18) 4-Methylphenol	9.099	108	6355	4.235	ng/ul	84
19) Hexachloroethane	9.446	117	2146	4.849	ng/ul	89
22) Nitrobenzene	9.557	77	6315	4.620	ng/ul#	96
23) Isophorone	10.080	82	15917	5.093	ng/ul	97
25) 2-Nitrophenol	10.280	139	3391	4.881	ng/ul	95
26) 2,4-Dimethylphenol	10.321	107	7958	4.841	ng/ul	97
27) Bis(2-Chloroethoxy)met...	10.550	93	8061	4.977	ng/ul#	84
29) 2,4-Dichlorophenol	10.809	162	7322	4.793	ng/ul	97
30) Naphthalene	11.232	128	18749	4.758	ng/ul	96
32) 4-Chloroaniline	11.320	127	8248	4.525	ng/ul#	89
33) Hexachlorobutadiene	11.502	225	7123	4.976	ng/ul	91
34) Caprolactam	12.060	113	2451	5.305	ng/ul#	84
35) 4-Chloro-3-methylphenol	12.413	107	6826	4.651	ng/ul	94
36) 2-Methylnaphthalene	12.807	142	15429	4.883	ng/ul	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122422\
 Data File : BG056089.D
 Acq On : 24 Dec 2022 14:24
 Operator : CG/JU
 Sample : SST00541
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :


Quant Time: Dec 24 21:47:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG122422.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Dec 24 18:26:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	13.030	142	15240	4.788	ng/ul#	82
39) 1,2,4,5-Tetrachloroben...	13.165	216	14974	4.808	ng/ul#	93
40) Hexachlorocyclopentadiene	13.147	237	7796	3.861	ng/ul#	95
41) 2,4,6-Trichlorophenol	13.394	196	8470	4.614	ng/ul	96
42) 2,4,5-Trichlorophenol	13.465	196	9456	4.751	ng/ul	92
43) 1,1'-Biphenyl	13.794	154	25037	4.690	ng/ul#	87
44) 2-Chloronaphthalene	13.841	162	21187	4.788	ng/ul	97
45) 2-Nitroaniline	14.029	65	2958	4.814	ng/ul#	75
47) Dimethylphthalate	14.387	163	28034	4.837	ng/ul#	97
48) 2,6-Dinitrotoluene	14.516	165	5256	4.595	ng/ul#	95
50) Acenaphthylene	14.681	152	31433	4.792	ng/ul	99
51) 3-Nitroaniline	14.845	138	4618	5.374	ng/ul#	64
52) Acenaphthene	15.016	153	21357	4.635	ng/ul	89
53) 2,4-Dinitrophenol	15.039	184	1933	2.171	ng/ul#	53
55) 4-Nitrophenol	15.127	109	3196	3.734	ng/ul#	77
56) Dibenzofuran	15.351	168	35818	4.966	ng/ul	100
57) 2,4-Dinitrotoluene	15.298	165	7552	4.761	ng/ul#	82
58) 2,3,4,6-Tetrachlorophenol	15.568	232	9374	4.765	ng/ul#	87
59) Diethylphthalate	15.738	149	24035	4.642	ng/ul	97
61) Fluorene	15.991	166	29866	5.041	ng/ul	94
62) 4-Chlorophenyl-phenyle...	15.979	204	19182	4.984	ng/ul	96
63) 4-Nitroaniline	15.997	138	3390	4.134	ng/ul#	94
66) 4,6-Dinitro-2-methylph...	16.056	198	6345	4.460	ng/ul#	94
67) N-Nitrosodiphenylamine	16.185	169	26811	4.584	ng/ul	94
68) 4-Bromophenyl-phenylether	16.872	248	13586	4.642	ng/ul	97
69) Hexachlorobenzene	16.996	284	13193	4.658	ng/ul	91
70) Atrazine	17.119	200	12153	4.358	ng/ul	97
71) Pentachlorophenol	17.331	266	7506	4.168	ng/ul	96
72) Phenanthrene	17.736	178	56778	4.845	ng/ul	100
74) Anthracene	17.824	178	60367	5.032	ng/ul	97
75) 1,2,3,4-Tetrachloroben...	13.764	216	15965	4.559	ng/uL	90
76) Pentachlorobenzene	15.268	250	15482	4.479	ng/uL	92
77) Carbazole	18.083	167	49462	5.043	ng/ul	96
78) Di-n-butylphthalate	18.617	149	39879	4.450	ng/ul	96
80) Fluoranthene	19.722	202	84510	4.710	ng/ul#	98
82) Pyrene	20.080	202	90001	4.972	ng/ul	98
83) Butylbenzylphthalate	20.944	149	16301	4.232	ng/ul	92
84) 3,3'-Dichlorobenzidine	21.872	252	31307	5.043	ng/ul#	95
85) Benzo(a)anthracene	21.972	228	91841	4.870	ng/ul	96
86) Bis(2-ethylhexyl)phtha...	21.843	149	23939	4.057	ng/ul#	92
87) Chrysene	22.037	228	86838	4.962	ng/ul	97
89) Di-n-octyl phthalate	23.142	149	43720	3.888	ng/ul	100
90) Benzo(b)fluoranthene	24.352	252	98062	4.889	ng/ul	95
91) Benzo(k)fluoranthene	24.352	252	98062	5.013	ng/ul	97
93) Benzo(a)pyrene	25.298	252	85552	4.894	ng/ul#	98
94) Indeno(1,2,3-cd)pyrene	29.446	276	55577	2.416	ng/ul	97
95) Dibenzo(a,h)anthracene	29.511	278	47454	2.464	ng/ul	98
96) Benzo(g,h,i)perylene	30.692	276	46804	2.534	ng/ul#	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Dec 24 21:47:50 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG122422.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Dec 24 18:26:52 2022
Response via : Initial Calibration








#2
1,4-Dioxane
Concen: 1.625 ng/uL
RT: 3.688 min Scan# 61
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24


Instrument :
BNA_G
ClientSampleId :





Tgt Ion: 88 Resp: 854
Ion Ratio Lower Upper
88 100
43 0.0 7.4 11.0#
58 0.0 19.1 28.7#




#5
Pyridine
Concen: 4.566 ng/uL
RT: 4.111 min Scan# 140
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24



Tgt Ion: 79 Resp: 7007
Ion Ratio Lower Upper
79 100
52 15.9 11.0 16.6
51 10.3 11.1 16.7#









#6
Benzaldehyde
Concen: 4.769 ng/ul
RT: 7.466 min Scan# 712
Delta R.T. -0.006 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24




Instrument :
BNA_G
ClientSampleId :


Tgt Ion: 77 Resp: 3246
Ion Ratio Lower Upper
77 100
105 93.4 72.6 109.0
106 90.1 76.8 115.2



#8
Phenol
Concen: 4.394 ng/ul
RT: 7.501 min Scan# 717
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24

Tgt Ion: 94 Resp: 8108
Ion Ratio Lower Upper
94 100
65 25.3 18.8 28.2
66 43.3 26.2 39.2#









#10
Bis(2-Chloroethyl)ether
Concen: 4.621 ng/ul
RT: 7.748 min Scan# 759
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24




Instrument :
BNA_G
ClientSampleId :


Tgt Ion: 93 Resp: 5612
Ion Ratio Lower Upper
93 100
63 56.3 37.4 56.2#
95 37.4 26.3 39.5



#12
2-Chlorophenol
Concen: 4.522 ng/ul
RT: 7.901 min Scan# 785
Delta R.T. -0.006 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24

Tgt Ion:128 Resp: 4977
Ion Ratio Lower Upper
128 100
64 40.2 24.8 37.2#
130 32.7 27.9 41.9





#13

2-Methylphenol

Concen: 4.678 ng/ul

RT: 8.770 min Scan# 91

Delta R.T. -0.006 min

Lab File: BG056089.D

Acq: 24 Dec 2022 14:24

Instrument :

BNA_G





ClientSampleId :

Tgt Ion:108 Resp: 6481

Ion Ratio Lower Upper

108 100

107 83.3 71.4 107.0



#14

2,2'-oxybis(1-Chloropropane)

Concen: 0.705 ng/ul

RT: 8.864 min Scan# 949

Delta R.T. 0.000 min

Lab File: BG056089.D

Acq: 24 Dec 2022 14:24




Tgt Ion: 45 Resp: 110


Ion Ratio Lower Upper

45 100

77 178.5 105.8 158.8#

79 103.2 114.2 171.4#









#16
Acetophenone
Concen: 5.026 ng/ul
RT: 9.170 min Scan# 1001
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:105 Resp: 11616
Ion Ratio Lower Upper
105 100
77 67.2 59.0 88.4
51 10.2 9.6 14.4



#17
N-Nitroso-di-n-propylamine
Concen: 5.057 ng/ul
RT: 9.146 min Scan# 997
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24

Tgt Ion: 70 Resp: 4942
Ion Ratio Lower Upper
70 100
42 20.0 16.1 24.1
101 14.0 8.4 12.6#
130 23.6 16.8 25.2









#18
4-Methylphenol
Concen: 4.235 ng/ul
RT: 9.099 min Scan# 91
Delta R.T. -0.006 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:108 Resp: 6355
Ion Ratio Lower Upper
108 100
107 124.8 86.2 129.2



#19
Hexachloroethane
Concen: 4.849 ng/ul
RT: 9.446 min Scan# 1048
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24


Tgt Ion:117 Resp: 2146
Ion Ratio Lower Upper
117 100
201 165.1 149.0 223.4
199 114.0 87.7 131.5








#22
Nitrobenzene
Concen: 4.620 ng/ul
RT: 9.557 min Scan# 1067
Delta R.T. -0.006 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24

Instrument :
BNA_G
ClientSampleId :







Tgt Ion: 77 Resp: 6315
Ion Ratio Lower Upper
77 100
123 46.5 36.0 54.0
65 5.9 7.6 11.4#



#23
Isophorone
Concen: 5.093 ng/ul
RT: 10.080 min Scan# 1156
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24

Tgt Ion: 82 Resp: 15917
Ion Ratio Lower Upper
82 100
95 11.1 7.9 11.9
138 18.3 15.8 23.6









#25
2-Nitrophenol
Concen: 4.881 ng/ul
RT: 10.280 min Scan# 1190
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:139 Resp: 3391
Ion Ratio Lower Upper
139 100
65 25.0 23.2 34.8
109 28.8 23.8 35.8



#26
2,4-Dimethylphenol
Concen: 4.841 ng/ul
RT: 10.321 min Scan# 1197
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24

Tgt Ion:107 Resp: 7958
Ion Ratio Lower Upper
107 100
121 48.8 36.5 54.7
122 82.3 64.4 96.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 4.977 ng/ul
 RT: 10.550 min Scan# 1236
 Delta R.T. -0.006 min
 Lab File: BG056089.D
 Acq: 24 Dec 2022 14:24




Instrument :
 BNA_G
 ClientSampleId :


Tgt Ion: 93 Resp: 8061
 Ion Ratio Lower Upper
 93 100
 95 23.9 26.8 40.2#
 123 14.3 7.4 11.0#



#29
 2,4-Dichlorophenol
 Concen: 4.793 ng/ul
 RT: 10.809 min Scan# 1280
 Delta R.T. -0.006 min
 Lab File: BG056089.D
 Acq: 24 Dec 2022 14:24


Tgt Ion: 162 Resp: 7322
 Ion Ratio Lower Upper
 162 100
 164 61.9 47.3 70.9
 98 22.1 17.8 26.8








#30
Naphthalene
Concen: 4.758 ng/ul
RT: 11.232 min Scan# 1352
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24


Instrument :
BNA_G
ClientSampleId :





Tgt Ion:128 Resp: 18749
Ion Ratio Lower Upper
128 100
129 9.2 8.4 12.6
127 16.5 11.6 17.4




#32
4-Chloroaniline
Concen: 4.525 ng/ul
RT: 11.320 min Scan# 1367
Delta R.T. -0.006 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24



Tgt Ion:127 Resp: 8248
Ion Ratio Lower Upper
127 100
129 35.3 23.4 35.0#









#33
Hexachlorobutadiene
Concen: 4.976 ng/ul
RT: 11.502 min Scan# 11
Delta R.T. -0.006 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:225 Resp: 7123
Ion Ratio Lower Upper
225 100
223 54.3 51.2 76.8
227 58.5 50.5 75.7



#34
Caprolactam
Concen: 5.305 ng/ul
RT: 12.060 min Scan# 1493
Delta R.T. -0.018 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24

Tgt Ion:113 Resp: 2451
Ion Ratio Lower Upper
113 100
55 29.7 28.0 42.0
56 21.0 27.7 41.5#









#35
4-Chloro-3-methylphenol
Concen: 4.651 ng/ul
RT: 12.413 min Scan# 1553
Delta R.T. -0.006 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:107 Resp: 6826
Ion Ratio Lower Upper
107 100
144 24.6 20.0 30.0
142 82.0 60.7 91.1



#36
2-Methylnaphthalene
Concen: 4.883 ng/ul
RT: 12.807 min Scan# 1620
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24

Tgt Ion:142 Resp: 15429
Ion Ratio Lower Upper
142 100
141 89.7 69.8 104.6





#37

1-Methylnaphthalene

Concen: 4.788 ng/ul

RT: 13.030 min Scan# 1

Delta R.T. 0.006 min

Lab File: BG056089.D

Acq: 24 Dec 2022 14:24

Instrument :

BNA_G

ClientSampleId :





Tgt Ion:142 Resp: 15240

Ion Ratio Lower Upper

142 100

141 77.4 76.7 115.1

116 2.6 3.8 5.8#



#39

1,2,4,5-Tetrachlorobenzene

Concen: 4.808 ng/ul

RT: 13.165 min Scan# 1681

Delta R.T. 0.000 min

Lab File: BG056089.D

Acq: 24 Dec 2022 14:24

Tgt Ion:216 Resp: 14974




Ion Ratio Lower Upper


216 100

214 82.3 61.0 91.4

179 11.1 13.4 20.0#

108 12.0 9.1 13.7









#40
Hexachlorocyclopentadiene
Concen: 3.861 ng/ul
RT: 13.147 min Scan# 1678
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:237 Resp: 7796
Ion Ratio Lower Upper
237 100
235 65.4 52.0 78.0
272 0.0 9.9 14.9#



#41
2,4,6-Trichlorophenol
Concen: 4.614 ng/ul
RT: 13.394 min Scan# 1720
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24

Tgt Ion:196 Resp: 8470
Ion Ratio Lower Upper
196 100
198 88.9 70.2 105.4
200 35.1 23.4 35.2









#42
2,4,5-Trichlorophenol
Concen: 4.751 ng/ul
RT: 13.465 min Scan# 1732
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:196 Resp: 9456
Ion Ratio Lower Upper
196 100
198 99.4 87.4 131.2
200 33.0 25.0 37.6



#43
1,1'-Biphenyl
Concen: 4.690 ng/ul
RT: 13.794 min Scan# 1788
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24

Tgt Ion:154 Resp: 25037
Ion Ratio Lower Upper
154 100
153 48.2 31.3 46.9#
76 7.8 7.4 11.0





#44

2-Chloronaphthalene

Concen: 4.788 ng/ul

RT: 13.841 min Scan# 1

Delta R.T. -0.006 min


Lab File: BG056089.D

Acq: 24 Dec 2022 14:24

Instrument :

BNA_G

ClientSampleId :






Tgt Ion:162 Resp: 21187

Ion Ratio Lower Upper

162 100

164 32.3 27.0 40.4

127 27.0 23.4 35.2



#45

2-Nitroaniline

Concen: 4.814 ng/ul

RT: 14.029 min Scan# 1828

Delta R.T. -0.006 min

Lab File: BG056089.D

Acq: 24 Dec 2022 14:24




Tgt Ion: 65 Resp: 2958


Ion Ratio Lower Upper

65 100

92 96.4 92.6 138.8

138 158.4 162.2 243.4#





#47

Dimethylphthalate

Concen: 4.837 ng/ul

RT: 14.387 min Scan# 1889

Delta R.T. -0.006 min


Lab File: BG056089.D

Acq: 24 Dec 2022 14:24




Instrument :

BNA_G

ClientSampleId :



Tgt Ion:	163	Resp:	28034
Ion Ratio	Lower	Upper	
163	100		
194	6.0	3.2	4.8#
164	10.5	7.9	11.9



#48

2,6-Dinitrotoluene

Concen: 4.595 ng/ul




RT: 14.516 min Scan# 1911


Delta R.T. 0.000 min

Lab File: BG056089.D

Acq: 24 Dec 2022 14:24


Tgt Ion:	165	Resp:	5256
Ion Ratio	Lower	Upper	
165	100		
89	44.8	37.0	55.6
121	11.6	13.6	20.4#








#50
Acenaphthylene
Concen: 4.792 ng/ul
RT: 14.681 min Scan# 1939
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24

Instrument :
BNA_G
ClientSampleId :







Tgt Ion:152 Resp: 31433
Ion Ratio Lower Upper
152 100
151 19.6 16.2 24.2
153 12.3 9.7 14.5



#51
3-Nitroaniline
Concen: 5.374 ng/ul
RT: 14.845 min Scan# 1967
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24


Tgt Ion:138 Resp: 4618
Ion Ratio Lower Upper
138 100
108 17.6 8.0 12.0#
92 68.5 86.3 129.5#








#52
 Acenaphthene
 Concen: 4.635 ng/ul
 RT: 15.016 min Scan# 1996
 Delta R.T. -0.006 min
 Lab File: BG056089.D
 Acq: 24 Dec 2022 14:24

Instrument :
 BNA_G
 ClientSampleId :







Tgt Ion:153 Resp: 21357
 Ion Ratio Lower Upper
 153 100
 152 50.3 34.5 51.7
 154 89.4 63.9 95.9



#53
 2,4-Dinitrophenol
 Concen: 2.171 ng/ul
 RT: 15.039 min Scan# 2000
 Delta R.T. -0.006 min
 Lab File: BG056089.D
 Acq: 24 Dec 2022 14:24

Tgt Ion:184 Resp: 1933
 Ion Ratio Lower Upper
 184 100
 63 31.1 16.8 25.2#
 154 101.1 47.1 70.7#









#55
4-Nitrophenol
Concen: 3.734 ng/ul
RT: 15.127 min Scan# 2015
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:109 Resp: 3196
Ion Ratio Lower Upper
109 100
139 125.5 79.5 119.3#
65 55.8 56.7 85.1#



#56
Dibenzofuran
Concen: 4.966 ng/ul
RT: 15.351 min Scan# 2053
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24

Tgt Ion:168 Resp: 35818
Ion Ratio Lower Upper
168 100
139 33.1 26.3 39.5





#57

2,4-Dinitrotoluene

Concen: 4.761 ng/ul

RT: 15.298 min Scan# 2044

Delta R.T. 0.000 min


Lab File: BG056089.D

Acq: 24 Dec 2022 14:24




Instrument :

BNA_G

ClientSampleId :



Tgt Ion:	165	Resp:	7552
Ion Ratio	Lower	Upper	
165	100		
63	16.6	20.9	31.3#
182	4.5	1.8	2.8#



#58

2,3,4,6-Tetrachlorophenol

Concen: 4.765 ng/ul




RT: 15.568 min Scan# 2090


Delta R.T. 0.000 min

Lab File: BG056089.D

Acq: 24 Dec 2022 14:24

Tgt Ion:	232	Resp:	9374
Ion Ratio	Lower	Upper	
232	100		
131	20.6	20.2	30.4
130	0.0	1.0	1.6#
166	13.6	17.8	26.8#





#59

Diethylphthalate

Concen: 4.642 ng/ul

RT: 15.738 min Scan# 21

Delta R.T. -0.006 min

Lab File: BG056089.D





Acq: 24 Dec 2022 14:24

Instrument :

BNA_G

ClientSampleId :

Tgt Ion:149	Resp:	24035
Ion Ratio	Lower	Upper
149	100	
177	26.5	20.5 30.7
150	14.1	9.8 14.8



#61

Fluorene

Concen: 5.041 ng/ul




RT: 15.991 min Scan# 2162


Delta R.T. -0.006 min

Lab File: BG056089.D

Acq: 24 Dec 2022 14:24

Tgt Ion:166	Resp:	29866
Ion Ratio	Lower	Upper
166	100	
165	96.2	81.8 122.8
167	12.4	10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 4.984 ng/ul
RT: 15.979 min Scan# 2160
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:204 Resp: 19182
Ion Ratio Lower Upper
204 100
206 34.0 26.3 39.5
141 35.0 30.5 45.7



#63
4-Nitroaniline
Concen: 4.134 ng/ul
RT: 15.997 min Scan# 2163
Delta R.T. -0.006 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24

Tgt Ion:138 Resp: 3390
Ion Ratio Lower Upper
138 100
92 49.2 32.6 49.0#
108 80.2 66.3 99.5









#66
 4,6-Dinitro-2-methylphenol
 Concen: 4.460 ng/ul
 RT: 16.056 min Scan# 2173
 Delta R.T. 0.000 min
 Lab File: BG056089.D
 Acq: 24 Dec 2022 14:24




Instrument :
 BNA_G
 ClientSampleId :


Tgt Ion:198 Resp: 6345
 Ion Ratio Lower Upper
 198 100
 182 0.0 2.1 3.1#
 77 12.2 8.2 12.2



#67
 N-Nitrosodiphenylamine
 Concen: 4.584 ng/ul
 RT: 16.185 min Scan# 2195
 Delta R.T. 0.000 min
 Lab File: BG056089.D
 Acq: 24 Dec 2022 14:24

Tgt Ion:169 Resp: 26811
 Ion Ratio Lower Upper
 169 100
 168 74.0 57.4 86.0
 167 45.8 31.1 46.7









#68
4-Bromophenyl-phenylether
Concen: 4.642 ng/ul
RT: 16.872 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:248 Resp: 13586
Ion Ratio Lower Upper
248 100
250 103.5 79.4 119.2
141 42.1 33.9 50.9



#69
Hexachlorobenzene
Concen: 4.658 ng/ul
RT: 16.996 min Scan# 2333
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24

Tgt Ion:284 Resp: 13193
Ion Ratio Lower Upper
284 100
142 25.7 18.1 27.1
249 37.5 25.2 37.8









#70
 Atrazine
 Concen: 4.358 ng/ul
 RT: 17.119 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BG056089.D
 Acq: 24 Dec 2022 14:24




Instrument :
 BNA_G
 ClientSampleId :


Tgt Ion:200 Resp: 12153
 Ion Ratio Lower Upper
 200 100
 173 18.7 16.9 25.3
 215 55.2 42.8 64.2



#71
 Pentachlorophenol
 Concen: 4.168 ng/ul
 RT: 17.331 min Scan# 2390
 Delta R.T. 0.000 min
 Lab File: BG056089.D
 Acq: 24 Dec 2022 14:24


Tgt Ion:266 Resp: 7506
 Ion Ratio Lower Upper
 266 100
 264 57.6 50.7 76.1
 268 72.1 58.3 87.5







#72
Phenanthrene
Concen: 4.845 ng/ul
RT: 17.736 min Scan# 2459
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24

Instrument :
BNA_G
ClientSampleId :







Tgt Ion:178 Resp: 56778
Ion Ratio Lower Upper
178 100
179 15.7 12.7 19.1
176 18.6 15.0 22.4



#74
Anthracene
Concen: 5.032 ng/ul
RT: 17.824 min Scan# 2474
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24

Tgt Ion:178 Resp: 60367
Ion Ratio Lower Upper
178 100
179 17.5 12.1 18.1
176 17.7 13.9 20.9





#75

1,2,3,4-Tetrachlorobenzene

Concen: 4.559 ng/uL

RT: 13.764 min Scan# 1

Delta R.T. 0.000 min

Lab File: BG056089.D





Acq: 24 Dec 2022 14:24

Instrument :

BNA_G

ClientSampleId :

Tgt Ion:	216	Resp:	15965
Ion Ratio	Lower	Upper	
216	100		
214	66.5	59.0	88.6
218	52.6	36.0	54.0



#76

Pentachlorobenzene

Concen: 4.479 ng/uL




RT: 15.268 min Scan# 2039


Delta R.T. 0.000 min

Lab File: BG056089.D

Acq: 24 Dec 2022 14:24

Tgt Ion:	250	Resp:	15482
Ion Ratio	Lower	Upper	
250	100		
248	64.4	54.4	81.6
252	64.8	59.1	88.7









#77
 Carbazole
 Concen: 5.043 ng/ul
 RT: 18.083 min Scan# 2519
 Delta R.T. -0.006 min
 Lab File: BG056089.D
 Acq: 24 Dec 2022 14:24




Instrument :
 BNA_G
 ClientSampleId :


Tgt Ion:167 Resp: 49462
 Ion Ratio Lower Upper
 167 100
 166 19.1 17.4 26.0
 139 10.9 8.3 12.5



#78
 Di-n-butylphthalate
 Concen: 4.450 ng/ul
 RT: 18.617 min Scan# 2609
 Delta R.T. 0.000 min
 Lab File: BG056089.D
 Acq: 24 Dec 2022 14:24


Tgt Ion:149 Resp: 39879
 Ion Ratio Lower Upper
 149 100
 150 8.3 7.8 11.6
 104 4.8 4.6 7.0








#80
Fluoranthene
Concen: 4.710 ng/ul
RT: 19.722 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24

Instrument :
BNA_G
ClientSampleId :







Tgt Ion:202 Resp: 84510
Ion Ratio Lower Upper
202 100
101 4.8 3.2 4.8#
100 4.3 3.0 4.4



#82
Pyrene
Concen: 4.972 ng/ul
RT: 20.080 min Scan# 2858
Delta R.T. -0.006 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24

Tgt Ion:202 Resp: 90001
Ion Ratio Lower Upper
202 100
101 4.2 3.6 5.4
100 4.9 3.3 4.9









#83
Butylbenzylphthalate
Concen: 4.232 ng/ul
RT: 20.944 min Scan# 3005
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:149 Resp: 16301
Ion Ratio Lower Upper
149 100
91 52.6 46.9 70.3
206 35.2 24.6 36.8



#84
3,3'-Dichlorobenzidine
Concen: 5.043 ng/ul
RT: 21.872 min Scan# 3163
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24

Tgt Ion:252 Resp: 31307
Ion Ratio Lower Upper
252 100
254 60.0 50.8 76.2
126 2.3 3.4 5.0#





#85

Benzo(a)anthracene

Concen: 4.870 ng/ul

RT: 21.972 min Scan# 3180

Delta R.T. 0.000 min


Lab File: BG056089.D

Acq: 24 Dec 2022 14:24

Instrument :

BNA_G

ClientSampleId :






Tgt Ion:228 Resp: 91841

Ion Ratio Lower Upper

228 100

229 16.5 15.5 23.3

226 26.2 22.3 33.5



#86

Bis(2-ethylhexyl)phthalate

Concen: 4.057 ng/ul

RT: 21.843 min Scan# 3158

Delta R.T. 0.000 min

Lab File: BG056089.D

Acq: 24 Dec 2022 14:24




Tgt Ion:149 Resp: 23939


Ion Ratio Lower Upper

149 100

167 34.4 23.7 35.5


279 8.1 5.4 8.0#








#87
Chrysene
Concen: 4.962 ng/ul
RT: 22.037 min Scan# 3192
Delta R.T. -0.006 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24

Instrument :
BNA_G
ClientSampleId :







Tgt Ion:228 Resp: 86838
Ion Ratio Lower Upper
228 100
226 28.9 23.9 35.9
229 17.9 16.2 24.4



#89
Di-n-octyl phthalate
Concen: 3.888 ng/ul
RT: 23.142 min Scan# 3379
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24

Tgt Ion:149 Resp: 43720









#90
Benzo(b)fluoranthene
Concen: 4.889 ng/ul
RT: 24.352 min Scan# 3585
Delta R.T. 0.000 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:252 Resp: 98062
Ion Ratio Lower Upper
252 100
253 20.7 18.5 27.7
125 2.3 2.2 3.4



#91
Benzo(k)fluoranthene
Concen: 5.013 ng/ul
RT: 24.352 min Scan# 3585
Delta R.T. -0.076 min
Lab File: BG056089.D
Acq: 24 Dec 2022 14:24

Tgt Ion:252 Resp: 98062
Ion Ratio Lower Upper
252 100
253 20.7 17.6 26.4
125 2.3 2.0 3.0





#93

Benzo(a)pyrene

Concen: 4.894 ng/ul

RT: 25.298 min Scan# 31

Delta R.T. 0.000 min


Lab File: BG056089.D

Acq: 24 Dec 2022 14:24

Instrument :

BNA_G

ClientSampleId :






Tgt Ion:252 Resp: 85552

Ion Ratio Lower Upper

252 100

253 20.5 17.0 25.6

125 1.6 2.3 3.5#



#94

Indeno(1,2,3-cd)pyrene

Concen: 2.416 ng/ul

RT: 29.446 min Scan# 4452

Delta R.T. -0.006 min

Lab File: BG056089.D

Acq: 24 Dec 2022 14:24




Tgt Ion:276 Resp: 55577


Ion Ratio Lower Upper

276 100

138 5.4 4.4 6.6


277 24.3 20.7 31.1








#95
 Dibenzo(a,h)anthracene
 Concen: 2.464 ng/ul
 RT: 29.511 min Scan# 4463
 Delta R.T. -0.018 min
 Lab File: BG056089.D
 Acq: 24 Dec 2022 14:24

Instrument :
 BNA_G
 ClientSampleId :



Tgt Ion:278 Resp: 47454
 Ion Ratio Lower Upper
 278 100
 139 4.6 4.3 6.5
 279 22.2 18.3 27.5



#96
 Benzo(g,h,i)perylene
 Concen: 2.534 ng/ul
 RT: 30.692 min Scan# 4664
 Delta R.T. -0.018 min
 Lab File: BG056089.D
 Acq: 24 Dec 2022 14:24

Tgt Ion:276 Resp: 46804
 Ion Ratio Lower Upper
 276 100
 138 4.7 5.1 7.7#
 277 23.6 18.5 27.7

