

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122422\
 Data File : BG056113.D
 Acq On : 25 Dec 2022 8:00
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :

Quant Time: Dec 25 22:36:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG122422.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Dec 24 22:54:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.350	152	47825	20.000	ng/ul	0.00
20) Naphthalene-d8	11.182	136	194884	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.954	164	169079	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.692	188	437678	20.000	ng/ul	0.00
79) Chrysene-d12	21.993	240	447358	20.000	ng/ul	# 0.00
88) Perylene-d12	25.459	264	480651	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.649	96	8685	7.602	ng/uL	0.00
4) Pyridine-d5	4.084	84	64930	17.430	ng/ul	0.00
7) Phenol-d5	7.474	99	85718	19.759	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.650	67	33708	19.714	ng/ul	0.00
11) 2-Chlorophenol-d4	7.868	132	54775	21.363	ng/ul	0.00
15) 4-Methylphenol-d8	9.037	113	66478	20.361	ng/ul	0.00
21) Nitrobenzene-d5	9.519	128	29022	20.446	ng/ul	0.00
24) 2-Nitrophenol-d4	10.242	143	37805	20.385	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.782	165	81495	19.775	ng/ul	0.00
31) 4-Chloroaniline-d4	11.305	131	85835	18.415	ng/ul	0.00
46) Dimethylphthalate-d6	14.343	166	255057	19.491	ng/ul	0.00
49) Acenaphthylene-d8	14.654	160	289301	19.576	ng/ul	0.00
54) 4-Nitrophenol-d4	15.118	143	40687	20.655	ng/ul	0.00
60) Fluorene-d10	15.941	176	249272	19.298	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	16.041	200	55472	19.716	ng/ul	0.00
73) Anthracene-d10	17.792	188	403434	19.947	ng/ul	0.00
81) Pyrene-d10	20.054	212	496933	20.113	ng/ul	0.00
92) Benzo(a)pyrene-d12	25.224	264	477180	19.532	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.685	88	8404	6.576	ng/uL#	91
5) Pyridine	4.102	79	64604	18.033	ng/ul#	94
6) Benzaldehyde	7.468	77	31031	19.484	ng/ul	93
8) Phenol	7.504	94	86169	19.958	ng/ul	97
10) Bis(2-Chloroethyl)ether	7.744	93	56069	19.729	ng/ul	98
12) 2-Chlorophenol	7.903	128	54787	21.274	ng/ul	97
13) 2-Methylphenol	8.773	108	64240	19.817	ng/ul	93
14) 2,2'-oxybis(1-Chloropr...	8.867	45	8778	24.047	ng/ul#	66
16) Acetophenone	9.172	105	103602	19.157	ng/ul	96
17) N-Nitroso-di-n-propyla...	9.149	70	46667	20.409	ng/ul	97
18) 4-Methylphenol	9.102	108	67118	19.114	ng/ul	85
19) Hexachloroethane	9.448	117	22136	21.378	ng/ul	94
22) Nitrobenzene	9.560	77	68494	19.447	ng/ul#	97
23) Isophorone	10.083	82	144556	17.951	ng/ul	99
25) 2-Nitrophenol	10.277	139	35585	19.878	ng/ul#	79
26) 2,4-Dimethylphenol	10.318	107	81485	19.236	ng/ul	96
27) Bis(2-Chloroethoxy)met...	10.559	93	77611	18.597	ng/ul#	98
29) 2,4-Dichlorophenol	10.811	162	78050	19.830	ng/ul#	91
30) Naphthalene	11.229	128	198645	19.562	ng/ul	99
32) 4-Chloroaniline	11.329	127	87906	18.715	ng/ul	96
33) Hexachlorobutadiene	11.511	225	67021	18.170	ng/ul	98
34) Caprolactam	12.069	113	24602	19.550	ng/ul#	85
35) 4-Chloro-3-methylphenol	12.415	107	74658	19.744	ng/ul	99
36) 2-Methylnaphthalene	12.809	142	156165	19.182	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122422\
 Data File : BG056113.D
 Acq On : 25 Dec 2022 8:00
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :

Quant Time: Dec 25 22:36:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG122422.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Dec 24 22:54:38 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	13.027	142	158102	19.278	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	13.168	216	131368	18.839	ng/ul	97
40) Hexachlorocyclopentadiene	13.144	237	80737	17.857	ng/ul	96
41) 2,4,6-Trichlorophenol	13.397	196	82947	20.182	ng/ul#	88
42) 2,4,5-Trichlorophenol	13.467	196	90913	20.401	ng/ul	90
43) 1,1'-Biphenyl	13.790	154	236383	19.775	ng/ul	98
44) 2-Chloronaphthalene	13.843	162	196855	19.869	ng/ul	98
45) 2-Nitroaniline	14.031	65	34942	25.396	ng/ul#	68
47) Dimethylphthalate	14.390	163	252495	19.458	ng/ul	98
48) 2,6-Dinitrotoluene	14.513	165	52561	20.521	ng/ul	93
50) Acenaphthylene	14.683	152	291396	19.841	ng/ul	98
51) 3-Nitroaniline	14.848	138	40231	20.911	ng/ul	99
52) Acenaphthene	15.018	153	202295	19.607	ng/ul	94
53) 2,4-Dinitrophenol	15.048	184	36024	18.070	ng/ul#	94
55) 4-Nitrophenol	15.136	109	38932	20.315	ng/ul	98
56) Dibenzofuran	15.347	168	313571	19.417	ng/ul	97
57) 2,4-Dinitrotoluene	15.294	165	72684	20.464	ng/ul#	95
58) 2,3,4,6-Tetrachlorophenol	15.565	232	86797	19.705	ng/ul#	99
59) Diethylphthalate	15.741	149	233048	20.102	ng/ul	98
61) Fluorene	15.994	166	251154	18.933	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.976	204	160202	18.589	ng/ul	97
63) 4-Nitroaniline	16.000	138	33167	18.067	ng/ul	88
66) 4,6-Dinitro-2-methylph...	16.058	198	54129	20.066	ng/ul#	94
67) N-Nitrosodiphenylamine	16.188	169	220833	19.914	ng/ul	98
68) 4-Bromophenyl-phenylether	16.869	248	106829	19.252	ng/ul	97
69) Hexachlorobenzene	16.992	284	106815	19.889	ng/ul	94
70) Atrazine	17.122	200	107259	20.285	ng/ul	96
71) Pentachlorophenol	17.333	266	65898	19.300	ng/ul	96
72) Phenanthrene	17.733	178	430067	19.355	ng/ul	98
74) Anthracene	17.827	178	447036	19.653	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.767	216	136049	20.493	ng/ul	98
76) Pentachlorobenzene	15.265	250	133054	20.302	ng/ul	89
77) Carbazole	18.085	167	374464	20.136	ng/ul	99
78) Di-n-butylphthalate	18.620	149	390411	22.978	ng/ul	99
80) Fluoranthene	19.719	202	579318	19.941	ng/ul#	98
82) Pyrene	20.083	202	572588	19.537	ng/ul#	97
83) Butylbenzylphthalate	20.947	149	169442	27.168	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.869	252	212321	21.124	ng/ul#	97
85) Benzo(a)anthracene	21.969	228	598462	19.600	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.840	149	245157	25.659	ng/ul	99
87) Chrysene	22.040	228	559762	19.753	ng/ul	98
89) Di-n-octyl phthalate	23.138	149	416412	23.411	ng/ul	100
90) Benzo(b)fluoranthene	24.349	252	620060	19.542	ng/ul	96
91) Benzo(k)fluoranthene	24.425	252	609743	19.692	ng/ul#	99
93) Benzo(a)pyrene	25.294	252	531774	19.233	ng/ul#	99
94) Indeno(1,2,3-cd)pyrene	29.454	276	691567	18.990	ng/ul#	96
95) Dibenzo(a,h)anthracene	29.525	278	581537	19.092	ng/ul	97
96) Benzo(g,h,i)perylene	30.712	276	563688	19.260	ng/ul	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122422\
Data File : BG056113.D
Acq On : 25 Dec 2022 8:00
Operator : CG/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :


Quant Time: Dec 25 22:36:31 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG122422.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Dec 24 22:54:38 2022
Response via : Initial Calibration








#2
1,4-Dioxane
Concen: 6.576 ng/uL
RT: 3.685 min Scan# 61
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Instrument :
BNA_G
ClientSampleId :







Tgt Ion: 88 Resp: 8404
Ion Ratio Lower Upper
88 100
43 10.1 7.4 11.0
58 29.4 19.1 28.7#



#5
Pyridine
Concen: 18.033 ng/uL
RT: 4.102 min Scan# 138
Delta R.T. -0.009 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Tgt Ion: 79 Resp: 64604
Ion Ratio Lower Upper
79 100
52 17.2 11.0 16.6#
51 12.4 11.1 16.7









#6
Benzaldehyde
Concen: 19.484 ng/ul
RT: 7.468 min Scan# 712
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion: 77 Resp: 31031
Ion Ratio Lower Upper
77 100
105 101.8 72.6 109.0
106 94.1 76.8 115.2



#8
Phenol
Concen: 19.958 ng/ul
RT: 7.504 min Scan# 717
Delta R.T. 0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Tgt Ion: 94 Resp: 86169
Ion Ratio Lower Upper
94 100
65 20.4 18.8 28.2
66 33.3 26.2 39.2









#10
Bis(2-Chloroethyl)ether
Concen: 19.729 ng/ul
RT: 7.744 min Scan# 759
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion	Ratio	Lower	Upper
93	100		
63	47.1	37.4	56.2
95	35.8	26.3	39.5



#12
2-Chlorophenol
Concen: 21.274 ng/ul
RT: 7.903 min Scan# 785
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Tgt Ion	Ratio	Lower	Upper
128	100		
64	32.7	24.8	37.2
130	32.8	27.9	41.9









#13
2-Methylphenol
Concen: 19.817 ng/ul
RT: 8.773 min Scan# 913
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:108 Resp: 64240
Ion Ratio Lower Upper
108 100
107 95.7 71.4 107.0



#14
2,2'-oxybis(1-Chloropropane)
Concen: 24.047 ng/ul
RT: 8.867 min Scan# 949
Delta R.T. 0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Tgt Ion: 45 Resp: 8778
Ion Ratio Lower Upper
45 100
77 110.8 105.8 158.8
79 82.3 114.2 171.4#









#16
Acetophenone
Concen: 19.157 ng/ul
RT: 9.172 min Scan# 1001
Delta R.T. 0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:105 Resp: 103602
Ion Ratio Lower Upper
105 100
77 70.0 59.0 88.4
51 12.3 9.6 14.4



#17
N-Nitroso-di-n-propylamine
Concen: 20.409 ng/ul
RT: 9.149 min Scan# 997
Delta R.T. 0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Tgt Ion: 70 Resp: 46667
Ion Ratio Lower Upper
70 100
42 19.4 16.1 24.1
101 11.8 8.4 12.6
130 22.7 16.8 25.2





#18

4-Methylphenol

Concen: 19.114 ng/ul

RT: 9.102 min Scan# 91

Delta R.T. -0.003 min

Lab File: BG056113.D

Acq: 25 Dec 2022 8:00

Instrument :

BNA_G





ClientSampleId :

Tgt Ion:108 Resp: 67118

Ion Ratio Lower Upper

108 100

107 123.2 86.2 129.2



#19

Hexachloroethane

Concen: 21.378 ng/ul

RT: 9.448 min Scan# 1048

Delta R.T. 0.003 min

Lab File: BG056113.D

Acq: 25 Dec 2022 8:00




Tgt Ion:117 Resp: 22136


Ion Ratio Lower Upper

117 100

201 173.1 149.0 223.4

199 110.9 87.7 131.5









#22
Nitrobenzene
Concen: 19.447 ng/ul
RT: 9.560 min Scan# 1068
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion: 77 Resp: 68494
Ion Ratio Lower Upper
77 100
123 46.5 36.0 54.0
65 11.5 7.6 11.4#



#23
Isophorone
Concen: 17.951 ng/ul
RT: 10.083 min Scan# 1156
Delta R.T. 0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Tgt Ion: 82 Resp: 144556
Ion Ratio Lower Upper
82 100
95 8.5 7.9 11.9
138 19.6 15.8 23.6









#25
2-Nitrophenol
Concen: 19.878 ng/ul
RT: 10.277 min Scan# 1189
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:139 Resp: 35585
Ion Ratio Lower Upper
139 100
65 41.9 23.2 34.8#
109 39.4 23.8 35.8#



#26
2,4-Dimethylphenol
Concen: 19.236 ng/ul
RT: 10.318 min Scan# 1196
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Tgt Ion:107 Resp: 81485
Ion Ratio Lower Upper
107 100
121 50.4 36.5 54.7
122 79.1 64.4 96.6





#27

Bis(2-Chloroethoxy)methane

Concen: 18.597 ng/ul

RT: 10.559 min Scan# 11

Delta R.T. 0.003 min

Lab File: BG056113.D

Acq: 25 Dec 2022 8:00

Instrument :

BNA_G

ClientSampleId :





Tgt Ion: 93 Resp: 77611

Ion Ratio Lower Upper

93 100

95 33.1 26.8 40.2

123 11.1 7.4 11.0#



#29

2,4-Dichlorophenol

Concen: 19.830 ng/ul

RT: 10.811 min Scan# 1280

Delta R.T. -0.003 min

Lab File: BG056113.D

Acq: 25 Dec 2022 8:00




Tgt Ion:162 Resp: 78050


Ion Ratio Lower Upper

162 100

164 65.0 47.3 70.9


98 27.5 17.8 26.8#








#30
Naphthalene
Concen: 19.562 ng/ul
RT: 11.229 min Scan# 1351
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Instrument :
BNA_G
ClientSampleId :







Tgt Ion:128 Resp: 198645
Ion Ratio Lower Upper
128 100
129 11.0 8.4 12.6
127 14.2 11.6 17.4



#32
4-Chloroaniline
Concen: 18.715 ng/ul
RT: 11.329 min Scan# 1368
Delta R.T. 0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00


Tgt Ion:127 Resp: 87906
Ion Ratio Lower Upper
127 100
129 31.5 23.4 35.0








#33
Hexachlorobutadiene
Concen: 18.170 ng/ul
RT: 11.511 min Scan# 11
Delta R.T. 0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Instrument :
BNA_G
ClientSampleId :







Tgt Ion:225 Resp: 67021
Ion Ratio Lower Upper
225 100
223 60.8 51.2 76.8
227 62.5 50.5 75.7



#34
Caprolactam
Concen: 19.550 ng/ul
RT: 12.069 min Scan# 1494
Delta R.T. -0.009 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Tgt Ion:113 Resp: 24602
Ion Ratio Lower Upper
113 100
55 44.6 28.0 42.0#
56 42.3 27.7 41.5#









#35
4-Chloro-3-methylphenol
Concen: 19.744 ng/ul
RT: 12.415 min Scan# 1554
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:107 Resp: 74658
Ion Ratio Lower Upper
107 100
144 26.3 20.0 30.0
142 76.5 60.7 91.1



#36
2-Methylnaphthalene
Concen: 19.182 ng/ul
RT: 12.809 min Scan# 1620
Delta R.T. 0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Tgt Ion:142 Resp: 156165
Ion Ratio Lower Upper
142 100
141 88.3 69.8 104.6





#37

1-Methylnaphthalene

Concen: 19.278 ng/ul

RT: 13.027 min Scan# 1

Delta R.T. 0.003 min

Lab File: BG056113.D

Acq: 25 Dec 2022 8:00

Instrument :

BNA_G

ClientSampleId :





Tgt Ion:142 Resp: 158102

Ion Ratio Lower Upper

142 100

141 94.6 76.7 115.1

116 3.9 3.8 5.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 18.839 ng/ul

RT: 13.168 min Scan# 1681

Delta R.T. 0.003 min

Lab File: BG056113.D

Acq: 25 Dec 2022 8:00

Tgt Ion:216 Resp: 131368




Ion Ratio Lower Upper


216 100

214 78.7 61.0 91.4

179 17.5 13.4 20.0

108 10.6 9.1 13.7









#40
Hexachlorocyclopentadiene
Concen: 17.857 ng/ul
RT: 13.144 min Scan# 1
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:237 Resp: 80737
Ion Ratio Lower Upper
237 100
235 62.0 52.0 78.0
272 10.4 9.9 14.9



#41
2,4,6-Trichlorophenol
Concen: 20.182 ng/ul
RT: 13.397 min Scan# 1720
Delta R.T. 0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Tgt Ion:196 Resp: 82947
Ion Ratio Lower Upper
196 100
198 99.0 70.2 105.4
200 35.5 23.4 35.2#









#42
2,4,5-Trichlorophenol
Concen: 20.401 ng/ul
RT: 13.467 min Scan# 1732
Delta R.T. 0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:196 Resp: 90913
Ion Ratio Lower Upper
196 100
198 96.6 87.4 131.2
200 32.5 25.0 37.6



#43
1,1'-Biphenyl
Concen: 19.775 ng/ul
RT: 13.790 min Scan# 1787
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Tgt Ion:154 Resp: 236383
Ion Ratio Lower Upper
154 100
153 40.7 31.3 46.9
76 9.2 7.4 11.0









#44
2-Chloronaphthalene
Concen: 19.869 ng/ul
RT: 13.843 min Scan# 1796
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:162 Resp: 196855
Ion Ratio Lower Upper
162 100
164 33.0 27.0 40.4
127 31.0 23.4 35.2



#45
2-Nitroaniline
Concen: 25.396 ng/ul
RT: 14.031 min Scan# 1828
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Tgt Ion: 65 Resp: 34942
Ion Ratio Lower Upper
65 100
92 92.7 92.6 138.8
138 145.7 162.2 243.4#









#47
Dimethylphthalate
Concen: 19.458 ng/ul
RT: 14.390 min Scan# 1890
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:163 Resp: 252495
Ion Ratio Lower Upper
163 100
194 4.2 3.2 4.8
164 10.6 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 20.521 ng/ul
RT: 14.513 min Scan# 1910
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Tgt Ion:165 Resp: 52561
Ion Ratio Lower Upper
165 100
89 51.8 37.0 55.6
121 18.5 13.6 20.4









#50
Acenaphthylene
Concen: 19.841 ng/ul
RT: 14.683 min Scan# 1939
Delta R.T. 0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:152 Resp: 291396
Ion Ratio Lower Upper
152 100
151 19.7 16.2 24.2
153 13.4 9.7 14.5



#51
3-Nitroaniline
Concen: 20.911 ng/ul
RT: 14.848 min Scan# 1967
Delta R.T. 0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Tgt Ion:138 Resp: 40231
Ion Ratio Lower Upper
138 100
108 8.5 8.0 12.0
92 108.6 86.3 129.5









#52
Acenaphthene
Concen: 19.607 ng/ul
RT: 15.018 min Scan# 1996
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:153 Resp: 202295
Ion Ratio Lower Upper
153 100
152 45.7 34.5 51.7
154 86.5 63.9 95.9



#53
2,4-Dinitrophenol
Concen: 18.070 ng/ul
RT: 15.048 min Scan# 2001
Delta R.T. 0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Tgt Ion:184 Resp: 36024
Ion Ratio Lower Upper
184 100
63 31.8 16.8 25.2#
154 58.9 47.1 70.7









#55
4-Nitrophenol
Concen: 20.315 ng/ul
RT: 15.136 min Scan# 2015
Delta R.T. 0.008 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:109 Resp: 38932
Ion Ratio Lower Upper
109 100
139 97.9 79.5 119.3
65 73.8 56.7 85.1



#56
Dibenzofuran
Concen: 19.417 ng/ul
RT: 15.347 min Scan# 2052
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Tgt Ion:168 Resp: 313571
Ion Ratio Lower Upper
168 100
139 31.2 26.3 39.5





#57

2,4-Dinitrotoluene

Concen: 20.464 ng/ul

RT: 15.294 min Scan# 2044

Delta R.T. -0.003 min


Lab File: BG056113.D

Acq: 25 Dec 2022 8:00




Instrument :

BNA_G

ClientSampleId :



Tgt Ion:	165	Resp:	72684
Ion Ratio	Lower	Upper	
165	100		
63	28.7	20.9	31.3
182	2.9	1.8	2.8#



#58

2,3,4,6-Tetrachlorophenol

Concen: 19.705 ng/ul




RT: 15.565 min Scan# 2089


Delta R.T. -0.003 min

Lab File: BG056113.D

Acq: 25 Dec 2022 8:00


Tgt Ion:	232	Resp:	86797
Ion Ratio	Lower	Upper	
232	100		
131	26.0	20.2	30.4
130	1.7	1.0	1.6#
166	22.2	17.8	26.8








#59
Diethylphthalate
Concen: 20.102 ng/ul
RT: 15.741 min Scan# 2119
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Instrument :
BNA_G
ClientSampleId :







Tgt Ion:149 Resp: 233048
Ion Ratio Lower Upper
149 100
177 26.5 20.5 30.7
150 13.2 9.8 14.8



#61
Fluorene
Concen: 18.933 ng/ul
RT: 15.994 min Scan# 2162
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Tgt Ion:166 Resp: 251154
Ion Ratio Lower Upper
166 100
165 99.7 81.8 122.8
167 14.9 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 18.589 ng/ul
RT: 15.976 min Scan# 2160
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:204 Resp: 160202
Ion Ratio Lower Upper
204 100
206 30.3 26.3 39.5
141 39.3 30.5 45.7



#63
4-Nitroaniline
Concen: 18.067 ng/ul
RT: 16.000 min Scan# 2163
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00


Tgt Ion:138 Resp: 33167
Ion Ratio Lower Upper
138 100
92 42.3 32.6 49.0
108 98.1 66.3 99.5








#66
 4,6-Dinitro-2-methylphenol
 Concen: 20.066 ng/ul
 RT: 16.058 min Scan# 2173
 Delta R.T. 0.003 min
 Lab File: BG056113.D
 Acq: 25 Dec 2022 8:00


Instrument :
 BNA_G
 ClientSampleId :





Tgt Ion:198 Resp: 54129
 Ion Ratio Lower Upper
 198 100
 182 3.4 2.1 3.1#
 77 12.9 8.2 12.2#




#67
 N-Nitrosodiphenylamine
 Concen: 19.914 ng/ul
 RT: 16.188 min Scan# 2195
 Delta R.T. 0.003 min
 Lab File: BG056113.D
 Acq: 25 Dec 2022 8:00



Tgt Ion:169 Resp: 220833
 Ion Ratio Lower Upper
 169 100
 168 69.6 57.4 86.0
 167 38.1 31.1 46.7









#68
4-Bromophenyl-phenylether
Concen: 19.252 ng/ul
RT: 16.869 min Scan# 21
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:248 Resp: 106829
Ion Ratio Lower Upper
248 100
250 98.7 79.4 119.2
141 47.6 33.9 50.9



#69
Hexachlorobenzene
Concen: 19.889 ng/ul
RT: 16.992 min Scan# 2332
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Tgt Ion:284 Resp: 106815
Ion Ratio Lower Upper
284 100
142 23.9 18.1 27.1
249 36.1 25.2 37.8





#70

Atrazine

Concen: 20.285 ng/ul

RT: 17.122 min Scan# 21

Delta R.T. 0.003 min

Lab File: BG056113.D

Acq: 25 Dec 2022 8:00

Instrument :

BNA_G

ClientSampleId :





Tgt Ion:200 Resp: 107259

Ion Ratio Lower Upper

200 100

173 21.4 16.9 25.3

215 49.5 42.8 64.2



#71

Pentachlorophenol

Concen: 19.300 ng/ul

RT: 17.333 min Scan# 2390

Delta R.T. 0.003 min

Lab File: BG056113.D

Acq: 25 Dec 2022 8:00




Tgt Ion:266 Resp: 65898


Ion Ratio Lower Upper

266 100

264 61.5 50.7 76.1

268 67.9 58.3 87.5





#72

Phenanthrene

Concen: 19.355 ng/ul

RT: 17.733 min Scan# 2459

Delta R.T. -0.003 min

Lab File: BG056113.D

Acq: 25 Dec 2022 8:00

Instrument :

BNA_G

ClientSampleId :





Tgt Ion:178 Resp: 430067

Ion Ratio Lower Upper

178 100

179 15.8 12.7 19.1

176 19.9 15.0 22.4



#74

Anthracene

Concen: 19.653 ng/ul

RT: 17.827 min Scan# 2474

Delta R.T. 0.003 min

Lab File: BG056113.D

Acq: 25 Dec 2022 8:00




Tgt Ion:178 Resp: 447036


Ion Ratio Lower Upper

178 100

179 15.5 12.1 18.1

176 17.5 13.9 20.9









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.493 ng/uL
RT: 13.767 min Scan# 11
Delta R.T. 0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:216 Resp: 136049
Ion Ratio Lower Upper
216 100
214 75.8 59.0 88.6
218 44.9 36.0 54.0



#76
Pentachlorobenzene
Concen: 20.302 ng/uL
RT: 15.265 min Scan# 2038
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Tgt Ion:250 Resp: 133054
Ion Ratio Lower Upper
250 100
248 62.2 54.4 81.6
252 62.4 59.1 88.7









#77
 Carbazole
 Concen: 20.136 ng/ul
 RT: 18.085 min Scan# 2519
 Delta R.T. -0.003 min
 Lab File: BG056113.D
 Acq: 25 Dec 2022 8:00




Instrument :
 BNA_G
 ClientSampleId :


Tgt Ion:167 Resp: 374464
 Ion Ratio Lower Upper
 167 100
 166 21.3 17.4 26.0
 139 10.3 8.3 12.5



#78
 Di-n-butylphthalate
 Concen: 22.978 ng/ul
 RT: 18.620 min Scan# 2609
 Delta R.T. 0.003 min
 Lab File: BG056113.D
 Acq: 25 Dec 2022 8:00


Tgt Ion:149 Resp: 390411
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.8 11.6
 104 5.6 4.6 7.0








#80
Fluoranthene
Concen: 19.941 ng/ul
RT: 19.719 min Scan# 21
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Instrument :
BNA_G
ClientSampleId :







Tgt Ion:202 Resp: 579318
Ion Ratio Lower Upper
202 100
101 4.9 3.2 4.8#
100 4.3 3.0 4.4



#82
Pyrene
Concen: 19.537 ng/ul
RT: 20.083 min Scan# 2858
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Tgt Ion:202 Resp: 572588
Ion Ratio Lower Upper
202 100
101 5.5 3.6 5.4#
100 5.0 3.3 4.9#





#83

Butylbenzylphthalate

Concen: 27.168 ng/ul

RT: 20.947 min Scan# 3005

Delta R.T. 0.003 min

Lab File: BG056113.D





Acq: 25 Dec 2022 8:00

Instrument :

BNA_G

ClientSampleId :

Tgt Ion:	149	Resp:	169442
Ion Ratio	Lower	Upper	
149	100		
91	64.5	46.9	70.3
206	32.6	24.6	36.8



#84

3,3'-Dichlorobenzidine

Concen: 21.124 ng/ul




RT: 21.869 min Scan# 3162


Delta R.T. -0.003 min

Lab File: BG056113.D

Acq: 25 Dec 2022 8:00

Tgt Ion:	252	Resp:	212321
Ion Ratio	Lower	Upper	
252	100		
254	61.2	50.8	76.2
126	6.0	3.4	5.0









#85
Benzo(a)anthracene
Concen: 19.600 ng/ul
RT: 21.969 min Scan# 3180
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:228 Resp: 598462
Ion Ratio Lower Upper
228 100
229 19.3 15.5 23.3
226 29.1 22.3 33.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 25.659 ng/ul
RT: 21.840 min Scan# 3157
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00


Tgt Ion:149 Resp: 245157
Ion Ratio Lower Upper
149 100
167 29.6 23.7 35.5
279 5.8 5.4 8.0








#87
Chrysene
Concen: 19.753 ng/ul
RT: 22.040 min Scan# 3192
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Instrument :
BNA_G
ClientSampleId :







Tgt Ion:228 Resp: 559762
Ion Ratio Lower Upper
228 100
226 30.9 23.9 35.9
229 19.8 16.2 24.4



#89
Di-n-octyl phthalate
Concen: 23.411 ng/ul
RT: 23.138 min Scan# 3378
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Tgt Ion:149 Resp: 416412









#90
Benzo(b)fluoranthene
Concen: 19.542 ng/ul
RT: 24.349 min Scan# 3584
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00




Instrument :
BNA_G
ClientSampleId :


Tgt Ion:252 Resp: 620060
Ion Ratio Lower Upper
252 100
253 21.3 18.5 27.7
125 3.3 2.2 3.4



#91
Benzo(k)fluoranthene
Concen: 19.692 ng/ul
RT: 24.425 min Scan# 3597
Delta R.T. -0.003 min
Lab File: BG056113.D
Acq: 25 Dec 2022 8:00

Tgt Ion:252 Resp: 609743
Ion Ratio Lower Upper
252 100
253 22.4 17.6 26.4
125 3.1 2.0 3.0#





#93

Benzo(a)pyrene

Concen: 19.233 ng/ul

RT: 25.294 min Scan# 31

Delta R.T. -0.003 min

Lab File: BG056113.D

Acq: 25 Dec 2022 8:00

Instrument :

BNA_G

ClientSampleId :





Tgt Ion:252 Resp: 531774

Ion Ratio Lower Upper

252 100

253 20.9 17.0 25.6

125 3.7 2.3 3.5#



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.990 ng/ul

RT: 29.454 min Scan# 4453

Delta R.T. 0.003 min

Lab File: BG056113.D

Acq: 25 Dec 2022 8:00




Tgt Ion:276 Resp: 691567


Ion Ratio Lower Upper

276 100

138 8.0 4.4 6.6#

277 24.1 20.7 31.1





#95

Dibenzo(a,h)anthracene

Concen: 19.092 ng/ul

RT: 29.525 min Scan# 4466

Delta R.T. -0.003 min

Lab File: BG056113.D

Acq: 25 Dec 2022 8:00

Instrument :

BNA_G

ClientSampleId :





Tgt Ion:278 Resp: 581537

Ion Ratio Lower Upper

278 100

139 5.4 4.3 6.5

279 24.5 18.3 27.5



#96

Benzo(g,h,i)perylene

Concen: 19.260 ng/ul

RT: 30.712 min Scan# 4667

Delta R.T. 0.003 min

Lab File: BG056113.D

Acq: 25 Dec 2022 8:00

Tgt Ion:276 Resp: 563688

Ion Ratio Lower Upper

276 100

138 7.2 5.1 7.7

277 24.4 18.5 27.7

