



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122422\
Data File : BG056088.D
Acq On : 24 Dec 2022 13:43
Operator : CG/JU
Sample : DFTPP485
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP485

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG122422.M
Title : SVOA CALIBRATION
Last Update : Sat Dec 24 22:54:38 2022



AutoFind: Scans 2519, 2520, 2521; Background Corrected with Scan 2513

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	17.6	15320	PASS
68	69	0.00	2	1.5	333	PASS
69	198	0.00	100	25.1	21823	PASS
70	69	0.00	2	0.3	60	PASS
127	198	10	80	35.6	30933	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	86877	PASS
199	198	5	9	7.2	6215	PASS
275	198	10	60	36.1	31389	PASS
365	198	1	100	5.3	4564	PASS
441	198	0.01	100	15.3	13293	PASS
442	198	50	100	87.9	76395	PASS
443	442	15	24	20.5	15676	PASS

Scan 2513 (18.056 min): BG056088.D\data.ms
DFTPP485

m/z	Abundance
36.20	187.0
206.90	533.0
344.60	210.0

Instrument :

BNA_G

ClientSampleId :

DFTPP485

m/z	Abundance
37.90	302.0
39.00	1241.0
50.00	6225.0
51.00	16416.0
52.00	941.0
54.90	163.0
55.90	588.0
57.00	1107.0
62.10	718.0
63.00	942.0
65.00	623.0
66.90	194.0
67.90	693.0
68.90	24592.0
73.10	324.0
74.00	3849.0
75.00	5279.0
76.10	1425.0
77.00	33560.0
78.00	2499.0
79.00	2687.0
79.90	1705.0
81.00	3176.0
81.90	825.0
83.00	537.0
85.00	429.0
85.90	863.0
86.80	358.0
87.70	171.0
90.90	671.0
91.90	622.0
92.90	5111.0
94.00	378.0
96.00	253.0
97.90	5291.0
99.00	3117.0
99.70	235.0
101.00	1552.0
101.90	193.0
102.80	771.0
103.90	1331.0
104.90	915.0
107.00	15244.0
107.80	1341.0
109.90	16488.0
111.00	2379.0
112.00	199.0
115.80	596.0
116.90	17776.0
118.00	1623.0
120.00	228.0
121.90	1058.0
122.90	1418.0
123.90	734.0
125.00	669.0
126.90	33488.0
127.90	2598.0
129.00	16472.0
130.00	1114.0
131.00	367.0
131.80	275.0
132.90	216.0
133.80	627.0
135.00	1215.0
135.90	615.0
136.90	802.0
137.90	339.0
139.20	188.0
139.90	438.0
140.90	2968.0
141.90	831.0
142.80	372.0
143.00	393.0
144.00	196.0
145.10	204.0
146.00	990.0
146.90	1497.0
148.00	4883.0
148.90	510.0
150.90	260.0
151.10	276.0
151.70	176.0
152.90	1156.0
153.80	550.0
155.00	1222.0
155.90	1714.0
156.90	362.0
157.90	418.0
158.90	453.0
159.90	1006.0
161.00	1189.0
161.70	430.0

Instrument :
BNA_G
ClientSampleId :
DFTPP485

163.60	217.0
164.90	1288.0
166.00	795.0
167.00	6637.0
168.00	2515.0
169.00	718.0
169.60	194.0
171.00	403.0
172.00	623.0
173.10	773.0
173.90	1031.0
175.00	2771.0
175.90	523.0
176.90	1147.0
177.90	471.0
178.90	4249.0
179.90	2935.0
180.90	1032.0
182.10	184.0
184.10	397.0
184.90	2095.0
186.00	13838.0
186.90	4475.0
188.00	573.0
188.90	1411.0
190.00	351.0
191.00	920.0
191.90	1770.0
192.90	1629.0
193.80	463.0
195.90	2913.0
197.90	89456.0
198.90	6371.0
199.90	527.0
201.10	438.0
201.70	240.0
202.90	1205.0
203.90	4353.0
205.00	7760.0
206.00	30536.0
207.00	4230.0
207.90	1255.0
208.90	681.0
209.90	728.0
211.00	1720.0
211.70	238.0
213.00	196.0
214.80	502.0
215.90	732.0
216.90	8772.0
218.00	1164.0
220.90	3903.0
221.80	1549.0
222.90	2514.0
224.00	20088.0
225.00	5347.0
226.00	521.0
227.00	8769.0
227.90	1152.0
228.90	1837.0
229.90	334.0
230.90	466.0
232.10	311.0
233.90	872.0
234.90	587.0
236.10	496.0
236.90	630.0
239.10	212.0
239.90	330.0
240.90	717.0
241.90	1409.0
243.10	912.0
243.90	12676.0
245.00	1545.0
245.90	3752.0
246.80	528.0
248.00	186.0
249.10	370.0
250.90	239.0
251.80	342.0
252.80	433.0
254.90	72048.0
255.90	10500.0
256.90	1016.0
257.90	5104.0
258.90	1091.0
264.90	1912.0
265.70	238.0
270.70	200.0
271.80	334.0
272.90	2325.0
273.90	5563.0
274.90	29240.0
275.90	4273.0
276.90	2975.0
278.10	463.0

Instrument :
BNA_G
ClientSampleId :
DFTPP485

282.80	497.0
283.80	175.0
285.00	681.0
289.00	262.0
291.00	200.0
292.00	196.0
292.90	642.0
293.80	318.0
294.70	174.0
295.90	13532.0
296.90	1888.0
300.80	172.0
301.80	203.0
303.00	1127.0
304.00	218.0
314.10	493.0
315.00	1282.0
315.90	790.0
321.20	203.0
323.00	2390.0
324.00	428.0
327.00	364.0
328.10	384.0
332.00	254.0
333.00	239.0
334.00	2086.0
334.90	320.0
335.10	319.0
340.80	406.0
345.90	731.0
351.90	611.0
353.10	468.0
353.90	1008.0
355.00	191.0
364.90	3830.0
365.90	569.0
371.00	357.0
371.90	1213.0
372.80	290.0
382.90	282.0
387.70	162.0
390.80	231.0
402.00	618.0
402.90	420.0
403.60	183.0
421.00	422.0
422.10	490.0
423.00	2515.0
424.10	610.0
441.00	11743.0
442.00	61816.0
443.00	12747.0
444.00	1367.0

Instrument :
BNA_G
ClientSampleId :
DFTPP485

m/z	Abundance
38.00	458.0
39.00	1097.0
40.00	162.0
48.80	271.0
50.00	5630.0
51.00	19408.0
51.90	1309.0
55.00	196.0
56.00	521.0
56.90	1299.0
61.00	298.0
61.90	238.0
63.00	1475.0
65.00	272.0
68.90	26224.0
72.80	235.0
74.00	4470.0
75.00	5051.0
76.00	2193.0
77.00	35792.0
78.00	2638.0
78.90	2492.0
79.90	1810.0
81.00	3228.0
82.00	761.0
83.00	367.0
85.00	391.0
85.90	1132.0
87.10	278.0
88.00	159.0
90.80	587.0
91.90	806.0
93.00	6948.0
94.00	315.0
95.80	328.0
98.00	4873.0
98.90	3148.0
101.00	1690.0
104.00	1083.0
104.90	1178.0
106.90	15934.0
107.90	1974.0
109.00	777.0
109.90	16600.0
111.00	3040.0
111.90	343.0
117.00	18048.0
118.00	795.0
119.00	238.0
119.90	167.0
122.00	1389.0
122.90	1562.0
123.80	898.0
124.80	891.0
126.90	37584.0
127.90	2860.0
128.90	20272.0
129.90	1567.0
131.00	520.0
131.80	156.0
132.90	180.0
133.80	772.0
134.80	1769.0
136.00	864.0
137.00	985.0
138.10	358.0
139.00	221.0
140.10	331.0
140.90	2799.0
141.90	925.0
142.90	535.0
144.00	187.0
145.90	892.0
147.00	1699.0
148.00	4319.0
148.80	771.0
149.90	160.0
151.00	280.0
152.00	242.0
152.90	887.0
153.90	767.0
154.90	1600.0
156.00	2812.0
156.90	390.0
157.80	918.0
159.00	547.0
159.80	944.0
161.00	1419.0
161.90	524.0
162.70	165.0
163.80	284.0
164.90	1399.0

Instrument :
BNA_G
ClientSampleId :
DFTPP485

166.00	870.0
166.90	6716.0
168.00	3141.0
168.90	391.0
170.00	275.0
170.90	415.0
171.90	682.0
172.90	745.0
173.90	1212.0
175.00	2901.0
176.00	846.0
176.90	1517.0
177.90	407.0
178.90	5515.0
179.90	3234.0
180.80	1554.0
181.80	361.0
182.90	159.0
183.90	680.0
185.00	2921.0
185.90	16440.0
187.00	4582.0
187.90	657.0
188.90	1590.0
190.00	464.0
190.90	851.0
191.90	1822.0
193.00	2600.0
193.90	681.0
194.90	241.0
195.90	3940.0
197.90	104424.0
198.90	7730.0
199.90	745.0
201.40	448.0
201.80	286.0
202.90	880.0
203.90	4308.0
204.90	10015.0
206.00	35800.0
207.00	4957.0
208.00	1639.0
208.90	535.0
209.80	336.0
211.00	1525.0
211.80	334.0
212.70	177.0
215.00	473.0
215.90	1575.0
216.90	12254.0
217.90	1435.0
220.90	4949.0
221.90	1644.0
223.00	2409.0
224.00	21480.0
225.00	6233.0
225.90	678.0
226.90	13311.0
227.90	1381.0
228.90	2467.0
229.90	366.0
231.00	1010.0
233.90	664.0
234.90	604.0
235.80	505.0
236.90	927.0
238.90	341.0
240.00	352.0
240.90	626.0
242.00	1154.0
243.00	1444.0
243.90	17848.0
244.90	2662.0
246.00	3993.0
246.90	629.0
247.80	216.0
248.90	771.0
252.70	482.0
254.90	92416.0
255.90	12883.0
256.90	1079.0
258.00	7096.0
258.90	1156.0
261.00	178.0
263.80	173.0
264.90	3778.0
266.00	340.0
269.70	211.0
270.10	227.0
271.00	292.0
271.70	204.0
272.90	2787.0
273.90	6890.0
274.90	38616.0
275.90	5091.0
277.00	4299.0

Instrument :
BNA_G
ClientSampleId :
DFTPP485

277.90	689.0
280.90	157.0
282.80	534.0
284.00	536.0
285.00	618.0
288.00	197.0
289.20	188.0
292.90	848.0
293.90	169.0
296.00	14010.0
296.90	1909.0
297.90	204.0
301.90	177.0
303.00	1216.0
304.10	185.0
309.00	151.0
312.30	207.0
313.90	554.0
314.90	1635.0
315.90	611.0
320.80	545.0
321.90	222.0
323.00	3604.0
324.00	537.0
327.00	779.0
327.90	319.0
332.00	409.0
332.80	625.0
333.90	2540.0
334.90	1037.0
340.80	363.0
342.00	252.0
345.90	737.0
347.00	234.0
351.90	958.0
352.80	585.0
353.90	611.0
354.90	301.0
364.90	5746.0
365.70	622.0
370.00	253.0
370.90	296.0
371.90	1681.0
373.00	366.0
382.90	336.0
383.90	163.0
390.80	272.0
401.90	461.0
403.00	648.0
403.90	189.0
420.90	489.0
421.80	657.0
423.00	3311.0
423.90	470.0
424.10	450.0
441.00	15634.0
442.00	90320.0
442.90	18680.0
444.10	1851.0
474.30	342.0

Instrument :
BNA_G
ClientSampleId :
DFTPP485

m/z	Abundance
39.00	676.0
50.10	3280.0
51.00	10136.0
52.00	559.0
55.10	154.0
55.90	337.0
57.00	803.0
61.00	293.0
61.90	461.0
63.00	775.0
64.00	157.0
64.80	233.0
68.00	306.0
68.90	14653.0
70.00	180.0
73.10	157.0
74.00	2105.0
75.00	2754.0
76.00	1100.0
77.00	20544.0
78.00	1838.0
78.90	1504.0
79.90	1135.0
80.90	1141.0
81.90	539.0
83.10	330.0
84.90	374.0
86.00	571.0
87.00	249.0
88.00	188.0
90.80	390.0
91.90	502.0
93.00	3618.0
94.00	251.0
95.80	288.0
97.90	3011.0
99.00	2277.0
99.80	183.0
100.90	1074.0
103.00	386.0
104.00	716.0
105.00	525.0
105.90	341.0
106.90	8900.0
107.90	1063.0
108.90	390.0
109.90	10867.0
110.90	1496.0
112.00	184.0
116.00	508.0
117.00	10691.0
118.00	1036.0
118.90	229.0
122.00	476.0
123.00	914.0
123.80	367.0
124.90	396.0
126.90	21728.0
128.00	1251.0
128.90	11350.0
129.90	710.0
133.90	571.0
134.90	881.0
135.90	352.0
137.00	575.0
140.00	214.0
140.90	2413.0
141.80	730.0
142.90	541.0
146.00	380.0
147.00	1054.0
147.90	2975.0
148.90	511.0
150.90	463.0
152.90	736.0
154.00	371.0
155.00	1415.0
156.00	1529.0
157.10	352.0
157.90	616.0
158.90	470.0
160.00	641.0
160.90	894.0
161.80	196.0
164.90	1094.0
165.90	506.0
166.90	4118.0
168.00	2068.0
168.90	441.0
169.90	197.0
170.90	323.0
171.90	525.0

Instrument :
BNA_G
ClientSampleId :
DFTPP485

172.90	441.0
174.00	1065.0
175.00	1786.0
175.70	316.0
176.90	854.0
177.50	167.0
178.90	3301.0
180.00	1845.0
180.90	875.0
182.10	251.0
183.60	206.0
183.90	179.0
185.00	1739.0
186.00	12655.0
187.00	3581.0
188.00	323.0
189.00	676.0
190.80	512.0
192.00	1144.0
193.10	1205.0
194.00	414.0
195.90	2691.0
197.90	66752.0
198.90	4545.0
199.90	385.0
201.30	276.0
201.90	320.0
203.00	931.0
204.00	3697.0
205.00	5715.0
206.00	23880.0
207.00	3576.0
207.90	1047.0
208.90	545.0
210.10	335.0
211.10	753.0
211.70	187.0
214.80	311.0
216.00	917.0
216.90	7271.0
217.80	898.0
220.90	3253.0
221.80	1123.0
223.00	1855.0
223.90	14140.0
224.90	4388.0
225.90	216.0
226.90	7850.0
227.80	1085.0
228.90	1239.0
229.90	327.0
230.90	602.0
233.30	215.0
234.00	390.0
234.80	527.0
235.70	241.0
236.90	635.0
238.80	327.0
239.90	244.0
240.80	677.0
241.90	422.0
243.10	1527.0
243.90	12182.0
245.00	1336.0
245.90	2337.0
247.00	597.0
247.90	280.0
248.80	479.0
252.00	198.0
252.90	334.0
254.90	63840.0
255.90	8587.0
257.10	642.0
257.90	4350.0
259.00	711.0
260.70	297.0
264.90	2284.0
265.90	296.0
270.90	505.0
271.90	247.0
272.90	2128.0
273.90	5462.0
274.90	26312.0
275.90	3221.0
276.90	3181.0
277.80	654.0
281.80	210.0
282.80	395.0
283.90	329.0
284.90	423.0
286.10	176.0
290.00	276.0
292.80	366.0
293.10	342.0
293.90	181.0
294.80	186.0

Instrument :
BNA_G
ClientSampleId :
DFTPP485

295.90	10688.0
297.00	1853.0
298.10	174.0
302.10	160.0
303.00	1460.0
303.70	171.0
304.10	160.0
313.80	517.0
314.90	1380.0
315.80	533.0
321.00	384.0
321.90	220.0
323.00	2582.0
324.10	538.0
325.80	200.0
326.80	989.0
327.80	220.0
331.60	255.0
333.00	186.0
334.00	1673.0
334.80	454.0
340.90	370.0
346.00	878.0
346.90	271.0
352.00	647.0
353.10	759.0
353.90	762.0
355.20	172.0
364.90	4116.0
366.00	521.0
370.20	164.0
370.80	205.0
371.90	1017.0
372.90	315.0
383.00	284.0
401.90	462.0
402.90	592.0
404.00	200.0
421.00	592.0
421.80	656.0
422.90	2498.0
424.00	737.0
441.00	12502.0
442.00	77048.0
443.00	15601.0
444.00	1371.0

Instrument :
BNA_G
ClientSampleId :
DFTPP485