



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122721\
 Data File : BG051794.D
 Acq On : 27 Dec 2021 9:03
 Operator : CG/JU
 Sample : DFTPP446
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP446

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121421.M
 Title : SVOA CALIBRATION
 Last Update : Fri Dec 31 00:13:36 2021



AutoFind: Scans 2483, 2484, 2485; Background Corrected with Scan 2477

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	43.6	35235	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	45.3	36589	PASS
70	69	0.00	2	0.4	132	PASS
127	198	10	80	49.0	39600	PASS
197	198	0.00	2	0.2	186	PASS
198	198	100	100	100.0	80755	PASS
199	198	5	9	7.1	5730	PASS
275	198	10	60	31.1	25075	PASS
365	198	1	100	3.8	3074	PASS
441	198	0.01	100	12.3	9967	PASS
442	198	50	100	78.2	63157	PASS
443	442	15	24	19.7	12461	PASS

m/z	Abundance
40.15	179.0
43.25	220.0
44.15	390.0
55.25	179.0
57.15	201.0
67.45	161.0
83.15	165.0
148.35	176.0
207.25	160.0

Instrument :
BNA_G
ClientSampleId :
DFTPP446

m/z	Abundance
38.05	462.0
39.15	2830.0
39.85	208.0
40.15	212.0
41.15	235.0
43.25	263.0
43.95	469.0
50.15	10343.0
51.15	34792.0
52.15	1580.0
55.25	166.0
56.05	1179.0
57.15	2154.0
61.05	545.0
62.15	594.0
63.15	1321.0
64.15	196.0
65.15	546.0
66.85	173.0
67.15	171.0
69.05	36392.0
73.35	347.0
74.15	3415.0
75.05	6116.0
76.15	2360.0
77.15	39792.0
78.15	2458.0
79.05	2575.0
80.05	2196.0
81.15	2957.0
82.05	681.0
83.05	733.0
85.15	368.0
85.95	804.0
86.95	523.0
87.95	186.0
91.05	625.0
92.15	834.0
93.05	5000.0
94.05	410.0
95.05	206.0
98.05	3198.0
99.05	2804.0
100.05	400.0
101.05	1358.0
103.05	474.0
104.05	1227.0
105.05	877.0
106.05	285.0
107.05	12396.0
108.05	1647.0
110.05	18304.0
111.05	3263.0
112.05	414.0
117.05	9644.0
118.05	703.0
120.15	226.0
121.95	984.0
123.05	1349.0
124.05	490.0
125.05	464.0
127.05	36904.0
128.05	3187.0
129.05	14743.0
130.05	1525.0
131.05	337.0
134.05	558.0
135.05	1208.0
136.15	585.0
136.95	636.0
141.05	2317.0
142.05	615.0
143.05	405.0
146.15	362.0
147.05	1433.0
148.05	2205.0
149.05	647.0
150.05	318.0
151.05	354.0
151.85	228.0
153.15	977.0
154.05	582.0
155.05	1358.0
156.05	1787.0
156.95	303.0
158.05	391.0
159.15	250.0
160.05	644.0
160.95	878.0
162.15	279.0
165.05	929.0
166.05	1048.0

Instrument :

BNA_G

ClientSampleId :

DFTPP446

167.05	3881.0
168.05	1924.0
168.85	259.0
170.05	193.0
170.75	211.0
171.95	413.0
172.95	472.0
174.05	663.0
175.15	1541.0
176.15	275.0
177.05	934.0
177.85	198.0
179.05	3160.0
180.05	1875.0
181.15	747.0
182.35	183.0
185.05	1057.0
186.05	10765.0
187.05	2900.0
188.05	318.0
189.15	819.0
191.05	297.0
192.05	1140.0
193.05	1117.0
194.05	287.0
195.05	205.0
196.05	2169.0
198.05	70480.0
198.95	4571.0
199.85	267.0
201.45	618.0
202.95	714.0
204.05	2833.0
205.05	4981.0
206.05	21160.0
207.05	2993.0
208.05	821.0
208.75	169.0
209.85	290.0
210.15	299.0
210.95	1102.0
211.95	186.0
214.95	398.0
215.95	828.0
216.95	5598.0
217.95	691.0
221.05	2500.0
222.05	830.0
222.95	1548.0
224.05	10929.0
225.05	3099.0
226.05	396.0
227.05	5164.0
228.05	878.0
228.95	1156.0
229.95	152.0
230.95	331.0
232.95	186.0
234.05	410.0
235.05	708.0
237.05	452.0
237.95	157.0
240.95	153.0
242.05	619.0
243.15	525.0
244.05	9005.0
245.05	1548.0
246.05	1803.0
247.15	465.0
248.95	376.0
250.05	154.0
252.95	302.0
255.05	43360.0
256.05	6943.0
257.05	522.0
258.05	2317.0
259.05	401.0
264.05	160.0
265.05	969.0
273.05	1454.0
273.95	3556.0
275.05	20616.0
275.95	3319.0
276.95	1751.0
278.15	183.0
284.05	208.0
285.05	238.0
292.95	271.0
294.05	224.0
296.05	5253.0
297.05	1090.0
303.05	639.0
304.05	203.0
308.85	193.0
313.95	481.0
315.05	645.0

Instrument :
BNA_G
ClientSampleId :
DFTPP446

315.95	276.0
320.95	252.0
323.05	1914.0
324.05	401.0
326.95	599.0
328.05	291.0
332.25	152.0
333.05	258.0
334.05	798.0
334.95	349.0
342.05	187.0
346.05	190.0
346.95	161.0
352.05	518.0
352.95	382.0
354.05	452.0
354.95	227.0
364.95	2110.0
365.95	399.0
371.05	210.0
372.05	662.0
372.95	205.0
382.75	254.0
389.95	198.0
402.05	274.0
402.95	624.0
421.05	457.0
421.95	349.0
423.05	3158.0
424.05	582.0
441.05	6199.0
442.05	43560.0
443.05	7873.0
444.05	645.0

Instrument :
BNA_G
ClientSampleId :
DFTPP446

m/z	Abundance
37.25	199.0
38.25	473.0
39.15	3682.0
40.15	344.0
41.25	332.0
43.15	188.0
44.05	653.0
50.15	10321.0
51.15	40936.0
52.15	2042.0
55.15	234.0
56.05	1237.0
57.15	3014.0
58.05	228.0
61.15	440.0
62.15	485.0
63.15	1757.0
64.15	477.0
65.05	511.0
69.05	39896.0
70.05	232.0
73.35	387.0
74.15	4841.0
75.05	7617.0
76.15	2807.0
77.15	43296.0
78.05	2581.0
79.15	3003.0
80.05	2723.0
81.05	3434.0
82.15	1058.0
83.05	987.0
85.15	709.0
86.15	1042.0
87.05	322.0
88.05	171.0
91.15	708.0
92.05	893.0
93.05	4966.0
94.15	394.0
96.15	402.0
97.15	271.0
98.15	5138.0
99.15	3150.0
99.95	305.0
101.15	1476.0
103.15	562.0
104.15	685.0
105.15	1039.0
106.05	556.0
107.05	13979.0
108.05	1857.0
108.95	522.0
110.05	22456.0
111.05	3608.0
112.05	761.0
115.25	216.0
116.15	1107.0
117.05	11695.0
118.15	1111.0
119.15	212.0
120.15	251.0
121.05	227.0
122.05	893.0
123.05	1551.0
124.05	670.0
125.05	401.0
127.05	46144.0
128.15	3584.0
129.05	19384.0
130.05	1597.0
131.05	437.0
132.85	178.0
134.05	640.0
135.05	1711.0
136.05	647.0
137.05	925.0
138.05	248.0
138.85	172.0
140.05	245.0
141.05	2747.0
142.15	768.0
142.95	577.0
144.95	238.0
145.95	617.0
147.05	1311.0
148.05	2938.0
149.05	450.0
151.05	415.0
153.05	720.0
154.05	556.0
155.05	1512.0

Instrument :
BNA_G
ClientSampleId :
DFTPP446

156.05	2256.0
157.05	333.0
158.05	435.0
158.95	293.0
159.15	286.0
160.05	853.0
161.05	1034.0
162.05	551.0
163.75	260.0
165.05	1108.0
166.05	698.0
167.05	3929.0
168.05	2305.0
169.05	280.0
170.25	267.0
171.05	351.0
171.95	433.0
173.05	738.0
174.15	1007.0
175.05	2378.0
176.05	555.0
176.95	1111.0
178.05	383.0
179.05	3851.0
180.05	2810.0
181.05	1282.0
181.95	167.0
182.25	167.0
183.95	594.0
185.05	2397.0
186.05	14766.0
187.05	3781.0
188.05	477.0
189.05	961.0
190.05	185.0
191.15	425.0
191.95	997.0
193.05	1405.0
194.05	229.0
194.95	254.0
195.25	200.0
196.05	2839.0
196.75	557.0
197.95	91296.0
199.05	6550.0
200.05	770.0
201.45	421.0
203.15	940.0
204.05	3486.0
205.05	7003.0
206.05	27344.0
207.05	2935.0
208.05	1077.0
209.05	324.0
210.55	703.0
210.75	701.0
211.15	1077.0
215.25	256.0
216.05	590.0
217.05	8153.0
218.05	1023.0
221.05	4046.0
221.95	1216.0
223.05	1942.0
224.05	16311.0
225.05	3809.0
226.15	336.0
227.05	6298.0
228.05	1138.0
229.05	1485.0
229.95	214.0
231.05	640.0
233.85	441.0
235.05	489.0
235.95	309.0
237.05	580.0
238.05	239.0
238.95	325.0
240.05	193.0
241.05	455.0
242.15	786.0
243.15	1401.0
244.05	12093.0
245.05	1790.0
246.05	2102.0
247.05	524.0
248.95	447.0
252.05	192.0
253.15	376.0
255.05	62840.0
256.05	10239.0
257.05	857.0
258.05	3599.0
259.05	540.0
259.95	196.0
265.05	1802.0

Instrument :
BNA_G
ClientSampleId :
DFTPP446

265.85	198.0
271.95	238.0
273.05	2400.0
274.05	4909.0
275.05	28936.0
276.05	4267.0
277.05	2776.0
278.05	255.0
282.95	327.0
284.95	348.0
293.05	652.0
296.05	8610.0
297.05	1124.0
302.05	177.0
303.05	1100.0
307.95	258.0
310.15	210.0
314.05	443.0
315.05	1009.0
316.15	480.0
320.95	303.0
322.15	175.0
322.95	2279.0
324.05	319.0
326.95	578.0
328.15	378.0
331.95	185.0
332.15	181.0
333.05	377.0
334.05	1985.0
335.15	508.0
340.95	405.0
341.85	166.0
346.05	617.0
351.95	729.0
352.95	434.0
353.95	750.0
355.05	228.0
359.15	179.0
364.95	3551.0
366.05	456.0
372.05	1678.0
373.05	294.0
382.95	346.0
389.95	254.0
390.95	155.0
401.05	188.0
401.95	407.0
402.95	646.0
404.15	259.0
421.05	673.0
421.95	674.0
422.95	3995.0
424.05	1199.0
441.05	11617.0
442.05	72448.0
443.05	14839.0
444.05	922.0

Instrument :
BNA_G
ClientSampleId :
DFTPP446

m/z	Abundance
37.05	200.0
38.15	498.0
39.15	2861.0
40.05	277.0
41.05	294.0
43.25	384.0
44.05	549.0
49.05	241.0
50.15	8449.0
51.15	29976.0
52.15	1296.0
55.15	353.0
56.05	798.0
57.05	2640.0
58.05	201.0
61.05	565.0
62.15	565.0
63.15	1397.0
64.05	173.0
65.15	662.0
67.35	179.0
69.05	33480.0
70.05	164.0
71.15	197.0
73.25	284.0
74.15	3725.0
75.15	5311.0
76.15	1738.0
77.15	35192.0
78.15	2921.0
79.05	2663.0
80.15	1515.0
81.05	2946.0
81.95	637.0
83.05	634.0
85.05	552.0
86.15	865.0
87.15	259.0
91.15	808.0
92.15	738.0
93.15	4837.0
94.15	394.0
95.95	229.0
97.15	322.0
98.15	3503.0
99.15	2845.0
101.05	936.0
103.05	472.0
104.15	1272.0
105.15	1047.0
106.05	307.0
107.05	11288.0
108.05	1766.0
109.05	559.0
110.05	18240.0
111.05	3419.0
112.05	444.0
116.05	927.0
117.05	10292.0
118.05	471.0
121.95	726.0
123.05	1230.0
124.05	530.0
125.15	543.0
127.05	35752.0
128.05	2666.0
129.05	15148.0
130.05	1402.0
130.95	385.0
132.05	179.0
133.15	226.0
134.05	514.0
135.05	1317.0
136.05	684.0
137.05	525.0
138.15	296.0
139.05	164.0
141.05	2768.0
142.05	715.0
143.05	507.0
144.05	157.0
145.15	154.0
146.05	612.0
147.15	1341.0
148.05	2910.0
149.05	604.0
150.15	163.0
150.95	334.0
151.45	302.0
151.95	344.0
153.05	778.0
154.15	628.0

Instrument :
BNA_G
ClientSampleId :
DFTPP446

155.05	1334.0
156.05	1415.0
156.95	310.0
158.15	312.0
159.05	583.0
160.15	630.0
161.15	981.0
162.15	298.0
163.05	217.0
164.05	157.0
165.05	949.0
166.05	995.0
167.05	4260.0
168.05	1726.0
169.15	363.0
169.95	158.0
171.05	192.0
171.95	415.0
173.15	611.0
174.05	1253.0
175.15	1450.0
175.95	506.0
177.05	902.0
178.05	330.0
179.05	3445.0
180.05	2080.0
181.05	980.0
182.15	320.0
184.05	261.0
185.05	1450.0
186.05	12331.0
187.05	3168.0
187.95	340.0
189.05	791.0
191.05	417.0
191.95	1321.0
193.05	1200.0
194.15	282.0
195.95	2334.0
197.95	80488.0
198.95	6070.0
200.05	504.0
201.15	344.0
201.75	378.0
202.95	743.0
204.05	3216.0
205.05	5627.0
206.05	24488.0
207.05	3637.0
208.15	1218.0
209.05	266.0
210.15	546.0
211.05	801.0
211.85	286.0
214.95	300.0
216.15	824.0
217.05	6815.0
217.95	782.0
221.05	3869.0
222.05	958.0
223.05	1756.0
224.05	15056.0
225.05	3728.0
226.05	631.0
227.05	7176.0
228.05	787.0
229.05	1304.0
229.85	221.0
231.15	586.0
234.05	324.0
234.95	510.0
236.05	254.0
236.95	518.0
238.95	342.0
240.95	501.0
242.05	716.0
243.05	731.0
244.05	10808.0
245.05	1248.0
246.05	2696.0
247.05	369.0
249.05	379.0
251.95	245.0
253.05	382.0
253.95	495.0
255.05	59000.0
256.05	8822.0
257.05	669.0
258.05	3259.0
259.05	404.0
265.05	1327.0
266.05	192.0
271.85	235.0
273.05	2043.0
274.05	4385.0
275.05	25672.0

Instrument :
BNA_G
ClientSampleId :
DFTPP446

276.05	3511.0
277.05	2379.0
278.15	293.0
282.15	198.0
283.05	247.0
283.95	239.0
285.05	410.0
293.05	721.0
293.95	203.0
296.05	7437.0
296.95	1278.0
297.85	178.0
301.05	212.0
302.05	194.0
303.05	1077.0
304.05	310.0
309.95	166.0
314.15	386.0
314.95	867.0
316.05	362.0
323.05	2704.0
324.05	285.0
326.95	384.0
327.95	151.0
331.95	207.0
333.05	207.0
334.05	1442.0
334.85	245.0
341.05	402.0
346.05	755.0
352.05	865.0
352.95	380.0
354.05	725.0
365.05	3562.0
365.95	307.0
371.05	285.0
371.95	1404.0
372.95	325.0
383.05	385.0
390.05	245.0
401.85	729.0
402.95	851.0
404.05	331.0
420.95	963.0
422.05	520.0
423.05	4530.0
423.95	662.0
441.05	12084.0
442.05	73464.0
443.05	14672.0
444.05	1267.0

Instrument :
BNA_G
ClientSampleId :
DFTPP446