

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122920\
 Data File : BG047835.D
 Acq On : 29 Dec 2020 14:15
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 29 15:10:48 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG121720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 17 15:55:16 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	122	0.00
2	1,4-Dioxane	40.000	28.979	27.6#	113	0.00
3	Pyridine	40.000	34.181	14.5	119	0.00
4	n-Nitrosodimethylamine	40.000	34.281	14.3	120	0.00
5 S	2-Fluorophenol	80.000	76.536	4.3	139	0.00
6	Aniline	40.000	42.156	-5.4	155	-0.01
7 S	Phenol-d6	80.000	78.897	1.4	140	0.00
8	2-Chlorophenol	40.000	38.075	4.8	136	0.00
9	Benzaldehyde	40.000	40.550	-1.4	156	0.00
10 C	Phenol	40.000	40.154	-0.4	141	0.00
11	bis(2-Chloroethyl)ether	40.000	38.049	4.9	140	0.00
12	1,3-Dichlorobenzene	40.000	36.171	9.6	130	0.00
13 C	1,4-Dichlorobenzene	40.000	36.382	9.0	133	0.00
14	1,2-Dichlorobenzene	40.000	36.498	8.8	134	0.00
15	Benzyl Alcohol	40.000	41.646	-4.1	145	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.884	7.8	136	0.00
17	2-Methylphenol	40.000	39.316	1.7	144	0.00
18	Hexachloroethane	40.000	35.861	10.3	126	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.587	3.5	140	0.00
20	3+4-Methylphenols	40.000	39.967	0.1	145	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	128	0.00
22	Acetophenone	40.000	40.025	-0.1	146	0.00
23 S	Nitrobenzene-d5	80.000	73.088	8.6	133	0.00
24	Nitrobenzene	40.000	38.352	4.1	140	0.00
25	Isophorone	40.000	41.439	-3.6	152	0.00
26 C	2-Nitrophenol	40.000	41.083	-2.7	148	0.00
27	2,4-Dimethylphenol	40.000	46.396	-16.0	173	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.667	3.3	148	0.00
29 C	2,4-Dichlorophenol	40.000	38.636	3.4	142	0.00
30	1,2,4-Trichlorobenzene	40.000	37.767	5.6	133	0.00
31	Naphthalene	40.000	39.227	1.9	141	0.00
32	Benzoic acid	40.000	26.104	34.7#	174	0.01
33	4-Chloroaniline	40.000	40.382	-1.0	143	0.00
34 C	Hexachlorobutadiene	40.000	35.135	12.2	128	0.00
35	Caprolactam	40.000	38.604	3.5	145	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.483	-3.7	149	0.00
37	2-Methylnaphthalene	40.000	39.712	0.7	145	0.00
38	1-Methylnaphthalene	40.000	38.665	3.3	141	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	135	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.232	1.9	147	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.919	-9.8	149	0.00
42 S	2,4,6-Tribromophenol	80.000	79.231	1.0	145	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.477	6.3	140	0.00
44	2,4,5-Trichlorophenol	40.000	39.644	0.9	146	0.00
45 S	2-Fluorobiphenyl	80.000	74.036	7.5	137	0.00
46	1,1'-Biphenyl	40.000	38.564	3.6	144	0.00
47	2-Chloronaphthalene	40.000	36.995	7.5	133	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122920\
 Data File : BG047835.D
 Acq On : 29 Dec 2020 14:15
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 29 15:10:48 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG121720.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 17 15:55:16 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	36.576	8.6	134	0.00
49	Acenaphthylene	40.000	38.937	2.7	143	0.00
50	Dimethylphthalate	40.000	36.335	9.2	135	0.00
51	2,6-Dinitrotoluene	40.000	37.974	5.1	142	0.00
52 C	Acenaphthene	40.000	38.145	4.6	138	0.00
53	3-Nitroaniline	40.000	36.547	8.6	133	0.00
54 P	2,4-Dinitrophenol	40.000	34.048	14.9	122	0.00
55	Dibenzofuran	40.000	37.938	5.2	143	0.00
56 P	4-Nitrophenol	40.000	41.302	-3.3	150	0.00
57	2,4-Dinitrotoluene	40.000	37.551	6.1	143	0.00
58	Fluorene	40.000	38.158	4.6	145	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	37.729	5.7	140	0.00
60	Diethylphthalate	40.000	36.735	8.2	138	0.00
61	4-Chlorophenyl-phenylether	40.000	36.491	8.8	136	0.00
62	4-Nitroaniline	40.000	34.716	13.2	127	0.00
63	Azobenzene	40.000	38.110	4.7	143	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	127	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.850	0.4	140	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.939	0.2	143	0.00
67	4-Bromophenyl-phenylether	40.000	37.508	6.2	132	0.00
68	Hexachlorobenzene	40.000	36.974	7.6	131	0.00
69	Atrazine	40.000	36.789	8.0	133	0.00
70 C	Pentachlorophenol	40.000	40.489	-1.2	140	0.00
71	Phenanthrene	40.000	38.287	4.3	141	0.00
72	Anthracene	40.000	40.208	-0.5	147	0.00
73	Carbazole	40.000	38.328	4.2	143	0.00
74	Di-n-butylphthalate	40.000	37.349	6.6	135	0.00
75 C	Fluoranthene	40.000	35.767	10.6	132	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	109	0.00
77	Benzydine	40.000	41.227	-3.1	116	0.00
78	Pyrene	40.000	41.335	-3.3	131	0.00
79 S	Terphenyl-d14	80.000	76.965	3.8	121	0.00
80	Butylbenzylphthalate	40.000	38.397	4.0	120	0.00
81	Benzo(a)anthracene	40.000	38.802	3.0	120	0.00
82	3,3'-Dichlorobenzidine	40.000	43.601	-9.0	133	0.00
83	Chrysene	40.000	39.082	2.3	123	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.122	2.2	121	0.00
85 c	Di-n-octyl phthalate	40.000	38.227	4.4	118	0.00
86 I	Perylene-d12	20.000	20.000	0.0	103	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.716	0.7	115	0.00
88	Benzo(b)fluoranthene	40.000	39.523	1.2	114	0.00
89	Benzo(k)fluoranthene	40.000	39.696	0.8	118	0.00
90 C	Benzo(a)pyrene	40.000	38.824	2.9	114	-0.01
91	Dibenzo(a,h)anthracene	40.000	40.336	-0.8	117	0.00
92	Benzo(g,h,i)perylene	40.000	41.323	-3.3	120	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG122920\
Data File : BG047835.D
Acq On : 29 Dec 2020 14:15
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Dec 29 15:10:48 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG121720.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 17 15:55:16 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0