


Data Path : Z:\SVOASRV\HPCHEM1\BNA_G\DATA\BG123019\
 Data File : BG043858.D
 Acq On : 30 Dec 2019 8:48
 Operator : JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG123019.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Dec 31 13:32:23 2019



AutoFind: Scans 2501, 2502, 2503; Background Corrected with Scan 2494

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.5	60909	PASS
68	69	0.00	2	0.6	395	PASS
69	198	0.00	100	37.7	61245	PASS
70	69	0.00	2	0.3	210	PASS
127	198	10	80	47.4	76928	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	162394	PASS
199	198	5	9	7.1	11462	PASS
275	198	10	60	24.8	40285	PASS
365	198	1	100	3.0	4949	PASS
441	198	0.01	100	10.4	16952	PASS
442	198	50	100	63.6	103328	PASS
443	442	15	24	20.2	20863	PASS