


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010322\
 Data File : BG051856.D
 Acq On : 3 Jan 2022 13:49
 Operator : CG/JU
 Sample : DFTPP448
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP448

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG121421.M
 Title : SVOA CALIBRATION
 Last Update : Fri Dec 31 00:13:36 2021



AutoFind: Scans 2481, 2482, 2483; Background Corrected with Scan 2475

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	45.4	32155	PASS
68	69	0.00	2	0.8	262	PASS
69	198	0.00	100	45.1	31931	PASS
70	69	0.00	2	0.5	154	PASS
127	198	10	80	47.6	33715	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	70819	PASS
199	198	5	9	7.0	4931	PASS
275	198	10	60	32.8	23248	PASS
365	198	1	100	4.5	3165	PASS
441	198	0.01	100	13.5	9529	PASS
442	198	50	100	85.8	60739	PASS
443	442	15	24	19.3	11712	PASS

m/z	Abundance
38.15	299.0
39.15	1928.0
40.05	358.0
44.05	749.0
49.05	152.0
50.15	4302.0
51.15	18512.0
52.05	810.0
55.05	170.0
56.05	716.0
57.05	1216.0
61.05	201.0
62.15	323.0
63.05	1078.0
64.05	185.0
65.15	290.0
69.05	18768.0
73.05	628.0
74.15	2356.0
75.05	3083.0
76.15	1235.0
77.15	21608.0
78.15	1148.0
79.05	1389.0
80.05	1057.0
81.05	1697.0
82.15	399.0
82.95	373.0
85.05	281.0
86.05	342.0
87.05	282.0
91.15	414.0
92.05	777.0
93.15	2372.0
94.25	279.0
98.05	2490.0
99.05	1612.0
101.05	999.0
102.85	188.0
103.25	247.0
104.15	646.0
105.05	412.0
106.25	324.0
107.05	7361.0
108.05	961.0
110.05	10021.0
111.05	2073.0
112.05	234.0
116.05	481.0
117.05	7160.0
117.95	454.0
122.05	704.0
123.05	702.0
123.95	434.0
125.15	272.0
127.05	20584.0
128.05	1572.0
129.05	9118.0
130.15	709.0
131.05	171.0
135.05	597.0
136.05	317.0
137.15	357.0
141.05	1377.0
141.95	486.0
143.15	255.0
144.95	184.0
146.05	353.0
147.05	615.0
148.05	1809.0
148.95	317.0
151.95	254.0
152.95	537.0
153.75	243.0
154.05	234.0
155.05	770.0
155.95	1349.0
157.05	161.0
158.05	215.0
160.05	436.0
161.05	827.0
162.05	237.0
165.05	737.0
165.95	422.0
167.05	2691.0
168.05	991.0
169.15	188.0
171.95	243.0
173.05	254.0
174.15	691.0
175.15	767.0
176.05	290.0

Instrument :
BNA_G
ClientSampleId :
DFTPP448

176.95	575.0
177.95	351.0
179.05	2132.0
180.05	1532.0
180.95	746.0
182.25	208.0
183.95	237.0
185.05	1298.0
186.05	7277.0
187.05	2645.0
188.05	197.0
189.05	376.0
191.05	334.0
192.05	566.0
193.05	773.0
194.05	241.0
194.85	177.0
195.95	1211.0
197.95	49536.0
198.95	2756.0
200.05	162.0
201.45	208.0
203.15	379.0
204.05	2498.0
205.05	4087.0
206.05	16126.0
207.05	2405.0
208.05	945.0
209.05	276.0
210.05	299.0
211.15	449.0
215.05	207.0
216.95	3970.0
218.05	447.0
221.05	2347.0
222.05	654.0
223.05	1025.0
224.05	8904.0
225.05	2694.0
226.05	286.0
227.05	5021.0
227.95	665.0
228.95	917.0
231.05	360.0
233.85	273.0
235.05	455.0
236.15	218.0
237.05	328.0
238.95	255.0
240.05	187.0
240.95	157.0
241.95	544.0
243.05	563.0
244.05	7016.0
244.95	868.0
246.05	1724.0
247.15	270.0
248.95	226.0
253.15	296.0
253.75	200.0
255.05	38800.0
256.05	5692.0
257.05	518.0
258.05	2550.0
259.05	377.0
265.05	1114.0
272.95	1150.0
274.05	3309.0
275.05	18512.0
275.95	2147.0
277.05	1815.0
277.95	268.0
280.95	155.0
282.85	312.0
285.15	251.0
292.85	421.0
295.05	167.0
296.05	5087.0
297.15	841.0
301.05	181.0
303.05	566.0
310.15	162.0
313.95	328.0
314.95	612.0
315.85	324.0
320.85	224.0
323.05	1245.0
323.95	492.0
326.85	350.0
328.05	171.0
331.95	231.0
334.05	835.0
335.05	551.0
340.85	179.0
346.05	299.0
351.95	443.0

Instrument :
BNA_G
ClientSampleId :
DFTPP448

352.85	280.0
354.05	453.0
364.95	2458.0
366.05	286.0
370.95	168.0
372.05	990.0
372.85	326.0
383.05	154.0
390.05	287.0
401.05	224.0
401.95	359.0
402.95	471.0
420.95	427.0
421.85	409.0
422.95	3615.0
424.05	787.0
440.95	8668.0
442.05	54744.0
443.05	9604.0
444.05	1133.0

Instrument :
BNA_G
ClientSampleId :
DFTPP448

m/z	Abundance
38.05	585.0
39.15	2886.0
40.15	433.0
44.05	788.0
49.15	313.0
50.15	8639.0
51.15	36480.0
52.15	1816.0
55.05	251.0
56.15	1163.0
57.05	2295.0
57.95	160.0
61.05	655.0
62.05	445.0
63.15	1337.0
64.05	247.0
65.15	750.0
69.05	37056.0
70.05	235.0
73.05	484.0
74.05	4133.0
75.15	6437.0
76.15	2098.0
77.15	40168.0
78.15	2683.0
79.15	3179.0
80.05	2201.0
81.05	2900.0
82.05	748.0
83.05	665.0
85.05	556.0
86.05	1214.0
87.05	482.0
88.15	195.0
91.05	633.0
92.05	797.0
93.05	5226.0
94.05	345.0
95.95	221.0
98.05	3922.0
99.05	2857.0
100.05	263.0
101.05	1448.0
103.05	662.0
104.15	1236.0
104.95	956.0
106.05	370.0
107.05	12520.0
108.05	2173.0
108.85	258.0
110.05	20784.0
111.05	2628.0
112.05	510.0
113.05	160.0
116.05	916.0
117.05	13149.0
118.05	973.0
122.05	882.0
123.05	1406.0
124.05	523.0
125.05	636.0
125.85	273.0
127.05	39528.0
128.05	3202.0
129.05	17072.0
130.05	1277.0
131.05	543.0
131.95	195.0
134.05	571.0
135.05	1214.0
136.05	631.0
137.05	679.0
139.95	336.0
140.95	2614.0
142.25	634.0
143.05	500.0
144.05	269.0
146.15	427.0
147.05	1555.0
148.05	2912.0
149.05	677.0
149.85	157.0
151.05	287.0
152.95	593.0
153.95	644.0
155.05	1417.0
156.05	1994.0
157.15	359.0
157.85	514.0
159.05	361.0
160.05	833.0
161.05	1128.0

Instrument :
BNA_G
ClientSampleId :
DFTPP448

162.05	345.0
164.95	998.0
165.95	759.0
167.05	4478.0
168.05	2156.0
169.05	335.0
171.15	293.0
172.15	403.0
173.05	550.0
174.05	1031.0
175.05	2594.0
175.95	472.0
176.95	800.0
177.95	380.0
179.05	3770.0
180.05	2642.0
181.05	1123.0
182.15	175.0
184.15	220.0
185.05	1770.0
186.05	13043.0
187.05	3974.0
188.05	588.0
189.05	828.0
190.05	153.0
190.95	578.0
192.05	1354.0
193.05	1373.0
194.05	521.0
195.05	249.0
196.15	2673.0
197.95	82608.0
198.95	6109.0
199.95	518.0
201.15	240.0
201.55	268.0
202.05	160.0
203.05	876.0
204.05	3807.0
205.05	5824.0
206.05	26248.0
207.05	3892.0
208.05	1027.0
209.05	316.0
210.05	464.0
211.05	1196.0
212.15	216.0
214.95	224.0
216.05	483.0
217.05	7617.0
218.05	791.0
221.05	3767.0
221.85	1037.0
223.05	1465.0
224.05	14723.0
225.05	4107.0
226.05	486.0
227.05	7895.0
227.95	927.0
228.95	1612.0
230.05	269.0
231.05	483.0
232.75	186.0
234.05	619.0
235.05	460.0
235.95	405.0
236.85	451.0
238.95	267.0
239.95	236.0
240.95	289.0
241.95	927.0
243.05	928.0
244.05	10804.0
245.05	1855.0
246.05	2747.0
247.05	484.0
248.95	298.0
252.95	222.0
253.45	356.0
255.05	63256.0
256.05	9176.0
257.05	868.0
258.05	3679.0
259.05	579.0
260.05	199.0
265.05	2158.0
265.85	181.0
270.95	263.0
272.95	2093.0
274.05	4522.0
275.05	28384.0
276.05	3614.0
276.95	2621.0
277.95	512.0
281.05	249.0
283.15	280.0

Instrument :
BNA_G
ClientSampleId :
DFTPP448

284.05	268.0
284.95	346.0
289.05	191.0
289.95	231.0
293.15	496.0
296.05	8628.0
297.05	1150.0
303.05	809.0
303.95	359.0
313.95	459.0
314.95	723.0
316.05	414.0
321.05	296.0
322.95	2702.0
323.95	302.0
326.95	437.0
327.95	285.0
331.95	221.0
332.95	227.0
334.05	1451.0
335.05	516.0
341.15	277.0
346.05	805.0
352.05	991.0
353.05	652.0
354.05	823.0
364.95	3472.0
365.85	415.0
370.95	261.0
371.95	1456.0
373.05	279.0
383.05	255.0
390.15	194.0
392.05	174.0
401.95	442.0
402.95	633.0
404.05	249.0
420.95	614.0
421.95	549.0
422.95	4534.0
423.95	1233.0
424.95	192.0
441.05	10617.0
441.95	74752.0
442.95	14915.0
444.05	853.0

Instrument :
BNA_G
ClientSampleId :
DFTPP448

m/z	Abundance
37.05	326.0
38.15	401.0
39.15	2943.0
40.05	256.0
41.25	182.0
44.05	731.0
49.15	312.0
50.15	11104.0
51.15	41472.0
52.15	2704.0
53.15	189.0
55.05	300.0
56.05	1169.0
57.05	2817.0
61.05	614.0
62.15	728.0
63.15	1653.0
64.15	297.0
65.05	482.0
68.05	785.0
69.05	39968.0
70.15	227.0
73.15	439.0
74.15	4742.0
75.05	7486.0
76.15	2572.0
77.15	44056.0
78.15	2593.0
79.05	3257.0
80.05	2332.0
81.15	3573.0
82.15	867.0
83.15	813.0
84.95	651.0
86.05	1394.0
87.05	333.0
88.05	205.0
91.05	793.0
92.05	792.0
93.05	6024.0
93.95	262.0
96.15	319.0
98.05	4316.0
99.05	3363.0
100.05	246.0
101.15	1901.0
101.95	197.0
102.55	191.0
103.15	542.0
104.05	1082.0
105.15	1348.0
106.05	658.0
107.05	13713.0
108.15	2054.0
110.05	21856.0
111.05	4234.0
112.05	456.0
112.95	168.0
116.05	1025.0
117.05	13211.0
118.05	1026.0
119.95	167.0
121.25	163.0
122.05	1084.0
123.05	1214.0
124.05	823.0
125.15	597.0
127.05	41032.0
128.05	3240.0
129.05	17720.0
130.05	1743.0
131.05	193.0
132.15	251.0
132.95	166.0
134.15	549.0
135.05	1330.0
136.05	580.0
137.15	751.0
138.05	276.0
139.15	209.0
141.05	2631.0
142.05	825.0
143.15	405.0
143.95	224.0
146.05	438.0
147.05	1508.0
148.05	3438.0
149.05	695.0
150.05	217.0
150.95	257.0
151.35	201.0
152.95	549.0

Instrument :
BNA_G
ClientSampleId :
DFTPP448

154.05	611.0
155.05	1583.0
156.05	1996.0
157.15	300.0
157.95	434.0
159.05	650.0
160.05	680.0
161.05	976.0
161.95	300.0
164.95	1261.0
166.05	701.0
167.05	4554.0
168.05	2646.0
168.95	575.0
169.95	290.0
172.05	523.0
173.05	562.0
174.05	1275.0
175.05	2171.0
176.05	283.0
176.95	769.0
177.95	281.0
179.05	3974.0
180.05	2599.0
181.05	797.0
182.15	469.0
182.95	198.0
184.05	307.0
185.05	2121.0
186.05	13458.0
187.05	3381.0
188.05	438.0
189.05	1170.0
189.95	262.0
191.05	807.0
192.05	1243.0
193.15	1488.0
193.95	191.0
196.05	2248.0
197.95	80312.0
198.95	5928.0
200.15	551.0
201.55	513.0
203.15	984.0
204.05	3413.0
205.05	5504.0
206.05	25312.0
207.05	4871.0
208.05	1066.0
209.05	448.0
210.05	354.0
210.95	1034.0
211.95	186.0
215.05	321.0
215.95	552.0
216.95	7538.0
218.05	1010.0
220.95	2901.0
222.05	725.0
223.05	1144.0
224.05	13196.0
225.05	4085.0
226.05	468.0
227.05	6754.0
228.05	879.0
228.95	1410.0
230.15	186.0
231.15	524.0
233.05	188.0
233.85	372.0
234.95	409.0
235.75	162.0
236.15	186.0
237.15	429.0
238.95	315.0
239.85	268.0
240.85	298.0
242.05	841.0
243.05	904.0
244.05	9821.0
244.95	1108.0
246.05	2316.0
247.05	363.0
248.95	408.0
250.05	215.0
250.75	227.0
251.95	183.0
253.25	354.0
255.05	55744.0
256.05	8054.0
256.95	805.0
258.05	3020.0
259.05	641.0
259.85	260.0
260.95	168.0
264.95	1407.0

Instrument :
BNA_G
ClientSampleId :
DFTPP448

266.05	316.0
271.95	229.0
272.95	1735.0
273.95	4164.0
275.05	22848.0
276.05	3810.0
277.05	2244.0
278.05	243.0
280.85	171.0
281.95	191.0
282.95	291.0
284.05	291.0
285.15	333.0
285.85	151.0
292.95	424.0
296.05	7460.0
296.95	943.0
302.95	899.0
304.05	222.0
307.95	186.0
314.05	343.0
315.05	887.0
315.95	350.0
321.05	486.0
322.95	1595.0
323.95	532.0
327.05	509.0
327.95	262.0
331.95	160.0
333.05	309.0
334.05	1508.0
335.05	274.0
340.95	483.0
345.95	489.0
351.95	905.0
352.95	638.0
354.05	613.0
365.05	3566.0
365.95	493.0
372.05	1605.0
372.95	317.0
382.85	290.0
402.05	385.0
403.05	689.0
404.05	203.0
420.85	581.0
422.15	468.0
423.05	3265.0
423.95	909.0
440.95	9302.0
442.05	52720.0
443.05	10617.0
443.95	1210.0

Instrument :
BNA_G
ClientSampleId :
DFTPP448

Scan 2475 (18.027 min): BG051856.D\data.ms
DFTPP448

m/z	Abundance
39.95	359.0
44.05	900.0
55.15	259.0
73.05	248.0
207.15	309.0

Instrument :
BNA_G
ClientSampleId :
DFTPP448