

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG020620\
 Data File : BG044326.D
 Acq On : 6 Feb 2020 17:14
 Operator : CG/JU
 Sample : L1408-02MSD
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SHORE-RD-PATH-BH-01-BMSD

Quant Time: Feb 06 21:47:41 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG013020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 30 16:05:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.91	152	90843	20.00	ng	0.00
21) Naphthalene-d8	10.71	136	393916	20.00	ng	0.00
39) Acenaphthene-d10	14.55	164	263560	20.00	ng	0.00
64) Phenanthrene-d10	17.29	188	545390	20.00	ng	0.00
76) Chrysene-d12	21.54	240	463067	20.00	ng	0.00
87) Perylene-d12	24.64	264	536690	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.50	112	740654	143.89	ng	0.00
7) Phenol-d6	7.09	99	993585	143.74	ng	0.00
23) Nitrobenzene-d5	9.07	82	594079	85.14	ng	0.00
42) 2,4,6-Tribromophenol	16.04	330	400189	129.34	ng	0.00
45) 2-Fluorobiphenyl	13.18	172	1340971	85.20	ng	0.00
79) Terphenyl-d14	19.91	244	1799239	79.92	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.38	88	98196	42.460	ng	99
3) Pyridine	3.77	79	249328	40.465	ng	97
4) n-Nitrosodimethylamine	3.69	42	122732	47.668	ng	97
6) Aniline	7.24	93	274772	32.026	ng	99
8) 2-Chlorophenol	7.48	128	331196	58.545	ng	98
9) Benzaldehyde	7.05	77	146074	37.106	ng	99
10) Phenol	7.11	94	409086	59.801	ng	98
11) bis(2-Chloroethyl)ether	7.33	93	292551	50.023	ng	97
12) 1,3-Dichlorobenzene	7.81	146	327764	48.526	ng	98
13) 1,4-Dichlorobenzene	7.95	146	331438	49.544	ng	98
14) 1,2-Dichlorobenzene	8.27	146	312042	49.349	ng	99
15) Benzyl Alcohol	8.15	79	274100	56.012	ng	98
16) 2,2'-oxybis(1-Chloropropan	8.45	45	379919	47.996	ng	99
17) 2-Methylphenol	8.36	107	280487	57.468	ng	99
18) Hexachloroethane	9.00	117	121555	48.421	ng	99
19) n-Nitroso-di-n-propylamine	8.72	70	235294	51.077	ng	98
20) 3+4-Methylphenols	8.69	107	386593	57.474	ng	98
22) Acetophenone	8.73	105	435976	45.496	ng	99
24) Nitrobenzene	9.11	77	329242	48.335	ng	100
25) Isophorone	9.64	82	648729	49.883	ng	99
26) 2-Nitrophenol	9.82	139	190502	53.898	ng	99
27) 2,4-Dimethylphenol	9.89	122	319919	64.121	ng	100
28) bis(2-Chloroethoxy)methane	10.12	93	413088	47.615	ng	99
29) 2,4-Dichlorophenol	10.37	162	336514	55.780	ng	99
30) 1,2,4-Trichlorobenzene	10.58	180	327492	47.342	ng	99
31) Naphthalene	10.76	128	942078	49.293	ng	99
32) Benzoic acid	10.06	122	169422	50.504	ng	91
33) 4-Chloroaniline	10.87	127	129087	14.851	ng	97
34) Hexachlorobutadiene	11.06	225	191895	45.082	ng	99
35) Caprolactam	11.65	113	65651	29.707	ng	93
36) 4-Chloro-3-methylphenol	12.01	107	343713	54.340	ng	99
37) 2-Methylnaphthalene	12.37	142	716644	50.504	ng	99
38) 1-Methylnaphthalene	12.59	142	682143	50.990	ng	98
40) 1,2,4,5-Tetrachlorobenzene	12.75	216	355930	45.146	ng	98


Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG020620\
 Data File : BG044326.D
 Acq On : 6 Feb 2020 17:14
 Operator : CG/JU
 Sample : L1408-02MSD
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SHORE-RD-PATH-BH-01-BMSD

Quant Time: Feb 06 21:47:41 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG013020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 30 16:05:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Hexachlorocyclopentadiene	12.73	237	431308	114.014	ng	98
43) 2,4,6-Trichlorophenol	12.99	196	267376	52.468	ng	97
44) 2,4,5-Trichlorophenol	13.06	196	280564	53.902	ng	97
46) 1,1'-Biphenyl	13.39	154	900492	47.367	ng	100
47) 2-Chloronaphthalene	13.42	162	705783	46.654	ng	99
48) 2-Nitroaniline	13.62	65	204068	45.709	ng	97
49) Acenaphthylene	14.27	152	1109712	50.013	ng	99
50) Dimethylphthalate	14.01	163	998307	52.694	ng	99
51) 2,6-Dinitrotoluene	14.12	165	211054	49.963	ng	98
52) Acenaphthene	14.62	154	698780	48.587	ng	97
53) 3-Nitroaniline	14.45	138	118386	25.332	ng	96
54) 2,4-Dinitrophenol	14.66	184	242440	95.252	ng	97
55) Dibenzofuran	14.95	168	1048498	48.152	ng	100
56) 4-Nitrophenol	14.77	139	368899	107.761	ng	96
57) 2,4-Dinitrotoluene	14.91	165	287906	48.628	ng	95
58) Fluorene	15.60	166	831897	48.505	ng	100
59) 2,3,4,6-Tetrachlorophenol	15.18	232	244314	51.965	ng	99
60) Diethylphthalate	15.37	149	892801	47.294	ng	100
61) 4-Chlorophenyl-phenylether	15.59	204	441114	47.436	ng	98
62) 4-Nitroaniline	15.62	138	201878	43.230	ng	98
63) Azobenzene	15.88	77	790646	47.788	ng	99
65) 4,6-Dinitro-2-methylphenol	15.68	198	177979	59.119	ng	99
66) n-Nitrosodiphenylamine	15.81	169	781349	51.121	ng	99
67) 4-Bromophenyl-phenylether	16.48	248	289090	48.652	ng	98
68) Hexachlorobenzene	16.61	284	315270	47.359	ng	97
69) Atrazine	16.76	200	289945	55.347	ng	100
70) Pentachlorophenol	16.95	266	356426	104.863	ng	96
71) Phenanthrene	17.34	178	1408439	53.330	ng	97
72) Anthracene	17.43	178	1373212	52.593	ng	98
73) Carbazole	17.69	167	1183366	49.162	ng	98
74) Di-n-butylphthalate	18.26	149	1431801	48.135	ng	99
75) Fluoranthene	19.35	202	1777909	58.195	ng	96
77) Benzidine	19.53	184	341774	26.609	ng	98
78) Pyrene	19.71	202	1792971	60.292	ng	96
80) Butylbenzylphthalate	20.60	149	663575	48.965	ng	100
81) Benzo(a)anthracene	21.52	228	1626268	56.984	ng	96
82) 3,3'-Dichlorobenzidine	21.44	252	326954	29.518	ng	99
83) Chrysene	21.59	228	1526574	55.339	ng	97
84) Bis(2-ethylhexyl)phthalate	21.44	149	942744	50.540	ng	99
85) Di-n-octyl phthalate	22.61	149	1640970	50.844	ng	100
86) Indeno(1,2,3-cd)pyrene	28.15	276	1902411	52.860	ng	99
88) Benzo(b)fluoranthene	23.66	252	1705078	55.134	ng	99
89) Benzo(k)fluoranthene	23.73	252	1575473	52.269	ng	98
90) Benzo(a)pyrene	24.50	252	1588336	54.320	ng	97
91) Dibenzo(a,h)anthracene	28.21	278	1491030	51.525	ng	98
92) Benzo(g,h,i)perylene	29.25	276	1611213	55.619	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed




#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.91 min Scan# 821
Delta R.T. -0.01 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Instrument :
BNA_G
ClientSampleId :
SHORE-RD-PATH-BH-01-BMSD

Tot Ion:152 Resp: 90843
Ion Ratio Lower Upper
152 100
150 155.1 122.6 183.8
115 58.5 46.0 69.0




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
Concen: 42.460 ng
RT: 3.38 min Scan# 49
Delta R.T. -0.01 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Tgt Ion: 88 Resp: 98196
Ion Ratio Lower Upper
88 100
58 70.1 55.8 83.6
43 26.0 21.6 32.4



Abundance Ion 88.00 (87.70 to 88.70): BGC
Ion 58.00 (57.70 to 58.70): BGC
Ion 43.00 (42.70 to 43.70): BGC





#3

Pvridine

Concen: 40.465 ng

RT: 3.77 min Scan# 116

Delta R.T. -0.01 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14

Instrument :

BNA_G

ClientSampleId :

SHORE-RD-PATH-BH-01-BMSD


Tot Ion: 79 Resp: 249328

Ion Ratio Lower Upper




79 100

52 67.8 51.9 77.9

51 32.5 25.3 37.9



Abundance Ion 79.00 (78.70 to 79.70): BGC
Ion 52.00 (51.70 to 52.70): BGC
Ion 51.00 (50.70 to 51.70): BGC



#4

n-Nitrosodimethylamine

Concen: 47.668 ng

RT: 3.69 min Scan# 102

Delta R.T. -0.01 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14


Tgt Ion: 42 Resp: 122732

Ion Ratio Lower Upper



42 100


74 147.5 115.1 172.7

44 7.4 5.6 8.4



Abundance Ion 42.00 (41.70 to 42.70): BGC
Ion 74.00 (73.70 to 74.70): BGC
Ion 44.00 (43.70 to 44.70): BGC





#5

2-Fluorophenol

Concen: 143.890 ng

RT: 5.50 min Scan# 410

Delta R.T. 0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14

Instrument :

BNA_G

ClientSampleId :

SHORE-RD-PATH-BH-01-BMSD


Tot Ion:112 Resp: 740654

Ion Ratio Lower Upper




112 100

64 55.6 45.4 68.2

63 29.5 22.9 34.3



Abundance Ion 112.00 (111.70 to 112.70): BGC
Ion 64.00 (63.70 to 64.70): BGC
Ion 63.00 (62.70 to 63.70): BGC



#6

Aniline

Concen: 32.026 ng

RT: 7.24 min Scan# 706

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14


Tgt Ion: 93 Resp: 274772

Ion Ratio Lower Upper



93 100


66 35.8 28.7 43.1

65 18.4 15.4 23.0



Abundance Ion 93.00 (92.70 to 93.70): BGC
Ion 66.00 (65.70 to 66.70): BGC
Ion 65.00 (64.70 to 65.70): BGC





#7

Phenol-d6

Concen: 143.745 ng

RT: 7.09 min Scan# 681

Delta R.T. 0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14

Instrument :

BNA_G

ClientSampleId :

SHORE-RD-PATH-BH-01-BMSD


Tot Ion: 99 Resp: 993585

Ion Ratio Lower Upper




99 100

42 17.7 13.8 20.6

71 31.6 25.5 38.3



Abundance Ion 99.00 (98.70 to 99.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 71.00 (70.70 to 71.70): BGC



#8

2-Chlorophenol

Concen: 58.545 ng

RT: 7.48 min Scan# 748

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14


Tgt Ion: 128 Resp: 331196

Ion Ratio Lower Upper



128 100


130 33.4 12.5 52.5

64 42.3 24.2 64.2



Abundance Ion 128.00 (127.70 to 128.70): BGC
Ion 130.00 (129.70 to 130.70): BGC
Ion 64.00 (63.70 to 64.70): BGC





#9

Benzaldehyde

Concen: 37.106 ng

RT: 7.05 min Scan# 674

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14

Instrument :

BNA_G

ClientSampleId :

SHORE-RD-PATH-BH-01-BMSD


Tot Ion: 77 Resp: 146074

Ion Ratio Lower Upper

77 100

105 104.1 82.0 122.0




106 100.0 80.7 120.7



Abundance Ion 77.00 (76.70 to 77.70): BGC

Ion 105.00 (104.70 to 105.70): BGC

Ion 106.00 (105.70 to 106.70): BGC



#10

Phenol

Concen: 59.801 ng

RT: 7.11 min Scan# 685

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14


Tgt Ion: 94 Resp: 409086

Ion Ratio Lower Upper

94 100

65 27.5 7.0 47.0



66 34.7 12.7 52.7




Abundance Ion 94.00 (93.70 to 94.70): BGC

Ion 65.00 (64.70 to 65.70): BGC

Ion 66.00 (65.70 to 66.70): BGC









#11
bis(2-Chloroethyl)ether
Concen: 50.023 ng
RT: 7.33 min Scan# 722
Delta R.T. -0.01 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Instrument :
BNA_G
ClientSampleId :
SHORE-RD-PATH-BH-01-BMSD

Tot Ion: 93 Resp: 292551
Ion Ratio Lower Upper
93 100
63 72.4 49.2 89.2
95 33.2 13.6 53.6




Abundance Ion 93.00 (92.70 to 93.70): BGC
Ion 63.00 (62.70 to 63.70): BGC
Ion 95.00 (94.70 to 95.70): BGC






#12
1,3-Dichlorobenzene
Concen: 48.526 ng
RT: 7.81 min Scan# 803
Delta R.T. -0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Tgt Ion: 146 Resp: 327764
Ion Ratio Lower Upper
146 100
148 64.6 51.8 77.8
75 26.8 23.7 35.5



Abundance Ion 146.00 (145.70 to 146.70): BGC
Ion 148.00 (147.70 to 148.70): BGC
Ion 75.00 (74.70 to 75.70): BGC









#13
 1,4-Dichlorobenzene
 Concen: 49.544 ng
 RT: 7.95 min Scan# 827
 Delta R.T. -0.00 min
 Lab File: BG044326.D
 Acq: 6 Feb 2020 17:14

Instrument :
 BNA_G
 ClientSampleId :
 SHORE-RD-PATH-BH-01-BMSD

Tot Ion:146 Resp: 331438
 Ion Ratio Lower Upper
 146 100
 148 62.9 51.4 77.0
 111 38.9 31.8 47.8




Abundance Ion 146.00 (145.70 to 146.70): E
 Ion 148.00 (147.70 to 148.70): E
 Ion 111.00 (110.70 to 111.70): E






#14
 1,2-Dichlorobenzene
 Concen: 49.349 ng
 RT: 8.27 min Scan# 882
 Delta R.T. -0.00 min
 Lab File: BG044326.D
 Acq: 6 Feb 2020 17:14

Tgt Ion:146 Resp: 312042
 Ion Ratio Lower Upper
 146 100
 148 65.7 51.2 76.8
 111 42.3 34.0 51.0



Abundance Ion 146.00 (145.70 to 146.70): E
 Ion 148.00 (147.70 to 148.70): E
 Ion 111.00 (110.70 to 111.70): E





#15

Benzyl Alcohol

Concen: 56.012 ng

RT: 8.15 min Scan# 862

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14

Instrument :

BNA_G

ClientSampleId :

SHORE-RD-PATH-BH-01-BMSD





Tot Ion: 79 Resp: 274100

Ion Ratio Lower Upper

79 100

108 82.2 68.3 102.5

77 65.0 51.8 77.8



#16

2,2'-oxybis(1-chloropropane)

Concen: 47.996 ng

RT: 8.45 min Scan# 912

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14




Tgt Ion: 45 Resp: 379919


Ion Ratio Lower Upper

45 100

77 15.1 0.0 35.2

79 11.6 0.0 31.1









#17
2-Methylphenol
Concen: 57.468 ng
RT: 8.36 min Scan# 898
Delta R.T. -0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14




Instrument :
BNA_G
ClientSampleId :
SHORE-RD-PATH-BH-01-BMSD


Tot Ion	Ratio	Lower	Upper
107	100		
108	109.5	86.4	129.6
77	43.4	34.0	51.0
79	42.2	32.9	49.3



#18
Hexachloroethane
Concen: 48.421 ng
RT: 9.00 min Scan# 1006
Delta R.T. -0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Tgt Ion	Ratio	Lower	Upper
117	100		
119	94.5	77.5	116.3
201	105.9	84.9	127.3






#19
n-Nitroso-di-n-propylamine
Concen: 51.077 ng
RT: 8.72 min Scan# 959
Delta R.T. -0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Instrument :
BNA_G
ClientSampleId :
SHORE-RD-PATH-BH-01-BMSD

Tot Ion:	70	Resp:	235294
Ion	Ratio	Lower	Upper
70	100		
42	55.6	46.2	69.2
101	9.5	8.4	12.6
130	22.8	17.8	26.8






Abundance

Ion 70.00 (69.70 to 70.70): BGC

Ion 42.00 (41.70 to 42.70): BGC


Ion 101.00 (100.70 to 101.70): BGC

Ion 130.00 (129.70 to 130.70): BGC



#20
3+4-Methylphenols
Concen: 57.474 ng
RT: 8.69 min Scan# 954
Delta R.T. 0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Tgt Ion:	107	Resp:	386593
Ion	Ratio	Lower	Upper
107	100		
108	90.3	68.5	108.5
77	31.0	10.9	50.9
79	28.0	7.1	47.1





Abundance


Ion 107.00 (106.70 to 107.70): BGC

Ion 108.00 (107.70 to 108.70): BGC

Ion 77.00 (76.70 to 77.70): BGC

Ion 79.00 (78.70 to 79.70): BGC









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.71 min Scan# 1298
Delta R.T. -0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Instrument :
BNA_G
ClientSampleId :
SHORE-RD-PATH-BH-01-BMSD

Tot Ion:136	Resp:	393916	
Ion Ratio	Lower	Upper	
136 100			
137 11.2	8.8	13.2	
54 7.2	5.8	8.8	
68 5.5	4.6	6.8	




Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BGC
Ion 68.00 (67.70 to 68.70): BGC






#22
Acetophenone
Concen: 45.496 ng
RT: 8.73 min Scan# 961
Delta R.T. -0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Tgt	Ion:105	Resp:	435976
Ion	Ratio	Lower	Upper
105	100		
71	4.5	3.4	5.0
51	24.1	20.2	30.4
120	24.0	19.1	28.7



Abundance Ion 105.00 (104.70 to 105.70): E
Ion 71.00 (70.70 to 71.70): BGC
Ion 51.00 (50.70 to 51.70): BGC
Ion 120.00 (119.70 to 120.70): E









#23
Nitrobenzene-d5
Concen: 85.144 ng
RT: 9.07 min Scan# 1018
Delta R.T. -0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14




Instrument :
BNA_G
ClientSampleId :
SHORE-RD-PATH-BH-01-BMSD


Tot Ion	82	Resp	594079
Ion Ratio	100	Lower	Upper
128	46.2	36.5	54.7
54	47.0	38.6	58.0



#24
Nitrobenzene
Concen: 48.335 ng
RT: 9.11 min Scan# 1025
Delta R.T. -0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Tgt Ion	77	Resp	329242
Ion Ratio	100	Lower	Upper
123	49.6	39.9	59.9
65	12.9	10.4	15.6





#25

Isophorone

Concen: 49.883 ng

RT: 9.64 min Scan# 1115

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14

Instrument :

BNA_G

ClientSampleId :

SHORE-RD-PATH-BH-01-BMSD


Tot Ion: 82 Resp: 648729

Ion Ratio Lower Upper




82 100

95 7.2 5.7 8.5

138 19.5 15.0 22.4



Abundance Ion 82.00 (81.70 to 82.70): BGC
Ion 95.00 (94.70 to 95.70): BGC
Ion 138.00 (137.70 to 138.70): BGC



#26

2-Nitrophenol

Concen: 53.898 ng

RT: 9.82 min Scan# 1146

Delta R.T. -0.01 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14


Tgt Ion: 139 Resp: 190502

Ion Ratio Lower Upper



139 100


109 25.0 21.0 31.4

65 43.2 35.0 52.4



Abundance Ion 139.00 (138.70 to 139.70): BGC
Ion 109.00 (108.70 to 109.70): BGC
Ion 65.00 (64.70 to 65.70): BGC





#27

2,4-Dimethylphenol

Concen: 64.121 ng

RT: 9.89 min Scan# 1158

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14

Instrument :

BNA_G

ClientSampleId :

SHORE-RD-PATH-BH-01-BMSD


Tot Ion:122 Resp: 319919

Ion Ratio Lower Upper

122 100

107 111.1 88.8 133.2




121 61.1 48.3 72.5



Abundance Ion 122.00 (121.70 to 122.70): E

Ion 107.00 (106.70 to 107.70): E

Ion 121.00 (120.70 to 121.70): E



#28

bis(2-Chloroethoxy)methane

Concen: 47.615 ng

RT: 10.12 min Scan# 1197

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14


Tgt Ion: 93 Resp: 413088

Ion Ratio Lower Upper

93 100

95 34.0 26.6 40.0



123 12.3 10.0 15.0




Abundance Ion 93.00 (92.70 to 93.70): BGC

Ion 95.00 (94.70 to 95.70): BGC

Ion 123.00 (122.70 to 123.70): E









#29
2,4-Dichlorophenol
Concen: 55.780 ng
RT: 10.37 min Scan# 1239
Delta R.T. 0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Instrument :
BNA_G
ClientSampleId :
SHORE-RD-PATH-BH-01-BMSD

Tot Ion	Ratio	Lower	Upper
162	100		
164	65.1	44.7	84.7
98	34.6	15.8	55.8




Abundance Ion 162.00 (161.70 to 162.70): E
Ion 164.00 (163.70 to 164.70): E
Ion 98.00 (97.70 to 98.70): BGC






#30
1,2,4-Trichlorobenzene
Concen: 47.342 ng
RT: 10.58 min Scan# 1275
Delta R.T. -0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Tgt Ion	Ratio	Lower	Upper
180	100		
182	93.6	75.5	113.3
145	28.4	22.6	34.0



Abundance Ion 180.00 (179.70 to 180.70): E
Ion 182.00 (181.70 to 182.70): E
Ion 145.00 (144.70 to 145.70): E





#31

Naphthalene

Concen: 49.293 ng

RT: 10.76 min Scan# 1306

Delta R.T. -0.01 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14

Instrument:

BNA_G

ClientSampleId:

SHORE-RD-PATH-BH-01-BMSD


Tot Ion:128 Resp: 942078

Ion Ratio Lower Upper

128 100

129 12.1 9.3 13.9




127 14.8 11.6 17.4



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 50.504 ng

RT: 10.06 min Scan# 1187

Delta R.T. 0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14


Tgt Ion:122 Resp: 169422

Ion Ratio Lower Upper

122 100

105 121.2 93.5 133.5



77 90.3 61.0 101.0




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BG









#33
4-Chloroaniline
Concen: 14.851 ng
RT: 10.87 min Scan# 1324
Delta R.T. -0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Instrument :
BNA_G
ClientSampleId :
SHORE-RD-PATH-BH-01-BMSD

Tot Ion:	127	Resp:	129087
Ion	Ratio	Lower	Upper
127	100		
129	32.4	26.6	40.0
65	26.0	22.5	33.7
92	18.5	14.1	21.1




Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 65.00 (64.70 to 65.70): BG
Ion 92.00 (91.70 to 92.70): BG






#34
Hexachlorobutadiene
Concen: 45.082 ng
RT: 11.06 min Scan# 1357
Delta R.T. -0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Tgt Ion:	225	Resp:	191895
Ion	Ratio	Lower	Upper
225	100		
223	63.1	50.2	75.4
227	64.2	50.8	76.2



Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E





#35

Caprolactam

Concen: 29.707 ng

RT: 11.65 min Scan# 1457

Delta R.T. 0.01 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14

Instrument :

BNA_G

ClientSampleId :

SHORE-RD-PATH-BH-01-BMSD


Tot Ion:113 Resp: 65651

Ion Ratio Lower Upper

113 100

55 140.0 129.9 169.9




56 106.5 92.5 132.5



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BGC

Ion 56.00 (55.70 to 56.70): BGC



#36

4-Chloro-3-methylphenol

Concen: 54.340 ng

RT: 12.01 min Scan# 1518

Delta R.T. 0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14


Tgt Ion:107 Resp: 343713

Ion Ratio Lower Upper

107 100

144 28.0 22.2 33.4



142 85.2 67.4 101.0




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E









#37
2-Methylnaphthalene
Concen: 50.504 ng
RT: 12.37 min Scan# 1580
Delta R.T. -0.01 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14




Instrument :
BNA_G
ClientSampleId :
SHORE-RD-PATH-BH-01-BMSD


Tot Ion:142 Resp: 716644
Ion Ratio Lower Upper
142 100
141 90.9 73.2 109.8
115 34.1 27.1 40.7



#38
1-Methylnaphthalene
Concen: 50.990 ng
RT: 12.59 min Scan# 1618
Delta R.T. -0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Tgt Ion:142 Resp: 682143
Ion Ratio Lower Upper
142 100
141 96.4 75.8 113.6
116 3.6 3.1 4.7





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.55 min Scan# 1951

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14

Instrument :

BNA_G

ClientSampleId :

SHORE-RD-PATH-BH-01-BMSD


Tot Ion:164 Resp: 263560

Ion Ratio Lower Upper

164 100

162 103.8 80.7 121.1

160 44.3 35.3 52.9






Abundance

Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#40

1,2,4,5-Tetrachlorobenzene

Concen: 45.146 ng

RT: 12.75 min Scan# 1644

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14

Tgt Ion:216 Resp: 355930


Ion Ratio Lower Upper

216 100

214 79.7 63.0 94.6

179 20.6 16.2 24.4

108 19.0 13.5 20.3





Abundance


Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E





#41

Hexachlorocyclopentadiene

Concen: 114.014 ng

RT: 12.73 min Scan# 1641

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14

Instrument :

BNA_G

ClientSampleId :

SHORE-RD-PATH-BH-01-BMSD


Tot Ion:237 Resp: 431308

Ion Ratio Lower Upper

237 100

235 63.3 41.2 81.2




272 13.0 0.0 32.4



Abundance Ion 236.80 (236.50 to 237.50): E

Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E



#42

2,4,6-Tribromophenol

Concen: 129.342 ng

RT: 16.04 min Scan# 2205

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14


Tgt Ion:330 Resp: 400189

Ion Ratio Lower Upper

330 100

332 96.7 77.2 115.8



141 34.9 28.3 42.5




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#43
 2,4,6-Trichlorophenol
 Concen: 52.468 ng
 RT: 12.99 min Scan# 1685
 Delta R.T. -0.00 min
 Lab File: BG044326.D
 Acq: 6 Feb 2020 17:14




Instrument :
 BNA_G
 ClientSampleId :
 SHORE-RD-PATH-BH-01-BMSD


Tot Ion:196 Resp: 267376
 Ion Ratio Lower Upper
 196 100
 198 95.8 79.3 118.9
 200 32.8 25.7 38.5



#44
 2,4,5-Trichlorophenol
 Concen: 53.902 ng
 RT: 13.06 min Scan# 1697
 Delta R.T. -0.00 min
 Lab File: BG044326.D
 Acq: 6 Feb 2020 17:14

Tgt Ion:196 Resp: 280564
 Ion Ratio Lower Upper
 196 100
 198 101.3 78.2 117.2
 97 45.3 36.2 54.4
 132 27.3 20.2 30.4









#45
2-Fluorobiphenyl
Concen: 85.200 ng
RT: 13.18 min Scan# 1717
Delta R.T. -0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Instrument :
BNA_G
ClientSampleId :
SHORE-RD-PATH-BH-01-BMSD

Tot Ion:172 Resp: 1340971
Ion Ratio Lower Upper
172 100
171 41.2 32.9 49.3
170 27.5 22.0 33.0




Abundance Ion 172.00 (171.70 to 172.70): E
Ion 171.00 (170.70 to 171.70): E
Ion 170.00 (169.70 to 170.70): E






#46
1,1'-Biphenyl
Concen: 47.367 ng
RT: 13.39 min Scan# 1753
Delta R.T. -0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Tgt Ion:154 Resp: 900492
Ion Ratio Lower Upper
154 100
153 42.5 22.7 62.7
76 12.6 0.0 32.9



Abundance Ion 154.00 (153.70 to 154.70): E
Ion 153.00 (152.70 to 153.70): E
Ion 76.00 (75.70 to 76.70): BG0









#47
 2-Chloronaphthalene
 Concen: 46.654 ng
 RT: 13.42 min Scan# 1759
 Delta R.T. -0.01 min
 Lab File: BG044326.D
 Acq: 6 Feb 2020 17:14

Instrument :
 BNA_G
 ClientSampleId :
 SHORE-RD-PATH-BH-01-BMSD

Tot Ion:162 Resp: 705783
 Ion Ratio Lower Upper
 162 100
 127 38.4 30.0 45.0
 164 34.8 27.2 40.8




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E






#48
 2-Nitroaniline
 Concen: 45.709 ng
 RT: 13.62 min Scan# 1793
 Delta R.T. -0.00 min
 Lab File: BG044326.D
 Acq: 6 Feb 2020 17:14

Tgt Ion: 65 Resp: 204068
 Ion Ratio Lower Upper
 65 100
 92 75.4 60.3 90.5
 138 119.8 91.9 137.9



Abundance Ion 65.00 (64.70 to 65.70): BGC
 Ion 92.00 (91.70 to 92.70): BGC
 Ion 138.00 (137.70 to 138.70): E





#49

Acenaphthylene

Concen: 50.013 ng

RT: 14.27 min Scan# 1904

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14

Instrument :

BNA_G

ClientSampleId :

SHORE-RD-PATH-BH-01-BMSD


Tot Ion:152 Resp: 1109712

Ion Ratio Lower Upper

152 100

151 23.3 18.1 27.1




153 14.8 11.5 17.3



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#50

Dimethylphthalate

Concen: 52.694 ng

RT: 14.01 min Scan# 1859

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14


Tgt Ion:163 Resp: 998307

Ion Ratio Lower Upper

163 100

194 4.6 3.6 5.4



164 11.4 8.7 13.1




Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E









#51
 2,6-Dinitrotoluene
 Concen: 49.963 ng
 RT: 14.12 min Scan# 1878
 Delta R.T. -0.00 min
 Lab File: BG044326.D
 Acq: 6 Feb 2020 17:14

Instrument :
 BNA_G
 ClientSampleId :
 SHORE-RD-PATH-BH-01-BMSD

Tot Ion:165 Resp: 211054
 Ion Ratio Lower Upper
 165 100
 63 46.9 38.7 58.1
 89 49.6 40.4 60.6




Abundance Ion 165.00 (164.70 to 165.70): E
 Ion 63.00 (62.70 to 63.70): BGC
 Ion 89.00 (88.70 to 89.70): BGC






#52
 Acenaphthene
 Concen: 48.587 ng
 RT: 14.62 min Scan# 1962
 Delta R.T. -0.00 min
 Lab File: BG044326.D
 Acq: 6 Feb 2020 17:14

Tgt Ion:154 Resp: 698780
 Ion Ratio Lower Upper
 154 100
 153 113.1 87.9 131.9
 152 55.1 43.1 64.7



Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 152.00 (151.70 to 152.70): E









#53
3-Nitroaniline
Concen: 25.332 ng
RT: 14.45 min Scan# 1934
Delta R.T. -0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14




Instrument :
BNA_G
ClientSampleId :
SHORE-RD-PATH-BH-01-BMSD


Tot Ion:138 Resp: 118386
Ion Ratio Lower Upper
138 100
108 10.2 7.9 11.9
92 108.5 91.0 136.4



#54
2,4-Dinitrophenol
Concen: 95.252 ng
RT: 14.66 min Scan# 1970
Delta R.T. 0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Tgt Ion:184 Resp: 242440
Ion Ratio Lower Upper
184 100
63 51.6 44.1 66.1
154 57.0 45.4 68.0









#55
Dibenzenofuran
Concen: 48.152 ng
RT: 14.95 min Scan# 2019
Delta R.T. -0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Instrument :
BNA_G
ClientSampleId :
SHORE-RD-PATH-BH-01-BMSD

Tot Ion:168 Resp: 1048498
Ion Ratio Lower Upper
168 100
139 38.4 30.5 45.7
169 14.7 11.7 17.5




Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E






#56
4-Nitrophenol
Concen: 107.761 ng
RT: 14.77 min Scan# 1989
Delta R.T. 0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Tgt Ion:139 Resp: 368899
Ion Ratio Lower Upper
139 100
109 63.5 47.5 87.5
65 82.0 64.9 104.9



Abundance Ion 139.00 (138.70 to 139.70): E
Ion 109.00 (108.70 to 109.70): E
Ion 65.00 (64.70 to 65.70): BGO









#57
2,4-Dinitrotoluene
Concen: 48.628 ng
RT: 14.91 min Scan# 2013
Delta R.T. 0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Instrument :
BNA_G
ClientSampleId :
SHORE-RD-PATH-BH-01-BMSD

Tot Ion:165	Resp:	287906
Ion	Ratio	Lower Upper
165	100	
63	35.2	30.3 45.5
89	61.9	53.4 80.2
182	2.7	2.1 3.1




Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BGC
Ion 89.00 (88.70 to 89.70): BGC
Ion 182.00 (181.70 to 182.70): E






#58
Fluorene
Concen: 48.505 ng
RT: 15.60 min Scan# 2130
Delta R.T. -0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Tgt Ion:166	Resp:	831897
Ion	Ratio	Lower Upper
166	100	
165	99.7	79.9 119.9
167	14.1	11.4 17.0



Abundance Ion 166.00 (165.70 to 166.70): E
Ion 165.00 (164.70 to 165.70): E
Ion 167.00 (166.70 to 167.70): E









#59
2,3,4,6-Tetrachlorophenol
Concen: 51.965 ng
RT: 15.18 min Scan# 2058
Delta R.T. -0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Instrument :
BNA_G
ClientSampleId :
SHORE-RD-PATH-BH-01-BMSD

Tot Ion:232 Resp: 244314
Ion Ratio Lower Upper
232 100
131 42.2 34.2 51.4
130 2.4 1.8 2.8
166 30.7 23.8 35.8




Abundance Ion 232.00 (231.70 to 232.70): E
250000
Ion 131.00 (130.70 to 131.70): E
200000
Ion 130.00 (129.70 to 130.70): E
150000
Ion 166.00 (165.70 to 166.70): E
100000
50000
0






#60
Diethylphthalate
Concen: 47.294 ng
RT: 15.37 min Scan# 2091
Delta R.T. -0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Tgt Ion:149 Resp: 892801
Ion Ratio Lower Upper
149 100
177 24.5 19.4 29.0
150 13.6 10.9 16.3



Abundance Ion 149.00 (148.70 to 149.70): E
800000
Ion 177.00 (176.70 to 177.70): E
600000
Ion 150.00 (149.70 to 150.70): E
400000
200000
0









#61
4-Chlorophenyl-phenylether
Concen: 47.436 ng
RT: 15.59 min Scan# 2128
Delta R.T. -0.01 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Instrument :
BNA_G
ClientSampleId :
SHORE-RD-PATH-BH-01-BMSD

Tot Ion:204 Resp: 441114
Ion Ratio Lower Upper
204 100
206 35.2 27.6 41.4
141 55.8 43.4 65.0




Abundance Ion 204.00 (203.70 to 204.70): E
Ion 206.00 (205.70 to 206.70): E
Ion 141.00 (140.70 to 141.70): E






#62
4-Nitroaniline
Concen: 43.230 ng
RT: 15.62 min Scan# 2133
Delta R.T. 0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Tgt Ion:138 Resp: 201878
Ion Ratio Lower Upper
138 100
92 49.5 31.3 71.3
108 70.6 52.6 92.6



Abundance Ion 138.00 (137.70 to 138.70): E
Ion 92.00 (91.70 to 92.70): BG
Ion 108.00 (107.70 to 108.70): E





#63

Azobenzene

Concen: 47.788 ng

RT: 15.88 min Scan# 2178

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14

Instrument :

BNA_G

ClientSampleId :

SHORE-RD-PATH-BH-01-BMSD

Tot Ion: 77 Resp: 790646


Ion Ratio Lower Upper

77 100

182 28.9 8.2 48.2

105 23.7 3.6 43.6

51 28.5 8.3 48.3






Abundance Ion 77.00 (76.70 to 77.70): BGC

Ion 182.00 (181.70 to 182.70): BGC

Ion 105.00 (104.70 to 105.70): BGC

Ion 51.00 (50.70 to 51.70): BGC



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.29 min Scan# 2418

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14


Tgt Ion: 188 Resp: 545390

Ion Ratio Lower Upper

188 100

94 9.0 7.4 11.2



80 9.6 7.7 11.5




Abundance Ion 188.00 (187.70 to 188.70): BGC

Ion 94.00 (93.70 to 94.70): BGC

Ion 80.00 (79.70 to 80.70): BGC





#65

4,6-Dinitro-2-methylphenol

Concen: 59.119 ng

RT: 15.68 min Scan# 2143

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14

Instrument :

BNA_G

ClientSampleId :

SHORE-RD-PATH-BH-01-BMSD


Tot Ion:198 Resp: 177979

Ion	Ratio	Lower	Upper
198	100		
51	35.7	16.6	56.6
105	40.0	20.6	60.6

198 100

51 35.7 16.6 56.6




105 40.0 20.6 60.6



Abundance Ion 198.00 (197.70 to 198.70): E

Ion 51.00 (50.70 to 51.70): BG

Ion 105.00 (104.70 to 105.70): E



#66

n-Nitrosodiphenylamine

Concen: 51.121 ng

RT: 15.81 min Scan# 2165

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14


Tgt Ion:169 Resp: 781349

Ion	Ratio	Lower	Upper
169	100		
168	72.8	57.9	86.9
167	39.0	30.7	46.1

169 100

168 72.8 57.9 86.9



167 39.0 30.7 46.1




Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E









#67
4-Bromophenyl-phenylether
Concen: 48.652 ng
RT: 16.48 min Scan# 2280
Delta R.T. -0.01 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Instrument :
BNA_G
ClientSampleId :
SHORE-RD-PATH-BH-01-BMSD

Tot Ion:248 Resp: 289090
Ion Ratio Lower Upper
248 100
250 98.5 76.6 114.8
141 69.0 54.7 82.1




Abundance Ion 248.00 (247.70 to 248.70): E
Ion 250.00 (249.70 to 250.70): E
Ion 141.00 (140.70 to 141.70): E






#68
Hexachlorobenzene
Concen: 47.359 ng
RT: 16.61 min Scan# 2301
Delta R.T. -0.01 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Tgt Ion:284 Resp: 315270
Ion Ratio Lower Upper
284 100
142 33.6 25.3 37.9
249 33.2 25.5 38.3



Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E








#69
 Atrazine
 Concen: 55.347 ng
 RT: 16.76 min Scan# 2327
 Delta R.T. -0.00 min
 Lab File: BG044326.D
 Acq: 6 Feb 2020 17:14

Instrument :
 BNA_G
 ClientSampleId :
 SHORE-RD-PATH-BH-01-BMSD

Tot Ion	Ratio	Lower	Upper
200	100		
173	26.6	6.6	46.6
215	51.0	31.0	71.0




Abundance Ion 200.00 (199.70 to 200.70): E
 Ion 173.00 (172.70 to 173.70): E
 Ion 215.00 (214.70 to 215.70): E






#70
 Pentachlorophenol
 Concen: 104.863 ng
 RT: 16.95 min Scan# 2360
 Delta R.T. -0.00 min
 Lab File: BG044326.D
 Acq: 6 Feb 2020 17:14

Tgt Ion	Ratio	Lower	Upper
266	100		
268	65.2	50.6	75.8
264	65.2	48.7	73.1



Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 268.00 (267.70 to 268.70): E
 Ion 264.00 (263.70 to 264.70): E









#71
Phenanthrene
Concen: 53.330 ng
RT: 17.34 min Scan# 2425
Delta R.T. -0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14




Instrument :
BNA_G
ClientSampleId :
SHORE-RD-PATH-BH-01-BMSD


Tot Ion:178 Resp: 1408439
Ion Ratio Lower Upper
178 100
176 23.3 17.7 26.5
179 18.4 13.7 20.5



#72
Anthracene
Concen: 52.593 ng
RT: 17.43 min Scan# 2441
Delta R.T. -0.00 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Tgt Ion:178 Resp: 1373212
Ion Ratio Lower Upper
178 100
176 22.6 17.0 25.4
179 18.1 14.4 21.6





#73

Carbazole

Concen: 49.162 ng

RT: 17.69 min Scan# 2486

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14

Instrument :

BNA_G

ClientSampleId :

SHORE-RD-PATH-BH-01-BMSD


Tot Ion:167 Resp: 1183366

Ion Ratio Lower Upper

167 100

166 26.2 19.8 29.6




139 13.9 11.1 16.7



Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E



#74

Di-n-butylphthalate

Concen: 48.135 ng

RT: 18.26 min Scan# 2583

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14


Tgt Ion:149 Resp: 1431801

Ion Ratio Lower Upper

149 100

150 11.1 8.7 13.1



104 5.5 4.3 6.5




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E





#75

Fluoranthene

Concen: 58.195 ng

RT: 19.35 min Scan# 2768

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14

Instrument :

BNA_G

ClientSampleId :

SHORE-RD-PATH-BH-01-BMSD


Tot Ion:202 Resp: 1777909

Ion Ratio Lower Upper

202 100

101 12.1 0.0 31.3




203 22.0 0.0 39.9



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.54 min Scan# 3141

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14


Tgt Ion:240 Resp: 463067

Ion Ratio Lower Upper

240 100

120 10.7 8.4 12.6



236 28.5 22.1 33.1




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#77

Benzidine

Concen: 26.609 ng

RT: 19.53 min Scan# 2798

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14

Instrument :

BNA_G

ClientSampleId :

SHORE-RD-PATH-BH-01-BMSD


Tot Ion:184 Resp: 341774

Ion Ratio Lower Upper

184 100

185 14.7 12.9 19.3




183 13.0 10.8 16.2



Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E



#78

Pyrene

Concen: 60.292 ng

RT: 19.71 min Scan# 2829

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14


Tgt Ion:202 Resp: 1792971

Ion Ratio Lower Upper

202 100

200 26.9 20.0 30.0



203 22.2 16.2 24.2




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E





#79

Terphenyl-d14

Concen: 79.919 ng

RT: 19.91 min Scan# 2863

Delta R.T. -0.01 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14

Instrument :

BNA_G

ClientSampleId :

SHORE-RD-PATH-BH-01-BMSD


Tot Ion:244 Resp: 1799239

Ion Ratio Lower Upper

244 100

212 9.3 7.4 11.2




122 12.2 9.8 14.6



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#80

Butylbenzylphthalate

Concen: 48.965 ng

RT: 20.60 min Scan# 2981

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14


Tgt Ion:149 Resp: 663575

Ion Ratio Lower Upper

149 100

91 68.2 54.5 81.7



206 23.8 19.1 28.7




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BG

Ion 206.00 (205.70 to 206.70): E





#81

Benzo(a)anthracene

Concen: 56.984 ng

RT: 21.52 min Scan# 3138

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14

Instrument :

BNA_G

ClientSampleId :

SHORE-RD-PATH-BH-01-BMSD


Tot Ion:228 Resp: 1626268

Ion Ratio Lower Upper

228 100

226 33.6 24.7 37.1

229 23.8 18.2 27.4






Abundance

Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#82

3,3'-Dichlorobenzidine

Concen: 29.518 ng

RT: 21.44 min Scan# 3124

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14


Tgt Ion:252 Resp: 326954

Ion Ratio Lower Upper

252 100

254 64.7 52.6 78.8

126 9.9 8.2 12.4






Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E





#83

Chrysene

Concen: 55.339 ng

RT: 21.59 min Scan# 3149

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14

Instrument :

BNA_G

ClientSampleId :

SHORE-RD-PATH-BH-01-BMSD


Tot Ion:228 Resp: 1526574

Ion Ratio Lower Upper

228 100

226 35.6 27.1 40.7




229 23.7 17.9 26.9



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#84

Bis(2-ethylhexyl)phthalate

Concen: 50.540 ng

RT: 21.44 min Scan# 3124

Delta R.T. -0.01 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14


Tgt Ion:149 Resp: 942744

Ion Ratio Lower Upper

149 100

167 30.7 24.0 36.0



279 4.8 3.8 5.8




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E





#85

Di-n-octyl phthalate

Concen: 50.844 ng

RT: 22.61 min Scan# 3323

Delta R.T. -0.01 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14

Instrument :

BNA_G

ClientSampleId :

SHORE-RD-PATH-BH-01-BMSD


Tot Ion:149 Resp: 1640970

Ion Ratio Lower Upper

149 100

167 1.7 1.3 1.9




43 9.0 7.4 11.0



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BGC



#86

Indeno(1,2,3-cd)pyrene

Concen: 52.860 ng

RT: 28.15 min Scan# 4265

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14


Tgt Ion:276 Resp: 1902411

Ion Ratio Lower Upper

276 100

138 24.5 19.4 29.0



277 25.9 21.0 31.6




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E





#87

Pervlene-d12

Concen: 20.000 ng

RT: 24.64 min Scan# 3668

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14

Instrument :

BNA_G

ClientSampleId :

SHORE-RD-PATH-BH-01-BMSD


Tot Ion:264 Resp: 536690

Ion Ratio Lower Upper

264 100

260 25.2 19.4 29.2




265 21.6 16.7 25.1



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E



#88

Benzo(b)fluoranthene

Concen: 55.134 ng

RT: 23.66 min Scan# 3502

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14


Tgt Ion:252 Resp: 1705078

Ion Ratio Lower Upper

252 100

253 24.2 19.2 28.8



125 11.0 8.4 12.6




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#89

Benzo(k)fluoranthene

Concen: 52.269 ng

RT: 23.73 min Scan# 3513

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14

Instrument :

BNA_G

ClientSampleId :

SHORE-RD-PATH-BH-01-BMSD


Tot Ion:252 Resp: 1575473

Ion Ratio Lower Upper

252 100

253 25.0 19.1 28.7




125 10.9 8.6 12.8



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Benzo(a)pyrene

Concen: 54.320 ng

RT: 24.50 min Scan# 3644

Delta R.T. -0.00 min

Lab File: BG044326.D

Acq: 6 Feb 2020 17:14


Tgt Ion:252 Resp: 1588336

Ion Ratio Lower Upper

252 100

253 24.5 18.1 27.1



125 11.5 8.6 13.0




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E






#91
Dibenzo(a,h)anthracene
Concen: 51.525 ng
RT: 28.21 min Scan# 4275
Delta R.T. -0.01 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14




Instrument :
BNA_G
ClientSampleId :
SHORE-RD-PATH-BH-01-BMSD

Tot Ion:278 Resp: 1491030

Ion	Ratio	Lower	Upper
278	100		
139	15.5	12.7	19.1
279	25.1	19.3	28.9




Abundance Ion 278.00 (277.70 to 278.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 279.00 (278.70 to 279.70): E



#92
Benzo(g,h,i)perylene
Concen: 55.619 ng
RT: 29.25 min Scan# 4452
Delta R.T. -0.01 min
Lab File: BG044326.D
Acq: 6 Feb 2020 17:14

Tgt Ion:276 Resp: 1611213

Ion	Ratio	Lower	Upper
276	100		
277	24.4	18.8	28.2
138	21.3	16.2	24.2



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 277.00 (276.70 to 277.70): E
Ion 138.00 (137.70 to 138.70): E

