

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062622\
 Data File : BG054072.D
 Acq On : 26 Jun 2022 14:36
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 27 00:28:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 00:27:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	103	0.00
2	1,4-Dioxane	40.000	40.123	-0.3	102	0.00
3	Pyridine	40.000	42.897	-7.2	107	0.00
4	n-Nitrosodimethylamine	40.000	43.003	-7.5	105	0.00
5 S	2-Fluorophenol	80.000	80.763	-1.0	102	0.00
6	Aniline	40.000	43.335	-8.3	109	0.00
7 S	Phenol-d6	80.000	88.728	-10.9	113	0.00
8	2-Chlorophenol	40.000	42.780	-7.0	109	0.00
9	Benzaldehyde	40.000	39.874	0.3	111	0.00
10 C	Phenol	40.000	45.412	-13.5	113	0.00
11	bis(2-Chloroethyl)ether	40.000	42.599	-6.5	110	0.00
12	1,3-Dichlorobenzene	40.000	38.228	4.4	100	0.00
13 C	1,4-Dichlorobenzene	40.000	39.455	1.4	105	0.00
14	1,2-Dichlorobenzene	40.000	38.901	2.7	100	0.00
15	Benzyl Alcohol	40.000	44.774	-11.9	111	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	46.972	-17.4	121	0.00
17	2-Methylphenol	40.000	44.366	-10.9	113	0.00
18	Hexachloroethane	40.000	40.445	-1.1	104	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	43.618	-9.0	111	0.00
20	3+4-Methylphenols	40.000	45.636	-14.1	116	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	116	0.00
22	Acetophenone	40.000	38.433	3.9	111	0.00
23 S	Nitrobenzene-d5	80.000	80.204	-0.3	114	0.00
24	Nitrobenzene	40.000	40.630	-1.6	116	0.00
25	Isophorone	40.000	39.077	2.3	113	0.00
26 C	2-Nitrophenol	40.000	42.351	-5.9	118	0.00
27	2,4-Dimethylphenol	40.000	40.615	-1.5	116	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.378	1.6	114	0.00
29 C	2,4-Dichlorophenol	40.000	41.818	-4.5	119	0.00
30	1,2,4-Trichlorobenzene	40.000	35.960	10.1	104	0.00
31	Naphthalene	40.000	38.866	2.8	112	0.00
32	Benzoic acid	40.000	42.250	-5.6	140	0.00
33	4-Chloroaniline	40.000	41.494	-3.7	117	0.00
34 C	Hexachlorobutadiene	40.000	32.854	17.9	93	0.00
35	Caprolactam	40.000	46.766	-16.9	128	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.981	-10.0	123	0.00
37	2-Methylnaphthalene	40.000	39.664	0.8	114	0.00
38	1-Methylnaphthalene	40.000	39.618	1.0	115	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	124	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	33.534	16.2	102	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.959	2.6	116	0.00
42 S	2,4,6-Tribromophenol	80.000	72.107	9.9	108	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.238	4.4	113	0.00
44	2,4,5-Trichlorophenol	40.000	38.720	3.2	115	0.00
45 S	2-Fluorobiphenyl	80.000	72.782	9.0	112	0.00
46	1,1'-Biphenyl	40.000	37.607	6.0	115	0.00
47	2-Chloronaphthalene	40.000	37.537	6.2	115	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062622\
 Data File : BG054072.D
 Acq On : 26 Jun 2022 14:36
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 27 00:28:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 00:27:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	45.943	-14.9	135	0.00
49	Acenaphthylene	40.000	39.124	2.2	119	0.00
50	Dimethylphthalate	40.000	37.944	5.1	116	0.00
51	2,6-Dinitrotoluene	40.000	41.645	-4.1	124	0.00
52 C	Acenaphthene	40.000	39.886	0.3	121	0.00
53	3-Nitroaniline	40.000	44.163	-10.4	129	0.00
54 P	2,4-Dinitrophenol	40.000	38.291	4.3	125	0.00
55	Dibenzofuran	40.000	38.577	3.6	118	0.00
56 P	4-Nitrophenol	40.000	43.165	-7.9	127	0.00
57	2,4-Dinitrotoluene	40.000	42.865	-7.2	126	0.00
58	Fluorene	40.000	39.089	2.3	120	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.363	4.1	114	0.00
60	Diethylphthalate	40.000	39.659	0.9	122	0.00
61	4-Chlorophenyl-phenylether	40.000	36.331	9.2	112	0.00
62	4-Nitroaniline	40.000	42.575	-6.4	128	0.00
63	Azobenzene	40.000	43.336	-8.3	134	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	127	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.388	-3.5	127	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.758	3.1	121	0.00
67	4-Bromophenyl-phenylether	40.000	35.495	11.3	111	0.00
68	Hexachlorobenzene	40.000	33.936	15.2	107	0.00
69	Atrazine	40.000	38.488	3.8	117	0.00
70 C	Pentachlorophenol	40.000	36.077	9.8	109	0.00
71	Phenanthrene	40.000	38.569	3.6	121	0.00
72	Anthracene	40.000	38.792	3.0	121	0.00
73	Carbazole	40.000	39.709	0.7	125	0.00
74	Di-n-butylphthalate	40.000	41.030	-2.6	127	0.00
75 C	Fluoranthene	40.000	37.428	6.4	118	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	119	0.00
77	Benzydine	40.000	33.140	17.1	95	0.00
78	Pyrene	40.000	40.609	-1.5	118	0.00
79 S	Terphenyl-d14	80.000	76.707	4.1	111	0.00
80	Butylbenzylphthalate	40.000	45.357	-13.4	130	0.00
81	Benzo(a)anthracene	40.000	39.401	1.5	116	0.00
82	3,3'-Dichlorobenzidine	40.000	38.413	4.0	110	0.00
83	Chrysene	40.000	38.805	3.0	115	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.866	-12.2	129	0.00
85 c	Di-n-octyl phthalate	40.000	44.949	-12.4	128	0.00
86 I	Perylene-d12	20.000	20.000	0.0	116	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.369	4.1	109	0.00
88	Benzo(b)fluoranthene	40.000	39.040	2.4	110	0.00
89	Benzo(k)fluoranthene	40.000	38.718	3.2	113	0.00
90 C	Benzo(a)pyrene	40.000	39.156	2.1	111	0.00
91	Dibenzo(a,h)anthracene	40.000	38.262	4.3	110	0.00
92	Benzo(g,h,i)perylene	40.000	37.560	6.1	107	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062622\
Data File : BG054072.D
Acq On : 26 Jun 2022 14:36
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jun 27 00:28:31 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 27 00:27:54 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0