

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG082725\
 Data File : BG064207.D
 Acq On : 27 Aug 2025 12:13
 Operator : RC/JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Quant Time: Aug 27 17:45:18 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG082725.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 27 17:42:35 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.859	152	27014	20.000	ng	0.00
21) Naphthalene-d8	10.644	136	127904	20.000	ng	# 0.00
39) Acenaphthene-d10	14.481	164	92630	20.000	ng	0.00
64) Phenanthrene-d10	17.219	188	216187	20.000	ng	0.00
76) Chrysene-d12	21.449	240	231921	20.000	ng	0.00
86) Perylene-d12	24.463	264	252321	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0d	0.000	ng	
7) Phenol-d6	0.000	99	0d	0.000	ng	
23) Nitrobenzene-d5	0.000	82	0d	0.000	ng	
42) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
45) 2-Fluorobiphenyl	0.000	172	0d	0.000	ng	
79) Terphenyl-d14	0.000	244	0d	0.000	ng	
Target Compounds						
19) n-Nitroso-di-n-propyla...	8.652	70	3684	2.335	ng	Qvalue # 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG082725\
Data File : BG064207.D
Acq On : 27 Aug 2025 12:13
Operator : RC/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Quant Time: Aug 27 17:45:18 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG082725.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 27 17:42:35 2025
Response via : Initial Calibration

