

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG090420\  
 Data File : BG046537.D  
 Acq On : 4 Sep 2020 12:26  
 Operator : CG/JU  
 Sample : PB131426BL  
 Misc :  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 PB131426BL

Quant Time: Sep 04 13:35:30 2020  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG082520.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Sep 02 15:06:53 2020  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 7.93  | 152  | 78098    | 20.00  | ng    | 0.00     |
| 21) Naphthalene-d8          | 10.73 | 136  | 314776   | 20.00  | ng    | 0.00     |
| 39) Acenaphthene-d10        | 14.56 | 164  | 215296   | 20.00  | ng    | 0.00     |
| 64) Phenanthrene-d10        | 17.30 | 188  | 505028   | 20.00  | ng    | 0.00     |
| 76) Chrysene-d12            | 21.55 | 240  | 513946   | 20.00  | ng    | 0.00     |
| 86) Perylene-d12            | 24.65 | 264  | 504996   | 20.00  | ng    | 0.00     |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.52  | 112  | 527622   | 119.66 | ng    | 0.00     |
| 7) Phenol-d6                | 7.11  | 99   | 683027   | 109.27 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 9.09  | 82   | 456074   | 82.81  | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 16.05 | 330  | 317390   | 108.80 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 13.18 | 172  | 1183420  | 83.91  | ng    | 0.00     |
| 79) Terphenyl-d14           | 19.92 | 244  | 1787454  | 74.02  | ng    | 0.00     |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG090420\  
Data File : BG046537.D  
Acq On : 4 Sep 2020 12:26  
Operator : CG/JU  
Sample : PB131426BL  
Misc :  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
BNA\_G  
ClientSampleId :  
PB131426BL

Quant Time: Sep 04 13:35:30 2020  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG082520.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Sep 02 15:06:53 2020  
Response via : Initial Calibration

