

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG092122\
 Data File : BG054993.D
 Acq On : 21 Sep 2022 12:10
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 22 04:39:35 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG091922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 19 17:41:40 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	117	0.00
2	1,4-Dioxane	40.000	38.693	3.3	120	0.00
3	Pyridine	40.000	41.234	-3.1	121	0.00
4	n-Nitrosodimethylamine	40.000	40.758	-1.9	122	0.00
5 S	2-Fluorophenol	80.000	83.282	-4.1	123	0.00
6	Aniline	40.000	40.333	-0.8	115	0.00
7 S	Phenol-d6	80.000	82.564	-3.2	117	0.00
8	2-Chlorophenol	40.000	41.299	-3.2	120	0.00
9	Benzaldehyde	40.000	38.380	4.0	115	0.00
10 C	Phenol	40.000	41.235	-3.1	117	0.00
11	bis(2-Chloroethyl)ether	40.000	39.578	1.1	116	0.00
12	1,3-Dichlorobenzene	40.000	41.143	-2.9	124	0.00
13 C	1,4-Dichlorobenzene	40.000	40.841	-2.1	122	0.00
14	1,2-Dichlorobenzene	40.000	40.813	-2.0	120	0.00
15	Benzyl Alcohol	40.000	40.886	-2.2	112	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	35.558	11.1	103	0.00
17	2-Methylphenol	40.000	41.436	-3.6	116	0.00
18	Hexachloroethane	40.000	40.837	-2.1	122	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.116	2.2	108	0.00
20	3+4-Methylphenols	40.000	41.043	-2.6	114	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	110	0.00
22	Acetophenone	40.000	40.302	-0.8	113	0.00
23 S	Nitrobenzene-d5	80.000	80.929	-1.2	113	0.00
24	Nitrobenzene	40.000	40.018	-0.0	111	0.00
25	Isophorone	40.000	39.453	1.4	107	0.00
26 C	2-Nitrophenol	40.000	41.449	-3.6	113	0.00
27	2,4-Dimethylphenol	40.000	41.538	-3.8	113	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.185	-0.5	111	0.00
29 C	2,4-Dichlorophenol	40.000	42.158	-5.4	114	0.00
30	1,2,4-Trichlorobenzene	40.000	41.091	-2.7	115	0.00
31	Naphthalene	40.000	40.694	-1.7	115	0.00
32	Benzoic acid	40.000	14.911	62.7#	85	0.03
33	4-Chloroaniline	40.000	40.903	-2.3	111	0.00
34 C	Hexachlorobutadiene	40.000	40.678	-1.7	116	0.00
35	Caprolactam	40.000	39.455	1.4	104	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.379	-0.9	107	0.00
37	2-Methylnaphthalene	40.000	40.563	-1.4	110	0.00
38	1-Methylnaphthalene	40.000	40.531	-1.3	111	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.130	-5.3	110	0.00
41 P	Hexachlorocyclopentadiene	40.000	23.774	40.6#	77	0.00
42 S	2,4,6-Tribromophenol	80.000	81.685	-2.1	100	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.918	-4.8	103	0.00
44	2,4,5-Trichlorophenol	40.000	41.759	-4.4	103	0.00
45 S	2-Fluorobiphenyl	80.000	83.360	-4.2	109	0.00
46	1,1'-Biphenyl	40.000	42.109	-5.3	109	0.00
47	2-Chloronaphthalene	40.000	41.653	-4.1	108	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG092122\
 Data File : BG054993.D
 Acq On : 21 Sep 2022 12:10
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 22 04:39:35 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG091922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 19 17:41:40 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.293	-0.7	100	0.00
49	Acenaphthylene	40.000	40.936	-2.3	106	0.00
50	Dimethylphthalate	40.000	40.006	-0.0	102	0.00
51	2,6-Dinitrotoluene	40.000	40.942	-2.4	103	0.00
52 C	Acenaphthene	40.000	41.184	-3.0	105	0.00
53	3-Nitroaniline	40.000	40.809	-2.0	101	0.00
54 P	2,4-Dinitrophenol	40.000	33.819	15.5	78	0.00
55	Dibenzofuran	40.000	41.041	-2.6	105	0.00
56 P	4-Nitrophenol	40.000	36.827	7.9	93	0.00
57	2,4-Dinitrotoluene	40.000	41.134	-2.8	101	0.00
58	Fluorene	40.000	40.335	-0.8	103	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	33.458	16.4	82	0.00
60	Diethylphthalate	40.000	39.873	0.3	102	0.00
61	4-Chlorophenyl-phenylether	40.000	40.382	-1.0	103	0.00
62	4-Nitroaniline	40.000	40.773	-1.9	100	0.00
63	Azobenzene	40.000	39.833	0.4	100	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	95	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.998	0.0	97	0.00
66 c	n-Nitrosodiphenylamine	40.000	42.198	-5.5	102	0.00
67	4-Bromophenyl-phenylether	40.000	41.777	-4.4	99	0.00
68	Hexachlorobenzene	40.000	41.718	-4.3	101	0.00
69	Atrazine	40.000	41.082	-2.7	99	0.00
70 C	Pentachlorophenol	40.000	34.623	13.4	83	0.00
71	Phenanthrene	40.000	41.126	-2.8	100	0.00
72	Anthracene	40.000	41.094	-2.7	99	0.00
73	Carbazole	40.000	41.615	-4.0	99	0.00
74	Di-n-butylphthalate	40.000	40.866	-2.2	97	0.00
75 C	Fluoranthene	40.000	40.625	-1.6	98	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	98	0.00
77	Benzidine	40.000	32.610	18.5	82	0.00
78	Pyrene	40.000	39.881	0.3	99	0.00
79 S	Terphenyl-d14	80.000	79.141	1.1	100	0.00
80	Butylbenzylphthalate	40.000	41.156	-2.9	103	0.00
81	Benzo(a)anthracene	40.000	40.951	-2.4	102	0.00
82	3,3'-Dichlorobenzidine	40.000	41.293	-3.2	103	0.00
83	Chrysene	40.000	40.735	-1.8	103	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.387	-6.0	106	0.00
85 c	Di-n-octyl phthalate	40.000	42.309	-5.8	106	0.00
86 I	Perylene-d12	20.000	20.000	0.0	100	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.427	-1.1	105	0.00
88	Benzo(b)fluoranthene	40.000	41.244	-3.1	106	0.00
89	Benzo(k)fluoranthene	40.000	40.733	-1.8	103	0.00
90 C	Benzo(a)pyrene	40.000	40.887	-2.2	105	0.00
91	Dibenzo(a,h)anthracene	40.000	40.249	-0.6	105	0.00
92	Benzo(g,h,i)perylene	40.000	40.054	-0.1	105	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG092122\
Data File : BG054993.D
Acq On : 21 Sep 2022 12:10
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Sep 22 04:39:35 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG091922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 19 17:41:40 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0