

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG092122\
 Data File : BG055005.D
 Acq On : 21 Sep 2022 22:52
 Operator : CG/JU
 Sample : N4724-04
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SB-10-COMP

Quant Time: Sep 22 04:45:54 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG091922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 19 17:41:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.058	152	35053	20.000	ng	0.00
21) Naphthalene-d8	10.884	136	140356	20.000	ng	# 0.00
39) Acenaphthene-d10	14.703	164	96975	20.000	ng	0.00
64) Phenanthrene-d10	17.447	188	237335	20.000	ng	0.00
76) Chrysene-d12	21.724	240	244612	20.000	ng	0.00
86) Perylene-d12	25.020	264	254658	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.596	112	235794	131.622	ng	0.00
7) Phenol-d6	7.218	99	309456	128.875	ng	0.00
23) Nitrobenzene-d5	9.233	82	176872	73.042	ng	0.00
42) 2,4,6-Tribromophenol	16.190	330	142197	125.167	ng	0.00
45) 2-Fluorobiphenyl	13.322	172	478829	77.816	ng	0.00
79) Terphenyl-d14	20.050	244	909020	77.736	ng	0.00


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG092122\
Data File : BG055005.D
Acq On : 21 Sep 2022 22:52
Operator : CG/JU
Sample : N4724-04
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SB-10-COMP

Quant Time: Sep 22 04:45:54 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG091922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 19 17:41:40 2022
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.058 min Scan# 812
Delta R.T. -0.003 min
Lab File: BG055005.D
Acq: 21 Sep 2022 22:52




Instrument :
BNA_G
ClientSampleId :
SB-10-COMP


Tgt Ion:152 Resp: 35053
Ion Ratio Lower Upper
152 100
150 153.9 125.2 187.8
115 55.7 47.9 71.9



#5
2-Fluorophenol
Concen: 131.622 ng
RT: 5.596 min Scan# 393
Delta R.T. -0.003 min
Lab File: BG055005.D
Acq: 21 Sep 2022 22:52

Tgt Ion:112 Resp: 235794
Ion Ratio Lower Upper
112 100
64 52.8 49.3 73.9
63 28.5 26.3 39.5









#7
Phenol-d6
Concen: 128.875 ng
RT: 7.218 min Scan# 669
Delta R.T. -0.003 min
Lab File: BG055005.D
Acq: 21 Sep 2022 22:52




Instrument :
BNA_G
ClientSampleId :
SB-10-COMP


Tgt Ion: 99 Resp: 309456
Ion Ratio Lower Upper
99 100
42 14.4 14.6 22.0#
71 29.5 24.6 37.0



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.884 min Scan# 1293
Delta R.T. -0.003 min
Lab File: BG055005.D
Acq: 21 Sep 2022 22:52

Tgt Ion: 136 Resp: 140356
Ion Ratio Lower Upper
136 100
137 10.4 9.2 13.8
54 6.2 7.7 11.5#
68 4.6 5.6 8.4#





#23

Nitrobenzene-d5

Concen: 73.042 ng

RT: 9.233 min Scan# 1012

Delta R.T. -0.009 min

Lab File: BG055005.D

Acq: 21 Sep 2022 22:52

Instrument :

BNA_G

ClientSampleId :

SB-10-COMP





Tgt Ion: 82 Resp: 176872

Ion Ratio Lower Upper

82 100

128 46.4 32.2 48.2

54 44.4 41.4 62.0



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.703 min Scan# 1943

Delta R.T. -0.003 min

Lab File: BG055005.D

Acq: 21 Sep 2022 22:52




Tgt Ion:164 Resp: 96975


Ion Ratio Lower Upper

164 100

162 104.5 83.0 124.4

160 46.9 38.0 57.0





#42

2,4,6-Tribromophenol

Concen: 125.167 ng

RT: 16.190 min Scan# 21

Delta R.T. -0.009 min

Lab File: BG055005.D


Acq: 21 Sep 2022 22:52

Instrument :

BNA_G

ClientSampleId :

SB-10-COMP






Tgt Ion:330 Resp: 142197

Ion Ratio Lower Upper

330 100

332 95.0 79.8 119.6

141 27.2 26.8 40.2



#45

2-Fluorobiphenyl

Concen: 77.816 ng

RT: 13.322 min Scan# 1708

Delta R.T. -0.003 min

Lab File: BG055005.D

Acq: 21 Sep 2022 22:52




Tgt Ion:172 Resp: 478829


Ion Ratio Lower Upper

172 100

171 36.8 31.2 46.8

170 24.3 20.6 30.8









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.447 min Scan# 2410
Delta R.T. -0.003 min
Lab File: BG055005.D
Acq: 21 Sep 2022 22:52




Instrument :
BNA_G
ClientSampleId :
SB-10-COMP


Tgt Ion:188 Resp: 237335
Ion Ratio Lower Upper
188 100
94 7.5 7.0 10.4
80 8.2 7.9 11.9



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.724 min Scan# 3138
Delta R.T. -0.009 min
Lab File: BG055005.D
Acq: 21 Sep 2022 22:52

Tgt Ion:240 Resp: 244612
Ion Ratio Lower Upper
240 100
120 8.8 8.2 12.4
236 26.0 21.2 31.8









#79
 Terphenyl-d14
 Concen: 77.736 ng
 RT: 20.050 min Scan# 2853
 Delta R.T. -0.003 min
 Lab File: BG055005.D
 Acq: 21 Sep 2022 22:52

Instrument :
 BNA_G
 ClientSampleId :
 SB-10-COMP

Tgt Ion:244 Resp: 909020
 Ion Ratio Lower Upper
 244 100
 212 7.1 6.2 9.2
 122 8.7 9.0 13.6#



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 25.020 min Scan# 3699
 Delta R.T. -0.014 min
 Lab File: BG055005.D
 Acq: 21 Sep 2022 22:52

Tgt Ion:264 Resp: 254658
 Ion Ratio Lower Upper
 264 100
 260 24.1 19.9 29.9
 265 21.6 16.9 25.3

