


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102121\
 Data File : BG050667.D
 Acq On : 21 Oct 2021 14:28
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Oct 14 10:56:22 2021






AutoFind: Scans 2472, 2473, 2474; Background Corrected with Scan 2465

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	45.6	30947	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	46.1	31312	PASS
70	69	0.00	2	0.7	222	PASS
127	198	10	80	51.0	34640	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	67925	PASS
199	198	5	9	6.8	4636	PASS
275	198	10	60	27.0	18325	PASS
365	198	1	100	3.7	2488	PASS
441	198	0.01	100	9.1	6212	PASS
442	442	50	100	100.0	37301	PASS
443	442	15	24	18.8	7012	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102121\
Data File : BG050667.D
Acq On : 21 Oct 2021 14:28
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Oct 21 15:31:21 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 14 10:56:22 2021
Response via : Initial Calibration



TIC: BG050667.D\data.ms

(70) Pentachlorophenol (C)

17.262min (-0.004) 18465.95 ng

response 61564

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.20	62.55
264.00	61.30	59.41
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/21/2021	BNA_G	BG050667.D
Compound Name	Response	Retention Time
DDT	204956	21.06
DDD	7503	20.61
DDE	1450	20.1
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
8953	213909	4.19

Instrument :
BNA_G
ClientSampleId :
DFTPP