

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102221\
 Data File : BG050683.D
 Acq On : 22 Oct 2021 10:28
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 22 11:59:25 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 14 10:56:22 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	93	0.00
2	1,4-Dioxane	40.000	38.938	2.7	91	0.00
3	Pyridine	40.000	36.541	8.6	86	0.00
4	n-Nitrosodimethylamine	40.000	38.659	3.4	94	0.00
5 S	2-Fluorophenol	80.000	77.444	3.2	93	0.00
6	Aniline	40.000	37.367	6.6	92	0.00
7 S	Phenol-d6	80.000	76.672	4.2	93	0.00
8	2-Chlorophenol	40.000	39.422	1.4	96	0.00
9	Benzaldehyde	40.000	40.058	-0.1	89	0.00
10 C	Phenol	40.000	38.272	4.3	94	0.00
11	bis(2-Chloroethyl)ether	40.000	38.244	4.4	93	0.00
12	1,3-Dichlorobenzene	40.000	38.134	4.7	94	0.00
13 C	1,4-Dichlorobenzene	40.000	38.207	4.5	94	0.00
14	1,2-Dichlorobenzene	40.000	37.825	5.4	93	0.00
15	Benzyl Alcohol	40.000	38.481	3.8	94	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.946	7.6	91	0.00
17	2-Methylphenol	40.000	38.423	3.9	95	0.00
18	Hexachloroethane	40.000	39.783	0.5	96	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	36.222	9.4	89	0.00
20	3+4-Methylphenols	40.000	38.150	4.6	92	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	97	0.00
22	Acetophenone	40.000	36.505	8.7	92	0.00
23 S	Nitrobenzene-d5	80.000	74.236	7.2	91	0.00
24	Nitrobenzene	40.000	36.715	8.2	91	0.00
25	Isophorone	40.000	35.368	11.6	88	0.00
26 C	2-Nitrophenol	40.000	41.285	-3.2	100	0.00
27	2,4-Dimethylphenol	40.000	36.564	8.6	90	0.00
28	bis(2-Chloroethoxy)methane	40.000	36.465	8.8	91	0.00
29 C	2,4-Dichlorophenol	40.000	37.600	6.0	92	0.00
30	1,2,4-Trichlorobenzene	40.000	37.601	6.0	94	0.00
31	Naphthalene	40.000	36.686	8.3	91	0.00
32	Benzoic acid	40.000	36.875	7.8	86	0.00
33	4-Chloroaniline	40.000	36.860	7.9	92	0.00
34 C	Hexachlorobutadiene	40.000	37.573	6.1	94	0.00
35	Caprolactam	40.000	35.426	11.4	86	0.00
36 C	4-Chloro-3-methylphenol	40.000	36.126	9.7	90	0.00
37	2-Methylnaphthalene	40.000	35.509	11.2	89	0.00
38	1-Methylnaphthalene	40.000	35.488	11.3	89	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	91	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.063	4.8	89	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.278	4.3	86	0.00
42 S	2,4,6-Tribromophenol	80.000	76.027	5.0	88	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.833	2.9	89	0.00
44	2,4,5-Trichlorophenol	40.000	39.325	1.7	91	0.00
45 S	2-Fluorobiphenyl	80.000	75.046	6.2	88	0.00
46	1,1'-Biphenyl	40.000	37.648	5.9	88	0.00
47	2-Chloronaphthalene	40.000	38.429	3.9	90	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102221\
 Data File : BG050683.D
 Acq On : 22 Oct 2021 10:28
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 22 11:59:25 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 14 10:56:22 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.213	-0.5	91	0.00
49	Acenaphthylene	40.000	37.499	6.3	88	0.00
50	Dimethylphthalate	40.000	37.039	7.4	87	0.00
51	2,6-Dinitrotoluene	40.000	38.726	3.2	89	0.00
52 C	Acenaphthene	40.000	34.056	14.9	79	0.00
53	3-Nitroaniline	40.000	38.795	3.0	89	0.00
54 P	2,4-Dinitrophenol	40.000	35.721	10.7	80	0.00
55	Dibenzofuran	40.000	36.943	7.6	88	0.00
56 P	4-Nitrophenol	40.000	36.900	7.8	85	0.00
57	2,4-Dinitrotoluene	40.000	39.582	1.0	91	0.00
58	Fluorene	40.000	36.641	8.4	86	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	35.866	10.3	82	0.00
60	Diethylphthalate	40.000	37.374	6.6	88	0.00
61	4-Chlorophenyl-phenylether	40.000	36.840	7.9	86	0.00
62	4-Nitroaniline	40.000	39.464	1.3	92	0.00
63	Azobenzene	40.000	36.453	8.9	86	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	91	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.708	3.2	89	0.00
66 c	n-Nitrosodiphenylamine	40.000	36.170	9.6	84	0.00
67	4-Bromophenyl-phenylether	40.000	37.777	5.6	88	0.00
68	Hexachlorobenzene	40.000	37.580	6.1	88	0.00
69	Atrazine	40.000	36.545	8.6	85	0.00
70 C	Pentachlorophenol	40.000	36.977	7.6	84	0.00
71	Phenanthrene	40.000	36.955	7.6	87	0.00
72	Anthracene	40.000	36.923	7.7	87	0.00
73	Carbazole	40.000	37.179	7.1	88	0.00
74	Di-n-butylphthalate	40.000	39.017	2.5	92	0.00
75 C	Fluoranthene	40.000	36.673	8.3	87	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	88	0.00
77	Benzydine	40.000	33.148	17.1	76	0.00
78	Pyrene	40.000	38.729	3.2	87	0.00
79 S	Terphenyl-d14	80.000	79.230	1.0	90	0.00
80	Butylbenzylphthalate	40.000	41.536	-3.8	93	0.00
81	Benzo(a)anthracene	40.000	38.957	2.6	89	0.00
82	3,3'-Dichlorobenzidine	40.000	38.908	2.7	88	0.00
83	Chrysene	40.000	38.751	3.1	88	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.842	0.4	89	0.00
85 c	Di-n-octyl phthalate	40.000	41.058	-2.6	91	0.00
86 I	Perylene-d12	20.000	20.000	0.0	89	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.467	3.8	88	0.00
88	Benzo(b)fluoranthene	40.000	38.616	3.5	88	0.00
89	Benzo(k)fluoranthene	40.000	38.789	3.0	89	0.00
90 C	Benzo(a)pyrene	40.000	38.992	2.5	89	0.00
91	Dibenzo(a,h)anthracene	40.000	38.438	3.9	88	0.00
92	Benzo(g,h,i)perylene	40.000	38.905	2.7	89	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102221\
Data File : BG050683.D
Acq On : 22 Oct 2021 10:28
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Oct 22 11:59:25 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 14 10:56:22 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0