

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102221\
 Data File : BG050690.D
 Acq On : 22 Oct 2021 15:15
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 22 16:05:08 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 14 10:56:22 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	107	0.00
2	1,4-Dioxane	40.000	37.979	5.1	103	0.00
3	Pyridine	40.000	36.288	9.3	99	0.00
4	n-Nitrosodimethylamine	40.000	37.622	5.9	105	0.00
5 S	2-Fluorophenol	80.000	75.945	5.1	105	0.00
6	Aniline	40.000	36.966	7.6	105	0.00
7 S	Phenol-d6	80.000	76.067	4.9	107	0.00
8	2-Chlorophenol	40.000	38.193	4.5	107	0.00
9	Benzaldehyde	40.000	38.972	2.6	100	0.00
10 C	Phenol	40.000	37.718	5.7	107	0.00
11	bis(2-Chloroethyl)ether	40.000	36.767	8.1	103	0.00
12	1,3-Dichlorobenzene	40.000	37.231	6.9	106	0.00
13 C	1,4-Dichlorobenzene	40.000	37.025	7.4	105	0.00
14	1,2-Dichlorobenzene	40.000	36.538	8.7	104	0.00
15	Benzyl Alcohol	40.000	38.594	3.5	109	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.311	9.2	103	0.00
17	2-Methylphenol	40.000	37.998	5.0	109	0.00
18	Hexachloroethane	40.000	37.945	5.1	106	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	35.813	10.5	101	0.00
20	3+4-Methylphenols	40.000	38.133	4.7	106	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	111	0.00
22	Acetophenone	40.000	35.904	10.2	103	0.00
23 S	Nitrobenzene-d5	80.000	74.051	7.4	104	0.00
24	Nitrobenzene	40.000	36.682	8.3	104	0.00
25	Isophorone	40.000	36.099	9.8	103	0.00
26 C	2-Nitrophenol	40.000	40.907	-2.3	113	0.00
27	2,4-Dimethylphenol	40.000	37.201	7.0	104	0.00
28	bis(2-Chloroethoxy)methane	40.000	36.444	8.9	104	0.00
29 C	2,4-Dichlorophenol	40.000	38.343	4.1	108	0.00
30	1,2,4-Trichlorobenzene	40.000	37.523	6.2	107	0.00
31	Naphthalene	40.000	36.887	7.8	105	0.00
32	Benzoic acid	40.000	37.037	7.4	99	0.02
33	4-Chloroaniline	40.000	35.967	10.1	102	0.00
34 C	Hexachlorobutadiene	40.000	37.502	6.2	107	0.00
35	Caprolactam	40.000	37.053	7.4	104	0.00
36 C	4-Chloro-3-methylphenol	40.000	36.810	8.0	105	0.00
37	2-Methylnaphthalene	40.000	36.344	9.1	104	0.00
38	1-Methylnaphthalene	40.000	36.134	9.7	104	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	108	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.285	6.8	103	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.164	4.6	101	0.00
42 S	2,4,6-Tribromophenol	80.000	73.011	8.7	100	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.189	4.5	104	0.00
44	2,4,5-Trichlorophenol	40.000	38.231	4.4	104	0.00
45 S	2-Fluorobiphenyl	80.000	73.149	8.6	101	0.00
46	1,1'-Biphenyl	40.000	36.887	7.8	103	0.00
47	2-Chloronaphthalene	40.000	37.340	6.6	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102221\
 Data File : BG050690.D
 Acq On : 22 Oct 2021 15:15
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 22 16:05:08 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 14 10:56:22 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.929	0.2	107	0.00
49	Acenaphthylene	40.000	36.686	8.3	102	0.00
50	Dimethylphthalate	40.000	36.354	9.1	101	0.00
51	2,6-Dinitrotoluene	40.000	38.380	4.0	104	0.00
52 C	Acenaphthene	40.000	36.544	8.6	100	0.00
53	3-Nitroaniline	40.000	37.890	5.3	103	0.00
54 P	2,4-Dinitrophenol	40.000	34.743	13.1	92	0.00
55	Dibenzofuran	40.000	36.409	9.0	102	0.00
56 P	4-Nitrophenol	40.000	36.625	8.4	100	0.00
57	2,4-Dinitrotoluene	40.000	38.417	4.0	105	0.00
58	Fluorene	40.000	36.370	9.1	101	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	35.599	11.0	97	0.00
60	Diethylphthalate	40.000	35.703	10.7	100	0.00
61	4-Chlorophenyl-phenylether	40.000	36.458	8.9	101	0.00
62	4-Nitroaniline	40.000	37.430	6.4	103	0.00
63	Azobenzene	40.000	35.655	10.9	100	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	105	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.496	3.8	101	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.029	7.4	99	0.00
67	4-Bromophenyl-phenylether	40.000	37.766	5.6	101	0.00
68	Hexachlorobenzene	40.000	37.751	5.6	102	0.00
69	Atrazine	40.000	37.231	6.9	99	0.00
70 C	Pentachlorophenol	40.000	37.345	6.6	97	0.00
71	Phenanthrene	40.000	36.729	8.2	99	0.00
72	Anthracene	40.000	36.916	7.7	99	0.00
73	Carbazole	40.000	37.163	7.1	101	0.00
74	Di-n-butylphthalate	40.000	38.624	3.4	104	0.00
75 C	Fluoranthene	40.000	36.429	8.9	99	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	105	0.00
77	Benzydine	40.000	35.202	12.0	96	0.00
78	Pyrene	40.000	37.220	7.0	99	0.00
79 S	Terphenyl-d14	80.000	75.859	5.2	102	0.00
80	Butylbenzylphthalate	40.000	40.990	-2.5	109	0.00
81	Benzo(a)anthracene	40.000	38.304	4.2	104	0.00
82	3,3'-Dichlorobenzidine	40.000	38.502	3.7	103	0.00
83	Chrysene	40.000	38.201	4.5	103	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.484	-1.2	108	0.00
85 c	Di-n-octyl phthalate	40.000	40.825	-2.1	107	0.00
86 I	Perylene-d12	20.000	20.000	0.0	104	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.724	3.2	104	0.00
88	Benzo(b)fluoranthene	40.000	39.216	2.0	104	0.00
89	Benzo(k)fluoranthene	40.000	37.641	5.9	101	0.00
90 C	Benzo(a)pyrene	40.000	38.545	3.6	102	0.00
91	Dibenzo(a,h)anthracene	40.000	38.760	3.1	104	0.00
92	Benzo(g,h,i)perylene	40.000	38.774	3.1	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG102221\
Data File : BG050690.D
Acq On : 22 Oct 2021 15:15
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Oct 22 16:05:08 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG100621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 14 10:56:22 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0