


Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG102519\
 Data File : BG043336.D
 Acq On : 25 Oct 2019 11:53
 Operator : JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG102119.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Oct 21 16:37:43 2019



AutoFind: Scans 2508, 2509, 2510; Background Corrected with Scan 2502

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.8	11862	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	31.8	12633	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	41.2	16379	PASS
197	198	0.00	2	0.4	168	PASS
198	198	100	100	100.0	39770	PASS
199	198	5	9	7.5	2966	PASS
275	198	10	60	35.8	14226	PASS
365	198	1	100	4.5	1782	PASS
441	198	0.01	100	13.5	5367	PASS
442	198	50	100	92.4	36736	PASS
443	442	15	24	18.3	6705	PASS