


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG120320\
Data File : BG047541.D
Acq On : 3 Dec 2020 16:43
Operator : CG/JU
Sample : L4865-02
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
C0B20

Manual Integrations
APPROVED

Quant Time: Dec 03 17:53:43 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG111920MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 03 15:14:44 2020
Response via : Initial Calibration






Quantitation Report (Qedit)

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG120320\
 Data File : BG047541.D
 Acq On : 3 Dec 2020 16:43
 Operator : CG/JU
 Sample : L4865-02
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C0B20

Manual Integrations
 APPROVED

Quant Time: Dec 03 17:43:39 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG111920MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 03 15:14:44 2020
 Response via : Initial Calibration



TIC: BG047541.D

(3) 1,4-Dioxane-d8 (S)
 3.581min (+0.002) 1.46ng/uL
 response 1261

Ion	Exp%	Act%
96.00	100	100
64.00	62.70	38.16#
34.00	0.00	0.00
0.00	0.00	0.00




Quantitation Report (Qedit)

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG120320\
 Data File : BG047541.D
 Acq On : 3 Dec 2020 16:43
 Operator : CG/JU
 Sample : L4865-02
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C0B20

Manual Integrations
 APPROVED

Quant Time: Dec 03 17:43:39 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG111920MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 03 15:14:44 2020
 Response via : Initial Calibration



TIC: BG047541.D

(3) 1,4-Dioxane-d8 (S)

3.581min (+0.002) 2.11ng/uL m 12/07/20 JU

response 1829

Ion	Exp%	Act%
96.00	100	100
64.00	62.70	38.16#
34.00	0.00	0.00
0.00	0.00	0.00




Quantitation Report (Qedit)

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG120320\
 Data File : BG047541.D
 Acq On : 3 Dec 2020 16:43
 Operator : CG/JU
 Sample : L4865-02
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C0B20

Manual Integrations
 APPROVED

Quant Time: Dec 03 17:43:39 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG111920MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 03 15:14:44 2020
 Response via : Initial Calibration



TIC: BG047541.D

(71) Anthracene-d10 (S)

17.600min (-0.098) 20.16ng/ul

response 240540

Ion	Exp%	Act%
188.00	100	100
94.00	4.00	3.22
80.00	4.20	3.51
0.00	0.00	0.00




Quantitation Report (Qedit)

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG120320\
 Data File : BG047541.D
 Acq On : 3 Dec 2020 16:43
 Operator : CG/JU
 Sample : L4865-02
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C0B20

Manual Integrations
 APPROVED

Quant Time: Dec 03 17:43:39 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG111920MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 03 15:14:44 2020
 Response via : Initial Calibration



TIC: BG047541.D

(71) Anthracene-d10 (S)

17.694min (-0.004) 12.39ng/ul m (2/07/2020)

response 147890

Ion	Exp%	Act%
188.00	100	100
94.00	4.00	5.44#
80.00	4.20	3.61
0.00	0.00	0.00




Quantitation Report (Qedit)

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG120320\
 Data File : BG047541.D
 Acq On : 3 Dec 2020 16:43
 Operator : CG/JU
 Sample : L4865-02
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C0B20

Manual Integrations
 APPROVED

Quant Time: Dec 03 17:43:39 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG111920MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 03 15:14:44 2020
 Response via : Initial Calibration



TIC: BG047541.D

(89) Benzo(k)fluoranthene
 24.180min (-0.068) 4.56ng/ul
 response 95324

Ion	Exp%	Act%
252.00	100	100
253.00	23.70	19.66
125.00	2.80	4.15#
0.00	0.00	0.00




Quantitation Report (Qedit)

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG120320\
 Data File : BG047541.D
 Acq On : 3 Dec 2020 16:43
 Operator : CG/JU
 Sample : L4865-02
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C0B20

Manual Integrations
 APPROVED

Quant Time: Dec 03 17:43:39 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG111920MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 03 15:14:44 2020
 Response via : Initial Calibration



TIC: BG047541.D

(89) Benzo(k)fluoranthene

24.245min (-0.004) 1.84ng/ul m 12/07/20 JU

response 38375

Ion	Exp%	Act%
252.00	100	100
253.00	23.70	30.02#
125.00	2.80	5.04#
0.00	0.00	0.00




Quantitation Report (Qedit)

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG120320\
 Data File : BG047541.D
 Acq On : 3 Dec 2020 16:43
 Operator : CG/JU
 Sample : L4865-02
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C0B20

Manual Integrations
 APPROVED

Quant Time: Dec 03 17:50:47 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG111920MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 03 15:14:44 2020
 Response via : Initial Calibration



TIC: BG047541.D

(94) Benzo(g,h,i)perylene
 30.344min (-0.021) 0.11ng/ul
 response 2011

Ion	Exp%	Act%
276.00	100	100
138.00	4.50	9.30#
277.00	21.90	25.90
0.00	0.00	0.00




Quantitation Report (Qedit)

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG120320\
Data File : BG047541.D
Acq On : 3 Dec 2020 16:43
Operator : CG/JU
Sample : L4865-02
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
C0B20

Manual Integrations
APPROVED

Quant Time: Dec 03 17:50:47 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG111920MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 03 15:14:44 2020
Response via : Initial Calibration



TIC: BG047541.D

(94) Benzo(g,h,i)perylene

30.344min (-0.021) 2.04ng/ul m 12/07/2020

response 38614

Ion	Exp%	Act%
276.00	100	100
138.00	4.50	5.90#
277.00	21.90	25.90
0.00	0.00	0.00




Quantitation Report (Qedit)

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG120320\
Data File : BG047541.D
Acq On : 3 Dec 2020 16:43
Operator : CG/JU
Sample : L4865-02
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
C0B20

Manual Integrations
APPROVED

Quant Time: Dec 03 17:50:47 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG111920MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 03 15:14:44 2020
Response via : Initial Calibration



TIC: BG047541.D

(92) Indeno(1,2,3-cd)pyrene

29.127min (-0.010) 1.87ng/ul

response 42959

Ion	Exp%	Act%
276.00	100	100
138.00	4.60	5.47
277.00	22.80	24.00
0.00	0.00	0.00




Quantitation Report (Qedit)

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG120320\
Data File : BG047541.D
Acq On : 3 Dec 2020 16:43
Operator : CG/JU
Sample : L4865-02
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
C0B20

Manual Integrations
APPROVED

Quant Time: Dec 03 17:50:47 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG111920MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 03 15:14:44 2020
Response via : Initial Calibration



TIC: BG047541.D

(92) Indeno(1,2,3-cd)pyrene

29.127min (-0.010) 1.97ng/ul m 12/07/2020

response 45299

Ion	Exp%	Act%
276.00	100	100
138.00	4.60	5.47
277.00	22.80	24.00
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG120320\
 Data File : BG047541.D
 Acq On : 3 Dec 2020 16:43
 Operator : CG/JU
 Sample : L4865-02
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 C0B20

Manual Integrations
 APPROVED

Quant Time: Dec 03 17:53:43 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG111920MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 03 15:14:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.25	152	31479	20.00	ng/ul	0.00
18) Naphthalene-d8	11.08	136	119558	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.86	164	89515	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.60	188	240540	20.00	ng/ul	0.00
78) Chrysene-d12	21.88	240	277213	20.00	ng/ul	0.00
86) Perylene-d12	25.26	264	300116	20.00	ng/ul	0.00

mohammad

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.58	96	1829m>	2.11	ng/uL>	0.00
5) Phenol-d5	7.38	99	29456	9.33	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.56	67	19008	10.56	ng/ul	0.00
9) 2-Chlorophenol-d4	7.78	132	25342	11.97	ng/ul	0.00
13) 4-Methylphenol-d8	8.95	113	17490	7.21	ng/ul	0.00
19) Nitrobenzene-d5	9.42	128	12218	12.37	ng/ul	0.00
22) 2-Nitrophenol-d4	10.14	143	15554	14.43	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.68	165	26977	10.93	ng/ul	0.00
29) 4-Chloroaniline-d4	11.20	131	26583	10.65	ng/ul	0.00
44) Dimethylphthalate-d6	14.26	166	98192	12.78	ng/ul	0.00
47) Acenaphthylene-d8	14.56	160	113415	12.77	ng/ul	0.00
52) 4-Nitrophenol-d4	15.03	143	10558	9.01	ng/ul	0.00
58) Fluorene-d10	15.85	176	92119	13.09	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.95	200	15708	11.47	ng/ul	0.00
71) Anthracene-d10	17.69	188	147890m>	12.39	ng/ul>	0.00
79) Pyrene-d10	19.96	212	201100	12.47	ng/ul	0.00
90) Benzo(a)pyrene-d12	25.02	264	229551	13.29	ng/ul	0.00

12/07/20 JU

12/4/2020 4:30:21 PM

12/07/20 JU

Target Compounds

					Ovalue	
6) Phenol	7.41	94	3603	1.191	ng/ul#	83
45) Dimethylphthalate	14.30	163	16810	2.145	ng/ul#	97
70) Phenanthrene	17.64	178	61109	4.704	ng/ul	96
77) Fluoranthene	19.63	202	187000	11.399	ng/ul	99
80) Pvrene	19.99	202	150517	7.665	ng/ul	98
83) Benzo(a)anthracene	21.85	228	73559	3.729	ng/ul#	90
85) Chrysene	21.92	228	82268	4.383	ng/ul	96
88) Benzo(b)fluoranthene	24.18	252	95324	4.499	ng/ul#	95
89) Benzo(k)fluoranthene	24.24	252	38375m>	1.838	ng/ul>	12/07/20 JU
91) Benzo(a)pyrene	25.10	252	67252	3.560	ng/ul	98
92) Indeno(1,2,3-cd)pyrene	29.13	276	45299m?	1.971	ng/ul?	
94) Benzo(a,h,i)perylene	30.34	276	38614m?	2.037	ng/ul?	12/07/20 JU

(#) = qualifier out of range (m) = manual integration (+) = signals summed