

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG122118\  
 Data File : BG038824.D  
 Acq On : 22 Dec 2018 19:12  
 Operator : JU/SJ  
 Sample : PB115917TB  
 Misc :  
 ALS Vial : 35 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 PB115917TB

Quant Time: Dec 24 04:26:58 2018  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG121218.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Dec 12 15:26:27 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 8.06  | 152  | 25835    | 20.00  | ng    | 0.00     |
| 21) Naphthalene-d8          | 10.88 | 136  | 110707   | 20.00  | ng    | 0.00     |
| 39) Acenaphthene-d10        | 14.70 | 164  | 76851    | 20.00  | ng    | 0.00     |
| 64) Phenanthrene-d10        | 17.44 | 188  | 180748   | 20.00  | ng    | 0.00     |
| 76) Chrysene-d12            | 21.72 | 240  | 180097   | 20.00  | ng    | 0.00     |
| 87) Perylene-d12            | 25.03 | 264  | 184525   | 20.00  | ng    | 0.00     |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 5.61  | 112  | 215181   | 147.73 | ng    | 0.01     |
| 7) Phenol-d6                | 7.22  | 99   | 302332   | 151.19 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 9.24  | 82   | 183641   | 84.74  | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 16.19 | 330  | 141614   | 133.71 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 13.32 | 172  | 526862   | 94.05  | ng    | 0.00     |
| 79) Terphenyl-d14           | 20.04 | 244  | 891317   | 107.71 | ng    | 0.00     |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG122118\  
Data File : BG038824.D  
Acq On : 22 Dec 2018 19:12  
Operator : JU/SJ  
Sample : PB115917TB  
Misc :  
ALS Vial : 35 Sample Multiplier: 1

Instrument :  
BNA\_G  
ClientSampleId :  
PB115917TB

Quant Time: Dec 24 04:26:58 2018  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG121218.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Dec 12 15:26:27 2018  
Response via : Initial Calibration

