

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010324\
 Data File : BG060235.D
 Acq On : 3 Jan 2024 20:21
 Operator : MA/JU
 Sample : 05994-07
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 WC-C-COMP

Quant Time: Jan 04 00:41:58 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG122923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 30 03:14:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.934	152	208326	20.000	ng	0.00
21) Naphthalene-d8	10.737	136	851487	20.000	ng	0.00
39) Acenaphthene-d10	14.574	164	661048	20.000	ng	0.00
64) Phenanthrene-d10	17.318	188	1610432	20.000	ng	0.00
76) Chrysene-d12	21.583	240	1467146	20.000	ng	0.00
86) Perylene-d12	24.762	264	1689607	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.508	112	1908367	138.263	ng	0.00
7) Phenol-d6	7.100	99	2805211	136.883	ng	0.00
23) Nitrobenzene-d5	9.098	82	1658487	89.967	ng	0.00
42) 2,4,6-Tribromophenol	16.066	330	1274197	143.996	ng	0.00
45) 2-Fluorobiphenyl	13.193	172	3830080	83.272	ng	0.00
79) Terphenyl-d14	19.944	244	5352669	70.139	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG010324\
Data File : BG060235.D
Acq On : 3 Jan 2024 20:21
Operator : MA/JU
Sample : 05994-07
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
WC-C-COMP

Quant Time: Jan 04 00:41:58 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG122923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 30 03:14:04 2023
Response via : Initial Calibration

