



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG052124\
Data File : BG061248.D
Acq On : 21 May 2024 9:51
Operator : MA/JU
Sample : DFTPP568
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP568

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG052124.MA.M
Title : SVOA CALIBRATION
Last Update : Tue May 21 15:23:39 2024



AutoFind: Scans 2505, 2506, 2507; Background Corrected with Scan 2499

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
51	198	10	80	45.3	36141	PASS
68	69	0.00	2	1.4	366	PASS
69	198	0.00	100	31.6	25243	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	38.2	30536	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	79848	PASS
199	198	5	9	6.7	5389	PASS
275	198	10	60	32.2	25747	PASS
365	198	1	100	3.5	2810	PASS
441	198	0.01	100	13.9	11132	PASS
442	198	50	100	85.4	68160	PASS
443	442	15	24	19.1	13017	PASS

m/z	Abundance
44.00	1723.0

Instrument :
BNA_G
ClientSampleId :
DFTPP568

m/z	Abundance
38.00	407.0
39.00	2579.0
40.00	433.0
44.00	1189.0
49.00	227.0
50.00	8589.0
51.00	36848.0
52.00	2011.0
55.00	191.0
56.00	886.0
57.00	2473.0
57.80	166.0
60.90	479.0
62.00	375.0
63.00	1305.0
64.10	182.0
65.00	291.0
68.00	587.0
69.00	25760.0
73.10	277.0
74.00	2326.0
75.00	4339.0
76.10	1533.0
77.00	28152.0
78.00	1809.0
79.00	2370.0
80.00	1771.0
81.00	2134.0
82.00	592.0
83.00	457.0
85.00	345.0
86.00	718.0
86.90	329.0
91.00	391.0
92.00	468.0
93.00	2744.0
94.10	232.0
96.10	155.0
97.90	2622.0
99.00	2081.0
101.00	1087.0
103.00	386.0
103.90	807.0
105.00	811.0
106.10	522.0
107.00	9718.0
108.00	1272.0
110.00	15672.0
111.00	2272.0
112.00	287.0
116.00	436.0
117.00	8006.0
118.00	625.0
118.80	155.0
122.10	506.0
123.00	1152.0
124.00	523.0
125.00	445.0
127.00	29472.0
128.10	2293.0
129.00	12644.0
130.00	841.0
131.00	210.0
133.80	303.0
134.10	281.0
135.00	1152.0
135.90	492.0
137.00	583.0
140.10	178.0
141.00	1996.0
142.00	587.0
143.00	222.0
143.90	156.0
146.00	371.0
146.90	1227.0
148.00	1869.0
149.00	479.0
149.80	152.0
150.90	331.0
152.90	702.0
154.00	585.0
155.00	1223.0
156.00	1457.0
157.10	291.0
158.10	239.0
159.10	297.0
160.00	624.0
160.90	756.0
162.00	365.0
165.00	642.0
166.10	678.0
167.00	3544.0

Instrument :
BNA_G
ClientSampleId :
DFTPP568

168.10	1652.0
169.00	461.0
170.80	153.0
171.90	294.0
173.10	474.0
174.00	672.0
175.00	1461.0
176.00	386.0
177.00	863.0
178.00	290.0
179.00	2563.0
180.00	2069.0
181.00	841.0
183.90	227.0
185.00	1525.0
186.00	11313.0
187.00	4024.0
188.10	297.0
189.00	671.0
191.10	415.0
192.00	926.0
193.00	1057.0
194.00	228.0
196.00	1972.0
198.00	72456.0
199.00	4879.0
200.00	481.0
201.50	363.0
202.10	185.0
203.10	631.0
204.00	3483.0
205.00	5281.0
206.00	21720.0
207.00	3411.0
208.00	875.0
209.10	336.0
210.00	333.0
211.00	1083.0
211.80	158.0
215.00	302.0
216.10	524.0
217.00	6206.0
218.00	807.0
219.90	203.0
221.10	2385.0
221.70	834.0
223.00	1324.0
224.00	12114.0
225.00	3212.0
226.00	364.0
227.00	5642.0
227.90	1111.0
229.00	928.0
230.10	323.0
231.10	592.0
232.00	197.0
233.90	658.0
235.00	549.0
236.00	307.0
237.00	512.0
239.00	186.0
240.10	177.0
240.90	293.0
242.10	570.0
243.10	819.0
244.10	9854.0
244.90	1104.0
246.00	2085.0
247.00	491.0
248.10	157.0
249.00	274.0
250.90	199.0
253.00	304.0
255.00	53192.0
256.00	8047.0
257.10	660.0
258.00	2944.0
259.00	698.0
259.90	180.0
265.10	1008.0
271.10	175.0
272.10	528.0
273.00	1847.0
274.00	3774.0
275.00	21816.0
276.00	2700.0
277.00	2519.0
277.90	458.0
283.00	445.0
284.90	479.0
286.10	158.0
289.90	183.0
292.90	437.0
296.00	6834.0
297.00	1180.0
303.10	944.0

Instrument :
BNA_G
ClientSampleId :
DFTPP568

314.00	223.0
315.00	706.0
316.10	524.0
317.00	181.0
321.00	194.0
323.10	1957.0
324.10	378.0
326.80	642.0
328.10	154.0
331.90	204.0
332.90	254.0
334.00	1106.0
335.00	402.0
341.00	183.0
345.90	485.0
352.00	685.0
353.00	483.0
354.10	654.0
355.00	180.0
364.90	2971.0
366.00	559.0
372.10	1042.0
373.00	392.0
383.00	360.0
384.00	176.0
402.00	470.0
403.00	618.0
421.20	449.0
421.90	727.0
423.00	2846.0
424.00	894.0
441.00	8738.0
442.00	48632.0
443.00	9257.0
444.00	933.0

Instrument :
BNA_G
ClientSampleId :
DFTPP568

m/z	Abundance
37.00	157.0
38.00	340.0
39.00	3070.0
39.90	401.0
44.00	1273.0
48.90	238.0
50.00	11291.0
51.00	41888.0
52.00	2195.0
55.00	388.0
56.00	1180.0
57.00	2412.0
58.00	223.0
61.00	477.0
62.00	576.0
63.00	1801.0
64.10	216.0
65.00	519.0
67.90	511.0
69.00	30208.0
73.00	265.0
74.00	3375.0
75.00	5420.0
76.10	1690.0
77.00	32656.0
78.10	1981.0
79.00	2315.0
80.00	1903.0
81.00	2383.0
82.10	704.0
83.00	555.0
85.00	405.0
86.00	710.0
87.00	301.0
91.00	698.0
92.00	558.0
93.00	4056.0
94.10	371.0
96.00	325.0
98.00	3288.0
99.00	2286.0
100.10	250.0
101.00	1566.0
102.10	184.0
103.00	487.0
103.90	934.0
105.00	713.0
105.90	339.0
107.00	11002.0
108.00	2032.0
110.00	18360.0
111.00	3160.0
111.90	312.0
112.80	179.0
116.10	815.0
117.00	9035.0
118.00	756.0
122.00	911.0
123.00	1120.0
124.00	772.0
125.00	388.0
127.00	34792.0
128.00	2596.0
129.00	15222.0
130.00	1435.0
131.00	239.0
133.90	457.0
135.00	1259.0
136.00	442.0
137.00	797.0
137.80	168.0
139.00	168.0
139.90	323.0
141.00	1864.0
141.90	724.0
142.90	388.0
146.00	546.0
147.00	1350.0
148.00	2618.0
149.00	487.0
151.10	393.0
153.00	932.0
154.00	588.0
155.00	1322.0
156.00	1449.0
157.00	374.0
158.00	497.0
159.00	425.0
160.10	685.0
161.10	876.0
162.00	200.0
165.00	964.0

Instrument :
BNA_G
ClientSampleId :
DFTPP568

165.90	786.0
167.00	4021.0
168.00	1903.0
169.00	449.0
170.00	188.0
171.00	205.0
171.90	363.0
173.10	631.0
174.00	1102.0
175.00	1673.0
176.10	450.0
176.90	879.0
178.00	489.0
179.00	3093.0
180.00	2412.0
181.00	1239.0
182.80	217.0
184.00	429.0
185.10	2307.0
186.00	14162.0
187.00	3520.0
188.00	418.0
189.00	988.0
190.10	244.0
191.00	457.0
192.00	1277.0
193.00	1425.0
194.10	244.0
196.00	2775.0
198.00	92592.0
198.90	6413.0
200.00	617.0
201.30	345.0
201.70	295.0
202.90	781.0
204.00	4081.0
205.10	6545.0
206.00	27448.0
207.10	4161.0
208.00	1010.0
208.90	373.0
209.80	267.0
210.30	394.0
210.80	1093.0
211.00	1125.0
215.00	570.0
216.10	612.0
217.00	8061.0
218.10	681.0
221.00	3063.0
221.70	1181.0
223.00	1926.0
224.00	16070.0
225.00	4351.0
226.00	382.0
227.00	7463.0
228.00	1061.0
229.00	1743.0
230.10	389.0
231.10	581.0
233.90	562.0
234.90	673.0
236.00	280.0
237.10	532.0
239.00	276.0
240.00	272.0
241.10	348.0
242.00	944.0
243.00	1052.0
244.00	11386.0
245.00	1745.0
246.10	3309.0
247.10	543.0
249.00	300.0
250.90	167.0
252.20	165.0
253.10	400.0
255.00	74968.0
256.00	10199.0
257.00	868.0
258.00	4249.0
259.00	682.0
261.10	180.0
265.10	2191.0
266.00	272.0
271.00	308.0
272.00	250.0
273.00	2340.0
274.10	5286.0
275.00	29664.0
276.00	4418.0
277.00	2833.0
278.10	456.0
283.00	338.0
284.10	205.0
285.10	410.0

Instrument :
BNA_G
ClientSampleId :
DFTPP568

288.90	151.0
290.10	187.0
292.10	161.0
293.10	498.0
294.00	213.0
296.00	9888.0
297.10	1242.0
301.90	327.0
303.00	1372.0
303.90	287.0
308.20	191.0
314.00	317.0
315.00	1196.0
316.10	661.0
320.80	408.0
321.80	266.0
323.00	2555.0
324.20	602.0
326.90	495.0
328.00	415.0
333.10	266.0
334.00	2183.0
335.10	385.0
340.90	229.0
346.20	524.0
352.00	751.0
353.10	636.0
353.90	694.0
355.00	254.0
365.10	4454.0
365.90	619.0
371.10	303.0
372.00	1633.0
373.10	446.0
383.00	518.0
384.00	172.0
390.00	257.0
391.00	184.0
391.90	179.0
402.10	563.0
403.00	900.0
403.80	221.0
404.30	273.0
421.00	645.0
421.90	616.0
423.00	4968.0
423.90	985.0
441.00	12738.0
442.00	78160.0
443.00	15150.0
444.00	1435.0

Instrument :
BNA_G
ClientSampleId :
DFTPP568

m/z	Abundance
38.00	221.0
39.00	2205.0
40.00	370.0
41.00	186.0
43.90	1505.0
44.90	152.0
50.00	7703.0
51.00	29688.0
52.00	1561.0
56.00	724.0
57.00	1774.0
58.00	188.0
61.00	239.0
62.00	658.0
63.00	904.0
65.00	458.0
69.00	19760.0
73.20	197.0
74.10	2329.0
75.00	3329.0
76.10	1338.0
77.00	21448.0
78.00	1573.0
79.00	1772.0
80.00	1009.0
81.00	1638.0
81.90	413.0
83.00	297.0
86.00	433.0
87.10	180.0
90.90	508.0
92.00	576.0
93.00	2786.0
98.00	2168.0
99.00	1325.0
100.00	194.0
101.00	808.0
103.10	317.0
103.90	778.0
104.90	787.0
105.90	222.0
107.00	8572.0
108.00	1034.0
110.00	13124.0
111.00	2011.0
112.00	316.0
116.00	644.0
117.00	6904.0
117.90	562.0
122.00	534.0
123.00	781.0
124.00	452.0
124.90	416.0
127.00	27344.0
128.00	2198.0
129.00	10970.0
130.00	773.0
134.00	456.0
135.00	722.0
136.00	396.0
137.00	422.0
140.10	207.0
141.00	1509.0
142.00	560.0
143.00	406.0
146.00	331.0
147.00	888.0
148.00	2135.0
149.00	516.0
151.10	169.0
151.60	221.0
153.00	445.0
154.00	530.0
155.00	1165.0
156.10	1174.0
157.00	510.0
157.90	228.0
159.00	217.0
160.00	673.0
161.00	763.0
162.10	242.0
165.10	691.0
166.10	554.0
167.00	2998.0
167.90	1278.0
168.90	400.0
169.90	213.0
172.00	278.0
173.00	534.0
173.90	463.0
175.00	1489.0
176.00	539.0

Instrument :
BNA_G
ClientSampleId :
DFTPP568

177.00	720.0
177.90	324.0
179.00	2754.0
180.00	1764.0
181.10	738.0
184.00	335.0
185.10	1525.0
186.00	11027.0
187.00	2954.0
188.00	328.0
189.00	621.0
191.20	481.0
192.00	900.0
193.00	1291.0
194.10	225.0
195.10	242.0
196.10	1616.0
198.00	74496.0
199.00	4875.0
199.90	317.0
201.60	337.0
202.90	737.0
204.00	3004.0
205.00	5316.0
206.10	22528.0
207.00	2906.0
208.00	694.0
209.10	201.0
209.90	318.0
210.90	710.0
214.90	257.0
215.90	624.0
217.00	6735.0
218.00	752.0
219.00	157.0
221.00	2571.0
221.90	1103.0
223.00	1651.0
224.00	14836.0
225.00	3844.0
227.00	6822.0
228.00	1083.0
229.00	1377.0
231.00	496.0
234.00	364.0
235.00	436.0
236.00	209.0
236.90	419.0
237.90	191.0
239.00	234.0
239.90	246.0
240.80	196.0
242.00	546.0
243.20	719.0
244.10	10341.0
245.00	1082.0
246.00	2598.0
246.90	416.0
249.00	656.0
250.90	161.0
251.80	168.0
253.10	241.0
255.00	60592.0
256.00	7624.0
257.10	468.0
258.00	4095.0
259.00	530.0
260.90	183.0
265.00	1642.0
266.00	238.0
271.00	221.0
273.10	1604.0
274.00	4421.0
275.00	25760.0
276.10	3140.0
277.00	2106.0
278.20	478.0
282.80	192.0
283.20	152.0
283.90	209.0
285.20	331.0
285.90	179.0
290.00	181.0
292.00	203.0
293.00	526.0
293.90	221.0
296.00	8903.0
297.00	849.0
302.10	152.0
303.10	871.0
307.90	220.0
314.10	643.0
314.90	1113.0
316.10	711.0
321.20	427.0
322.20	165.0

Instrument :
BNA_G
ClientSampleId :
DFTPP568

323.10	2178.0
324.00	397.0
326.90	384.0
328.10	261.0
331.90	215.0
333.00	384.0
334.10	1710.0
335.00	353.0
341.10	268.0
346.10	405.0
346.90	162.0
352.00	934.0
353.00	526.0
354.00	659.0
355.20	222.0
365.00	3977.0
365.90	586.0
370.80	165.0
372.00	1382.0
372.90	341.0
383.10	511.0
391.20	176.0
402.00	790.0
402.90	952.0
403.90	169.0
421.00	913.0
422.00	583.0
423.00	4765.0
424.00	928.0
441.00	11921.0
442.00	77688.0
443.00	14644.0
444.00	1374.0

Instrument :
BNA_G
ClientSampleId :
DFTPP568